

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: Weston Solutions, Inc.

SDG No.: Q3176

Contract: ROYF02

Lab Code: ACE

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB169833BS							
Aluminum	mg/Kg	100	94.3		94	80 - 120	P
Antimony	mg/Kg	40.0	38.8		97	80 - 120	P
Arsenic	mg/Kg	40.0	37.2		93	80 - 120	P
Barium	mg/Kg	10.0	8.95		90	80 - 120	P
Beryllium	mg/Kg	10.0	9.33		93	80 - 120	P
Cadmium	mg/Kg	10.0	9.01		90	80 - 120	P
Calcium	mg/Kg	50.0	48.3	J	97	80 - 120	P
Chromium	mg/Kg	20.0	19.0		95	80 - 120	P
Cobalt	mg/Kg	10.0	9.12		91	80 - 120	P
Copper	mg/Kg	15.0	14.5		97	80 - 120	P
Iron	mg/Kg	150	142		95	80 - 120	P
Lead	mg/Kg	50.0	44.6		89	80 - 120	P
Magnesium	mg/Kg	100	94.8	J	95	80 - 120	P
Manganese	mg/Kg	10.0	9.42		94	80 - 120	P
Nickel	mg/Kg	25.0	23.0		92	80 - 120	P
Potassium	mg/Kg	500	480		96	80 - 120	P
Selenium	mg/Kg	100	95.0		95	80 - 120	P
Silver	mg/Kg	3.8	3.38		89	80 - 120	P
Sodium	mg/Kg	150	143		95	80 - 120	P
Thallium	mg/Kg	100	95.1		95	80 - 120	P
Vanadium	mg/Kg	15.0	14.3		95	80 - 120	P
Zinc	mg/Kg	10.0	9.79		98	80 - 120	P

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: Weston Solutions, Inc.

SDG No.: Q3176

Contract: ROYF02

Lab Code: ACE

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB169842BS Mercury	mg/Kg	0.28	0.33		119	80 - 120	CV