

Continuing Calibration Report for SequenceID : FC092625AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FC069853.D

Aliphatic C9-C12	8368697.000	60.000	3.300	6.593	139478.283	133523.254	-4.460
Aliphatic C12-C16	6681561.000	40.000	6.594	9.993	167039.025	156673.343	-6.616
Aliphatic C16-C21	11414829.000	60.000	9.994	13.360	190247.150	179548.834	-5.958
Aliphatic C21-C28	15556389.000	80.000	13.361	17.023	194454.863	180941.740	-7.468
Aliphatic C28-C40	17104989.000	120.000	17.024	21.984	142541.575	128677.038	-10.775
Aliphatic EPH	59126465.000	360.000	3.300	21.984	164240.181	152688.452	-7.566