

Continuing Calibration Report for SequenceID : FC092625AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FC069858.D

Aliphatic C9-C12	7318913.000	60.000	3.300	6.593	121981.883	133523.254	8.644
Aliphatic C12-C16	5773350.000	40.000	6.594	9.993	144333.750	156673.343	7.876
Aliphatic C16-C21	9454936.000	60.000	9.994	13.360	157582.267	179548.834	12.234
Aliphatic C21-C28	12400776.000	80.000	13.361	17.023	155009.700	180941.740	14.332
Aliphatic C28-C40	13628539.000	120.000	17.024	21.984	113571.158	128677.038	11.739
Aliphatic EPH	48576514.000	360.000	3.300	21.984	134934.761	152688.452	11.627