

Continuing Calibration Report for SequenceID : FD092625AR

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FD049765.D

Aromatic C10-C12	4686196.000	40.000	4.389	6.132	117154.900	116056.185	-0.947
Aromatic C12-C16	8384234.000	60.000	6.133	8.761	139737.233	135865.661	-2.850
Aromatic C16-C21	11343302.000	80.000	8.762	13.054	141791.275	135770.533	-4.434
Aromatic C21-C36	22066183.000	180.000	13.055	18.477	122589.906	118321.901	-3.607
Aromatic EPH	46479915.000	360.000	4.389	18.477	129110.875	124871.589	-3.395