

Continuing Calibration Report for SequenceID : FD092625AR

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FD049771.D

Aromatic C10-C12	4880274.000	40.000	4.389	6.132	122006.850	116056.185	-5.127
Aromatic C12-C16	8569712.000	60.000	6.133	8.761	142828.533	135865.661	-5.125
Aromatic C16-C21	11430190.000	80.000	8.762	13.054	142877.375	135770.533	-5.234
Aromatic C21-C36	22186612.000	180.000	13.055	18.477	123258.956	118321.901	-4.173
Aromatic EPH	47066788.000	360.000	4.389	18.477	130741.078	124871.589	-4.700