

Continuing Calibration Report for SequenceID : FD100125AR

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FD049783.D

Aromatic C12-C16	8582976.000	60.000	6.132	8.760	143049.600	135865.661	-5.288
Aromatic C10-C12	4896792.000	40.000	4.388	6.131	122419.800	116056.185	-5.483
Aromatic C16-C21	11236332.000	80.000	8.761	13.050	140454.150	135770.533	-3.450
Aromatic C21-C36	21941719.000	180.000	13.051	18.471	121898.439	118321.901	-3.023
Aromatic EPH	46657819.000	360.000	4.388	18.471	129605.053	124871.589	-3.791