

Continuing Calibration Report for SequenceID : FF092525AR

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FF016462.D

Aromatic C10-C12	4862477.000	40.000	4.521	6.398	121561.925	143322.298	15.183
Aromatic C12-C16	7309251.000	60.000	6.399	9.102	121820.850	145185.730	16.093
Aromatic C16-C21	9311721.000	80.000	9.103	13.422	116396.513	144173.003	19.266
Aromatic C21-C36	19005003.000	180.000	13.423	18.876	105583.350	130834.072	19.300
Aromatic EPH	40488452.000	360.000	4.521	18.876	112467.922	137577.803	18.251