

Continuing Calibration Report for SequenceID : FG092525AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FG016647.D

Aliphatic C9-C12	6666765.000	60.000	3.279	6.920	111112.750	115503.733	3.802
Aliphatic C12-C16	4693819.000	40.000	6.921	10.379	117345.475	120612.787	2.709
Aliphatic C16-C21	7426900.000	60.000	10.380	13.768	123781.667	129289.357	4.260
Aliphatic C21-C28	9882490.000	80.000	13.769	17.453	123531.125	128839.727	4.120
Aliphatic C28-C40	12025117.000	120.000	17.454	22.488	100209.308	101419.097	1.193
Aliphatic EPH	40695091.000	360.000	3.279	22.488	113041.919	116637.685	3.083