

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ROYF02
Lab Code: ACE **SDG NO.:** Q3176
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 09/15/2025 09/15/2025

Client Sample No.: CCAL01 **Date Analyzed:** 09/25/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO114029.D **Time Analyzed:** 15:17

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.740	4.643	4.843	499.890	500.000	0.0
Aroclor-1016-2	4.759	4.661	4.861	510.650	500.000	2.1
Aroclor-1016-3	4.816	4.718	4.918	487.530	500.000	-2.5
Aroclor-1016-4	4.935	4.837	5.037	482.840	500.000	-3.4
Aroclor-1016-5	5.192	5.093	5.293	548.620	500.000	9.7
Aroclor-1260-1	6.226	6.129	6.329	479.650	500.000	-4.1
Aroclor-1260-2	6.416	6.319	6.519	438.800	500.000	-12.2
Aroclor-1260-3	6.781	6.683	6.883	425.960	500.000	-14.8
Aroclor-1260-4	7.040	6.941	7.141	459.650	500.000	-8.1
Aroclor-1260-5	7.283	7.185	7.385	474.860	500.000	-5.0
Decachlorobiphenyl	8.666	8.570	8.770	43.160	50.000	-13.7
Tetrachloro-m-xylene	3.659	3.560	3.760	54.530	50.000	9.1

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ROYF02
Lab Code: ACE **SDG NO.:** Q3176
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 09/15/2025 09/15/2025

Client Sample No.: CCAL01 **Date Analyzed:** 09/25/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO114029.D **Time Analyzed:** 15:17

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.721	4.623	4.823	497.510	500.000	-0.5
Aroclor-1016-2	4.739	4.641	4.841	495.760	500.000	-0.8
Aroclor-1016-3	4.913	4.816	5.016	490.970	500.000	-1.8
Aroclor-1016-4	4.956	4.858	5.058	493.200	500.000	-1.4
Aroclor-1016-5	5.167	5.070	5.270	509.920	500.000	2.0
Aroclor-1260-1	6.193	6.096	6.296	440.480	500.000	-11.9
Aroclor-1260-2	6.382	6.284	6.484	422.710	500.000	-15.5
Aroclor-1260-3	6.532	6.435	6.635	406.260	500.000	-18.7
Aroclor-1260-4	7.001	6.903	7.103	410.090	500.000	-18.0
Aroclor-1260-5	7.244	7.146	7.346	411.330	500.000	-17.7
Decachlorobiphenyl	8.612	8.514	8.714	46.290	50.000	-7.4
Tetrachloro-m-xylene	3.651	3.551	3.751	53.260	50.000	6.5