

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ROYF02
Lab Code: ACE **SDG NO.:** Q3176
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 09/15/2025 09/15/2025

Client Sample No.: CCAL04 **Date Analyzed:** 09/26/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO114064.D **Time Analyzed:** 10:55

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.738	4.643	4.843	507.400	500.000	1.5
Aroclor-1016-2	4.756	4.661	4.861	525.760	500.000	5.2
Aroclor-1016-3	4.814	4.718	4.918	508.410	500.000	1.7
Aroclor-1016-4	4.933	4.837	5.037	499.770	500.000	0.0
Aroclor-1016-5	5.189	5.093	5.293	573.470	500.000	14.7
Aroclor-1260-1	6.224	6.129	6.329	541.950	500.000	8.4
Aroclor-1260-2	6.415	6.319	6.519	522.520	500.000	4.5
Aroclor-1260-3	6.779	6.683	6.883	520.950	500.000	4.2
Aroclor-1260-4	7.038	6.941	7.141	579.540	500.000	15.9
Aroclor-1260-5	7.281	7.185	7.385	584.750	500.000	17.0
Decachlorobiphenyl	8.666	8.570	8.770	54.370	50.000	8.7
Tetrachloro-m-xylene	3.657	3.560	3.760	55.150	50.000	10.3

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ROYF02
Lab Code: ACE **SDG NO.:** Q3176
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 09/15/2025 09/15/2025

Client Sample No.: CCAL04 **Date Analyzed:** 09/26/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO114064.D **Time Analyzed:** 10:55

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.720	4.623	4.823	483.610	500.000	-3.3
Aroclor-1016-2	4.738	4.641	4.841	483.920	500.000	-3.2
Aroclor-1016-3	4.912	4.816	5.016	491.790	500.000	-1.6
Aroclor-1016-4	4.955	4.858	5.058	489.380	500.000	-2.1
Aroclor-1016-5	5.166	5.070	5.270	509.580	500.000	1.9
Aroclor-1260-1	6.192	6.096	6.296	463.880	500.000	-7.2
Aroclor-1260-2	6.380	6.284	6.484	453.620	500.000	-9.3
Aroclor-1260-3	6.531	6.435	6.635	428.570	500.000	-14.3
Aroclor-1260-4	7.000	6.903	7.103	439.450	500.000	-12.1
Aroclor-1260-5	7.242	7.146	7.346	439.760	500.000	-12.0
Decachlorobiphenyl	8.610	8.514	8.714	51.810	50.000	3.6
Tetrachloro-m-xylene	3.649	3.551	3.751	51.790	50.000	3.6