

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ROYF02
Lab Code: ACE **SDG NO.:** Q3176
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 09/23/2025 09/23/2025

Client Sample No.: CCAL05 **Date Analyzed:** 09/25/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP075326.D **Time Analyzed:** 14:57

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.813	5.708	5.908	506.760	500.000	1.4
Aroclor-1016-2	5.834	5.729	5.929	520.930	500.000	4.2
Aroclor-1016-3	5.897	5.792	5.992	501.030	500.000	0.2
Aroclor-1016-4	5.994	5.889	6.089	539.190	500.000	7.8
Aroclor-1016-5	6.287	6.181	6.381	537.320	500.000	7.5
Aroclor-1260-1	7.404	7.299	7.499	513.420	500.000	2.7
Aroclor-1260-2	7.657	7.551	7.751	508.080	500.000	1.6
Aroclor-1260-3	8.015	7.909	8.109	547.330	500.000	9.5
Aroclor-1260-4	8.243	8.137	8.337	587.530	500.000	17.5
Aroclor-1260-5	8.569	8.464	8.664	568.980	500.000	13.8
Decachlorobiphenyl	10.442	10.334	10.534	51.250	50.000	2.5
Tetrachloro-m-xylene	4.661	4.555	4.755	54.900	50.000	9.8

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ROYF02
Lab Code: ACE **SDG NO.:** Q3176
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 09/23/2025 09/23/2025

Client Sample No.: CCAL05 **Date Analyzed:** 09/25/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP075326.D **Time Analyzed:** 14:57

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.893	4.793	4.993	525.810	500.000	5.2
Aroclor-1016-2	4.950	4.851	5.051	534.450	500.000	6.9
Aroclor-1016-3	5.070	4.971	5.171	484.000	500.000	-3.2
Aroclor-1016-4	5.111	5.011	5.211	514.670	500.000	2.9
Aroclor-1016-5	5.324	5.225	5.425	505.590	500.000	1.1
Aroclor-1260-1	6.353	6.255	6.455	455.220	500.000	-9.0
Aroclor-1260-2	6.540	6.441	6.641	470.480	500.000	-5.9
Aroclor-1260-3	6.693	6.595	6.795	455.300	500.000	-8.9
Aroclor-1260-4	7.163	7.065	7.265	438.480	500.000	-12.3
Aroclor-1260-5	7.402	7.304	7.504	467.780	500.000	-6.4
Decachlorobiphenyl	8.802	8.703	8.903	58.350	50.000	16.7
Tetrachloro-m-xylene	3.796	3.696	3.896	54.620	50.000	9.2