

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ROYF02
Lab Code: ACE **SDG NO.:** Q3176
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 09/23/2025 09/23/2025

Client Sample No.: CCAL07 **Date Analyzed:** 09/26/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP075347.D **Time Analyzed:** 09:42

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.808	5.708	5.908	510.750	500.000	2.2
Aroclor-1016-2	5.829	5.729	5.929	518.630	500.000	3.7
Aroclor-1016-3	5.892	5.792	5.992	505.170	500.000	1.0
Aroclor-1016-4	5.989	5.889	6.089	540.290	500.000	8.1
Aroclor-1016-5	6.281	6.181	6.381	543.910	500.000	8.8
Aroclor-1260-1	7.399	7.299	7.499	518.520	500.000	3.7
Aroclor-1260-2	7.652	7.551	7.751	513.110	500.000	2.6
Aroclor-1260-3	8.010	7.909	8.109	550.690	500.000	10.1
Aroclor-1260-4	8.238	8.137	8.337	587.010	500.000	17.4
Aroclor-1260-5	8.565	8.464	8.664	571.960	500.000	14.4
Decachlorobiphenyl	10.438	10.334	10.534	52.570	50.000	5.1
Tetrachloro-m-xylene	4.655	4.555	4.755	53.580	50.000	7.2

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ROYF02
Lab Code: ACE **SDG NO.:** Q3176
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 09/23/2025 09/23/2025

Client Sample No.: CCAL07 **Date Analyzed:** 09/26/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP075347.D **Time Analyzed:** 09:42

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.893	4.793	4.993	532.150	500.000	6.4
Aroclor-1016-2	4.952	4.851	5.051	532.240	500.000	6.4
Aroclor-1016-3	5.072	4.971	5.171	494.880	500.000	-1.0
Aroclor-1016-4	5.112	5.011	5.211	544.360	500.000	8.9
Aroclor-1016-5	5.325	5.225	5.425	539.770	500.000	8.0
Aroclor-1260-1	6.355	6.255	6.455	500.150	500.000	0.0
Aroclor-1260-2	6.542	6.441	6.641	514.090	500.000	2.8
Aroclor-1260-3	6.696	6.595	6.795	494.630	500.000	-1.1
Aroclor-1260-4	7.165	7.065	7.265	472.680	500.000	-5.5
Aroclor-1260-5	7.404	7.304	7.504	501.220	500.000	0.2
Decachlorobiphenyl	8.803	8.703	8.903	60.660	50.000	21.3
Tetrachloro-m-xylene	3.797	3.696	3.896	54.630	50.000	9.3