



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.:	<u>Q3176</u>	Analytical Method:	<u>8082A</u>
Client:	<u>Weston Solutions, Inc.</u>	Datafile :	<u>PP075349.D</u>

Lab Sample ID	Parameter	Spike	Result	Units	Rec	RPD	Qual	RPD		Limits		RPD
								Qual	Low	High		
PB169825BS (Column 1)	AR1016	166.5	138	ug/kg	83				71	120		
	AR1260	166.5	147	ug/kg	88				65	130		
PB169825BS (Column 2)	AR1016	166.5	141	ug/kg	85				71	120		
	AR1260	166.5	142	ug/kg	85				65	130		