

CALIBRATION FACTOR OF INITIAL CALIBRATION

Lab Name: Alliance
Lab Code: ACE
Instrument ID: ECD_L

Contract: ROYF02
SDG NO.: Q3176

Calibration Date(s): 09/18/2025 09/18/2025
Calibration Times: 17:06 18:00

GC Column: ZB-MR1 ID: 0.32 (mm)

LAB FILE ID:		CF 100 =	<u>PL097271.D</u>	CF 075 =	<u>PL097272.D</u>		
CF 050 =		<u>PL097273.D</u>	CF 025 =	<u>PL097274.D</u>	CF 005 =	<u>PL097275.D</u>	
COMPOUND	CF 100	CF 075	CF 050	CF 025	CF 005	CF	% RSD
4,4'-DDD	3890970000	3788420000	3762720000	3580660000	3829940000	3770540000	3
4,4'-DDE	4963980000	4819990000	4737070000	4512820000	4789690000	4764710000	3
4,4'-DDT	4402840000	4274830000	4245340000	4064150000	4148830000	4227200000	3
Aldrin	6552430000	6368880000	6295000000	6072640000	6380380000	6333870000	3
alpha-BHC	7433460000	7150430000	7020140000	6635730000	6697400000	6987430000	5
alpha-Chlordane	5728330000	5597580000	5593620000	5491270000	5988720000	5679910000	3
beta-BHC	2691590000	2641310000	2679400000	2673590000	2945030000	2726180000	5
Decachlorobiphenyl	3151670000	3121880000	3187620000	3242730000	3657150000	3272210000	7
delta-BHC	6408970000	6173000000	6129940000	5874830000	5754960000	6068340000	4
Dieldrin	5667980000	5504780000	5450620000	5257570000	5529790000	5482150000	3
Endosulfan I	5248570000	5119040000	5121050000	5001710000	5416320000	5181340000	3
Endosulfan II	4694590000	4637630000	4774680000	4729590000	5755250000	4918350000	10
Endosulfan sulfate	4084400000	3990510000	4020090000	3958010000	4360620000	4082720000	4
Endrin	4777540000	4640970000	4618310000	4356250000	4672620000	4613140000	3
Endrin aldehyde	3035530000	2997840000	3059200000	3036470000	3285020000	3082810000	4
Endrin ketone	4296460000	4191230000	4224850000	4117820000	4391400000	4244350000	2
gamma-BHC (Lindane)	6940420000	6699850000	6593990000	6320490000	6510990000	6613150000	3
gamma-Chlordane	5793040000	5651990000	5610580000	5462080000	5749930000	5653520000	2
Heptachlor	6809100000	6621120000	6608040000	6426970000	6885270000	6670100000	3
Heptachlor epoxide	5764160000	5655320000	5756800000	5829540000	5728960000	5746960000	1
Methoxychlor	2052960000	2027370000	2062620000	2055200000	2228740000	2085380000	4
Tetrachloro-m-xylene	4554070000	4459430000	4465020000	4402740000	4754560000	4527160000	3

CALIBRATION FACTOR OF INITIAL CALIBRATION

Lab Name: Alliance
Lab Code: ACE
Instrument ID: ECD_L

Contract: ROYF02
SDG NO.: Q3176

Calibration Date(s): 09/18/2025 09/18/2025
Calibration Times: 17:06 18:00

GC Column: ZB-MR2 ID: 0.32 (mm)

LAB FILE ID:		CF 100 =	<u>PL097271.D</u>	CF 075 =	<u>PL097272.D</u>		
CF 050 =		<u>PL097273.D</u>	CF 025 =	<u>PL097274.D</u>	CF 005 =	<u>PL097275.D</u>	
COMPOUND	CF 100	CF 075	CF 050	CF 025	CF 005	CF	% RSD
4,4'-DDD	6589410000	6442790000	6358010000	6264340000	6590870000	6449080000	2
4,4'-DDE	7911410000	7746610000	7699440000	7406140000	7660040000	7684730000	2
4,4'-DDT	7190090000	7022180000	6972860000	6696570000	6720750000	6920490000	3
Aldrin	8860980000	8673260000	8661660000	8404070000	8796570000	8679310000	2
alpha-BHC	10445300000	10099800000	10014300000	9556100000	9672320000	9957580000	4
alpha-Chlordane	8274780000	8189370000	8245060000	8242520000	8550750000	8300500000	2
beta-BHC	3970220000	3899800000	3948240000	3983880000	4508040000	4062030000	6
Decachlorobiphenyl	5360550000	5307930000	5459740000	5623180000	6706450000	5691570000	10
delta-BHC	9586460000	9293680000	9232620000	8834360000	9014090000	9192240000	3
Dieldrin	8251310000	8119720000	8132170000	7909230000	8471500000	8176790000	3
Endosulfan I	7387320000	7345100000	7430480000	7454690000	8800570000	7683630000	8
Endosulfan II	6977690000	6911390000	6956920000	6938260000	7761900000	7109230000	5
Endosulfan sulfate	6664750000	6623680000	6716400000	6733120000	7510980000	6849790000	5
Endrin	7450260000	7335750000	7338920000	7289700000	8137930000	7510510000	5
Endrin aldehyde	4761890000	4756960000	4914450000	5156210000	7097510000	5337410000	19
Endrin ketone	7223190000	7153630000	7391340000	7391360000	8491940000	7530290000	7
gamma-BHC (Lindane)	9632950000	9373760000	9300450000	8991300000	9274840000	9314660000	2
gamma-Chlordane	8553830000	8490290000	8565810000	8624220000	9435120000	8733850000	5
Heptachlor	9299100000	9162550000	9216060000	9078210000	9769790000	9305140000	3
Heptachlor epoxide	7825700000	7749420000	7831840000	7763530000	8487220000	7931540000	4
Methoxychlor	3580590000	3582010000	3698660000	3768400000	4217720000	3769480000	7
Tetrachloro-m-xylene	6546800000	6457210000	6449880000	6384190000	7039690000	6575560000	4