

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: ROYF02

Lab Code: ACE

SDG NO.: Q3176

Continuing Calib Date: 09/30/2025

Initial Calibration Date(s): 09/18/2025

09/18/2025

Continuing Calib Time: 20:21

Initial Calibration Time(s): 17:06

18:00

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Decachlorobiphenyl	9.00	9.00	8.90	9.10	0.00
Tetrachloro-m-xylene	3.53	3.53	3.43	3.63	0.00
alpha-BHC	3.97	3.98	3.88	4.08	0.01
beta-BHC	4.49	4.49	4.39	4.59	0.00
delta-BHC	4.73	4.74	4.64	4.84	0.01
gamma-BHC (Lindane)	4.30	4.30	4.20	4.40	0.00
Heptachlor	4.89	4.89	4.79	4.99	0.00
Aldrin	5.23	5.23	5.13	5.33	0.00
Heptachlor epoxide	5.65	5.65	5.55	5.75	0.00
Endosulfan I	6.03	6.03	5.93	6.13	0.00
Dieldrin	6.30	6.31	6.21	6.41	0.01
4,4'-DDE	6.15	6.16	6.06	6.26	0.01
Endrin	6.53	6.53	6.43	6.63	0.00
Endosulfan II	6.74	6.75	6.65	6.85	0.01
4,4'-DDD	6.66	6.67	6.57	6.77	0.01
Endosulfan sulfate	7.11	7.11	7.01	7.21	0.00
4,4'-DDT	6.98	6.98	6.88	7.08	0.00
Methoxychlor	7.45	7.45	7.35	7.55	0.00
Endrin ketone	7.59	7.59	7.49	7.69	0.00
Endrin aldehyde	6.87	6.87	6.77	6.97	0.00
alpha-Chlordane	5.98	5.99	5.89	6.09	0.01
gamma-Chlordane	5.90	5.91	5.81	6.01	0.01

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: ROYF02

Lab Code: ACE

SDG NO.: Q3176

Continuing Calib Date: 09/30/2025

Initial Calibration Date(s): 09/18/2025

09/18/2025

Continuing Calib Time: 20:21

Initial Calibration Time(s): 17:06

18:00

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Decachlorobiphenyl	7.98	7.98	7.88	8.08	0.00
Tetrachloro-m-xylene	2.82	2.82	2.72	2.92	0.00
alpha-BHC	3.33	3.33	3.23	3.43	0.00
beta-BHC	3.96	3.96	3.86	4.06	0.00
delta-BHC	4.19	4.19	4.09	4.29	0.00
gamma-BHC (Lindane)	3.66	3.66	3.56	3.76	0.00
Heptachlor	4.01	4.01	3.91	4.11	0.00
Aldrin	4.29	4.29	4.19	4.39	0.00
Heptachlor epoxide	4.79	4.79	4.69	4.89	0.00
Endosulfan I	5.16	5.16	5.06	5.26	0.00
Dieldrin	5.43	5.43	5.33	5.53	0.00
4,4'-DDE	5.30	5.30	5.20	5.40	0.00
Endrin	5.70	5.70	5.60	5.80	0.00
Endosulfan II	5.99	6.00	5.90	6.10	0.01
4,4'-DDD	5.85	5.85	5.75	5.95	0.00
Endosulfan sulfate	6.39	6.40	6.30	6.50	0.01
4,4'-DDT	6.10	6.10	6.00	6.20	0.00
Methoxychlor	6.67	6.67	6.57	6.77	0.00
Endrin ketone	6.90	6.90	6.80	7.00	0.00
Endrin aldehyde	6.17	6.17	6.07	6.27	0.00
alpha-Chlordane	5.11	5.11	5.01	5.21	0.00
gamma-Chlordane	5.04	5.05	4.95	5.15	0.01