

Surrogate Summary

SDG No.: **Q3176**

Client: **Weston Solutions, Inc.**

Analytical Method: **SW8260D**

Lab Sample ID	Client ID	Parameter	Spike	Result	Recovery (%)	Qual	Limits (%)	
							Low	High
Q3176-01	EME-TS12-01	1,2-Dichloroethane-d4	50	65.0	130		63	155
		Dibromofluoromethane	50	56.5	113		70	134
		Toluene-d8	50	53.0	106		74	123
		4-Bromofluorobenzene	50	45.8	92		17	146
Q3176-02	EME-TS13-01	1,2-Dichloroethane-d4	50	66.5	133		63	155
		Dibromofluoromethane	50	56.9	114		70	134
		Toluene-d8	50	54.5	109		74	123
		4-Bromofluorobenzene	50	46.0	92		17	146
Q3176-03	EME-TS14-02	1,2-Dichloroethane-d4	50	68.0	136		63	155
		Dibromofluoromethane	50	55.5	111		70	134
		Toluene-d8	50	50.7	101		74	123
		4-Bromofluorobenzene	50	41.6	83		17	146
VW0924SBL01	VW0924SBL01	1,2-Dichloroethane-d4	50	58.3	117		63	155
		Dibromofluoromethane	50	54.6	109		70	134
		Toluene-d8	50	52.4	105		74	123
		4-Bromofluorobenzene	50	46.4	93		17	146
VW0924SBS01	VW0924SBS01	1,2-Dichloroethane-d4	50	49.6	99		63	155
		Dibromofluoromethane	50	51.3	103		70	134
		Toluene-d8	50	49.2	98		74	123
		4-Bromofluorobenzene	50	49.3	99		17	146
VW0924SBSD0	VW0924SBSD01	1,2-Dichloroethane-d4	50	50.4	101		63	155
		Dibromofluoromethane	50	49.9	100		70	134
		Toluene-d8	50	49.6	99		74	123
		4-Bromofluorobenzene	50	50.4	101		17	146