

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: ROYF02
 Lab Code: ACE SDG NO.: Q3176
 Lab File ID: VW032241.D Date Analyzed: 09/24/2025
 Instrument ID: MSVOA_W Time Analyzed: 12:59
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

	IS4 AREA #	RT #				
12 HOUR STD	147797	13.556				
UPPER LIMIT	295594	14.056				
LOWER LIMIT	73898.5	13.056				
EPA SAMPLE NO.						
EME-TS12-01	126415	13.56				
EME-TS13-01	140662	13.56				
EME-TS14-02	124449	13.56				
VW0924SBL01	144992	13.56				
VW0924SBS01	167865	13.56				
VW0924SBSD01	165345	13.56				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.