



284 Sheffield Street, Mountainside, NJ 07092  
Phone: 908 789 8900 Fax: 908 789 8922

## **CASE NARRATIVE**

**Weston Solutions, Inc.**

**Project Name: RFP 916**

**Project # N/A**

**Order ID # Q3177**

**Test Name: SPLP Pesticide**

### **A. Number of Samples and Date of Receipt:**

3 Solid samples were received on 09/23/2025.

### **B. Parameters**

According to the Chain of Custody document, the following analyses were requested: SPLP Pesticide. This data package contains results for SPLP Pesticide.

### **C. Analytical Techniques:**

The analysis was performed on instrument ECD\_D. The front column is ZB-MR1 which is 30 meters, 0.32 mm ID, 0.5 um df.; Catalog # 7HM-G016-17. The rear column is ZB-MR2 which is 30 meters, 0.32 mm ID, 0.25 um df, Catalog #: 7HMG017- 11. The analysis of SPLP Pesticides was based on method 8081B and extraction was done based on method 3541.

### **D. QA/ QC Samples:**

The Holding Times were met for all analysis.

The Surrogate recoveries were met for all analysis.

The Retention Times were met for all analysis.

The MS {Q3179-02MS} with File ID: PD090429.D recoveries met the requirements.

The MSD {Q3179-03MSD} with File ID: PD090430.D recoveries met the requirements.

The RPD were met for all analysis.

The Blank Spike met requirements for all compounds.

The Blank analysis did not indicate the presence of lab contamination.

The Initial Calibration met the requirements.

The Continuous Calibration File ID PD090418.D met the requirements.

### **E. Additional Comments:**

### **F. Calculation for Concentration in Water Samples:**

Concentration ug/L = (Ax) (Vt) (DF) (GPC)  
(CF) (Vo) (Vi)

Where,

Ax = Response (peak area or height) of the compound to be measured.

CF = Mean Calibration Factor from the initial calibration (area/ng).

Vo = Volume of water extracted in mL.



284 Sheffield Street, Mountainside, NJ 07092  
Phone: 908 789 8900 Fax: 908 789 8922

$V_i$  = Volume of extract injected in uL.

$V_t$  = Volume of the concentrated extract in uL

GPC =  $V_{in}$  = GPC factor (If no GPC is performed, GPC=1)

$V_{out}$

$V_{in}$  = Volume of extract loaded onto GPC column.

$V_{out}$  = Volume of extract collected after GPC cleanup.

DF = Dilution Factor.

**G. Manual Integration Comments:**

Please refer to the Manual integration Report included with the Run Logs for information on the manual integrations performed.

---

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. The laboratory manager or his designee, as verified by the following signature has authorized release of the data contained in this hard copy data package.

Signature\_\_\_\_\_