



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Initial and Continuing Calibration Verification

Client: Weston Solutions, Inc.

SDG No.: Q3177

Project: RFP 916

RunNo.: LB137352

Analyte		Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: Cyanide	ICV1	mg/L	0.099	0.099	100	90-110	09/29/2025
Sample ID: Cyanide	CCV1	mg/L	0.24	0.25	96	90-110	09/29/2025
Sample ID: Cyanide	CCV2	mg/L	0.24	0.25	96	90-110	09/29/2025