

Laboratory Control Sample Summary

Client: Weston Solutions, Inc.

SDG No.: Q3177

Project: RFP 916

Run No.: LB137352

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	PB169878BS							
Cyanide	mg/L	0.1	0.095		95	1	85-115	09/29/2025