

LB137307


7470A

INSTRUMENT-ID:CV1

Linear

μ Abs. :

40543



A= 0.0000e+000

B= 2.4935e-004

C= -4.1363e-002

Rho= 0.9995962

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.030	-0.030	44	0.000	44	0			
0.2	0.200	0.172	-0.028	854	0.0 %	854	0			
2.5	2.500	2.577	0.077	10500	0.0 %	10500	0			
5.0	5.000	5.115	0.115	20679	0.0 %	20679				
7.5	7.500	7.299	-0.201	29437	0.0 %	29437				
10.0	10.000	10.068	0.068	40543	0.0 %	40543				

LB137307 INSTRUMENT ID:CV1

Method: 7470A Operator: Admin Date of Analysis: 25 Sep 2025 14:26:47

Type	Sample ID	Extended ID	Conc	µ Abs	Units	Type	Std Cond	Method	Date
S	0.0 - 1	50	-	44	PPB	Std	0.0000	7470A	25 Sep 2025 14:33:13
S	0.2 - 1	50.2	-	854	PPB	Std	0.2000	7470A	25 Sep 2025 14:35:30
S	2.5 - 1	50.5	-	10500	PPB	Std	2.5000	7470A	25 Sep 2025 14:37:47
S	5.0 - 1	55.0	-	20679	PPB	Std	5.0000	7470A	25 Sep 2025 14:40:04
S	7.5 - 1	57.5	-	29437	PPB	Std	7.5000	7470A	25 Sep 2025 14:42:22
S	10.0 - 1	510.0	-	40543	PPB	Std	10.0000	7470A	25 Sep 2025 14:48:13
U	ICV47 - 1	ICV47	3.8603	15647	PPB	SMPL	-	7470A	25 Sep 2025 14:52:19
U	ICB47 - 1	ICB47	-0.0867	-182	PPB	SMPL	-	7470A	25 Sep 2025 14:54:33
U	CCV42 - 1	CCV42	5.0330	20350	PPB	SMPL	-	7470A	25 Sep 2025 14:58:45
U	CCB42 - 1	CCB42	-0.1005	-237	PPB	SMPL	-	7470A	25 Sep 2025 15:01:00
U	CRA - 1	CRA	0.1546	786	PPB	SMPL	-	7470A	25 Sep 2025 15:03:19
U	HighStd - 1	HighStd	9.8330	39600	PPB	SMPL	-	7470A	25 Sep 2025 15:05:34
U	ChkStd - 1	ChkStd	7.0173	28308	PPB	SMPL	-	7470A	25 Sep 2025 15:07:49
U	PB169846BL - 1	PBW	-0.0296	47	PPB	SMPL	-	7470A	25 Sep 2025 15:12:59
U	PB169846BS - 1	LCSW	4.1423	16778	PPB	SMPL	-	7470A	25 Sep 2025 15:17:46
U	Q3168-04 - 1	ARS20-0010-Oily Soil-Sludge	-0.0855	-177	PPB	SMPL	-	7470A	25 Sep 2025 15:20:02
U	Q3168-04DUP - 1	ARS20-0010-Oily Soil-SludgeDUP	-0.0389	10	PPB	SMPL	-	7470A	25 Sep 2025 15:22:22
U	Q3168-04MS - 1	ARS20-0010-Oily Soil-SludgeMS	4.5659	18477	PPB	SMPL	-	7470A	25 Sep 2025 15:24:39
U	Q3168-04MSD - 1	ARS20-0010-Oily Soil-SludgeMSD	4.3460	17595	PPB	SMPL	-	7470A	25 Sep 2025 15:26:56
U	Q3170-02 - 1	MOD-25-0271	0.0579	398	PPB	SMPL	-	7470A	25 Sep 2025 15:29:14
U	CCV43 - 1	CCV43	4.8858	19760	PPB	SMPL	-	7470A	25 Sep 2025 15:31:32
U	CCB43 - 1	CCB43	-0.0962	-220	PPB	SMPL	-	7470A	25 Sep 2025 15:33:47
U	PB169847BL - 1	PBW	-0.0306	43	PPB	SMPL	-	7470A	25 Sep 2025 15:36:06
U	PB169847BS - 1	LCSW	4.3213	17496	PPB	SMPL	-	7470A	25 Sep 2025 15:38:21
U	Q3177-01 - 1	EME-TS12-01	-0.0915	-201	PPB	SMPL	-	7470A	25 Sep 2025 15:40:37
U	Q3177-02 - 1	EME-TS13-01	-0.0414	0	PPB	SMPL	-	7470A	25 Sep 2025 15:42:56
U	Q3177-02 - 1	EME-TS13-01	-0.0436	-9	PPB	SMPL	-	7470A	25 Sep 2025 15:45:44
U	Q3177-03 - 1	EME-TS14-02	-0.0321	37	PPB	SMPL	-	7470A	25 Sep 2025 15:48:07
U	Q3179-01 - 1	EME-TS14-01	-0.0371	17	PPB	SMPL	-	7470A	25 Sep 2025 15:50:24
U	Q3179-01DUP - 1	EME-TS14-01DUP	-0.0326	35	PPB	SMPL	-	7470A	25 Sep 2025 15:52:40
U	Q3179-02 - 1	EME-TS14-01MS	4.4213	17897	PPB	SMPL	-	7470A	25 Sep 2025 15:54:57
U	Q3179-03 - 1	EME-TS14-01MSD	3.9707	16090	PPB	SMPL	-	7470A	25 Sep 2025 15:57:14
U	CCV44 - 1	CCV44	4.7741	19312	PPB	SMPL	-	7470A	25 Sep 2025 15:59:32
U	CCB44 - 1	CCB44	-0.0937	-210	PPB	SMPL	-	7470A	25 Sep 2025 16:01:57
U	PB169816TB - 1	PB169816TB	-0.0406	3	PPB	SMPL	-	7470A	25 Sep 2025 16:04:15
U	PB169847TB - 1	PB169847TB	-0.0394	8	PPB	SMPL	-	7470A	25 Sep 2025 16:06:31
U	Q3168-04LX5 - 1		-0.0306	43	PPB	SMPL	-	7470A	25 Sep 2025 16:08:53
U	Q3168-04A - 1		3.7154	15066	PPB	SMPL	-	7470A	25 Sep 2025 16:11:09
U	Q3179-01LX5 - 1		-0.0685	-109	PPB	SMPL	-	7470A	25 Sep 2025 16:13:24
U	Q3179-01A - 1		3.7201	15085	PPB	SMPL	-	7470A	25 Sep 2025 16:15:56
U	CCV45 - 1	CCV45	4.8741	19713	PPB	SMPL	-	7470A	25 Sep 2025 16:21:02
U	CCB45 - 1	CCB45	-0.1139	-291	PPB	SMPL	-	7470A	25 Sep 2025 16:23:18