

Continuing Calibration Report for SequenceID : FC100125AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FC069874.D

Aliphatic C9-C12	8434215.000	60.000	3.284	6.578	140570.250	133523.254	-5.278
Aliphatic C12-C16	6489004.000	40.000	6.579	9.980	162225.100	156673.343	-3.544
Aliphatic C16-C21	10057972.000	60.000	9.981	13.346	167632.867	179548.834	6.637
Aliphatic C21-C28	12806937.000	80.000	13.347	17.008	160086.713	180941.740	11.526
Aliphatic C28-C40	14869752.000	120.000	17.009	21.963	123914.600	128677.038	3.701
Aliphatic EPH	52657880.000	360.000	3.284	21.963	146271.889	152688.452	4.202