

Continuing Calibration Report for SequenceID : FC100125AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FC069882.D

Aliphatic C9-C12	8667968.000	60.000	3.284	6.578	144466.133	133523.254	-8.195
Aliphatic C12-C16	6739650.000	40.000	6.579	9.980	168491.250	156673.343	-7.543
Aliphatic C16-C21	10280195.000	60.000	9.981	13.346	171336.583	179548.834	4.574
Aliphatic C21-C28	12718137.000	80.000	13.347	17.008	158976.713	180941.740	12.139
Aliphatic C28-C40	14047588.000	120.000	17.009	21.963	117063.233	128677.038	9.026
Aliphatic EPH	52453538.000	360.000	3.284	21.963	145704.272	152688.452	4.574