

Continuing Calibration Report for SequenceID : FC100125AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FC069891.D

Aliphatic C9-C12	8525121.000	60.000	3.284	6.578	142085.350	133523.254	-6.412
Aliphatic C12-C16	6596209.000	40.000	6.579	9.980	164905.225	156673.343	-5.254
Aliphatic C16-C21	10060683.000	60.000	9.981	13.346	167678.050	179548.834	6.611
Aliphatic C21-C28	12282745.000	80.000	13.347	17.008	153534.313	180941.740	15.147
Aliphatic C28-C40	13044458.000	120.000	17.009	21.963	108703.817	128677.038	15.522
Aliphatic EPH	50509216.000	360.000	3.284	21.963	140303.378	152688.452	8.111