

Continuing Calibration Report for SequenceID : FD093025AR

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FD049777.D

Aromatic C10-C12	4919271.000	40.000	4.384	6.128	122981.775	116056.185	-5.967
Aromatic C12-C16	8626440.000	60.000	6.129	8.757	143774.000	135865.661	-5.821
Aromatic C16-C21	11198720.000	80.000	8.758	13.048	139984.000	135770.533	-3.103
Aromatic C21-C36	21751349.000	180.000	13.049	18.469	120840.828	118321.901	-2.129
Aromatic EPH	46495780.000	360.000	4.384	18.469	129154.944	124871.589	-3.430