

Continuing Calibration Report for SequenceID : FD093025AR

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FD049780.D

Aromatic C10-C12	4941191.000	40.000	4.384	6.128	123529.775	116056.185	-6.440
Aromatic C12-C16	8656586.000	60.000	6.129	8.757	144276.433	135865.661	-6.191
Aromatic C16-C21	11260309.000	80.000	8.758	13.048	140753.863	135770.533	-3.670
Aromatic C21-C36	21968202.000	180.000	13.049	18.469	122045.567	118321.901	-3.147
Aromatic EPH	46826288.000	360.000	4.384	18.469	130073.022	124871.589	-4.165