

Continuing Calibration Report for SequenceID : FE092925AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FE056035.D

Aliphatic C12-C16	5565780.000	40.000	6.947	10.396	139144.500	153376.491	9.279
Aliphatic C16-C21	9381446.000	60.000	10.397	13.773	156357.433	165879.737	5.740
Aliphatic C21-C28	12153072.000	80.000	13.774	17.443	151913.400	148527.151	-2.280
Aliphatic C28-C40	17347687.000	120.000	17.444	22.449	144564.058	136251.923	-6.101
Aliphatic EPH	52077379.000	360.000	3.315	22.449	144659.386	146449.789	1.223
Aliphatic C9-C12	7629394.000	60.000	3.315	6.946	127156.567	140027.957	9.192