

Continuing Calibration Report for SequenceID : FE092925AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FE056045.D

Aliphatic C12-C16	5652856.000	40.000	6.947	10.396	141321.400	153376.491	7.860
Aliphatic C16-C21	9543106.000	60.000	10.397	13.773	159051.767	165879.737	4.116
Aliphatic C21-C28	11632420.000	80.000	13.774	17.443	145405.250	148527.151	2.102
Aliphatic C28-C40	15735127.000	120.000	17.444	22.449	131126.058	136251.923	3.762
Aliphatic EPH	49968663.000	360.000	3.315	22.449	138801.842	146449.789	5.222
Aliphatic C9-C12	7405154.000	60.000	3.315	6.946	123419.233	140027.957	11.861