

Continuing Calibration Report for SequenceID : FE092925AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FE056059.D

Aliphatic C12-C16	6419588.000	40.000	6.947	10.396	160489.700	153376.491	-4.638
Aliphatic C16-C21	10175583.000	60.000	10.397	13.773	169593.050	165879.737	-2.239
Aliphatic C21-C28	11474864.000	80.000	13.774	17.443	143435.800	148527.151	3.428
Aliphatic C28-C40	16792154.000	120.000	17.444	22.449	139934.617	136251.923	-2.703
Aliphatic EPH	53005288.000	360.000	3.315	22.449	147236.911	146449.789	-0.537
Aliphatic C9-C12	8143099.000	60.000	3.315	6.946	135718.317	140027.957	3.078