

Continuing Calibration Report for SequenceID : FF092925AR

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FF016480.D

Aromatic C10-C12	5842008.000	40.000	4.521	6.397	146050.200	143322.298	-1.903
Aromatic C12-C16	8833480.000	60.000	6.398	9.101	147224.667	145185.730	-1.404
Aromatic C16-C21	10927914.000	80.000	9.102	13.420	136598.925	144173.003	5.253
Aromatic C21-C36	22884163.000	180.000	13.421	18.876	127134.239	130834.072	2.828
Aromatic EPH	48487565.000	360.000	4.521	18.876	134687.681	137577.803	2.101