

WATER EPH MATRIX SPIKE/MATRIX SPIKE DUPLICATE RECOVERY

Lab Name:	<u>Alliance</u>	Client:	<u>Weston Solutions, Inc.</u>
Lab Code:	<u>ACE</u>	SDG No:	<u>Q3177</u>
Sample No :	<u>Q3179-02</u>	Datafile:	<u>FE056056.D</u>
		Client ID :	<u>Q3179-02</u>

COMPOUND	SPIKE ADDED ug/l	SAMPLE CONCENTRATION ug/l	MS/MSD CONCENTRATION ug/l	% REC	Qual	QC LIMITS (%)
Aliphatic C9-C12	300	8.68	333	108		(40-140)
Aliphatic C12-C16	200	17.2	293	137		(40-140)
Aliphatic C16-C21	300	100	404	101		(40-140)
Aliphatic C21-C28	400	0	486	121		(40-140)
Aliphatic C28-C40	600	403	1130	121		(40-140)