

WATER EPH MATRIX SPIKE/MATRIX SPIKE DUPLICATE RECOVERY

Lab Name: Alliance **Client:** Weston Solutions, Inc.
Lab Code: ACE **SDG No:** Q3177
Sample No : Q3179-03 **Datafile:** FE056057.D **Client ID :** Q3179-03

COMPOUND	SPIKE ADDED ug/l	SAMPLE CONCENTRATION ug/l	MS/MSD CONCENTRATION ug/l	% REC	Qual	RPD QC LIMITS (%) QC Limit Of RPD
Aliphatic C9-C12	300	8.68	329	106		1.87 (40-140) 50
Aliphatic C12-C16	200	17.2	286	134		2.21 (40-140) 50
Aliphatic C16-C21	300	100	401	100		1 (40-140) 50
Aliphatic C21-C28	400	0	491	122		0.82 (40-140) 50
Aliphatic C28-C40	600	403	1120	119		1.67 (40-140) 50