

### CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: ROYF02

Lab Code: ACE

SDG NO.: Q3177

Continuing Calib Date: 09/29/2025

Initial Calibration Date(s): 09/23/2025

09/23/2025

Continuing Calib Time: 19:27

Initial Calibration Time(s): 08:45

17:14

GC Column: ZB-MR1 ID: 0.32 (mm)

| COMPOUND             | CCAL<br>RT | AVG<br>RT | RT WINDOW |       | DIFF<br>RT |
|----------------------|------------|-----------|-----------|-------|------------|
|                      |            |           | FROM      | TO    |            |
| Aroclor-1248-1 (1)   | 5.80       | 5.81      | 5.71      | 5.91  | 0.01       |
| Aroclor-1248-2 (2)   | 6.08       | 6.08      | 5.98      | 6.18  | 0.00       |
| Aroclor-1248-3 (3)   | 6.28       | 6.29      | 6.19      | 6.39  | 0.01       |
| Aroclor-1248-4 (4)   | 6.68       | 6.68      | 6.58      | 6.78  | 0.00       |
| Aroclor-1248-5 (5)   | 6.72       | 6.72      | 6.62      | 6.82  | 0.00       |
| Tetrachloro-m-xylene | 4.65       | 4.66      | 4.56      | 4.76  | 0.01       |
| Decachlorobiphenyl   | 10.43      | 10.44     | 10.34     | 10.54 | 0.01       |

### CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: ROYF02

Lab Code: ACE

SDG NO.: Q3177

Continuing Calib Date: 09/29/2025

Initial Calibration Date(s): 09/23/2025

09/23/2025

Continuing Calib Time: 19:27

Initial Calibration Time(s): 08:45

17:14

GC Column: ZB-MR2 ID: 0.32 (mm)

| COMPOUND             | CCAL<br>RT | AVG<br>RT | RT WINDOW |      | DIFF<br>RT |
|----------------------|------------|-----------|-----------|------|------------|
|                      |            |           | FROM      | TO   |            |
| Aroclor-1248-1 (1)   | 4.89       | 4.89      | 4.79      | 4.99 | 0.00       |
| Aroclor-1248-2 (2)   | 5.11       | 5.11      | 5.01      | 5.21 | 0.00       |
| Aroclor-1248-3 (3)   | 5.15       | 5.15      | 5.05      | 5.25 | 0.00       |
| Aroclor-1248-4 (4)   | 5.32       | 5.33      | 5.23      | 5.43 | 0.01       |
| Aroclor-1248-5 (5)   | 5.71       | 5.71      | 5.61      | 5.81 | 0.00       |
| Tetrachloro-m-xylene | 3.79       | 3.80      | 3.70      | 3.90 | 0.01       |
| Decachlorobiphenyl   | 8.80       | 8.80      | 8.70      | 8.90 | 0.00       |