

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ROYF02
Lab Code: ACE **SDG NO.:** Q3177
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 09/15/2025 09/15/2025

Client Sample No.: CCAL02 **Date Analyzed:** 09/30/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO114127.D **Time Analyzed:** 09:23

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.739	4.643	4.843	513.430	500.000	2.7
Aroclor-1016-2	4.757	4.661	4.861	510.770	500.000	2.2
Aroclor-1016-3	4.814	4.718	4.918	496.110	500.000	-0.8
Aroclor-1016-4	4.933	4.837	5.037	496.980	500.000	-0.6
Aroclor-1016-5	5.189	5.093	5.293	564.690	500.000	12.9
Aroclor-1260-1	6.223	6.129	6.329	513.360	500.000	2.7
Aroclor-1260-2	6.413	6.319	6.519	504.540	500.000	0.9
Aroclor-1260-3	6.778	6.683	6.883	478.480	500.000	-4.3
Aroclor-1260-4	7.036	6.941	7.141	510.820	500.000	2.2
Aroclor-1260-5	7.281	7.185	7.385	511.030	500.000	2.2
Decachlorobiphenyl	8.663	8.570	8.770	45.380	50.000	-9.2
Tetrachloro-m-xylene	3.658	3.560	3.760	52.740	50.000	5.5

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ROYF02
Lab Code: ACE **SDG NO.:** Q3177
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 09/15/2025 09/15/2025

Client Sample No.: CCAL02 **Date Analyzed:** 09/30/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO114127.D **Time Analyzed:** 09:23

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.718	4.623	4.823	500.270	500.000	0.1
Aroclor-1016-2	4.736	4.641	4.841	460.210	500.000	-8.0
Aroclor-1016-3	4.911	4.816	5.016	473.360	500.000	-5.3
Aroclor-1016-4	4.953	4.858	5.058	474.460	500.000	-5.1
Aroclor-1016-5	5.164	5.070	5.270	499.500	500.000	-0.1
Aroclor-1260-1	6.190	6.096	6.296	441.060	500.000	-11.8
Aroclor-1260-2	6.379	6.284	6.484	434.580	500.000	-13.1
Aroclor-1260-3	6.529	6.435	6.635	404.450	500.000	-19.1
Aroclor-1260-4	6.998	6.903	7.103	409.700	500.000	-18.1
Aroclor-1260-5	7.240	7.146	7.346	404.420	500.000	-19.1
Decachlorobiphenyl	8.607	8.514	8.714	46.790	50.000	-6.4
Tetrachloro-m-xylene	3.648	3.551	3.751	49.450	50.000	-1.1