



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q3177

Analytical Method: 8082A

Client: Weston Solutions, Inc.

Datafile : PP075439.D

Lab Sample ID	Parameter	Spike	Result	Units	Rec	RPD	Qual	RPD	Low	Limits	RPD
								Qual		High	
PB169881BS (Column 1)	AR1016	5	5.20	ug/L	104				77	107	
	AR1260	5	5.00	ug/L	100				66	113	
PB169881BS (Column 2)	AR1016	5	5.10	ug/L	102				77	107	
	AR1260	5	4.90	ug/L	98				66	113	