

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: ROYF02  
 Lab Code: ACE SDG NO.: Q3177  
 Lab File ID: VN087948.D Date Analyzed: 09/25/2025  
 Instrument ID: MSVOA\_N Time Analyzed: 12:04  
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

|                | IS4<br>AREA # | RT #  |  |  |  |  |
|----------------|---------------|-------|--|--|--|--|
| 12 HOUR STD    | 240530        | 13.77 |  |  |  |  |
| UPPER LIMIT    | 481060        | 14.27 |  |  |  |  |
| LOWER LIMIT    | 120265        | 13.27 |  |  |  |  |
| EPA SAMPLE NO. |               |       |  |  |  |  |
| EME-TS12-01    | 179435        | 13.77 |  |  |  |  |
| EME-TS13-01    | 173273        | 13.77 |  |  |  |  |
| EME-TS14-02    | 168811        | 13.77 |  |  |  |  |
| EME-TS14-01MS  | 221292        | 13.77 |  |  |  |  |
| EME-TS14-01MSD | 236701        | 13.77 |  |  |  |  |
| VN0925WBL01    | 171513        | 13.77 |  |  |  |  |
| VN0925WBS01    | 244468        | 13.77 |  |  |  |  |

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area  
 AREA LOWER LIMIT = -50% of internal standard area  
 RT UPPER LIMIT = +0.50 minutes of internal standard RT  
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

# Column used to flag values outside QC limits with an asterisk.  
 \* Values outside of QC limits.