

Initial and Continuing Calibration Verification

Client: Weston Solutions, Inc.

SDG No.: Q3178

Project: RFP 916

RunNo.: LB137340

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV1 Cyanide	mg/L	0.095	0.099	96	90-110	09/26/2025
Sample ID: CCV1 Cyanide	mg/L	0.24	0.25	96	90-110	09/26/2025
Sample ID: CCV2 Cyanide	mg/L	0.24	0.25	96	90-110	09/26/2025
Sample ID: CCV3 Cyanide	mg/L	0.24	0.25	96	90-110	09/26/2025