

Laboratory Control Sample Summary

Client: Weston Solutions, Inc.

SDG No.: Q3178

Project: RFP 916

Run No.: LB137340

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	PB169871BS							
Cyanide	mg/Kg	5	4.80		96	1	85-115	09/26/2025