

Continuing Calibration Report for SequenceID : FG092525AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FG016658.D

Aliphatic C9-C12	6204342.000	60.000	3.279	6.920	103405.700	115503.733	10.474
Aliphatic C12-C16	4278189.000	40.000	6.921	10.379	106954.725	120612.787	11.324
Aliphatic C16-C21	6517273.000	60.000	10.380	13.768	108621.217	129289.357	15.986
Aliphatic C21-C28	8123119.000	80.000	13.769	17.453	101538.988	128839.727	21.190
Aliphatic C28-C40	10096505.000	120.000	17.454	22.488	84137.542	101419.097	17.040
Aliphatic EPH	35219428.000	360.000	3.279	22.488	97831.744	116637.685	16.123