

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ROYF02
Lab Code: ACE **SDG NO.:** Q3178
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 09/15/2025 09/15/2025

Client Sample No.: CCAL02 **Date Analyzed:** 09/25/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO114043.D **Time Analyzed:** 20:06

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.740	4.643	4.843	526.480	500.000	5.3
Aroclor-1016-2	4.758	4.661	4.861	528.990	500.000	5.8
Aroclor-1016-3	4.816	4.718	4.918	512.580	500.000	2.5
Aroclor-1016-4	4.935	4.837	5.037	514.840	500.000	3.0
Aroclor-1016-5	5.192	5.093	5.293	573.940	500.000	14.8
Aroclor-1260-1	6.227	6.129	6.329	551.050	500.000	10.2
Aroclor-1260-2	6.417	6.319	6.519	523.960	500.000	4.8
Aroclor-1260-3	6.781	6.683	6.883	522.580	500.000	4.5
Aroclor-1260-4	7.040	6.941	7.141	570.000	500.000	14.0
Aroclor-1260-5	7.284	7.185	7.385	578.240	500.000	15.6
Decachlorobiphenyl	8.669	8.570	8.770	52.050	50.000	4.1
Tetrachloro-m-xylene	3.658	3.560	3.760	55.090	50.000	10.2

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ROYF02
Lab Code: ACE **SDG NO.:** Q3178
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 09/15/2025 09/15/2025

Client Sample No.: CCAL02 **Date Analyzed:** 09/25/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO114043.D **Time Analyzed:** 20:06

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.720	4.623	4.823	508.250	500.000	1.7
Aroclor-1016-2	4.738	4.641	4.841	511.250	500.000	2.3
Aroclor-1016-3	4.913	4.816	5.016	516.910	500.000	3.4
Aroclor-1016-4	4.955	4.858	5.058	512.890	500.000	2.6
Aroclor-1016-5	5.167	5.070	5.270	528.350	500.000	5.7
Aroclor-1260-1	6.193	6.096	6.296	491.140	500.000	-1.8
Aroclor-1260-2	6.381	6.284	6.484	473.990	500.000	-5.2
Aroclor-1260-3	6.532	6.435	6.635	460.910	500.000	-7.8
Aroclor-1260-4	7.001	6.903	7.103	475.260	500.000	-4.9
Aroclor-1260-5	7.244	7.146	7.346	469.730	500.000	-6.1
Decachlorobiphenyl	8.613	8.514	8.714	51.700	50.000	3.4
Tetrachloro-m-xylene	3.649	3.551	3.751	53.920	50.000	7.8