

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ROYF02
Lab Code: ACE **SDG NO.:** Q3178
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 09/23/2025 09/23/2025

Client Sample No.: CCAL04 **Date Analyzed:** 09/25/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP075341.D **Time Analyzed:** 20:13

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.812	5.708	5.908	527.230	500.000	5.4
Aroclor-1016-2	5.834	5.729	5.929	521.840	500.000	4.4
Aroclor-1016-3	5.896	5.792	5.992	515.760	500.000	3.2
Aroclor-1016-4	5.993	5.889	6.089	549.760	500.000	10.0
Aroclor-1016-5	6.286	6.181	6.381	547.120	500.000	9.4
Aroclor-1260-1	7.404	7.299	7.499	522.730	500.000	4.5
Aroclor-1260-2	7.656	7.551	7.751	512.050	500.000	2.4
Aroclor-1260-3	8.014	7.909	8.109	558.190	500.000	11.6
Aroclor-1260-4	8.243	8.137	8.337	591.030	500.000	18.2
Aroclor-1260-5	8.569	8.464	8.664	580.640	500.000	16.1
Decachlorobiphenyl	10.443	10.334	10.534	52.690	50.000	5.4
Tetrachloro-m-xylene	4.659	4.555	4.755	54.990	50.000	10.0

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ROYF02
Lab Code: ACE **SDG NO.:** Q3178
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 09/23/2025 09/23/2025

Client Sample No.: CCAL04 **Date Analyzed:** 09/25/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP075341.D **Time Analyzed:** 20:13

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.894	4.793	4.993	544.560	500.000	8.9
Aroclor-1016-2	4.952	4.851	5.051	549.510	500.000	9.9
Aroclor-1016-3	5.071	4.971	5.171	506.710	500.000	1.3
Aroclor-1016-4	5.112	5.011	5.211	520.430	500.000	4.1
Aroclor-1016-5	5.327	5.225	5.425	532.260	500.000	6.5
Aroclor-1260-1	6.355	6.255	6.455	521.820	500.000	4.4
Aroclor-1260-2	6.542	6.441	6.641	524.370	500.000	4.9
Aroclor-1260-3	6.696	6.595	6.795	502.110	500.000	0.4
Aroclor-1260-4	7.164	7.065	7.265	496.480	500.000	-0.7
Aroclor-1260-5	7.404	7.304	7.504	526.380	500.000	5.3
Decachlorobiphenyl	8.804	8.703	8.903	62.490	50.000	25.0
Tetrachloro-m-xylene	3.796	3.696	3.896	56.810	50.000	13.6