

Surrogate Summary

SDG No.: Q3178

Client: Weston Solutions, Inc.

Analytical Method: 8082A

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Recovery(%)	Qual	Limits(%)	
								Low	High
I.BLK-PO113719.D	PIBLK-PO113719.D	Tetrachloro-m-xyl	1	20	17.0	85		60	140
		Decachlorobiphen	1	20	17.3	86		60	140
		Tetrachloro-m-xyl	2	20	17.0	85		60	140
		Decachlorobiphen	2	20	18.1	91		60	140
I.BLK-PO114033.D	PIBLK-PO114033.D	Tetrachloro-m-xyl	1	20	18.8	94		60	140
		Decachlorobiphen	1	20	16.8	84		60	140
		Tetrachloro-m-xyl	2	20	18.0	90		60	140
		Decachlorobiphen	2	20	18.0	90		60	140
Q3178-01	EME-TS14-01	Tetrachloro-m-xyl	1	20	15.3	76		30	150
		Decachlorobiphen	1	20	7.77	39		30	150
		Tetrachloro-m-xyl	2	20	15.0	75		30	150
		Decachlorobiphen	2	20	8.25	41		30	150
Q3178-02MS	EME-TS14-01MS	Tetrachloro-m-xyl	1	20	17.1	85		32	144
		Decachlorobiphen	1	20	10.6	53		32	175
		Tetrachloro-m-xyl	2	20	16.5	82		32	144
		Decachlorobiphen	2	20	11.5	58		32	175
Q3178-03MSD	EME-TS14-01MSD	Tetrachloro-m-xyl	1	20	17.2	86		32	144
		Decachlorobiphen	1	20	10.7	53		32	175
		Tetrachloro-m-xyl	2	20	16.4	82		32	144
		Decachlorobiphen	2	20	11.7	58		32	175
I.BLK-PO114047.D	PIBLK-PO114047.D	Tetrachloro-m-xyl	1	20	18.5	93		60	140
		Decachlorobiphen	1	20	18.7	94		60	140
		Tetrachloro-m-xyl	2	20	17.9	89		60	140
		Decachlorobiphen	2	20	18.6	93		60	140
I.BLK-PP075246.D	PIBLK-PP075246.D	Tetrachloro-m-xyl	1	20	19.9	100		60	140
		Decachlorobiphen	1	20	19.4	97		60	140
		Tetrachloro-m-xyl	2	20	17.9	90		60	140
		Decachlorobiphen	2	20	20.0	100		60	140
I.BLK-PP075330.D	PIBLK-PP075330.D	Tetrachloro-m-xyl	1	20	19.3	96		60	140
		Decachlorobiphen	1	20	19.3	97		60	140
		Tetrachloro-m-xyl	2	20	16.2	81		60	140
		Decachlorobiphen	2	20	20.4	102		60	140
PB169825BL	PB169825BL	Tetrachloro-m-xyl	1	20	22.5	113		32	144
		Decachlorobiphen	1	20	22.1	111		32	175
		Tetrachloro-m-xyl	2	20	19.8	99		32	144
		Decachlorobiphen	2	20	24.4	122		32	175
I.BLK-PP075345.D	PIBLK-PP075345.D	Tetrachloro-m-xyl	1	20	19.4	97		60	140
		Decachlorobiphen	1	20	19.4	97		60	140
		Tetrachloro-m-xyl	2	20	16.7	83		60	140
		Decachlorobiphen	2	20	20.4	102		60	140
I.BLK-PP075348.D	PIBLK-PP075348.D	Tetrachloro-m-xyl	1	20	22.0	110		60	140

Surrogate Summary

SDG No.: Q3178

Client: Weston Solutions, Inc.

Analytical Method: 8082A

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Recovery(%)	Qual	Limits(%)	
								Low	High
I.BLK-PP075348.D	PIBLK-PP075348.D	Decachlorobiphen	1	20	21.9	109		60	140
		Tetrachloro-m-xyl	2	20	18.8	94		60	140
		Decachlorobiphen	2	20	22.9	114		60	140
PB169825BS	PB169825BS	Tetrachloro-m-xyl	1	20	15.1	76		32	144
		Decachlorobiphen	1	20	14.4	72		32	175
		Tetrachloro-m-xyl	2	20	9.85	49		32	144
		Decachlorobiphen	2	20	14.8	74		32	175
		Tetrachloro-m-xyl	1	20	19.6	98		60	140
I.BLK-PP075356.D	PIBLK-PP075356.D	Decachlorobiphen	1	20	19.9	99		60	140
		Tetrachloro-m-xyl	2	20	16.9	85		60	140
		Decachlorobiphen	2	20	20.9	104		60	140
		Tetrachloro-m-xyl	1	20	19.9	99		60	140