

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW092525\
 Data File : VW032262.D
 Acq On : 25 Sep 2025 12:46
 Operator : SY/MD
 Sample : VSTDICC100
 Misc : 5.00g/5mL/MSVOA_W/SOIL
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VSTDICC100

Manual Integrations
 APPROVED

Reviewed By :Mahesh Dadoda 09/29/2025
 Supervised By :Semsettin Yesilyurt 09/29/2025

Quant Time: Sep 26 06:50:34 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\82W092525S.M
 Quant Title : SW846 8260
 QLast Update : Fri Sep 26 06:41:54 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.965	168	270060	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.849	114	469392	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.635	117	452792	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.556	152	225531	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.319	65	379535	89.005	ug/l	0.00
Spiked Amount 50.000	Range 63 - 155		Recovery = 178.020%#			
35) Dibromofluoromethane	7.904	113	320954	95.077	ug/l	0.00
Spiked Amount 50.000	Range 70 - 134		Recovery = 190.160%#			
50) Toluene-d8	10.325	98	1195947	98.302	ug/l	0.00
Spiked Amount 50.000	Range 74 - 123		Recovery = 196.600%#			
62) 4-Bromofluorobenzene	12.617	95	456586	99.494	ug/l	0.00
Spiked Amount 50.000	Range 17 - 146		Recovery = 198.980%#			
Target Compounds					Qvalue	
2) Dichlorodifluoromethane	2.046	85	159187	85.317	ug/l	98
3) Chloromethane	2.253	50	236913	88.725	ug/l	93
4) Vinyl Chloride	2.405	62	277540	88.959	ug/l	96
5) Bromomethane	2.820	94	194678	91.381	ug/l	89
6) Chloroethane	2.972	64	192034	90.836	ug/l	98
7) Trichlorofluoromethane	3.301	101	236771	100.722	ug/l	97
8) Diethyl Ether	3.722	74	197656	88.924	ug/l	99
9) 1,1,2-Trichlorotrifluo...	4.106	101	267494	89.149	ug/l	99
10) Methyl Iodide	4.319	142	389172	95.893	ug/l	97
11) Tert butyl alcohol	5.228	59	132448	447.465	ug/l	98
12) 1,1-Dichloroethene	4.082	96	296922	92.573	ug/l	97
13) Acrolein	3.929	56	228335	433.414	ug/l	95
14) Allyl chloride	4.704	41	475222	95.469	ug/l	99
15) Acrylonitrile	5.405	53	531479	464.265	ug/l	99
16) Acetone	4.161	43	436453	398.550	ug/l	100
17) Carbon Disulfide	4.423	76	681761	92.447	ug/l	100
18) Methyl Acetate	4.716	43	404654	107.440	ug/l	99
19) Methyl tert-butyl Ether	5.466	73	572270	96.254	ug/l	98
20) Methylene Chloride	4.954	84	356808	97.287	ug/l	97
21) trans-1,2-Dichloroethene	5.453	96	325375	94.076	ug/l	95
22) Diisopropyl ether	6.344	45	1074108	97.109	ug/l	100
23) Vinyl Acetate	6.283	43	3409679	479.176	ug/l	98
24) 1,1-Dichloroethane	6.246	63	648524	92.880	ug/l	99
25) 2-Butanone	7.197	43	701086	459.305	ug/l	100
26) 2,2-Dichloropropane	7.191	77	322157	93.629	ug/l	99
27) cis-1,2-Dichloroethene	7.191	96	410677	95.469	ug/l	99
28) Bromochloromethane	7.532	49	316521	95.000	ug/l #	98
29) Tetrahydrofuran	7.557	42	467041	471.309	ug/l	100
30) Chloroform	7.697	83	662900	90.992	ug/l	98
31) Cyclohexane	7.971	56	492754	89.496	ug/l	99
32) 1,1,1-Trichloroethane	7.886	97	460229	93.772	ug/l	100
36) 1,1-Dichloropropene	8.099	75	449330	98.728	ug/l	99
37) Ethyl Acetate	7.276	43	294907	97.703	ug/l	98
38) Carbon Tetrachloride	8.087	117	424225	99.075	ug/l	100
39) Methylcyclohexane	9.343	83	548695	104.193	ug/l	96
40) Benzene	8.337	78	1422633	99.306	ug/l	98

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW092525\
 Data File : VW032262.D
 Acq On : 25 Sep 2025 12:46
 Operator : SY/MD
 Sample : VSTDICC100
 Misc : 5.00g/5mL/MSVOA_W/SOIL
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VSTDICC100

Manual Integrations
 APPROVED

Reviewed By :Mahesh Dadoda 09/29/2025
 Supervised By :Semsettin Yesilyurt 09/29/2025

Quant Time: Sep 26 06:50:34 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\82W092525S.M
 Quant Title : SW846 8260
 QLast Update : Fri Sep 26 06:41:54 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.502	41	166119	96.094	ug/l	98
42) 1,2-Dichloroethane	8.416	62	457918	96.165	ug/l	100
43) Isopropyl Acetate	8.441	43	542177	100.345	ug/l	99
44) Trichloroethene	9.105	130	336963	100.631	ug/l	87
45) 1,2-Dichloropropane	9.373	63	360778	97.236	ug/l	100
46) Dibromomethane	9.465	93	221052	96.902	ug/l	97
47) Bromodichloromethane	9.654	83	522751	100.472	ug/l	96
48) Methyl methacrylate	9.441	41	270335	106.349	ug/l	100
49) 1,4-Dioxane	9.465	88	63276	2071.095	ug/l #	100
51) 4-Methyl-2-Pentanone	10.215	43	1586527	495.237	ug/l	99
52) Toluene	10.392	92	904359	101.852	ug/l	97
53) t-1,3-Dichloropropene	10.611	75	519907	109.064	ug/l	100
54) cis-1,3-Dichloropropene	10.081	75	576534	104.759	ug/l	98
55) 1,1,2-Trichloroethane	10.788	97	296466	95.211	ug/l	96
56) Ethyl methacrylate	10.648	69	445747	108.677	ug/l	98
57) 1,3-Dichloropropane	10.934	76	527430	98.049	ug/l	99
58) 2-Chloroethyl Vinyl ether	9.928	63	1240539	550.975	ug/l	100
59) 2-Hexanone	10.971	43	1108386	506.646	ug/l	100
60) Dibromochloromethane	11.129	129	353508	103.376	ug/l	97
61) 1,2-Dibromoethane	11.239	107	289333	98.079	ug/l	98
64) Tetrachloroethene	10.867	164	260649	94.460	ug/l #	89
65) Chlorobenzene	11.660	112	1003942	96.257	ug/l	96
66) 1,1,1,2-Tetrachloroethane	11.733	131	331436	100.481	ug/l	99
67) Ethyl Benzene	11.733	91	1744976	102.215	ug/l	95
68) m/p-Xylenes	11.836	106	1319678	201.657	ug/l	95
69) o-Xylene	12.166	106	640114	104.080	ug/l	99
70) Styrene	12.178	104	1107088	102.647	ug/l	99
71) Bromoform	12.349	173	193866	100.857	ug/l #	96
73) Isopropylbenzene	12.464	105	1594475	101.576	ug/l	99
74) N-amyl acetate	12.269	43	489840	99.091	ug/l	99
75) 1,1,2,2-Tetrachloroethane	12.714	83	380449	92.439	ug/l	99
76) 1,2,3-Trichloropropane	12.763	75	284709m	80.269	ug/l	
77) Bromobenzene	12.745	156	377479	97.769	ug/l	93
78) n-propylbenzene	12.800	91	1997481	101.306	ug/l	99
79) 2-Chlorotoluene	12.891	91	1185558	99.258	ug/l	99
80) 1,3,5-Trimethylbenzene	12.940	105	1364777	101.693	ug/l	99
81) trans-1,4-Dichloro-2-b...	12.513	75	125300	104.893	ug/l	97
82) 4-Chlorotoluene	12.989	91	1241413	98.344	ug/l	99
83) tert-Butylbenzene	13.202	119	1144817	101.999	ug/l	99
84) 1,2,4-Trimethylbenzene	13.245	105	1402022	102.892	ug/l	97
85) sec-Butylbenzene	13.379	105	1686020	100.034	ug/l	99
86) p-Isopropyltoluene	13.495	119	1405656	100.702	ug/l	98
87) 1,3-Dichlorobenzene	13.495	146	747138	97.831	ug/l	100
88) 1,4-Dichlorobenzene	13.574	146	735529	94.276	ug/l	94
89) n-Butylbenzene	13.818	91	1403783	100.210	ug/l	97
90) Hexachloroethane	14.092	117	261127	100.372	ug/l	98
91) 1,2-Dichlorobenzene	13.867	146	660780	92.492	ug/l	97
92) 1,2-Dibromo-3-Chloropr...	14.482	75	66815	94.913	ug/l	98
93) 1,2,4-Trichlorobenzene	15.129	180	433343	102.243	ug/l	99
94) Hexachlorobutadiene	15.232	225	182643	93.830	ug/l	90
95) Naphthalene	15.360	128	1068790	105.067	ug/l	99
96) 1,2,3-Trichlorobenzene	15.549	180	397191	100.785	ug/l	98

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW092525\
Data File : VW032262.D
Acq On : 25 Sep 2025 12:46
Operator : SY/MD
Sample : VSTDICC100
Misc : 5.00g/5mL/MSVOA_W/SOIL
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
VSTDICC100

Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 09/29/2025
Supervised By :Semsettin Yesilyurt 09/29/2025

Quant Time: Sep 26 06:50:34 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\82W092525S.M
Quant Title : SW846 8260
QLast Update : Fri Sep 26 06:41:54 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW092525\
Data File : VW032262.D
Acq On : 25 Sep 2025 12:46
Operator : SY/MD
Sample : VSTDICC100
Misc : 5.00g/5mL/MSVOA_W/SOIL
ALS Vial : 8 Sample Multiplier: 1


Instrument :
MSVOA_W
ClientSampleId :
VSTDICC100

Quant Time: Sep 26 06:50:34 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\82W092525S.M
Quant Title : SW846 8260
QLast Update : Fri Sep 26 06:41:54 2025
Response via : Initial Calibration

Manual Integrations APPROVED


Reviewed By :Mahesh Dadoda 09/29/2025
Supervised By :Semsettin Yesilyurt 09/29/2025





#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 7.965 min Scan# 99
Delta R.T. 0.006 min
Lab File: VW032262.D
Acq: 25 Sep 2025 12:46




Instrument :
MSVOA_W
ClientSampleId :
VSTDICC100



Tgt Ion:168 Resp: 270060
Ion Ratio Lower Upper
168 100
99 63.5 55.0 82.4




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 09/29/2025
Supervised By :Semsettin Yesilyurt 09/29/2025



#2
Dichlorodifluoromethane
Concen: 85.317 ug/l
RT: 2.046 min Scan# 26
Delta R.T. 0.006 min
Lab File: VW032262.D
Acq: 25 Sep 2025 12:46


Tgt Ion: 85 Resp: 159187
Ion Ratio Lower Upper
85 100
87 32.3 16.7 50.1





#3
Chloromethane
Concen: 88.725 ug/l
RT: 2.253 min Scan# 60
Delta R.T. -0.000 min
Lab File: VW032262.D
Acq: 25 Sep 2025 12:46




Instrument :
MSVOA_W
ClientSampleId :
VSTDIC100




Tgt Ion: 50 Resp: 23691
Ion Ratio Lower Upper
50 100
52 32.1 28.9 43.3

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 09/29/2025
Supervised By :Semsettin Yesilyurt 09/29/2025




#4
Vinyl Chloride
Concen: 88.959 ug/l
RT: 2.405 min Scan# 85
Delta R.T. -0.000 min
Lab File: VW032262.D
Acq: 25 Sep 2025 12:46




Tgt Ion: 62 Resp: 277540
Ion Ratio Lower Upper
62 100
64 30.5 26.2 39.4





#5
Bromomethane
Concen: 91.381 ug/l
RT: 2.820 min Scan# 11
Delta R.T. -0.000 min
Lab File: VW032262.D
Acq: 25 Sep 2025 12:46




Instrument :
MSVOA_W
ClientSampleId :
VSTDICC100




Tgt Ion: 94 Resp: 194678
Ion Ratio Lower Upper
94 100
96 97.3 70.0 105.0

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 09/29/2025
Supervised By :Semsettin Yesilyurt 09/29/2025




#6
Chloroethane
Concen: 90.836 ug/l
RT: 2.972 min Scan# 178
Delta R.T. 0.006 min
Lab File: VW032262.D
Acq: 25 Sep 2025 12:46




Tgt Ion: 64 Resp: 192034
Ion Ratio Lower Upper
64 100
66 32.8 25.5 38.3





#7
 Trichlorofluoromethane
 Concen: 100.722 ug/l
 RT: 3.301 min Scan# 21
 Delta R.T. -0.000 min
 Lab File: VW032262.D
 Acq: 25 Sep 2025 12:46




Instrument :
 MSVOA_W
 ClientSampleId :
 VSTDIC100




Tgt Ion:101 Resp: 23677
 Ion Ratio Lower Upper
 101 100
 103 60.4 50.3 75.5

Manual Integrations
 APPROVED



Reviewed By :Mahesh Dadoda 09/29/2025
 Supervised By :Semsettin Yesilyurt 09/29/2025




#8
 Diethyl Ether
 Concen: 88.924 ug/l
 RT: 3.722 min Scan# 301
 Delta R.T. 0.006 min
 Lab File: VW032262.D
 Acq: 25 Sep 2025 12:46




Tgt Ion: 74 Resp: 197656
 Ion Ratio Lower Upper
 74 100
 45 97.0 49.0 147.2





#9
1,1,2-Trichlorotrifluoroethane
Concen: 89.149 ug/l
RT: 4.106 min Scan# 363
Delta R.T. 0.006 min
Lab File: VW032262.D
Acq: 25 Sep 2025 12:46




Instrument :
MSVOA_W
ClientSampleId :
VSTDICC100




Tgt Ion:101 Resp: 267494
Ion Ratio Lower Upper
101 100
85 43.9 33.7 50.5
151 69.8 55.7 83.5

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 09/29/2025
Supervised By :Semsettin Yesilyurt 09/29/2025




#10
Methyl Iodide
Concen: 95.893 ug/l
RT: 4.319 min Scan# 399
Delta R.T. 0.018 min
Lab File: VW032262.D
Acq: 25 Sep 2025 12:46




Tgt Ion:142 Resp: 389172
Ion Ratio Lower Upper
142 100
127 44.2 36.9 55.3
141 14.1 12.0 18.0





#11
Tert butyl alcohol
Concen: 447.465 ug/l
RT: 5.228 min Scan# 54
Delta R.T. 0.012 min
Lab File: VW032262.D
Acq: 25 Sep 2025 12:46




Instrument :
MSVOA_W
ClientSampleId :
VSTDICC100




Tgt Ion: 59 Resp: 132448
Ion Ratio Lower Upper
59 100
57 10.5 8.9 13.3

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 09/29/2025
Supervised By :Semsettin Yesilyurt 09/29/2025




#12
1,1-Dichloroethene
Concen: 92.573 ug/l
RT: 4.082 min Scan# 360
Delta R.T. 0.006 min
Lab File: VW032262.D
Acq: 25 Sep 2025 12:46




Tgt Ion: 96 Resp: 296922
Ion Ratio Lower Upper
96 100
61 164.2 136.3 204.5
98 63.9 50.7 76.1





#13
Acrolein
Concen: 433.414 ug/l
RT: 3.929 min Scan# 335
Delta R.T. -0.000 min
Lab File: VW032262.D
Acq: 25 Sep 2025 12:46




Instrument :
MSVOA_W
ClientSampleId :
VSTDICC100




Tgt Ion: 56 Resp: 228335
Ion Ratio Lower Upper
56 100
55 70.9 53.2 79.8

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 09/29/2025
Supervised By :Semsettin Yesilyurt 09/29/2025




#14
Allyl chloride
Concen: 95.469 ug/l
RT: 4.704 min Scan# 462
Delta R.T. 0.006 min
Lab File: VW032262.D
Acq: 25 Sep 2025 12:46




Tgt Ion: 41 Resp: 475222
Ion Ratio Lower Upper
41 100
39 66.7 54.3 81.5
76 36.4 28.7 43.1





#15
Acrylonitrile
Concen: 464.265 ug/l
RT: 5.405 min Scan# 51
Delta R.T. 0.006 min
Lab File: VW032262.D
Acq: 25 Sep 2025 12:46




Instrument :
MSVOA_W
ClientSampleId :
VSTDIC100



Tgt Ion: 53 Resp: 531479
Ion Ratio Lower Upper
53 100
52 83.3 66.3 99.5
51 36.1 29.6 44.4




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 09/29/2025
Supervised By :Semsettin Yesilyurt 09/29/2025



#16
Acetone
Concen: 398.550 ug/l
RT: 4.161 min Scan# 373
Delta R.T. -0.000 min
Lab File: VW032262.D
Acq: 25 Sep 2025 12:46

Tgt Ion: 43 Resp: 436453
Ion Ratio Lower Upper
43 100
58 31.9 25.4 38.0





#17

Carbon Disulfide

Concen: 92.447 ug/l

RT: 4.423 min Scan# 415

Delta R.T. 0.006 min

Lab File: VW032262.D

Acq: 25 Sep 2025 12:46

Instrument :

MSVOA_W

ClientSampleId :

VSTDICC100

Tgt Ion: 76 Resp: 68176

Ion Ratio Lower Upper

76 100





78 8.9 7.2 10.8

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 09/29/2025

Supervised By :Semsettin Yesilyurt 09/29/2025



#18

Methyl Acetate

Concen: 107.440 ug/l

RT: 4.716 min Scan# 464

Delta R.T. 0.006 min

Lab File: VW032262.D




Acq: 25 Sep 2025 12:46


Tgt Ion: 43 Resp: 404654

Ion Ratio Lower Upper

43 100

74 24.0 18.8 28.2





#19
Methyl tert-butyl Ether
Concen: 96.254 ug/l
RT: 5.466 min Scan# 586
Delta R.T. 0.006 min
Lab File: VW032262.D
Acq: 25 Sep 2025 12:46

Instrument :

MSVOA_W

ClientSampleId :





VSTDIC100

Tgt Ion: 73 Resp: 572270
Ion Ratio Lower Upper
73 100
57 22.6 17.3 25.9

Manual Integrations




APPROVED


Reviewed By :Mahesh Dadoda 09/29/2025
Supervised By :Semsettin Yesilyurt 09/29/2025



#20
Methylene Chloride
Concen: 97.287 ug/l
RT: 4.954 min Scan# 503
Delta R.T. -0.000 min
Lab File: VW032262.D
Acq: 25 Sep 2025 12:46


Tgt Ion: 84 Resp: 356808
Ion Ratio Lower Upper
84 100
49 120.0 97.8 146.6
51 35.3 29.9 44.9
86 61.2 51.3 76.9





#21
trans-1,2-Dichloroethene
Concen: 94.076 ug/l
RT: 5.453 min Scan# 585
Delta R.T. -0.000 min
Lab File: VW032262.D
Acq: 25 Sep 2025 12:46




Instrument :
MSVOA_W
ClientSampleId :
VSTDIC100



Tgt Ion: 96 Resp: 325379
Ion Ratio Lower Upper
96 100
61 142.5 119.2 178.8
98 61.2 52.5 78.7




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 09/29/2025
Supervised By :Semsettin Yesilyurt 09/29/2025



#22
Diisopropyl ether
Concen: 97.109 ug/l
RT: 6.344 min Scan# 731
Delta R.T. 0.006 min
Lab File: VW032262.D
Acq: 25 Sep 2025 12:46


Tgt Ion: 45 Resp: 1074108
Ion Ratio Lower Upper
45 100
43 58.2 46.2 69.4
87 27.6 22.2 33.2
59 11.4 9.5 14.3





#23
 Vinyl Acetate
 Concen: 479.176 ug/l
 RT: 6.283 min Scan# 721
 Delta R.T. -0.000 min
 Lab File: VW032262.D
 Acq: 25 Sep 2025 12:46




Instrument :
 MSVOA_W
 ClientSampleId :
 VSTDIC100




Tgt Ion: 43 Resp: 3409679
 Ion Ratio Lower Upper
 43 100
 86 10.7 9.0 13.6

Manual Integrations
 APPROVED



Reviewed By :Mahesh Dadoda 09/29/2025
 Supervised By :Semsettin Yesilyurt 09/29/2025




#24
 1,1-Dichloroethane
 Concen: 92.880 ug/l
 RT: 6.246 min Scan# 715
 Delta R.T. -0.000 min
 Lab File: VW032262.D
 Acq: 25 Sep 2025 12:46




Tgt Ion: 63 Resp: 648524
 Ion Ratio Lower Upper
 63 100
 98 6.6 3.6 10.9
 100 4.6 2.4 7.1





#25
2-Butanone
Concen: 459.305 ug/l
RT: 7.197 min Scan# 81
Delta R.T. 0.006 min
Lab File: VW032262.D
Acq: 25 Sep 2025 12:46




Instrument :
MSVOA_W
ClientSampleId :
VSTDICC100



Tgt Ion: 43 Resp: 70108
Ion Ratio Lower Upper
43 100
72 28.0 22.3 33.5




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 09/29/2025
Supervised By :Semsettin Yesilyurt 09/29/2025



#26
2,2-Dichloropropane
Concen: 93.629 ug/l
RT: 7.191 min Scan# 870
Delta R.T. 0.006 min
Lab File: VW032262.D
Acq: 25 Sep 2025 12:46


Tgt Ion: 77 Resp: 322157
Ion Ratio Lower Upper
77 100
97 23.9 11.7 35.0





#27
 cis-1,2-Dichloroethene
 Concen: 95.469 ug/l
 RT: 7.191 min Scan# 870
 Delta R.T. -0.000 min
 Lab File: VW032262.D
 Acq: 25 Sep 2025 12:46




Instrument :
 MSVOA_W
 ClientSampleId :
 VSTDIC100




Tgt Ion: 96 Resp: 41067
 Ion Ratio Lower Upper
 96 100
 61 141.3 0.0 281.8
 98 64.7 0.0 130.8

Manual Integrations
 APPROVED



Reviewed By :Mahesh Dadoda 09/29/2025
 Supervised By :Semsettin Yesilyurt 09/29/2025




#28
 Bromochloromethane
 Concen: 95.000 ug/l
 RT: 7.532 min Scan# 926
 Delta R.T. 0.006 min
 Lab File: VW032262.D
 Acq: 25 Sep 2025 12:46




Tgt Ion: 49 Resp: 316521
 Ion Ratio Lower Upper
 49 100
 129 2.0 0.0 2.2
 130 72.9 0.0 0.0#





#29
Tetrahydrofuran
Concen: 471.309 ug/l
RT: 7.557 min Scan# 91
Delta R.T. 0.006 min
Lab File: VW032262.D
Acq: 25 Sep 2025 12:46




Instrument :
MSVOA_W
ClientSampleId :
VSTDICC100



Tgt Ion: 42 Resp: 46704
Ion Ratio Lower Upper
42 100
72 43.8 35.1 52.7
71 41.3 33.3 49.9




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 09/29/2025
Supervised By :Semsettin Yesilyurt 09/29/2025



#30
Chloroform
Concen: 90.992 ug/l
RT: 7.697 min Scan# 953
Delta R.T. 0.006 min
Lab File: VW032262.D
Acq: 25 Sep 2025 12:46


Tgt Ion: 83 Resp: 662900
Ion Ratio Lower Upper
83 100
85 64.8 53.3 79.9





#31
Cyclohexane
Concen: 89.496 ug/l
RT: 7.971 min Scan# 998
Delta R.T. -0.000 min
Lab File: VW032262.D
Acq: 25 Sep 2025 12:46




Instrument :
MSVOA_W
ClientSampleId :
VSTDIC100



Tgt Ion: 56 Resp: 492754
Ion Ratio Lower Upper
56 100
69 33.1 28.2 42.2
84 89.8 71.8 107.6




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 09/29/2025
Supervised By :Semsettin Yesilyurt 09/29/2025



#32
1,1,1-Trichloroethane
Concen: 93.772 ug/l
RT: 7.886 min Scan# 984
Delta R.T. -0.000 min
Lab File: VW032262.D
Acq: 25 Sep 2025 12:46


Tgt Ion: 97 Resp: 460229
Ion Ratio Lower Upper
97 100
99 63.0 50.6 76.0
61 45.0 36.2 54.4





#33
 1,2-Dichloroethane-d4
 Concen: 89.005 ug/l
 RT: 8.319 min Scan# 1055
 Delta R.T. -0.000 min
 Lab File: VW032262.D
 Acq: 25 Sep 2025 12:46




Instrument : MSVOA_W
 ClientSampleId : VSTDIC100




Tgt Ion: 65 Resp: 379539
 Ion Ratio Lower Upper
 65 100
 67 52.5 0.0 105.8

Manual Integrations
 APPROVED



Reviewed By :Mahesh Dadoda 09/29/2025
 Supervised By :Semsettin Yesilyurt 09/29/2025




#34
 1,4-Difluorobenzene
 Concen: 50.000 ug/l
 RT: 8.849 min Scan# 1142
 Delta R.T. -0.006 min
 Lab File: VW032262.D
 Acq: 25 Sep 2025 12:46




Tgt Ion:114 Resp: 469392
 Ion Ratio Lower Upper
 114 100
 63 20.8 0.0 40.2
 88 15.5 0.0 32.0





#35
Dibromofluoromethane
Concen: 95.077 ug/l
RT: 7.904 min Scan# 91
Delta R.T. -0.000 min
Lab File: VW032262.D
Acq: 25 Sep 2025 12:46




Instrument :
MSVOA_W
ClientSampleId :
VSTDIC100




Tgt Ion: 113 Resp: 320954
Ion Ratio Lower Upper
113 100
111 104.0 81.8 122.8
192 16.4 13.0 19.4

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 09/29/2025
Supervised By :Semsettin Yesilyurt 09/29/2025




#36
1,1-Dichloropropene
Concen: 98.728 ug/l
RT: 8.099 min Scan# 1019
Delta R.T. 0.006 min
Lab File: VW032262.D
Acq: 25 Sep 2025 12:46




Tgt Ion: 75 Resp: 449330
Ion Ratio Lower Upper
75 100
110 34.0 16.7 50.1
77 31.6 25.2 37.8





#37
Ethyl Acetate
Concen: 97.703 ug/l
RT: 7.276 min Scan# 884
Delta R.T. -0.000 min
Lab File: VW032262.D
Acq: 25 Sep 2025 12:46




Instrument :
MSVOA_W
ClientSampleId :
VSTDICC100



Tgt Ion: 43 Resp: 29490
Ion Ratio Lower Upper
43 100
61 14.1 10.5 15.7
70 10.3 8.5 12.7




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 09/29/2025
Supervised By :Semsettin Yesilyurt 09/29/2025



#38
Carbon Tetrachloride
Concen: 99.075 ug/l
RT: 8.087 min Scan# 1017
Delta R.T. -0.000 min
Lab File: VW032262.D
Acq: 25 Sep 2025 12:46


Tgt Ion:117 Resp: 424225
Ion Ratio Lower Upper
117 100
119 92.5 73.9 110.9
121 28.5 23.0 34.6





#39
Methylcyclohexane
Concen: 104.193 ug/l
RT: 9.343 min Scan# 11
Delta R.T. -0.000 min
Lab File: VW032262.D
Acq: 25 Sep 2025 12:46




Instrument :
MSVOA_W
ClientSampleId :
VSTDICC100



Tgt Ion: 83 Resp: 54869
Ion Ratio Lower Upper
83 100
55 76.6 65.0 97.6
98 49.6 40.8 61.2




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 09/29/2025
Supervised By :Semsettin Yesilyurt 09/29/2025



#40
Benzene
Concen: 99.306 ug/l
RT: 8.337 min Scan# 1058
Delta R.T. 0.006 min
Lab File: VW032262.D
Acq: 25 Sep 2025 12:46


Tgt Ion: 78 Resp: 1422633
Ion Ratio Lower Upper
78 100
77 24.2 18.7 28.1





#41
Methacrylonitrile
Concen: 96.094 ug/l
RT: 7.502 min Scan# 91
Delta R.T. -0.000 min
Lab File: VW032262.D
Acq: 25 Sep 2025 12:46




Instrument :
MSVOA_W
ClientSampleId :
VSTDICC100



Tgt Ion: 41 Resp: 166119
Ion Ratio Lower Upper
41 100
39 58.2 46.9 70.3
67 77.3 59.0 88.4
52 32.0 24.9 37.3




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 09/29/2025
Supervised By :Semsettin Yesilyurt 09/29/2025



#42
1,2-Dichloroethane
Concen: 96.165 ug/l
RT: 8.416 min Scan# 1071
Delta R.T. 0.006 min
Lab File: VW032262.D
Acq: 25 Sep 2025 12:46


Tgt Ion: 62 Resp: 457918
Ion Ratio Lower Upper
62 100
98 9.3 0.0 18.4





#43
Isopropyl Acetate
Concen: 100.345 ug/l
RT: 8.441 min Scan# 1075
Delta R.T. 0.006 min
Lab File: VW032262.D
Acq: 25 Sep 2025 12:46




Instrument :
MSVOA_W
ClientSampleId :
VSTDICC100



Tgt Ion: 43 Resp: 54217
Ion Ratio Lower Upper
43 100
61 29.8 24.2 36.2
87 13.7 10.6 16.0




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 09/29/2025
Supervised By :Semsettin Yesilyurt 09/29/2025



#44
Trichloroethene
Concen: 100.631 ug/l
RT: 9.105 min Scan# 1184
Delta R.T. 0.006 min
Lab File: VW032262.D
Acq: 25 Sep 2025 12:46


Tgt Ion:130 Resp: 336963
Ion Ratio Lower Upper
130 100
95 94.0 0.0 215.2





#45
1,2-Dichloropropane
Concen: 97.236 ug/l
RT: 9.373 min Scan# 1228
Delta R.T. -0.000 min
Lab File: VW032262.D
Acq: 25 Sep 2025 12:46




Instrument :
MSVOA_W
ClientSampleId :
VSTDICC100




Tgt Ion: 63 Resp: 360778
Ion Ratio Lower Upper
63 100
65 31.3 24.9 37.3

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 09/29/2025
Supervised By :Semsettin Yesilyurt 09/29/2025




#46
Dibromomethane
Concen: 96.902 ug/l
RT: 9.465 min Scan# 1243
Delta R.T. -0.000 min
Lab File: VW032262.D
Acq: 25 Sep 2025 12:46




Tgt Ion: 93 Resp: 221052
Ion Ratio Lower Upper
93 100
95 82.0 66.6 99.8
174 89.9 69.0 103.6





#47
Bromodichloromethane
Concen: 100.472 ug/l
RT: 9.654 min Scan# 1274
Delta R.T. 0.006 min
Lab File: VW032262.D
Acq: 25 Sep 2025 12:46




Instrument :
MSVOA_W
ClientSampleId :
VSTDIC100




Tgt Ion: 83 Resp: 52275
Ion Ratio Lower Upper
83 100
85 65.9 50.3 75.5
127 7.8 5.6 8.4

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 09/29/2025
Supervised By :Semsettin Yesilyurt 09/29/2025




#48
Methyl methacrylate
Concen: 106.349 ug/l
RT: 9.441 min Scan# 1239
Delta R.T. -0.000 min
Lab File: VW032262.D
Acq: 25 Sep 2025 12:46




Tgt Ion: 41 Resp: 270335
Ion Ratio Lower Upper
41 100
69 89.0 71.4 107.0
39 58.3 46.9 70.3





#49
1,4-Dioxane
Concen: 2071.095 ug/l
RT: 9.465 min Scan# 1243
Delta R.T. -0.000 min
Lab File: VW032262.D
Acq: 25 Sep 2025 12:46




Instrument :
MSVOA_W
ClientSampleId :
VSTDICC100



Tgt Ion: 88 Resp: 63270
Ion Ratio Lower Upper
88 100
43 21.2 0.0 0.0
58 78.3 62.9 94.3




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 09/29/2025
Supervised By :Semsettin Yesilyurt 09/29/2025



#50
Toluene-d8
Concen: 98.302 ug/l
RT: 10.325 min Scan# 1384
Delta R.T. -0.000 min
Lab File: VW032262.D
Acq: 25 Sep 2025 12:46

Tgt Ion: 98 Resp: 1195947
Ion Ratio Lower Upper
98 100
100 65.1 53.0 79.4









#51
4-Methyl-2-Pentanone
Concen: 495.237 ug/l
RT: 10.215 min Scan# 1366
Delta R.T. -0.000 min
Lab File: VW032262.D
Acq: 25 Sep 2025 12:46

Instrument :
MSVOA_W
ClientSampleId :
VSTDICC100

Tgt Ion: 43 Resp: 158652
Ion Ratio Lower Upper
43 100
58 40.0 32.3 48.5




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 09/29/2025
Supervised By :Semsettin Yesilyurt 09/29/2025



#52
Toluene
Concen: 101.852 ug/l
RT: 10.392 min Scan# 1395
Delta R.T. 0.006 min
Lab File: VW032262.D
Acq: 25 Sep 2025 12:46


Tgt Ion: 92 Resp: 904359
Ion Ratio Lower Upper
92 100
91 169.3 139.2 208.8





#53
t-1,3-Dichloropropene
Concen: 109.064 ug/l
RT: 10.611 min Scan# 1431
Delta R.T. -0.000 min
Lab File: VW032262.D
Acq: 25 Sep 2025 12:46




Instrument :
MSVOA_W
ClientSampleId :
VSTDIC100




Tgt Ion: 75 Resp: 51990
Ion Ratio Lower Upper
75 100
77 31.1 24.6 37.0

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 09/29/2025
Supervised By :Semsettin Yesilyurt 09/29/2025




#54
cis-1,3-Dichloropropene
Concen: 104.759 ug/l
RT: 10.081 min Scan# 1344
Delta R.T. 0.006 min
Lab File: VW032262.D
Acq: 25 Sep 2025 12:46




Tgt Ion: 75 Resp: 576534
Ion Ratio Lower Upper
75 100
77 31.8 26.2 39.4
39 46.8 39.0 58.6





#55
1,1,2-Trichloroethane
Concen: 95.211 ug/l
RT: 10.788 min Scan# 1460
Delta R.T. -0.000 min
Lab File: VW032262.D
Acq: 25 Sep 2025 12:46

Instrument :
MSVOA_W
ClientSampleId :
VSTDICC100






Tgt Ion: 97 Resp: 296460


Ion	Ratio	Lower	Upper
97	100		
83	88.1	67.4	101.2
85	56.3	42.0	63.0
99	64.1	49.0	73.6

Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 09/29/2025
Supervised By :Semsettin Yesilyurt 09/29/2025






#56
Ethyl methacrylate
Concen: 108.677 ug/l
RT: 10.648 min Scan# 1437
Delta R.T. -0.000 min
Lab File: VW032262.D
Acq: 25 Sep 2025 12:46



Tgt Ion: 69 Resp: 445747


Ion	Ratio	Lower	Upper
69	100		
41	66.4	54.7	82.1
39	32.0	26.3	39.5





#57
1,3-Dichloropropane
Concen: 98.049 ug/l
RT: 10.934 min Scan# 1484
Delta R.T. -0.000 min
Lab File: VW032262.D
Acq: 25 Sep 2025 12:46




Instrument :
MSVOA_W
ClientSampleId :
VSTDICC100




Tgt Ion: 76 Resp: 527430
Ion Ratio Lower Upper
76 100
78 32.1 25.3 37.9

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 09/29/2025
Supervised By :Semsettin Yesilyurt 09/29/2025




#58
2-Chloroethyl Vinyl ether
Concen: 550.975 ug/l
RT: 9.928 min Scan# 1319
Delta R.T. -0.000 min
Lab File: VW032262.D
Acq: 25 Sep 2025 12:46




Tgt Ion: 63 Resp: 1240539
Ion Ratio Lower Upper
63 100
106 28.1 22.5 33.7





#59
2-Hexanone
Concen: 506.646 ug/l
RT: 10.971 min Scan# 1490
Delta R.T. -0.000 min
Lab File: VW032262.D
Acq: 25 Sep 2025 12:46




Instrument :
MSVOA_W
ClientSampleId :
VSTDIC100



Tgt Ion: 43 Resp: 110838
Ion Ratio Lower Upper
43 100
58 55.4 27.9 83.6




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 09/29/2025
Supervised By :Semsettin Yesilyurt 09/29/2025



#60
Dibromochloromethane
Concen: 103.376 ug/l
RT: 11.129 min Scan# 1516
Delta R.T. -0.000 min
Lab File: VW032262.D
Acq: 25 Sep 2025 12:46


Tgt Ion:129 Resp: 353508
Ion Ratio Lower Upper
129 100
127 75.6 39.3 117.8





#61
1,2-Dibromoethane
Concen: 98.079 ug/l
RT: 11.239 min Scan# 1534
Delta R.T. -0.000 min
Lab File: VW032262.D
Acq: 25 Sep 2025 12:46




Instrument :
MSVOA_W
ClientSampleId :
VSTDICC100




Tgt Ion:107 Resp: 28933
Ion Ratio Lower Upper
107 100
109 94.8 74.1 111.1

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 09/29/2025
Supervised By :Semsettin Yesilyurt 09/29/2025




#62
4-Bromofluorobenzene
Concen: 99.494 ug/l
RT: 12.617 min Scan# 1760
Delta R.T. -0.000 min
Lab File: VW032262.D
Acq: 25 Sep 2025 12:46




Tgt Ion: 95 Resp: 456586
Ion Ratio Lower Upper
95 100
174 65.9 0.0 130.6
176 62.2 0.0 127.0





#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.635 min Scan# 1599
Delta R.T. 0.006 min
Lab File: VW032262.D
Acq: 25 Sep 2025 12:46




Instrument :
MSVOA_W
ClientSampleId :
VSTDIC100



Tgt Ion:117 Resp: 452792
Ion Ratio Lower Upper
117 100
82 54.0 46.9 70.3
119 30.5 24.9 37.3




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 09/29/2025
Supervised By :Semsettin Yesilyurt 09/29/2025



#64
Tetrachloroethene
Concen: 94.460 ug/l
RT: 10.867 min Scan# 1473
Delta R.T. -0.000 min
Lab File: VW032262.D
Acq: 25 Sep 2025 12:46


Tgt Ion:164 Resp: 260649
Ion Ratio Lower Upper
164 100
166 122.5 96.9 145.3
129 103.6 75.0 112.4
131 106.1 67.4 101.0#





#65
Chlorobenzene
Concen: 96.257 ug/l
RT: 11.660 min Scan# 1603
Delta R.T. -0.000 min
Lab File: VW032262.D
Acq: 25 Sep 2025 12:46




Instrument :
MSVOA_W
ClientSampleId :
VSTDICC100




Tgt Ion:112 Resp: 100394
Ion Ratio Lower Upper
112 100
114 33.1 24.8 37.2

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 09/29/2025
Supervised By :Semsettin Yesilyurt 09/29/2025




#66
1,1,1,2-Tetrachloroethane
Concen: 100.481 ug/l
RT: 11.733 min Scan# 1615
Delta R.T. 0.006 min
Lab File: VW032262.D
Acq: 25 Sep 2025 12:46




Tgt Ion:131 Resp: 331436
Ion Ratio Lower Upper
131 100
133 94.0 47.1 141.3
119 66.0 33.9 101.7





#67
Ethyl Benzene
Concen: 102.215 ug/l
RT: 11.733 min Scan# 1615
Delta R.T. 0.006 min
Lab File: VW032262.D
Acq: 25 Sep 2025 12:46




Instrument :
MSVOA_W
ClientSampleId :
VSTDIC100



Tgt Ion: 91 Resp: 1744970
Ion Ratio Lower Upper
91 100
106 28.3 24.7 37.1




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 09/29/2025
Supervised By :Semsettin Yesilyurt 09/29/2025



#68
m/p-Xylenes
Concen: 201.657 ug/l
RT: 11.836 min Scan# 1632
Delta R.T. -0.000 min
Lab File: VW032262.D
Acq: 25 Sep 2025 12:46

Tgt Ion:106 Resp: 1319678
Ion Ratio Lower Upper
106 100
91 204.1 169.0 253.4









#69
o-Xylene
Concen: 104.080 ug/l
RT: 12.166 min Scan# 1686
Delta R.T. -0.000 min
Lab File: VW032262.D
Acq: 25 Sep 2025 12:46

Instrument :
MSVOA_W
ClientSampleId :
VSTDIC100

Tgt Ion:106 Resp: 640114
Ion Ratio Lower Upper
106 100
91 214.3 107.8 323.6




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 09/29/2025
Supervised By :Semsettin Yesilyurt 09/29/2025



#70
Styrene
Concen: 102.647 ug/l
RT: 12.178 min Scan# 1688
Delta R.T. -0.000 min
Lab File: VW032262.D
Acq: 25 Sep 2025 12:46

Tgt Ion:104 Resp: 1107088
Ion Ratio Lower Upper
104 100
78 51.8 40.8 61.2
103 55.4 43.4 65.2





#71

Bromoform

Concen: 100.857 ug/l

RT: 12.349 min Scan# 1716

Delta R.T. -0.000 min

Lab File: VW032262.D

Acq: 25 Sep 2025 12:46

Instrument :

MSVOA_W

ClientSampleId :

VSTDICC100

Tgt Ion:173 Resp: 193860

Ion Ratio Lower Upper

173 100

175 48.7 23.1 69.2





254 0.0 0.1 0.1

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 09/29/2025

Supervised By :Semsettin Yesilyurt 09/29/2025



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 13.556 min Scan# 1914

Delta R.T. -0.000 min

Lab File: VW032262.D

Acq: 25 Sep 2025 12:46




Tgt Ion:152 Resp: 225531


Ion Ratio Lower Upper

152 100

115 61.2 32.5 97.5


150 181.9 0.0 370.2





#73
Isopropylbenzene
Concen: 101.576 ug/l
RT: 12.464 min Scan# 1735
Delta R.T. -0.000 min
Lab File: VW032262.D
Acq: 25 Sep 2025 12:46




Instrument :
MSVOA_W
ClientSampleId :
VSTDICC100




Tgt Ion:105 Resp: 1594475
Ion Ratio Lower Upper
105 100
120 25.9 13.1 39.3

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 09/29/2025
Supervised By :Semsettin Yesilyurt 09/29/2025




#74
N-ethyl acetate
Concen: 99.091 ug/l
RT: 12.269 min Scan# 1703
Delta R.T. -0.000 min
Lab File: VW032262.D
Acq: 25 Sep 2025 12:46




Tgt Ion: 43 Resp: 489840
Ion Ratio Lower Upper
43 100
70 44.8 35.0 52.4
55 25.2 20.6 31.0
61 25.2 20.0 30.0





#75
1,1,2,2-Tetrachloroethane
Concen: 92.439 ug/l
RT: 12.714 min Scan# 1776
Delta R.T. -0.000 min
Lab File: VW032262.D
Acq: 25 Sep 2025 12:46




Instrument :
MSVOA_W
ClientSampleId :
VSTDICC100




Tgt Ion: 83 Resp: 380449
Ion Ratio Lower Upper
83 100
131 10.0 4.9 14.7
85 64.4 31.9 95.7

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 09/29/2025
Supervised By :Semsettin Yesilyurt 09/29/2025




#76
1,2,3-Trichloropropane
Concen: 80.269 ug/l m
RT: 12.763 min Scan# 1784
Delta R.T. -0.000 min
Lab File: VW032262.D
Acq: 25 Sep 2025 12:46




Tgt Ion: 75 Resp: 284709
Ion Ratio Lower Upper
75 100
77 288.8 140.9 422.7





#77
Bromobenzene
Concen: 97.769 ug/l
RT: 12.745 min Scan# 1781
Delta R.T. -0.000 min
Lab File: VW032262.D
Acq: 25 Sep 2025 12:46




Instrument :
MSVOA_W
ClientSampleId :
VSTDICC100




Tgt Ion:156 Resp: 377479
Ion Ratio Lower Upper
156 100
77 217.8 103.3 309.8
158 97.5 45.4 136.2

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 09/29/2025
Supervised By :Semsettin Yesilyurt 09/29/2025




#78
n-propylbenzene
Concen: 101.306 ug/l
RT: 12.800 min Scan# 1790
Delta R.T. -0.000 min
Lab File: VW032262.D
Acq: 25 Sep 2025 12:46



Tgt Ion: 91 Resp: 1997481
Ion Ratio Lower Upper
91 100
120 22.4 10.9 32.9









#79
2-Chlorotoluene
Concen: 99.258 ug/l
RT: 12.891 min Scan# 1805
Delta R.T. -0.000 min
Lab File: VW032262.D
Acq: 25 Sep 2025 12:46

Instrument :
MSVOA_W
ClientSampleId :
VSTDICC100

Tgt Ion: 91 Resp: 1185558
Ion Ratio Lower Upper
91 100
126 31.9 16.3 48.8




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 09/29/2025
Supervised By :Semsettin Yesilyurt 09/29/2025



#80
1,3,5-Trimethylbenzene
Concen: 101.693 ug/l
RT: 12.940 min Scan# 1813
Delta R.T. -0.000 min
Lab File: VW032262.D
Acq: 25 Sep 2025 12:46


Tgt Ion:105 Resp: 1364777
Ion Ratio Lower Upper
105 100
120 47.5 24.1 72.2





#81
trans-1,4-Dichloro-2-butene
Concen: 104.893 ug/l
RT: 12.513 min Scan# 1743
Delta R.T. 0.006 min
Lab File: VW032262.D
Acq: 25 Sep 2025 12:46




Instrument :
MSVOA_W
ClientSampleId :
VSTDICC100



Tgt Ion: 75 Resp: 125300
Ion Ratio Lower Upper
75 100
53 99.5 81.4 122.2
89 47.9 35.4 53.2




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 09/29/2025
Supervised By :Semsettin Yesilyurt 09/29/2025



#82
4-Chlorotoluene
Concen: 98.344 ug/l
RT: 12.989 min Scan# 1821
Delta R.T. 0.006 min
Lab File: VW032262.D
Acq: 25 Sep 2025 12:46


Tgt Ion: 91 Resp: 1241413
Ion Ratio Lower Upper
91 100
126 32.6 16.5 49.5





#83
tert-Butylbenzene
Concen: 101.999 ug/l
RT: 13.202 min Scan# 1856
Delta R.T. -0.000 min
Lab File: VW032262.D
Acq: 25 Sep 2025 12:46




Instrument :
MSVOA_W
ClientSampleId :
VSTDICC100




Tgt Ion:119 Resp: 114481
Ion Ratio Lower Upper
119 100
91 70.7 35.5 106.5
134 25.3 13.5 40.4

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 09/29/2025
Supervised By :Semsettin Yesilyurt 09/29/2025




#84
1,2,4-Trimethylbenzene
Concen: 102.892 ug/l
RT: 13.245 min Scan# 1863
Delta R.T. -0.000 min
Lab File: VW032262.D
Acq: 25 Sep 2025 12:46




Tgt Ion:105 Resp: 1402022
Ion Ratio Lower Upper
105 100
120 43.7 22.8 68.3





#85
 sec-Butylbenzene
 Concen: 100.034 ug/l
 RT: 13.379 min Scan# 1885
 Delta R.T. -0.000 min
 Lab File: VW032262.D
 Acq: 25 Sep 2025 12:46




Instrument :
 MSVOA_W
 ClientSampleId :
 VSTDICC100



Tgt Ion:105 Resp: 1686020
 Ion Ratio Lower Upper
 105 100
 134 20.0 9.8 29.5




Manual Integrations
 APPROVED


Reviewed By :Mahesh Dadoda 09/29/2025
 Supervised By :Semsettin Yesilyurt 09/29/2025



#86
 p-Isopropyltoluene
 Concen: 100.702 ug/l
 RT: 13.495 min Scan# 1904
 Delta R.T. -0.000 min
 Lab File: VW032262.D
 Acq: 25 Sep 2025 12:46


Tgt Ion:119 Resp: 1405656
 Ion Ratio Lower Upper
 119 100
 134 26.8 12.5 37.5
 91 24.5 12.1 36.3





#87
1,3-Dichlorobenzene
Concen: 97.831 ug/l
RT: 13.495 min Scan# 1904
Delta R.T. -0.000 min
Lab File: VW032262.D
Acq: 25 Sep 2025 12:46




Instrument :
MSVOA_W
ClientSampleId :
VSTDIC100




Tgt Ion:146 Resp: 747138
Ion Ratio Lower Upper
146 100
111 43.9 21.9 65.5
148 64.2 31.9 95.9

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 09/29/2025
Supervised By :Semsettin Yesilyurt 09/29/2025




#88
1,4-Dichlorobenzene
Concen: 94.276 ug/l
RT: 13.574 min Scan# 1917
Delta R.T. -0.000 min
Lab File: VW032262.D
Acq: 25 Sep 2025 12:46




Tgt Ion:146 Resp: 735529
Ion Ratio Lower Upper
146 100
111 41.8 22.4 67.2
148 62.5 34.4 103.1





#89
n-Butylbenzene
Concen: 100.210 ug/l
RT: 13.818 min Scan# 1957
Delta R.T. -0.000 min
Lab File: VW032262.D
Acq: 25 Sep 2025 12:46




Instrument :
MSVOA_W
ClientSampleId :
VSTDICC100



Tgt Ion: 91 Resp: 140378
Ion Ratio Lower Upper
91 100
92 56.0 26.7 80.1
134 25.3 12.3 36.8




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 09/29/2025
Supervised By :Semsettin Yesilyurt 09/29/2025



#90
Hexachloroethane
Concen: 100.372 ug/l
RT: 14.092 min Scan# 2002
Delta R.T. -0.000 min
Lab File: VW032262.D
Acq: 25 Sep 2025 12:46


Tgt Ion:117 Resp: 261127
Ion Ratio Lower Upper
117 100
201 61.2 31.4 94.0





#91
1,2-Dichlorobenzene
Concen: 92.492 ug/l
RT: 13.867 min Scan# 1965
Delta R.T. -0.000 min
Lab File: VW032262.D
Acq: 25 Sep 2025 12:46




Instrument :
MSVOA_W
ClientSampleId :
VSTDICC100




Tgt Ion:146 Resp: 660780
Ion Ratio Lower Upper
146 100
111 46.7 21.9 65.8
148 64.7 31.9 95.5

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 09/29/2025
Supervised By :Semsettin Yesilyurt 09/29/2025




#92
1,2-Dibromo-3-Chloropropane
Concen: 94.913 ug/l
RT: 14.482 min Scan# 2066
Delta R.T. 0.006 min
Lab File: VW032262.D
Acq: 25 Sep 2025 12:46




Tgt Ion: 75 Resp: 66815
Ion Ratio Lower Upper
75 100
155 74.1 38.2 114.6
157 92.8 47.0 141.0





#93
1,2,4-Trichlorobenzene
Concen: 102.243 ug/l
RT: 15.129 min Scan# 2172
Delta R.T. -0.000 min
Lab File: VW032262.D
Acq: 25 Sep 2025 12:46




Instrument :
MSVOA_W
ClientSampleId :
VSTDICC100




Tgt Ion:180 Resp: 43334
Ion Ratio Lower Upper
180 100
182 95.3 48.4 145.3
145 33.6 16.9 50.7

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 09/29/2025
Supervised By :Semsettin Yesilyurt 09/29/2025




#94
Hexachlorobutadiene
Concen: 93.830 ug/l
RT: 15.232 min Scan# 2189
Delta R.T. -0.000 min
Lab File: VW032262.D
Acq: 25 Sep 2025 12:46




Tgt Ion:225 Resp: 182643
Ion Ratio Lower Upper
225 100
223 62.1 36.4 109.2
227 66.8 36.4 109.2





#95
Naphthalene
Concen: 105.067 ug/l
RT: 15.360 min Scan# 2210
Delta R.T. -0.000 min
Lab File: VW032262.D
Acq: 25 Sep 2025 12:46




Instrument :
MSVOA_W
ClientSampleId :
VSTDICC100



Tgt Ion:128 Resp: 1068790
Ion Ratio Lower Upper
128 100
127 13.0 10.6 15.8
129 10.8 9.0 13.4

Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 09/29/2025
Supervised By :Semsettin Yesilyurt 09/29/2025



#96
1,2,3-Trichlorobenzene
Concen: 100.785 ug/l
RT: 15.549 min Scan# 2241
Delta R.T. 0.006 min
Lab File: VW032262.D
Acq: 25 Sep 2025 12:46

Tgt Ion:180 Resp: 397191
Ion Ratio Lower Upper
180 100
182 93.7 45.5 136.4
145 36.2 18.1 54.1

