



Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW092425\
Data File : VW032240.D
Acq On : 24 Sep 2025 12:29
Operator : SY/MD
Sample : BFB
Misc : 5.00g/5mL/MSVOA_W/SOIL
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\82W082625S.M
Title : SW846 8260
Last Update : Wed Aug 27 04:13:40 2025



AutoFind: Scans 1759, 1760, 1761; Background Corrected with Scan 1751

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	20.5	15061	PASS
75	95	30	60	52.6	38539	PASS
95	95	100	100	100.0	73309	PASS
96	95	5	9	6.5	4763	PASS
173	174	0.00	2	0.0	0	PASS
174	95	50	100	75.9	55619	PASS
175	174	5	9	7.3	4070	PASS
176	174	95	101	97.2	54048	PASS
177	176	5	9	6.2	3357	PASS