


Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW092625\
Data File : VW032266.D
Acq On : 26 Sep 2025 08:13
Operator : SY/MD
Sample : BFB
Misc : 5.00g/5mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\82W092525S.M
Title : SW846 8260
Last Update : Fri Sep 26 07:11:42 2025



AutoFind: Scans 1759, 1760, 1761; Background Corrected with Scan 1750

Target	Rel. to	Lower	Upper	Rel.	Raw	Result
Mass	Mass	Limit%	Limit%	Abn%	Abn	Pass/Fail
50	95	15	40	19.0	36931	PASS
75	95	30	60	50.9	98688	PASS
95	95	100	100	100.0	194027	PASS
96	95	5	9	6.7	13073	PASS
173	174	0.00	2	0.0	0	PASS
174	95	50	100	64.6	125301	PASS
175	174	5	9	8.2	10245	PASS
176	174	95	101	97.8	122568	PASS
177	176	5	9	6.8	8275	PASS