

Surrogate Summary

SDG No.: **Q3178**

Client: **Weston Solutions, Inc.**

Analytical Method: **SW8260D**

Lab Sample ID	Client ID	Parameter	Spike	Result	Recovery (%)	Qual	Limits (%)	
							Low	High
Q3178-01	EME-TS14-01	1,2-Dichloroethane-d4	50	64.8	130		63	155
		Dibromofluoromethane	50	57.6	115		70	134
		Toluene-d8	50	53.8	108		74	123
		4-Bromofluorobenzene	50	48.5	97		17	146
Q3178-02MS	EME-TS14-01MS	1,2-Dichloroethane-d4	50	59.0	118		63	155
		Dibromofluoromethane	50	50.7	101		70	134
		Toluene-d8	50	39.4	79		74	123
		4-Bromofluorobenzene	50	23.0	46		17	146
Q3178-03MSD	EME-TS14-01MSD	1,2-Dichloroethane-d4	50	65.3	131		63	155
		Dibromofluoromethane	50	60.0	120		70	134
		Toluene-d8	50	51.0	102		74	123
		4-Bromofluorobenzene	50	37.1	74		17	146
VW0924SBL01	VW0924SBL01	1,2-Dichloroethane-d4	50	58.3	117		63	155
		Dibromofluoromethane	50	54.6	109		70	134
		Toluene-d8	50	52.4	105		74	123
		4-Bromofluorobenzene	50	46.4	93		17	146
VW0924SBS01	VW0924SBS01	1,2-Dichloroethane-d4	50	49.6	99		63	155
		Dibromofluoromethane	50	51.3	103		70	134
		Toluene-d8	50	49.2	98		74	123
		4-Bromofluorobenzene	50	49.3	99		17	146
VW0926SBL01	VW0926SBL01	1,2-Dichloroethane-d4	50	63.5	127		63	155
		Dibromofluoromethane	50	53.0	106		70	134
		Toluene-d8	50	53.6	107		74	123
		4-Bromofluorobenzene	50	47.5	95		17	146
VW0926SBS01	VW0926SBS01	1,2-Dichloroethane-d4	50	47.1	94		63	155
		Dibromofluoromethane	50	46.3	93		70	134
		Toluene-d8	50	46.9	94		74	123
		4-Bromofluorobenzene	50	48.4	97		17	146