

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: ROYF02
 Lab Code: ACE SDG NO.: Q3178
 Lab File ID: VW032267.D Date Analyzed: 09/26/2025
 Instrument ID: MSVOA_W Time Analyzed: 08:43
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	273367	7.97	482464	8.85	450604	11.63
UPPER LIMIT	546734	8.465	964928	9.349	901208	12.129
LOWER LIMIT	136684	7.465	241232	8.349	225302	11.129
EPA SAMPLE NO.						
EME-TS14-01MS	141001	7.96	266280	8.85	156298 *	11.63
EME-TS14-01MSD	143487	7.96	267581	8.85	203835 *	11.64
VW0926SBL01	193650	7.96	436371	8.85	462353	11.63
VW0926SBS01	258288	7.96	477257	8.85	426681	11.63

IS1 = Pentafluorobenzene

IS2 = 1,4-Difluorobenzene

IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.