

DATA PACKAGE
GC SEMI-VOLATILES

PROJECT NAME : RFP 916

WESTON SOLUTIONS, INC.
1090 King Georges Post Road
Suite 201
Edison, NJ - 08837-3703
Phone No: 732-585-4410

ORDER ID : Q3179
ATTENTION : Smita Sumbaly



Laboratory Certification ID # 20012



1) SPLP PCB Data	2
2) Signature Page	4
3) Case Narrative	5
4) Qualifier Page	7
5) Conformance/Non Conformance	8
6) QA Checklist	10
7) Chronicle	11
8) Hit Summary	12
9) QC Data Summary For SPLP PCB	13
9.1) Deuterated Monitoring Compound Summary	14
9.2) MS/MSD Summary	16
9.3) LCS/LCSD Summary	18
9.4) Method Blank Summary	19
10) Sample Data	20
10.1) PB169881TB	21
10.2) EME-TS14-01	25
11) Calibration Data Summary	29
11.1) Initial Calibration Data	30
11.1.1) PP092325	30
11.2) Continued Calibration Data	249
11.2.1) PP075413.D	249
11.2.2) PP075428.D	261
11.2.3) PP075434.D	273
11.2.4) PP075449.D	285
11.3) Analytical Seq	297
12) Compound Detection Summary	301
13) QC Sample Data	304
13.1) Method Blank Data	305
13.2) PIBLK Data	309
13.3) LCS Data	329
13.4) MS Data	338
13.5) MSD Data	347
14) Manual Integration	356
15) Analytical Runlogs	364
16) Extraction Logs	379

Table Of Contents for Q3179

16.1) PB169818.pdf	379
16.2) PB169818IC.pdf	383
16.3) PB169881.pdf	384
16.4) PB169881IC.pdf	386
17) Standard Prep Logs	387
18) Shipping Document	484
18.1) Chain Of Custody	485
18.2) Lab Certificate	486

1
2
3
4
5
6
7
8
9
10
11
12
13
14
15
16
17
18

Cover Page

Order ID : Q3179

Project ID : RFP 916

Client : Weston Solutions, Inc.

Lab Sample Number

Q3179-01
Q3179-02
Q3179-03

Client Sample Number

EME-TS14-01
EME-TS14-01MS
EME-TS14-01MSD

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. Release of the data contained in this hard copy data package has been authorized by the laboratory manager or his designee, as verified by the following signature.

Signature : _____

Date: 10/2/2025

NYDOH CERTIFICATION NO - 11376

NJDEP CERTIFICATION NO - 20012

CASE NARRATIVE

Weston Solutions, Inc.
Project Name: RFP 916
Project # N/A
Order ID # Q3179
Test Name: SPLP PCB

A. Number of Samples and Date of Receipt:

3 Solid samples were received on 09/23/2025.

B. Parameters

According to the Chain of Custody document, the following analyses were requested: EPH, SPLP BNA, SPLP Cyanide, SPLP Extraction, SPLP ICP Metals, SPLP Mercury, SPLP Metals, SPLP PCB, SPLP Pesticide, SPLP VOA and SPLP ZHE Ext. This data package contains results for SPLP PCB.

C. Analytical Techniques:

The analyses were performed on instrument GCECD_P. The front column is ZB-MR1 which is 30 meters, 0.32 mm ID, 0.5 um df, Catalogue # 7HM-G016-17. The rear column is ZB-MR2 which is 30 meters, 0.32 mm ID, 0.25 µm; Catalogue # 7HM-G017-11. The analysis of SPLP PCBs was based on method 8082A and extraction was done based on method 3541.

D. QA/ QC Samples:

The Holding Times were met for all analysis.
The Surrogate recoveries met the acceptable criteria.

The Retention Times were acceptable for all samples.
The MS recoveries met the requirements for all compounds.
The MSD recoveries met the acceptable requirements.
The RPD met criteria.
The Blank Spike met requirements for all samples.
The Blank analysis did not indicate the presence of lab contamination.
The Initial Calibration met the requirements.
The Continuous Calibration File ID PP075428.D met the requirements except for Decachlorobiphenyl is failing in 2nd column.
The Continuous Calibration File ID PP075434.D met the requirements except for Decachlorobiphenyl is failing in 2nd column.

E. Additional Comments:



284 Sheffield Street, Mountainside, NJ 07092
Phone: 908 789 8900 Fax: 908 789 8922

F. Manual Integration Comments:

Please refer to the Manual integration Report included with the Run Logs for information on the manual integrations performed.

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. The laboratory manager or his designee, as verified by the following signature has authorized release of the data contained in this hard copy data package.

Signature_____

1
2
3
4
5
6
7
8
9
10
11
12
13
14
15
16
17
18

DATA REPORTING QUALIFIERS- ORGANIC

For reporting results, the following “ Results Qualifiers” are used:

Value	If the result is a value greater than or equal to the detection limit, report the value
U	Indicates the compound was analyzed for but was not detected. Report the minimum detection limit for the sample with the U, i.e. “10 U”. This is not necessarily the instrument detection limit attainable for this particular sample based on any concentration or dilution that may have been required.
ND	Indicates the analyte was analyzed for, but not detected
J	Indicates an estimated value. This flag is used: (1) When estimating a concentration for a tentatively identified compound (library search hits, where a 1:1 response is assumed.) (2) When the mass spectral data indicated the identification, however the result was less than the specified detection limit greater than zero. If the detection limit was 10ug/L and a concentration of 3 ug/L was calculated report as 3 J. This flag is used when similar situation arise on any organic parameter i.e. Pest, PCB and others.
B	Indicates the analyte was found in the blank as well as the sample report as “12 B”.
E	Indicates the analyte ‘s concentration exceeds the calibrated range of the instrument for that specific analysis.
D	This flag identifies all compounds identified in an analysis at a secondary dilution factor.
P	This flag is used for Pesticide/PCB target analyte when there is >25% difference for detected concentrations between the two GC columns. The lower of the two values is reported on Form 1 and flagged with a “P”.
N	This flag indicates presumptive evidence of a compound. This is only used for tentatively identified compounds (TICs), where the identification is based on a mass spectral library search. It applies to all TIC results. For generic characterization of a TIC, such as chlorinated hydrocarbon, the flag is not used.
A	This flag indicates that a Tentatively Identified Compound is a suspected aldol-condensation product.
Q	Indicates the LCS did not meet the control limits requirements



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

GC ANALYSIS CONFORMANCE/NON-CONFORMANCE SUMMARY

ORDER ID: Q3179

MATRIX: Solid

METHOD: 8082A/3541

	NA	NO	YES
1. Chromatograms Labeled/Compounds Identified.			✓
2. Standard Summary Submitted.			✓
3. Calibration - Initial Calibration performed within 30 days before sample analysis and continuing calibration performed within 24 hours of sample analysis, 12 HOURS IF 8000 SERIES METHOD. The Initial Calibration met the requirements. The Continuous Calibration File ID PP075428.D met the requirements except for Decachlorobiphenyl is failing in 2nd column. The Continuous Calibration File ID PP075434.D met the requirements except for Decachlorobiphenyl is failing in 2nd column.			✓
4. Blank Contamination - If yes, list compounds and concentrations in each blank:		✓	
5. Surrogate Recoveries Meet Criteria If not met, list those compounds and their recoveries which fall outside the acceptable ranges.			✓
6. Matrix Spike/Matrix Spike Duplicate Recoveries Meet Criteria If not met, list those compounds and their recoveries which fall outside the acceptable range. The MS recoveries met the requirements for all compounds. The MSD recoveries met the acceptable requirements. The Blank Spike met requirements for all samples. The RPD met criteria.			✓
7. Retention Time Shift Meet Criteria (if applicable) Comments:			✓
8. Extraction Holding Time Met If not met, list number of days exceeded for each sample:			✓
9. Analysis Holding Time Met If not met, list those compounds and their recoveries which fall outside the acceptable range.			✓



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

GC ANALYSIS CONFORMANCE/NON-CONFORMANCE SUMMARY (CONTINUED)

NA NO YES

ADDITIONAL COMMENTS:

QA REVIEW

Date

APPENDIX A

QA REVIEW GENERAL DOCUMENTATION

Project #: Q3179

Completed

For thorough review, the report must have the following:

GENERAL:

Are all original paperwork present (chain of custody, record of communication,airbill, sample management lab chronicle, login page)

✓

Check chain-of-custody for proper relinquish/return of samples

✓

Is the chain of custody signed and complete

✓

Check internal chain-of-custody for proper relinquish/return of samples /sample extracts

✓

Collect information for each project id from server. Were all requirements followed

✓

COVER PAGE:

Do numbers of samples correspond to the number of samples in the Chain of Custody on login page

✓

Do lab numbers and client Ids on cover page agree with the Chain of Custody

✓

CHAIN OF CUSTODY:

Do requested analyses on Chain of Custody agree with form I results

✓

Do requested analyses on Chain of Custody agree with the log-in page

✓

Were the correct method log-in for analysis according to the Analytical Request and Chain of Custody

✓

Were the samples received within hold time

✓

Were any problems found with the samples at arrival recorded in the Sample Management Laboratory Chronicle

✓

ANALYTICAL:

Was method requirement followed?

✓

Was client requirement followed?

✓

Does the case narrative summarize all QC failure?

✓

All runlogs and manual integration are reviewed for requirements

✓

All manual calculations and /or hand notations verified

✓

QA Review Signature: SOHIL JODHANI

Date: 10/02/2025

LAB CHRONICLE

OrderID:	Q3179	OrderDate:	9/24/2025 11:08:00 AM
Client:	Weston Solutions, Inc.	Project:	RFP 916
Contact:	Smita Sumbaly	Location:	J42

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
Q3179-01	EME-TS14-01	Water			09/23/25			09/23/25
			EPH	NJEPH		09/29/25	09/29/25	
			SPLP PCB	8082A		09/29/25	09/29/25	
			SPLP Pesticide	8081B		09/29/25	09/29/25	

Hit Summary Sheet
SW-846

SDG No.: Q3179

Order ID: Q3179

Client: Weston Solutions, Inc.

Project ID: RFP 916

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	RDL	Units
Client ID :								

Total Concentration: 0.000



QC SUMMARY

- 1
- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13
- 14
- 15
- 16
- 17
- 18

Surrogate Summary

SDG No.: Q3179

Client: Weston Solutions, Inc.

Analytical Method: 8082A

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Recovery(%)	Qual	Limits(%)	
								Low	High
I.BLK-PP075246.D	PIBLK-PP075246.D	Tetrachloro-m-xyl	1	20	19.9	100		60	140
		Decachlorobiphen	1	20	19.4	97		60	140
		Tetrachloro-m-xyl	2	20	17.9	90		60	140
		Decachlorobiphen	2	20	20.0	100		60	140
I.BLK-PP075417.D	PIBLK-PP075417.D	Tetrachloro-m-xyl	1	20	19.3	97		60	140
		Decachlorobiphen	1	20	21.1	106		60	140
		Tetrachloro-m-xyl	2	20	17.0	85		60	140
		Decachlorobiphen	2	20	22.3	111		60	140
PB169881BL	PB169881BL	Tetrachloro-m-xyl	1	20	19.5	97		30	173
		Decachlorobiphen	1	20	21.4	107		10	173
		Tetrachloro-m-xyl	2	20	17.6	88		30	173
		Decachlorobiphen	2	20	21.8	109		10	173
PB169881TB	PB169881TB	Tetrachloro-m-xyl	1	20	20.2	101		30	173
		Decachlorobiphen	1	20	22.1	111		10	173
		Tetrachloro-m-xyl	2	20	17.5	87		30	173
		Decachlorobiphen	2	20	23.1	115		10	173
Q3179-01	EME-TS14-01	Tetrachloro-m-xyl	1	20	22.2	111		30	173
		Decachlorobiphen	1	20	23.6	118		10	173
		Tetrachloro-m-xyl	2	20	29.6	148		30	173
		Decachlorobiphen	2	20	24.7	123		10	173
Q3179-02MS	EME-TS14-01MS	Tetrachloro-m-xyl	1	20	23.4	117		30	173
		Decachlorobiphen	1	20	22.5	113		10	173
		Tetrachloro-m-xyl	2	20	14.5	73		30	173
		Decachlorobiphen	2	20	23.5	118		10	173
Q3179-03MSD	EME-TS14-01MSD	Tetrachloro-m-xyl	1	20	23.1	116		30	173
		Decachlorobiphen	1	20	22.7	114		10	173
		Tetrachloro-m-xyl	2	20	21.0	105		30	173
		Decachlorobiphen	2	20	23.8	119		10	173
I.BLK-PP075432.D	PIBLK-PP075432.D	Tetrachloro-m-xyl	1	20	19.6	98		60	140
		Decachlorobiphen	1	20	21.6	108		60	140
		Tetrachloro-m-xyl	2	20	18.3	91		60	140
		Decachlorobiphen	2	20	22.9	114		60	140
I.BLK-PP075438.D	PIBLK-PP075438.D	Tetrachloro-m-xyl	1	20	20.5	102		60	140
		Decachlorobiphen	1	20	22.4	112		60	140
		Tetrachloro-m-xyl	2	20	17.6	88		60	140
		Decachlorobiphen	2	20	24.2	121		60	140
PB169881BS	PB169881BS	Tetrachloro-m-xyl	1	20	18.5	93		30	173
		Decachlorobiphen	1	20	19.4	97		10	173
		Tetrachloro-m-xyl	2	20	13.0	65		30	173
		Decachlorobiphen	2	20	20.2	101		10	173
I.BLK-PP075453.D	PIBLK-PP075453.D	Tetrachloro-m-xyl	1	20	17.7	88		60	140

Surrogate Summary

SDG No.: Q3179

Client: Weston Solutions, Inc.

Analytical Method: 8082A

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Recovery(%)	Qual	Limits(%)	
								Low	High
I.BLK-PP075453.D	PIBLK-PP075453.D	Decachlorobiphen	1	20	19.6	98		60	140
		Tetrachloro-m-xyl	2	20	16.3	81		60	140
		Decachlorobiphen	2	20	18.8	94		60	140

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.:	<u>Q3179</u>	Analytical Method:	<u>8082A</u>
Client:	<u>Weston Solutions, Inc.</u>	DataFile :	<u>PP075422.D</u>

		Sample					Rec	RPD		Limits		
	Parameter	Spike	Result	Result	Units	Rec	Qual	RPD	Qual	Low	High	RPD
Lab Sample ID:	Q3179-02MS	Client Sample ID:		EME-TS14-01MS								
	(Column 1)											
	AR1016	5	0	6.90	ug/L	138				66	146	
	AR1260	5	0	7.10	ug/L	142				49	147	
Lab Sample ID:	Q3179-02MS	Client Sample ID:		EME-TS14-01MS								
	(Column 2)											
	AR1016	5	0	7.10	ug/L	142				66	146	
	AR1260	5	0	6.90	ug/L	138				49	147	

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.:	<u>Q3179</u>	Analytical Method:	<u>8082A</u>
Client:	<u>Weston Solutions, Inc.</u>	DataFile :	<u>PP075423.D</u>

		Sample					Rec	RPD		Limits		
	Parameter	Spike	Result	Result	Units	Rec	Qual	RPD	Qual	Low	High	RPD
Lab Sample ID:	Q3179-03MSD (Column 1)	Client Sample ID:		EME-TS14-01MSD								
	AR1016	5	0	6.70	ug/L	134		3		66	146	20
	AR1260	5	0	6.90	ug/L	138		3		49	147	20
Lab Sample ID:	Q3179-03MSD (Column 2)	Client Sample ID:		EME-TS14-01MSD								
	AR1016	5	0	6.80	ug/L	136		4		66	146	20
	AR1260	5	0	6.50	ug/L	130		6		49	147	20

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.:	<u>Q3179</u>	Analytical Method:	<u>8082A</u>
Client:	<u>Weston Solutions, Inc.</u>	Datafile :	<u>PP075439.D</u>

Lab Sample ID	Parameter	Spike	Result	Units	Rec	RPD	Qual	RPD	Limits		RPD
								Qual	Low	High	
PB169881BS (Column 1)	AR1016	5	5.20	ug/L	104				77	107	
	AR1260	5	5.00	ug/L	100				66	113	
PB169881BS (Column 2)	AR1016	5	5.10	ug/L	102				77	107	
	AR1260	5	4.90	ug/L	98				66	113	

4C

PESTICIDE METHOD BLANK SUMMARY

Client ID

PB169881BL

Lab Name: Alliance

Contract: ROYF02

Lab Code: ACE

SDG NO.: Q3179

Lab Sample ID: PB169881BL

Lab File ID: PP075418.D

Matrix: (soil/water) WATER

Extraction: (Type) SEPF

Sulfur Cleanup: (Y/N) N

Date Extracted: 09/29/2025

Date Analyzed (1): 09/29/2025

Date Analyzed (2): 09/29/2025

Time Analyzed (1): 20:15

Time Analyzed (2): 20:15

Instrument ID (1): ECD_P

Instrument ID (2): ECD_P

GC Column (1): ZB-MR1 ID: 0.32 (mm)

GC Column (2): ZB-MR2 ID: 0.32 (mm)

THIS METHOD BLANK APPLIES TO THE FOLLOWING SAMPLES, MS AND MSD:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED 1	DATE ANALYZED 2
PB169881TB	PB169881TB	PP075420.D	09/29/2025	09/29/2025
EME-TS14-01	Q3179-01	PP075421.D	09/29/2025	09/29/2025
EME-TS14-01MS	Q3179-02MS	PP075422.D	09/29/2025	09/29/2025
EME-TS14-01MSD	Q3179-03MSD	PP075423.D	09/29/2025	09/29/2025
PB169881BS	PB169881BS	PP075439.D	09/30/2025	09/30/2025

COMMENTS: _____



SAMPLE DATA

1
2
3
4
5
6
7
8
9
10
11
12
13
14
15
16
17
18

Report of Analysis

Client:	Weston Solutions, Inc.	Date Collected:	
Project:	RFP 916	Date Received:	09/29/25
Client Sample ID:	PB169881TB	SDG No.:	Q3179
Lab Sample ID:	PB169881TB	Matrix:	WATER
Analytical Method:	8082A	% Solid:	0
Sample Wt/Vol:	1000	Units:	mL
Soil Aliquot Vol:			uL
Extraction Type:		Test:	SPLP PCB
GPC Factor :	1.0	PH :	
Prep Method :	3510C	Injection Volume :	

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PP075420.D	1	09/29/25 09:15	09/29/25 20:48	PB169881

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
12674-11-2	Aroclor-1016	0.097	U	0.097	0.50	ug/L
11104-28-2	Aroclor-1221	0.13	U	0.13	0.50	ug/L
11141-16-5	Aroclor-1232	0.096	U	0.096	0.50	ug/L
53469-21-9	Aroclor-1242	0.12	U	0.12	0.50	ug/L
12672-29-6	Aroclor-1248	0.071	U	0.071	0.50	ug/L
11097-69-1	Aroclor-1254	0.094	U	0.094	0.50	ug/L
37324-23-5	Aroclor-1262	0.14	U	0.14	0.50	ug/L
11100-14-4	Aroclor-1268	0.11	U	0.11	0.50	ug/L
11096-82-5	Aroclor-1260	0.081	U	0.081	0.50	ug/L
SURROGATES						
877-09-8	Tetrachloro-m-xylene	20.2		30 - 173	101%	SPK: 20
2051-24-3	Decachlorobiphenyl	23.1		10 - 173	115%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP092925\
Data File : PP075420.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 29 Sep 2025 20:48
Operator : YP\AJ
Sample : PB169881TB
Misc :
ALS Vial : 27 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
PB169881TB

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 30 03:03:11 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP092325.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Sep 24 10:55:35 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.654	3.795	26448503	106.5E6	20.213	17.461
2) SA Decachlor...	10.428	8.796	22072467	139.1E6	22.127	23.060

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP092925\
Data File : PP075420.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 29 Sep 2025 20:48
Operator : YP\AJ
Sample : PB169881TB
Misc :
ALS Vial : 27 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
PB169881TB

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 30 03:03:11 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP092325.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Sep 24 10:55:35 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.654 min
Delta R.T.: -0.001 min
Response: 26448503
Conc: 20.21 ng/ml

Instrument :
ECD_P
ClientSampleId :
PB169881TB




#1 Tetrachloro-m-xylene

R.T.: 3.795 min
Delta R.T.: 0.000 min
Response: 106529918
Conc: 17.46 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.428 min
Delta R.T.: -0.006 min
Response: 22072467
Conc: 22.13 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.796 min
Delta R.T.: -0.008 min
Response: 139106924
Conc: 23.06 ng/ml

Report of Analysis

Client:	Weston Solutions, Inc.	Date Collected:	09/23/25
Project:	RFP 916	Date Received:	09/23/25
Client Sample ID:	EME-TS14-01	SDG No.:	Q3179
Lab Sample ID:	Q3179-01	Matrix:	WATER
Analytical Method:	8082A	% Solid:	0
Sample Wt/Vol:	1000	Units:	mL
Soil Aliquot Vol:			uL
Extraction Type:		Test:	SPLP PCB
GPC Factor :	1.0	PH :	
Prep Method :	3510C	Injection Volume :	

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PP075421.D	1	09/29/25 09:15	09/29/25 21:04	PB169881

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
12674-11-2	Aroclor-1016	0.097	U	0.097	0.50	ug/L
11104-28-2	Aroclor-1221	0.13	U	0.13	0.50	ug/L
11141-16-5	Aroclor-1232	0.096	U	0.096	0.50	ug/L
53469-21-9	Aroclor-1242	0.12	U	0.12	0.50	ug/L
12672-29-6	Aroclor-1248	0.071	U	0.071	0.50	ug/L
11097-69-1	Aroclor-1254	0.094	U	0.094	0.50	ug/L
37324-23-5	Aroclor-1262	0.14	U	0.14	0.50	ug/L
11100-14-4	Aroclor-1268	0.11	U	0.11	0.50	ug/L
11096-82-5	Aroclor-1260	0.081	U	0.081	0.50	ug/L
SURROGATES						
877-09-8	Tetrachloro-m-xylene	29.6		30 - 173	148%	SPK: 20
2051-24-3	Decachlorobiphenyl	24.7		10 - 173	123%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP092925\
Data File : PP075421.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 29 Sep 2025 21:04
Operator : YP\AJ
Sample : Q3179-01
Misc :
ALS Vial : 28 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
EME-TS14-01

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 09/30/2025
Supervised By :mohammad ahmed 10/01/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 30 03:03:24 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP092325.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Sep 24 10:55:35 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.654	3.794	29088723	180.4E6	22.231	29.572m#
2) SA Decachlor...	10.429	8.796	23527813	148.8E6	23.586	24.673

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP092925\
Data File : PP075421.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 29 Sep 2025 21:04
Operator : YP\AJ
Sample : Q3179-01
Misc :
ALS Vial : 28 Sample Multiplier: 1

Instrument :

ECD_P

ClientSampleId :

EME-TS14-01

Manual Integrations



APPROVED


Reviewed By :Yogesh Patel 09/30/2025

Supervised By :mohammad ahmed 10/01/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 30 03:03:24 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP092325.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Sep 24 10:55:35 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.654 min
Delta R.T.: 0.000 min
Response: 29088723
Conc: 22.23 ng/ml

Instrument :
ECD_P
ClientSampleId :
EME-TS14-01


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 09/30/2025
Supervised By :mohammad ahmed 10/01/2025




#1 Tetrachloro-m-xylene

R.T.: 3.794 min
Delta R.T.: -0.002 min
Response: 180415212
Conc: 29.57 ng/ml m



#2 Decachlorobiphenyl

R.T.: 10.429 min
Delta R.T.: -0.006 min
Response: 23527813
Conc: 23.59 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.796 min
Delta R.T.: -0.007 min
Response: 148838813
Conc: 24.67 ng/ml



CALIBRATION SUMMARY

1
2
3
4
5
6
7
8
9
10
11
12
13
14
15
16
17
18

RETENTION TIMES OF INITIAL CALIBRATION

Lab Name:	<u>Alliance</u>	Contract:	<u>ROYF02</u>
Lab Code:	<u>ACE</u>	SDG NO.:	<u>Q3179</u>
Instrument ID:	<u>ECD_P</u>	Calibration Date(s):	<u>09/23/2025</u> <u>09/23/2025</u>
		Calibration Times:	<u>08:45</u> <u>17:14</u>

 GC Column: ZB-MR1 ID: 0.32 (mm)

LAB FILE ID:	RT 1000 = <u>PP075247.D</u>	RT 750 = <u>PP075248.D</u>
RT 500 = <u>PP075249.D</u>	RT 250 = <u>PP075250.D</u>	RT 050 = <u>PP075251.D</u>

COMPOUND		RT 1000	RT 750	RT 500	RT 250	RT 050	MEAN RT	RT WINDOW	
								FROM	TO
Aroclor-1016-1	(1)	5.81	5.81	5.81	5.81	5.82	5.81	5.71	5.91
Aroclor-1016-2	(2)	5.83	5.83	5.83	5.83	5.84	5.83	5.73	5.93
Aroclor-1016-3	(3)	5.89	5.90	5.89	5.89	5.90	5.89	5.79	5.99
Aroclor-1016-4	(4)	5.99	5.99	5.99	5.99	6.00	5.99	5.89	6.09
Aroclor-1016-5	(5)	6.28	6.29	6.28	6.28	6.29	6.28	6.18	6.38
Aroclor-1260-1	(1)	7.40	7.40	7.40	7.40	7.41	7.40	7.30	7.50
Aroclor-1260-2	(2)	7.65	7.66	7.65	7.65	7.66	7.65	7.55	7.75
Aroclor-1260-3	(3)	8.01	8.01	8.01	8.01	8.02	8.01	7.91	8.11
Aroclor-1260-4	(4)	8.24	8.24	8.24	8.24	8.25	8.24	8.14	8.34
Aroclor-1260-5	(5)	8.56	8.57	8.56	8.56	8.57	8.57	8.47	8.67
Decachlorobiphenyl		10.43	10.44	10.43	10.43	10.45	10.44	10.34	10.54
Tetrachloro-m-xylene		4.65	4.66	4.66	4.66	4.66	4.66	4.56	4.76
Aroclor-1242-1	(1)	5.81	5.81	5.81	5.81	5.81	5.81	5.71	5.91
Aroclor-1242-2	(2)	5.83	5.83	5.83	5.83	5.83	5.83	5.73	5.93
Aroclor-1242-3	(3)	5.90	5.89	5.90	5.90	5.90	5.90	5.80	6.00
Aroclor-1242-4	(4)	5.99	5.99	5.99	5.99	5.99	5.99	5.89	6.09
Aroclor-1242-5	(5)	6.72	6.72	6.72	6.72	6.72	6.72	6.62	6.82
Decachlorobiphenyl		10.44	10.44	10.44	10.44	10.44	10.44	10.34	10.54
Tetrachloro-m-xylene		4.66	4.66	4.66	4.66	4.66	4.66	4.56	4.76
Aroclor-1248-1	(1)	5.81	5.81	5.81	5.81	5.81	5.81	5.71	5.91
Aroclor-1248-2	(2)	6.08	6.08	6.08	6.08	6.08	6.08	5.98	6.18
Aroclor-1248-3	(3)	6.28	6.28	6.29	6.28	6.28	6.28	6.18	6.38
Aroclor-1248-4	(4)	6.68	6.68	6.68	6.68	6.68	6.68	6.58	6.78
Aroclor-1248-5	(5)	6.72	6.72	6.72	6.72	6.72	6.72	6.62	6.82
Decachlorobiphenyl		10.43	10.44	10.44	10.44	10.44	10.44	10.34	10.54
Tetrachloro-m-xylene		4.66	4.66	4.66	4.66	4.66	4.66	4.56	4.76
Aroclor-1254-1	(1)	6.66	6.66	6.66	6.66	6.66	6.66	6.56	6.76
Aroclor-1254-2	(2)	6.88	6.87	6.88	6.88	6.87	6.87	6.77	6.97
Aroclor-1254-3	(3)	7.24	7.24	7.24	7.24	7.23	7.24	7.14	7.34
Aroclor-1254-4	(4)	7.52	7.52	7.52	7.52	7.52	7.52	7.42	7.62
Aroclor-1254-5	(5)	7.94	7.93	7.94	7.94	7.93	7.93	7.83	8.03
Decachlorobiphenyl		10.44	10.44	10.44	10.44	10.43	10.44	10.34	10.54
Tetrachloro-m-xylene		4.66	4.66	4.66	4.66	4.66	4.66	4.56	4.76
Aroclor-1268-1	(1)	8.89	8.89	8.89	8.89	8.89	8.89	8.79	8.99
Aroclor-1268-2	(2)	8.99	8.99	8.99	8.98	8.98	8.99	8.89	9.09
Aroclor-1268-3	(3)	9.23	9.23	9.23	9.23	9.22	9.23	9.13	9.33
Aroclor-1268-4	(4)	9.66	9.66	9.66	9.65	9.65	9.65	9.55	9.75
Aroclor-1268-5	(5)	10.09	10.09	10.09	10.09	10.08	10.09	9.99	10.19

RETENTION TIMES OF INITIAL CALIBRATION

Decachlorobiphenyl	10.44	10.44	10.44	10.44	10.44	10.44	10.34	10.54
Tetrachloro-m-xylene	4.66	4.66	4.66	4.66	4.66	4.66	4.56	4.76

RETENTION TIMES OF INITIAL CALIBRATION

Lab Name:	<u>Alliance</u>	Contract:	<u>ROYF02</u>
Lab Code:	<u>ACE</u>	SDG NO.:	<u>Q3179</u>
Instrument ID:	<u>ECD_P</u>	Calibration Date(s):	<u>09/23/2025</u> <u>09/23/2025</u>
		Calibration Times:	<u>08:45</u> <u>17:14</u>

 GC Column: ZB-MR2 ID: 0.32 (mm)

LAB FILE ID:	RT 1000 = <u>PP075247.D</u>	RT 750 = <u>PP075248.D</u>
RT 500 = <u>PP075249.D</u>	RT 250 = <u>PP075250.D</u>	RT 050 = <u>PP075251.D</u>

COMPOUND		RT 1000	RT 750	RT 500	RT 250	RT 050	MEAN RT	RT WINDOW	
								FROM	TO
Aroclor-1016-1	(1)	4.90	4.89	4.89	4.89	4.90	4.90	4.80	5.00
Aroclor-1016-2	(2)	4.95	4.95	4.95	4.95	4.95	4.95	4.85	5.05
Aroclor-1016-3	(3)	5.07	5.07	5.07	5.07	5.07	5.07	4.97	5.17
Aroclor-1016-4	(4)	5.11	5.11	5.11	5.11	5.12	5.11	5.01	5.21
Aroclor-1016-5	(5)	5.33	5.33	5.33	5.33	5.33	5.33	5.23	5.43
Aroclor-1260-1	(1)	6.36	6.36	6.36	6.35	6.36	6.36	6.26	6.46
Aroclor-1260-2	(2)	6.54	6.54	6.54	6.54	6.55	6.54	6.44	6.64
Aroclor-1260-3	(3)	6.70	6.70	6.70	6.70	6.70	6.70	6.60	6.80
Aroclor-1260-4	(4)	7.17	7.17	7.17	7.16	7.17	7.17	7.07	7.27
Aroclor-1260-5	(5)	7.41	7.40	7.40	7.40	7.41	7.41	7.31	7.51
Decachlorobiphenyl		8.81	8.80	8.80	8.80	8.81	8.81	8.71	8.91
Tetrachloro-m-xylene		3.80	3.80	3.80	3.80	3.80	3.80	3.70	3.90
Aroclor-1242-1	(1)	4.90	4.90	4.90	4.89	4.89	4.90	4.80	5.00
Aroclor-1242-2	(2)	4.95	4.96	4.95	4.95	4.95	4.95	4.85	5.05
Aroclor-1242-3	(3)	5.07	5.07	5.07	5.07	5.08	5.07	4.97	5.17
Aroclor-1242-4	(4)	5.16	5.16	5.16	5.15	5.16	5.16	5.06	5.26
Aroclor-1242-5	(5)	5.68	5.68	5.68	5.68	5.68	5.68	5.58	5.78
Decachlorobiphenyl		8.81	8.81	8.81	8.81	8.80	8.81	8.71	8.91
Tetrachloro-m-xylene		3.80	3.80	3.80	3.80	3.80	3.80	3.70	3.90
Aroclor-1248-1	(1)	4.89	4.90	4.89	4.90	4.90	4.90	4.80	5.00
Aroclor-1248-2	(2)	5.11	5.12	5.11	5.12	5.11	5.11	5.01	5.21
Aroclor-1248-3	(3)	5.15	5.16	5.15	5.16	5.15	5.16	5.06	5.26
Aroclor-1248-4	(4)	5.33	5.33	5.33	5.33	5.33	5.33	5.23	5.43
Aroclor-1248-5	(5)	5.72	5.72	5.71	5.72	5.73	5.72	5.62	5.82
Decachlorobiphenyl		8.80	8.81	8.80	8.81	8.80	8.81	8.71	8.91
Tetrachloro-m-xylene		3.80	3.80	3.80	3.80	3.80	3.80	3.70	3.90
Aroclor-1254-1	(1)	5.68	5.68	5.68	5.68	5.68	5.68	5.58	5.78
Aroclor-1254-2	(2)	5.83	5.83	5.83	5.83	5.83	5.83	5.73	5.93
Aroclor-1254-3	(3)	6.23	6.23	6.23	6.23	6.23	6.23	6.13	6.33
Aroclor-1254-4	(4)	6.45	6.45	6.46	6.45	6.45	6.45	6.35	6.55
Aroclor-1254-5	(5)	6.87	6.87	6.87	6.87	6.87	6.87	6.77	6.97
Decachlorobiphenyl		8.81	8.81	8.81	8.81	8.81	8.81	8.71	8.91
Tetrachloro-m-xylene		3.80	3.80	3.80	3.80	3.80	3.80	3.70	3.90
Aroclor-1268-1	(1)	7.69	7.69	7.69	7.69	7.69	7.69	7.59	7.79
Aroclor-1268-2	(2)	7.75	7.75	7.75	7.75	7.75	7.75	7.65	7.85
Aroclor-1268-3	(3)	7.96	7.96	7.96	7.96	7.96	7.96	7.86	8.06
Aroclor-1268-4	(4)	8.25	8.25	8.25	8.25	8.25	8.25	8.15	8.35
Aroclor-1268-5	(5)	8.55	8.55	8.55	8.54	8.54	8.54	8.44	8.64

RETENTION TIMES OF INITIAL CALIBRATION

Decachlorobiphenyl	8.81	8.81	8.81	8.80	8.81	8.81	8.71	8.91
Tetrachloro-m-xylene	3.80	3.80	3.80	3.80	3.80	3.80	3.70	3.90

1
2
3
4
5
6
7
8
9
10
11
12
13
14
15
16
17
18



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION FACTOR OF INITIAL CALIBRATION

Lab Name: Alliance
Lab Code: ACE
Instrument ID: ECD_P

Contract: ROYF02
SDG NO.: Q3179

Calibration Date(s): 09/23/2025 09/23/2025
Calibration Times: 08:45 17:14

GC Column: ZB-MR1 ID: 0.32 (mm)

LAB FILE ID:		CF 1000 = <u>PP075247.D</u>		CF 750 = <u>PP075248.D</u>			
CF 500 = <u>PP075249.D</u>		CF 250 = <u>PP075250.D</u>		CF 050 = <u>PP075251.D</u>			
COMPOUND		CF 1000	CF 750	CF 500	CF 250	CF 050	% RSD
Aroclor-1016-1	(1)	42347817	45007967	46870170	50153844	56144100	48104780
Aroclor-1016-2	(2)	64402738	67720917	69660408	74265036	82593880	71728596
Aroclor-1016-3	(3)	39431965	41872236	43807668	47351972	60190040	46530776
Aroclor-1016-4	(4)	32893250	34833364	35556514	38336304	36563880	35636662
Aroclor-1016-5	(5)	31586228	33455985	34268374	37039376	32256580	33721309
Aroclor-1260-1	(1)	52823399	56388379	60551008	64105520	66729180	60119497
Aroclor-1260-2	(2)	63303692	67485644	69648556	75704048	87848060	72798000
Aroclor-1260-3	(3)	48177125	50964003	52841238	56234364	62508960	54145138
Aroclor-1260-4	(4)	52102929	55158468	56921418	59245152	59257840	56537161
Aroclor-1260-5	(5)	102126116	107524053	110369268	115809508	113212520	109808293
Decachlorobiphenyl		920246130	962712813	986576680	1032362440	1085842800	997548173
Tetrachloro-m-xylene		1241637260	1308534467	1312330580	1382545720	1297359400	1308481485
Aroclor-1242-1	(1)	36886529	39218005	41223848	44446876	46523260	41659704
Aroclor-1242-2	(2)	55040068	57977600	60249464	64089044	60830740	59637383
Aroclor-1242-3	(3)	34279044	36622237	38652038	41896284	50681480	40426217
Aroclor-1242-4	(4)	28461551	29972721	31322910	33558448	33854640	31434054
Aroclor-1242-5	(5)	31568865	33442463	35656256	39523980	35586800	35155673
Decachlorobiphenyl		926627030	968517160	1003394580	1047152200	937997000	976737594
Tetrachloro-m-xylene		1236238500	1295575613	1320312660	1382228640	1251655600	1297202203
Aroclor-1248-1	(1)	29172565	31397577	32249970	33755340	31223340	31559758
Aroclor-1248-2	(2)	39359802	41783284	43459792	45079104	47158200	43368036
Aroclor-1248-3	(3)	44196144	46704360	48478236	50173996	49240700	47758687
Aroclor-1248-4	(4)	53562132	56623921	59266248	61341900	64765060	59111852
Aroclor-1248-5	(5)	51867384	54812452	57671502	61547912	69597980	59099446
Decachlorobiphenyl		936214370	980173387	1009425800	1018487400	982414600	985343111
Tetrachloro-m-xylene		1235222620	1285172653	1318552840	1328682760	1313272800	1296180735
Aroclor-1254-1	(1)	61475917	62734304	65986682	62998532	67924500	64223987
Aroclor-1254-2	(2)	77552440	82328021	86739650	88146820	94902000	85933786
Aroclor-1254-3	(3)	82845377	87949261	92224268	94074964	105017660	92422306
Aroclor-1254-4	(4)	64880438	68693089	72666206	72469616	78036900	71349250
Aroclor-1254-5	(5)	77310574	81941501	85874638	85948608	77961320	81807328
Decachlorobiphenyl		940196120	995283773	1040085460	1002732880	956220600	986903767
Tetrachloro-m-xylene		1243265480	1304021133	1335815700	1326930840	1339353000	1309877231
Aroclor-1268-1	(1)	152883178	159692704	169160672	171922300	156133040	161958379

CALIBRATION FACTOR OF INITIAL CALIBRATION

Aroclor-1268-2	(2)	141133811	146979519	155825750	156513748	141348060	148360178	5
Aroclor-1268-3	(3)	118887581	123478560	131058116	130049472	116969640	124088674	5
Aroclor-1268-4	(4)	55874289	58464284	61978114	60989252	60829860	59627160	4
Aroclor-1268-5	(5)	373677572	388367939	407123198	405183456	366620440	388194521	5
Decachlorobiphenyl		1657406580	1735675800	1838831580	1851125440	1610932200	1738794320	6
Tetrachloro-m-xylene		1279137450	1334222493	1390576680	1375184440	1232599000	1322344013	5



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION FACTOR OF INITIAL CALIBRATION

Lab Name: Alliance
Lab Code: ACE
Instrument ID: ECD_P

Contract: ROYF02
SDG NO.: Q3179

Calibration Date(s): 09/23/2025 09/23/2025
Calibration Times: 08:45 17:14

GC Column: ZB-MR2 ID: 0.32 (mm)

LAB FILE ID:								
CF 1000 =		PP075247.D		CF 750 =		PP075248.D		
CF 500 =		PP075249.D		CF 250 =		PP075250.D		
CF 050 =		PP075251.D						
COMPOUND		CF 1000	CF 750	CF 500	CF 250	CF 050	CF	% RSD
Aroclor-1016-1	(1)	550669557	575041595	570968608	557638892	550748680	561013466	2
Aroclor-1016-2	(2)	253692272	266987116	261503738	252594292	250233300	257002144	3
Aroclor-1016-3	(3)	139883237	146298759	147016140	140568364	202575320	155268364	17
Aroclor-1016-4	(4)	139175266	149989423	152767546	146969012	145938660	146967981	3
Aroclor-1016-5	(5)	157135257	166127832	167159970	171673236	154814600	163382179	4
Aroclor-1260-1	(1)	389924896	398807288	400126940	426602072	414215600	405935359	4
Aroclor-1260-2	(2)	565343775	583502056	582903442	580470456	566929720	575829890	2
Aroclor-1260-3	(3)	422351862	433041807	427033404	426203564	431794540	428085035	1
Aroclor-1260-4	(4)	428778820	432299260	434882920	423190072	436968900	431223994	1
Aroclor-1260-5	(5)	1056009939	1104491933	1106785704	1094138908	898515660	1051988429	8
Decachlorobiphenyl		6122260390	6260409867	6205942400	6354732320	5219165200	6032502035	8
Tetrachloro-m-xylene		6636388100	6400989987	6258078260	6278037480	4930931000	6100884965	11
Aroclor-1242-1	(1)	472236904	485846165	486239260	487321948	403955360	467119927	8
Aroclor-1242-2	(2)	217510086	216447451	216513930	219383960	174933800	208957845	9
Aroclor-1242-3	(3)	123250580	122980740	121647250	125144732	116794760	121963612	3
Aroclor-1242-4	(4)	159270704	162580923	163635768	175639772	134330060	159091445	10
Aroclor-1242-5	(5)	192075351	196719551	194226854	201700156	188695940	194683570	3
Decachlorobiphenyl		6210283890	6223405693	6336701840	6098511680	5039253200	5981631261	9
Tetrachloro-m-xylene		6371057840	6261042827	6076223420	6032393800	4834060000	5914955577	10
Aroclor-1248-1	(1)	313001525	314287516	316712298	313219832	245110400	300466314	10
Aroclor-1248-2	(2)	194621063	204942157	200519274	206042640	206270120	202479051	2
Aroclor-1248-3	(3)	238833016	244338553	263622092	258775616	243161480	249746151	4
Aroclor-1248-4	(4)	233390044	236006324	231247510	230022068	233051700	232743529	1
Aroclor-1248-5	(5)	417516536	433645501	453462226	437040196	301192600	408571412	15
Decachlorobiphenyl		6280619880	6334319187	6392002300	6030990520	5377561400	6083098657	7
Tetrachloro-m-xylene		6342135760	6351132560	6173502940	5553544360	4399164000	5763895924	14
Aroclor-1254-1	(1)	503004292	535011767	531987056	521951800	402057780	498802539	11
Aroclor-1254-2	(2)	378219544	403583375	399300676	392234176	422583680	399184290	4
Aroclor-1254-3	(3)	715208754	722755991	718569236	662347256	504737840	664723815	14
Aroclor-1254-4	(4)	505627774	538597859	544931308	492275784	386803020	493647149	13
Aroclor-1254-5	(5)	551845651	631515513	645579268	578311224	446211920	570692715	14
Decachlorobiphenyl		6237706050	6279124973	6507474640	6157241280	5410407600	6118390909	7
Tetrachloro-m-xylene		6240840860	6469002347	6413333980	5502278160	4576318000	5840354669	14
Aroclor-1268-1	(1)	1460168699	1513699808	1559348234	1460709316	1219243940	1442633999	9

CALIBRATION FACTOR OF INITIAL CALIBRATION

Aroclor-1268-2	(2)	1471677824	1500245056	1520481630	1382554016	1099105140	1394812733	12
Aroclor-1268-3	(3)	1142468076	1153335345	1195510260	1093929288	869367940	1090922182	12
Aroclor-1268-4	(4)	470517486	472924412	482908940	460922856	362513880	449957515	11
Aroclor-1268-5	(5)	3043336629	3184766664	3278330982	3285110680	2399968480	3038302687	12
Decachlorobiphenyl		12094690940	12187875973	12233092380	12124440600	9791304400	11686280859	9
Tetrachloro-m-xylene		6567329620	6628773733	6640722000	6057317560	4103278800	5999484343	18

1
2
3
4
5
6
7
8
9
10
11
12
13
14
15
16
17
18



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

INITIAL CALIBRATION OF MULTICOMPONENT ANALYTES

Lab Name: Alliance Contract: ROYF02

Lab Code: ACE SDG NO.: Q3179

Instrument ID: ECD_P Date(s) Analyzed: 09/23/2025 09/23/2025

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND	AMOUNT (ng)	PEAK	RT	RT WINDOW		CALIBRATION FACTOR
				FROM	TO	
Aroclor-1221	500	1	4.86	4.76	4.96	18490700
		2	4.95	4.85	5.05	13697900
		3	5.02	4.92	5.12	41397800
		4	0.00			0
		5	0.00			0
Aroclor-1232	500	1	5.02	4.92	5.12	32115600
		2	5.54	5.44	5.64	16774900
		3	5.83	5.73	5.93	33440800
		4	5.99	5.89	6.09	17299100
		5	6.08	5.98	6.18	12004400
Aroclor-1262	500	1	8.24	8.14	8.34	75955200
		2	8.57	8.47	8.67	135361000
		3	8.90	8.80	9.00	93451000
		4	8.98	8.88	9.08	72237200
		5	9.66	9.56	9.76	52135400



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

INITIAL CALIBRATION OF MULTICOMPONENT ANALYTES

Lab Name: Alliance Contract: ROYF02

Lab Code: ACE SDG NO.: Q3179

Instrument ID: ECD_P Date(s) Analyzed: 09/23/2025 09/23/2025

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND	AMOUNT (ng)	PEAK	RT	RT WINDOW		CALIBRATION FACTOR
				FROM	TO	
Aroclor-1221	500	1	4.01	3.91	4.11	76477600
		2	4.09	3.99	4.19	52263200
		3	4.17	4.07	4.27	190406000
		4	0.00			0
		5	0.00			0
Aroclor-1232	500	1	4.17	4.07	4.27	146566000
		2	4.89	4.79	4.99	252198000
		3	5.07	4.97	5.17	63015000
		4	5.15	5.05	5.25	103164000
		5	5.33	5.23	5.43	64046600
Aroclor-1262	500	1	6.91	6.81	7.01	791116000
		2	7.17	7.07	7.27	552368000
		3	7.69	7.59	7.79	499974000
		4	7.75	7.65	7.85	908084000
		5	8.25	8.15	8.35	429906000

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP092325\
Data File : PP075247.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Sep 2025 08:45
Operator : YP\AJ
Sample : AR1660ICC1000
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1660ICC1000

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 23 12:10:15 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP092325.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Sep 23 12:07:08 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.654	3.798	124.2E6	663.6E6	97.232	102.934
2) SA Decachlor...	10.434	8.805	92024613	612.2E6	96.521	99.321
Target Compounds						
3) L1 AR-1016-1	5.806	4.895	42347817	550.7E6	949.311	981.902
4) L1 AR-1016-2	5.828	4.954	64402738	253.7E6	960.782	984.838
5) L1 AR-1016-3	5.890	5.073	39431965	139.9E6	947.432	975.138
6) L1 AR-1016-4	5.988	5.114	32893250	139.2E6	961.092	953.442
7) L1 AR-1016-5	6.280	5.328	31586228	157.1E6	959.272	969.088
31) L7 AR-1260-1	7.397	6.356	52823399	389.9E6	931.840	987.087
32) L7 AR-1260-2	7.650	6.543	63303692	565.3E6	952.277	984.707
33) L7 AR-1260-3	8.009	6.696	48177125	422.4E6	953.829	994.488
34) L7 AR-1260-4	8.236	7.166	52102929	428.8E6	955.804	992.932
35) L7 AR-1260-5	8.562	7.406	102.1E6	1056.0E6	961.208	976.523



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP092325\
Data File : PP075247.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Sep 2025 08:45
Operator : YP\AJ
Sample : AR1660ICC1000
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1660ICC1000

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 23 12:10:15 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP092325.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Sep 23 12:07:08 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Response_ Signal: PP075247.D\ECD1A.ch

#1 Tetrachloro-m-xylene

R.T.: 4.654 min
Delta R.T.: 0.000 min
Response: 124163726
Conc: 97.23 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC1000

Time

Response_ Signal: PP075247.D\ECD2B.ch

#1 Tetrachloro-m-xylene

R.T.: 3.798 min
Delta R.T.: 0.000 min
Response: 663638810
Conc: 102.93 ng/ml

Time

Response_ Signal: PP075247.D\ECD1A.ch

#2 Decachlorobiphenyl

R.T.: 10.434 min
Delta R.T.: 0.000 min
Response: 92024613
Conc: 96.52 ng/ml

Time

Response_ Signal: PP075247.D\ECD2B.ch

#2 Decachlorobiphenyl

R.T.: 8.805 min
Delta R.T.: 0.000 min
Response: 612226039
Conc: 99.32 ng/ml

Time

Response_ Signal: PP075247.D\ECD1A.ch

#3 AR-1016-1

R.T.: 5.806 min
Delta R.T.: 0.000 min
Response: 42347817
Conc: 949.31 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC1000

Time

Response_ Signal: PP075247.D\ECD2B.ch

#3 AR-1016-1

R.T.: 4.895 min
Delta R.T.: 0.000 min
Response: 550669557
Conc: 981.90 ng/ml

Time

Response_ Signal: PP075247.D\ECD1A.ch

#4 AR-1016-2

R.T.: 5.828 min
Delta R.T.: 0.000 min
Response: 64402738
Conc: 960.78 ng/ml


Time

Response_ Signal: PP075247.D\ECD2B.ch

#4 AR-1016-2

R.T.: 4.954 min
Delta R.T.: 0.000 min
Response: 253692272
Conc: 984.84 ng/ml


Time



#5 AR-1016-3


R.T.: 5.890 min
Delta R.T.: 0.000 min
Response: 39431965
Conc: 947.43 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC1000




#5 AR-1016-3

R.T.: 5.073 min
Delta R.T.: 0.000 min
Response: 139883237
Conc: 975.14 ng/ml




#6 AR-1016-4

R.T.: 5.988 min
Delta R.T.: 0.000 min
Response: 32893250
Conc: 961.09 ng/ml



#6 AR-1016-4


R.T.: 5.114 min
Delta R.T.: 0.000 min
Response: 139175266
Conc: 953.44 ng/ml



#7 AR-1016-5


R.T.: 6.280 min
Delta R.T.: 0.000 min
Response: 31586228
Conc: 959.27 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC1000




#7 AR-1016-5

R.T.: 5.328 min
Delta R.T.: 0.000 min
Response: 157135257
Conc: 969.09 ng/ml




#31 AR-1260-1

R.T.: 7.397 min
Delta R.T.: 0.000 min
Response: 52823399
Conc: 931.84 ng/ml



#31 AR-1260-1


R.T.: 6.356 min
Delta R.T.: 0.000 min
Response: 389924896
Conc: 987.09 ng/ml



#32 AR-1260-2


R.T.: 7.650 min
Delta R.T.: 0.000 min
Response: 63303692
Conc: 952.28 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC1000




#32 AR-1260-2

R.T.: 6.543 min
Delta R.T.: 0.000 min
Response: 565343775
Conc: 984.71 ng/ml




#33 AR-1260-3

R.T.: 8.009 min
Delta R.T.: 0.000 min
Response: 48177125
Conc: 953.83 ng/ml



#33 AR-1260-3


R.T.: 6.696 min
Delta R.T.: 0.000 min
Response: 422351862
Conc: 994.49 ng/ml



#34 AR-1260-4


R.T.: 8.236 min
Delta R.T.: 0.000 min
Response: 52102929
Conc: 955.80 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC1000




#34 AR-1260-4

R.T.: 7.166 min
Delta R.T.: 0.000 min
Response: 428778820
Conc: 992.93 ng/ml



#35 AR-1260-5

R.T.: 8.562 min
Delta R.T.: 0.000 min
Response: 102126116
Conc: 961.21 ng/ml



#35 AR-1260-5

R.T.: 7.406 min
Delta R.T.: 0.000 min
Response: 1056009939
Conc: 976.52 ng/ml

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP092325\
Data File : PP075248.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Sep 2025 09:01
Operator : YP\AJ
Sample : AR1660ICC750
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1660ICC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 23 12:21:51 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP092325.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Sep 23 12:07:08 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.659	3.796	98140085	480.1E6	76.225	74.641
2) SA Decachlor...	10.438	8.803	72203461	469.5E6	75.486	75.777
Target Compounds						
3) L1 AR-1016-1	5.811	4.894	33755975	431.3E6	754.459	762.574
4) L1 AR-1016-2	5.832	4.952	50790688	200.2E6	755.124	768.006
5) L1 AR-1016-3	5.895	5.072	31404177	109.7E6	753.026	759.865
6) L1 AR-1016-4	5.992	5.111	26125023	112.5E6	758.837	763.638
7) L1 AR-1016-5	6.285	5.326	25091989	124.6E6	757.985	762.174
31) L7 AR-1260-1	7.402	6.355	42291284	299.1E6	747.360	754.771
32) L7 AR-1260-2	7.655	6.541	50614233	437.6E6	757.555	758.123
33) L7 AR-1260-3	8.013	6.695	38223002	324.8E6	754.489	759.766
34) L7 AR-1260-4	8.241	7.165	41368851	324.2E6	755.905	750.542
35) L7 AR-1260-5	8.567	7.404	80643040	828.4E6	755.983	760.602



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP092325\
Data File : PP075248.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Sep 2025 09:01
Operator : YP\AJ
Sample : AR1660ICC750
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1660ICC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 23 12:21:51 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP092325.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Sep 23 12:07:08 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm




Response_ Signal: PP075248.D\ECD1A.ch

#1 Tetrachloro-m-xylene

R.T.: 4.659 min
Delta R.T.: 0.000 min
Response: 98140085
Conc: 76.23 ng/ml


Instrument :
ECD_P
ClientSampleId :
AR1660ICC750



Time Response_ Signal: PP075248.D\ECD2B.ch

#1 Tetrachloro-m-xylene


R.T.: 3.796 min
Delta R.T.: 0.000 min
Response: 480074249
Conc: 74.64 ng/ml



Time Response_ Signal: PP075248.D\ECD1A.ch

#2 Decachlorobiphenyl


R.T.: 10.438 min
Delta R.T.: 0.000 min
Response: 72203461
Conc: 75.49 ng/ml



Time Response_ Signal: PP075248.D\ECD2B.ch

#2 Decachlorobiphenyl

R.T.: 8.803 min
Delta R.T.: 0.000 min
Response: 469530740
Conc: 75.78 ng/ml



Response_ Signal: PP075248.D\ECD1A.ch

#3 AR-1016-1

R.T.: 5.811 min
Delta R.T.: 0.000 min
Response: 33755975
Conc: 754.46 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC750

Time

Response_ Signal: PP075248.D\ECD2B.ch

#3 AR-1016-1

R.T.: 4.894 min
Delta R.T.: 0.000 min
Response: 431281196
Conc: 762.57 ng/ml

Time

Response_ Signal: PP075248.D\ECD1A.ch

#4 AR-1016-2

R.T.: 5.832 min
Delta R.T.: 0.000 min
Response: 50790688
Conc: 755.12 ng/ml

Time

Response_ Signal: PP075248.D\ECD2B.ch

#4 AR-1016-2

R.T.: 4.952 min
Delta R.T.: 0.000 min
Response: 200240337
Conc: 768.01 ng/ml


Time

Response_ Signal: PP075248.D\ECD1A.ch

#5 AR-1016-3

R.T.: 5.895 min
Delta R.T.: 0.000 min
Response: 31404177
Conc: 753.03 ng/ml


Instrument :
ECD_P
ClientSampleId :
AR1660ICC750



Response_ Signal: PP075248.D\ECD2B.ch

#5 AR-1016-3


R.T.: 5.072 min
Delta R.T.: 0.000 min
Response: 109724069
Conc: 759.87 ng/ml



Response_ Signal: PP075248.D\ECD1A.ch

#6 AR-1016-4


R.T.: 5.992 min
Delta R.T.: 0.000 min
Response: 26125023
Conc: 758.84 ng/ml




Response_ Signal: PP075248.D\ECD2B.ch

#6 AR-1016-4

R.T.: 5.111 min
Delta R.T.: 0.000 min
Response: 112492067
Conc: 763.64 ng/ml






#7 AR-1016-5


R.T.: 6.285 min
Delta R.T.: 0.000 min
Response: 25091989
Conc: 757.99 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC750




#7 AR-1016-5

R.T.: 5.326 min
Delta R.T.: 0.000 min
Response: 124595874
Conc: 762.17 ng/ml



#31 AR-1260-1


R.T.: 7.402 min
Delta R.T.: 0.000 min
Response: 42291284
Conc: 747.36 ng/ml



#31 AR-1260-1

R.T.: 6.355 min
Delta R.T.: 0.000 min
Response: 299105466
Conc: 754.77 ng/ml

Response_ Signal: PP075248.D\ECD1A.ch




#32 AR-1260-2

R.T.: 7.655 min
Delta R.T.: 0.000 min
Response: 50614233
Conc: 757.55 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC750


Time Response_ Signal: PP075248.D\ECD2B.ch



#32 AR-1260-2

R.T.: 6.541 min
Delta R.T.: 0.000 min
Response: 437626542
Conc: 758.12 ng/ml


Time Response_ Signal: PP075248.D\ECD1A.ch



#33 AR-1260-3

R.T.: 8.013 min
Delta R.T.: 0.000 min
Response: 38223002
Conc: 754.49 ng/ml

Time Response_ Signal: PP075248.D\ECD2B.ch




#33 AR-1260-3

R.T.: 6.695 min
Delta R.T.: 0.000 min
Response: 324781355
Conc: 759.77 ng/ml

Response_ Signal: PP075248.D\ECD1A.ch

#34 AR-1260-4




R.T.: 8.241 min
Delta R.T.: 0.000 min
Response: 41368851
Conc: 755.90 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC750

Time Response_ Signal: PP075248.D\ECD2B.ch


#34 AR-1260-4



R.T.: 7.165 min
Delta R.T.: 0.000 min
Response: 324224445
Conc: 750.54 ng/ml

Time Response_ Signal: PP075248.D\ECD1A.ch


#35 AR-1260-5



R.T.: 8.567 min
Delta R.T.: 0.000 min
Response: 80643040
Conc: 755.98 ng/ml

Time Response_ Signal: PP075248.D\ECD2B.ch

#35 AR-1260-5



R.T.: 7.404 min
Delta R.T.: 0.000 min
Response: 828368950
Conc: 760.60 ng/ml

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP092325\
Data File : PP075249.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Sep 2025 09:18
Operator : YP\AJ
Sample : AR1660ICC500
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1660ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 23 12:07:32 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP092325.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Sep 23 12:07:08 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.655	3.796	65616529	312.9E6	50.000	50.000
2) SA Decachlor...	10.434	8.803	49328834	310.3E6	50.000	50.000
Target Compounds						
3) L1 AR-1016-1	5.808	4.893	23435085	285.5E6	500.000	500.000
4) L1 AR-1016-2	5.829	4.951	34830204	130.8E6	500.000	500.000
5) L1 AR-1016-3	5.892	5.071	21903834	73508070	500.000	500.000
6) L1 AR-1016-4	5.989	5.111	17778257	76383773	500.000	500.000
7) L1 AR-1016-5	6.281	5.325	17134187	83579985	500.000	500.000
31) L7 AR-1260-1	7.399	6.355	30275504	200.1E6	500.000	500.000
32) L7 AR-1260-2	7.651	6.541	34824278	291.5E6	500.000	500.000
33) L7 AR-1260-3	8.009	6.695	26420619	213.5E6	500.000	500.000
34) L7 AR-1260-4	8.237	7.165	28460709	217.4E6	500.000	500.000
35) L7 AR-1260-5	8.564	7.404	55184634	553.4E6	500.000	500.000



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP092325\
Data File : PP075249.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Sep 2025 09:18
Operator : YP\AJ
Sample : AR1660ICC500
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1660ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 23 12:07:32 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP092325.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Sep 23 12:07:08 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.655 min
Delta R.T.: 0.000 min
Response: 65616529
Conc: 50.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC500




#1 Tetrachloro-m-xylene

R.T.: 3.796 min
Delta R.T.: 0.000 min
Response: 312903913
Conc: 50.00 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.434 min
Delta R.T.: 0.000 min
Response: 49328834
Conc: 50.00 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.803 min
Delta R.T.: 0.000 min
Response: 310297120
Conc: 50.00 ng/ml



#3 AR-1016-1


R.T.: 5.808 min
Delta R.T.: 0.000 min
Response: 23435085
Conc: 500.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC500




#3 AR-1016-1

R.T.: 4.893 min
Delta R.T.: 0.000 min
Response: 285484304
Conc: 500.00 ng/ml




#4 AR-1016-2

R.T.: 5.829 min
Delta R.T.: 0.000 min
Response: 34830204
Conc: 500.00 ng/ml



#4 AR-1016-2


R.T.: 4.951 min
Delta R.T.: 0.000 min
Response: 130751869
Conc: 500.00 ng/ml



#5 AR-1016-3


R.T.: 5.892 min
Delta R.T.: 0.000 min
Response: 21903834
Conc: 500.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC500




#5 AR-1016-3

R.T.: 5.071 min
Delta R.T.: 0.000 min
Response: 73508070
Conc: 500.00 ng/ml



#6 AR-1016-4

R.T.: 5.989 min
Delta R.T.: 0.000 min
Response: 17778257
Conc: 500.00 ng/ml



#6 AR-1016-4

R.T.: 5.111 min
Delta R.T.: 0.000 min
Response: 76383773
Conc: 500.00 ng/ml

Response_ Signal: PP075249.D\ECD1A.ch

#7 AR-1016-5

R.T.: 6.281 min
Delta R.T.: 0.000 min
Response: 17134187
Conc: 500.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC500

Time

Response_ Signal: PP075249.D\ECD2B.ch

#7 AR-1016-5

R.T.: 5.325 min
Delta R.T.: 0.000 min
Response: 83579985
Conc: 500.00 ng/ml

Time

Response_ Signal: PP075249.D\ECD1A.ch

#31 AR-1260-1

R.T.: 7.399 min
Delta R.T.: 0.000 min
Response: 30275504
Conc: 500.00 ng/ml


Time

Response_ Signal: PP075249.D\ECD2B.ch

#31 AR-1260-1

R.T.: 6.355 min
Delta R.T.: 0.000 min
Response: 200063470
Conc: 500.00 ng/ml


Time



#32 AR-1260-2


R.T.: 7.651 min
Delta R.T.: 0.000 min
Response: 34824278
Conc: 500.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC500




#32 AR-1260-2

R.T.: 6.541 min
Delta R.T.: 0.000 min
Response: 291451721
Conc: 500.00 ng/ml



#33 AR-1260-3


R.T.: 8.009 min
Delta R.T.: 0.000 min
Response: 26420619
Conc: 500.00 ng/ml



#33 AR-1260-3

R.T.: 6.695 min
Delta R.T.: 0.000 min
Response: 213516702
Conc: 500.00 ng/ml

Response_ Signal: PP075249.D\ECD1A.ch




#34 AR-1260-4

R.T.: 8.237 min
Delta R.T.: 0.000 min
Response: 28460709
Conc: 500.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC500


Time Response_ Signal: PP075249.D\ECD2B.ch



#34 AR-1260-4

R.T.: 7.165 min
Delta R.T.: 0.000 min
Response: 217441460
Conc: 500.00 ng/ml


Time Response_ Signal: PP075249.D\ECD1A.ch



#35 AR-1260-5

R.T.: 8.564 min
Delta R.T.: 0.000 min
Response: 55184634
Conc: 500.00 ng/ml

Time Response_ Signal: PP075249.D\ECD2B.ch



#35 AR-1260-5

R.T.: 7.404 min
Delta R.T.: 0.000 min
Response: 553392852
Conc: 500.00 ng/ml

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP092325\
Data File : PP075250.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Sep 2025 09:34
Operator : YP\AJ
Sample : AR1660ICC250
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1660ICC250

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 09/24/2025
Supervised By :mohammad ahmed 09/25/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 23 12:25:02 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP092325.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Sep 23 12:07:08 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.655	3.796	34563643	157.0E6	26.359	24.549
2) SA Decachlor...	10.434	8.803	25809061	158.9E6	26.458	25.477
Target Compounds						
3) L1 AR-1016-1	5.807	4.894	12538461	139.4E6	272.014	247.365
4) L1 AR-1016-2	5.828	4.952	18566259	63148573	269.028	244.105
5) L1 AR-1016-3	5.891	5.072	11837993	35142091	274.562	244.992
6) L1 AR-1016-4	5.988	5.113	9584076	36742253	270.699	249.565
7) L1 AR-1016-5	6.280	5.326	9259844	42918309	271.649	259.287
31) L7 AR-1260-1	7.398	6.354	16026380	106.7E6	274.109	264.074
32) L7 AR-1260-2	7.651	6.542	18926012	145.1E6	274.149	251.045
33) L7 AR-1260-3	8.009	6.695	14058591	106.6E6	270.076	249.442
34) L7 AR-1260-4	8.236	7.164	14811288	105.8E6	254.816m	246.162
35) L7 AR-1260-5	8.563	7.404	28952377	273.5E6	265.722	250.867

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP092325\
Data File : PP075250.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Sep 2025 09:34
Operator : YP\AJ
Sample : AR1660ICC250
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :

ECD_P

ClientSampleId :

AR1660ICC250

Manual Integrations



APPROVED


Reviewed By :Yogesh Patel 09/24/2025

Supervised By :mohammad ahmed 09/25/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 23 12:25:02 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP092325.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Sep 23 12:07:08 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.655 min
Delta R.T.: 0.000 min
Response: 34563643
Conc: 26.36 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC250


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 09/24/2025
Supervised By :mohammad ahmed 09/25/2025




#1 Tetrachloro-m-xylene

R.T.: 3.796 min
Delta R.T.: 0.000 min
Response: 156950937
Conc: 24.55 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.434 min
Delta R.T.: 0.000 min
Response: 25809061
Conc: 26.46 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.803 min
Delta R.T.: 0.000 min
Response: 158868308
Conc: 25.48 ng/ml




#3 AR-1016-1

R.T.: 5.807 min
Delta R.T.: 0.000 min
Response: 12538461
Conc: 272.01 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC250


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 09/24/2025
Supervised By :mohammad ahmed 09/25/2025




#3 AR-1016-1

R.T.: 4.894 min
Delta R.T.: 0.000 min
Response: 139409723
Conc: 247.36 ng/ml




#4 AR-1016-2

R.T.: 5.828 min
Delta R.T.: 0.000 min
Response: 18566259
Conc: 269.03 ng/ml



#4 AR-1016-2

R.T.: 4.952 min
Delta R.T.: 0.000 min
Response: 63148573
Conc: 244.10 ng/ml




#5 AR-1016-3

R.T.: 5.891 min
Delta R.T.: 0.000 min
Response: 11837993
Conc: 274.56 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC250


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 09/24/2025
Supervised By :mohammad ahmed 09/25/2025




#5 AR-1016-3

R.T.: 5.072 min
Delta R.T.: 0.000 min
Response: 35142091
Conc: 244.99 ng/ml




#6 AR-1016-4

R.T.: 5.988 min
Delta R.T.: 0.000 min
Response: 9584076
Conc: 270.70 ng/ml



#6 AR-1016-4

R.T.: 5.113 min
Delta R.T.: 0.000 min
Response: 36742253
Conc: 249.56 ng/ml




#7 AR-1016-5

R.T.: 6.280 min
Delta R.T.: 0.000 min
Response: 9259844
Conc: 271.65 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC250


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 09/24/2025
Supervised By :mohammad ahmed 09/25/2025




#7 AR-1016-5

R.T.: 5.326 min
Delta R.T.: 0.000 min
Response: 42918309
Conc: 259.29 ng/ml




#31 AR-1260-1

R.T.: 7.398 min
Delta R.T.: 0.000 min
Response: 16026380
Conc: 274.11 ng/ml



#31 AR-1260-1

R.T.: 6.354 min
Delta R.T.: 0.000 min
Response: 106650518
Conc: 264.07 ng/ml




#32 AR-1260-2

R.T.: 7.651 min
Delta R.T.: 0.000 min
Response: 18926012
Conc: 274.15 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC250


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 09/24/2025
Supervised By :mohammad ahmed 09/25/2025




#32 AR-1260-2

R.T.: 6.542 min
Delta R.T.: 0.000 min
Response: 145117614
Conc: 251.04 ng/ml




#33 AR-1260-3

R.T.: 8.009 min
Delta R.T.: 0.000 min
Response: 14058591
Conc: 270.08 ng/ml



#33 AR-1260-3

R.T.: 6.695 min
Delta R.T.: 0.000 min
Response: 106550891
Conc: 249.44 ng/ml




#34 AR-1260-4

R.T.: 8.236 min
Delta R.T.: -0.001 min
Response: 14811288
Conc: 254.82 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC250


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 09/24/2025
Supervised By :mohammad ahmed 09/25/2025




#34 AR-1260-4

R.T.: 7.164 min
Delta R.T.: 0.000 min
Response: 105797518
Conc: 246.16 ng/ml



#35 AR-1260-5

R.T.: 8.563 min
Delta R.T.: 0.000 min
Response: 28952377
Conc: 265.72 ng/ml



#35 AR-1260-5

R.T.: 7.404 min
Delta R.T.: 0.000 min
Response: 273534727
Conc: 250.87 ng/ml

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP092325\
Data File : PP075251.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Sep 2025 10:11
Operator : YP\AJ
Sample : AR1660ICC050
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1660ICC050

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 09/24/2025
Supervised By :mohammad ahmed 09/25/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 23 12:30:34 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP092325.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Sep 23 12:07:08 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.662	3.796	6486797	24654655	4.958	4.041
2) SA Decachlor...	10.447	8.810	5429214	26095826	5.432m	4.326
Target Compounds						
3) L1 AR-1016-1	5.815	4.897	2807205	27537434	58.356	49.085
4) L1 AR-1016-2	5.837	4.954	4129694	12511665	57.574	48.683
5) L1 AR-1016-3	5.898	5.071	3009502	10128766	65.402m	65.234
6) L1 AR-1016-4	5.997	5.115	1828194	7296933	51.301	49.650
7) L1 AR-1016-5	6.288	5.331	1612829	7740730	48.062m	47.378
31) L7 AR-1260-1	7.406	6.357	3336459	20710780	56.486m	51.020
32) L7 AR-1260-2	7.660	6.545	4392403	28346486	62.449	49.227
33) L7 AR-1260-3	8.018	6.695	3125448	21589727	57.724	50.432m
34) L7 AR-1260-4	8.246	7.168	2962892	21848445	51.832m	52.185
35) L7 AR-1260-5	8.573	7.405	5660626	44925783	51.550	43.361m

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP092325\
Data File : PP075251.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Sep 2025 10:11
Operator : YP\AJ
Sample : AR1660ICC050
Misc :
ALS Vial : 7 Sample Multiplier: 1


Instrument :
ECD_P
ClientSampleId :
AR1660ICC050


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 09/24/2025
Supervised By :mohammad ahmed 09/25/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 23 12:30:34 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP092325.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Sep 23 12:07:08 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.662 min
Delta R.T.: 0.000 min
Response: 6486797
Conc: 4.96 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC050


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 09/24/2025
Supervised By :mohammad ahmed 09/25/2025




#1 Tetrachloro-m-xylene

R.T.: 3.796 min
Delta R.T.: 0.000 min
Response: 24654655
Conc: 4.04 ng/ml



#2 Decachlorobiphenyl


R.T.: 10.447 min
Delta R.T.: 0.000 min
Response: 5429214
Conc: 5.43 ng/ml m



#2 Decachlorobiphenyl

R.T.: 8.810 min
Delta R.T.: 0.000 min
Response: 26095826
Conc: 4.33 ng/ml

Response_ Signal: PP075251.D\ECD1A.ch



#3 AR-1016-1


R.T.: 5.815 min
Delta R.T.: 0.000 min
Response: 2807205
Conc: 58.36 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC050

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 09/24/2025
Supervised By :mohammad ahmed 09/25/2025


Time Response_ Signal: PP075251.D\ECD2B.ch



#3 AR-1016-1

R.T.: 4.897 min
Delta R.T.: 0.000 min
Response: 27537434
Conc: 49.09 ng/ml


Time Response_ Signal: PP075251.D\ECD1A.ch



#4 AR-1016-2

R.T.: 5.837 min
Delta R.T.: 0.000 min
Response: 4129694
Conc: 57.57 ng/ml


Time Response_ Signal: PP075251.D\ECD2B.ch



#4 AR-1016-2

R.T.: 4.954 min
Delta R.T.: 0.000 min
Response: 12511665
Conc: 48.68 ng/ml

Response_ Signal: PP075251.D\ECD1A.ch



#5 AR-1016-3


R.T.: 5.898 min
Delta R.T.: 0.000 min
Response: 3009502
Conc: 65.40 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC050

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 09/24/2025
Supervised By :mohammad ahmed 09/25/2025


Time Response_ Signal: PP075251.D\ECD2B.ch



#5 AR-1016-3

R.T.: 5.071 min
Delta R.T.: 0.000 min
Response: 10128766
Conc: 65.23 ng/ml


Response_ Signal: PP075251.D\ECD1A.ch



#6 AR-1016-4


R.T.: 5.997 min
Delta R.T.: 0.000 min
Response: 1828194
Conc: 51.30 ng/ml

Time Response_ Signal: PP075251.D\ECD2B.ch



#6 AR-1016-4

R.T.: 5.115 min
Delta R.T.: 0.000 min
Response: 7296933
Conc: 49.65 ng/ml




#7 AR-1016-5

R.T.: 6.288 min
Delta R.T.: 0.000 min
Response: 1612829
Conc: 48.06 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC050


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 09/24/2025
Supervised By :mohammad ahmed 09/25/2025




#7 AR-1016-5

R.T.: 5.331 min
Delta R.T.: 0.000 min
Response: 7740730
Conc: 47.38 ng/ml




#31 AR-1260-1

R.T.: 7.406 min
Delta R.T.: 0.000 min
Response: 3336459
Conc: 56.49 ng/ml m



#31 AR-1260-1

R.T.: 6.357 min
Delta R.T.: 0.000 min
Response: 20710780
Conc: 51.02 ng/ml




#32 AR-1260-2

R.T.: 7.660 min
Delta R.T.: 0.001 min
Response: 4392403
Conc: 62.45 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC050


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 09/24/2025
Supervised By :mohammad ahmed 09/25/2025




#32 AR-1260-2

R.T.: 6.545 min
Delta R.T.: 0.000 min
Response: 28346486
Conc: 49.23 ng/ml




#33 AR-1260-3

R.T.: 8.018 min
Delta R.T.: 0.000 min
Response: 3125448
Conc: 57.72 ng/ml



#33 AR-1260-3

R.T.: 6.695 min
Delta R.T.: 0.000 min
Response: 21589727
Conc: 50.43 ng/ml m




#34 AR-1260-4

R.T.: 8.246 min
Delta R.T.: 0.000 min
Response: 2962892
Conc: 51.83 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC050


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 09/24/2025
Supervised By :mohammad ahmed 09/25/2025




#34 AR-1260-4

R.T.: 7.168 min
Delta R.T.: 0.001 min
Response: 21848445
Conc: 52.18 ng/ml



#35 AR-1260-5

R.T.: 8.573 min
Delta R.T.: 0.000 min
Response: 5660626
Conc: 51.55 ng/ml



#35 AR-1260-5

R.T.: 7.405 min
Delta R.T.: 0.000 min
Response: 44925783
Conc: 43.36 ng/ml m

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP092325\
Data File : PP075252.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Sep 2025 10:27
Operator : YP\AJ
Sample : AR1221ICC500
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1221ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 23 12:52:20 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP092325.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Sep 23 12:52:01 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.660	3.797	66333410	311.2E6	50.000	50.000
2) SA Decachlor...	10.443	8.805	49168629	311.8E6	50.000	50.000
Target Compounds						
8) L2 AR-1221-1	4.860	4.008	9245327	38238805	500.000	500.000
9) L2 AR-1221-2	4.946	4.093	6848929	26131566	500.000	500.000
10) L2 AR-1221-3	5.021	4.168	20698858	95202751	500.000	500.000



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP092325\
Data File : PP075252.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Sep 2025 10:27
Operator : YP\AJ
Sample : AR1221ICC500
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1221ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 23 12:52:20 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP092325.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Sep 23 12:52:01 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.660 min
Delta R.T.: 0.000 min
Response: 66333410
Conc: 50.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1221ICC500




#1 Tetrachloro-m-xylene

R.T.: 3.797 min
Delta R.T.: 0.000 min
Response: 311163134
Conc: 50.00 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.443 min
Delta R.T.: 0.000 min
Response: 49168629
Conc: 50.00 ng/ml




#2 Decachlorobiphenyl

R.T.: 8.805 min
Delta R.T.: 0.000 min
Response: 311812597
Conc: 50.00 ng/ml

Signal: PP075252.D\ECD1A.ch

#8 AR-1221-1




R.T.: 4.860 min
Delta R.T.: 0.000 min
Response: 9245327
Conc: 500.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1221ICC500

Signal: PP075252.D\ECD2B.ch


#8 AR-1221-1



R.T.: 4.008 min
Delta R.T.: 0.000 min
Response: 38238805
Conc: 500.00 ng/ml

Signal: PP075252.D\ECD1A.ch


#9 AR-1221-2




R.T.: 4.946 min
Delta R.T.: 0.000 min
Response: 6848929
Conc: 500.00 ng/ml

Signal: PP075252.D\ECD2B.ch

#9 AR-1221-2




R.T.: 4.093 min
Delta R.T.: 0.000 min
Response: 26131566
Conc: 500.00 ng/ml



#10 AR-1221-3

R.T.: 5.021 min
Delta R.T.: 0.000 min
Response: 20698858
Conc: 500.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1221ICC500



#10 AR-1221-3

R.T.: 4.168 min
Delta R.T.: 0.000 min
Response: 95202751
Conc: 500.00 ng/ml

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP092325\
Data File : PP075253.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Sep 2025 10:43
Operator : YP\AJ
Sample : AR1232ICC500
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1232ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 23 13:16:56 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP092325.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Sep 23 13:16:43 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.657	3.797	66756778	323.1E6	50.000	50.000
2) SA Decachlor...	10.440	8.805	50078510	307.1E6	50.000	50.000
Target Compounds						
11) L3 AR-1232-1	5.018	4.168	16057787	73283142	500.000	500.000
12) L3 AR-1232-2	5.544	4.894	8387436	126.1E6	500.000	500.000
13) L3 AR-1232-3	5.831	5.071	16720373	31507455	500.000	500.000
14) L3 AR-1232-4	5.990	5.154	8649563	51581797	500.000	500.000
15) L3 AR-1232-5	6.080	5.326	6002207	32023272	500.000	500.000


(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP092325\
Data File : PP075253.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Sep 2025 10:43
Operator : YP\AJ
Sample : AR1232ICC500
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1232ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 23 13:16:56 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP092325.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Sep 23 13:16:43 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.657 min
Delta R.T.: 0.000 min
Response: 66756778
Conc: 50.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1232ICC500




#1 Tetrachloro-m-xylene

R.T.: 3.797 min
Delta R.T.: 0.000 min
Response: 323108376
Conc: 50.00 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.440 min
Delta R.T.: 0.000 min
Response: 50078510
Conc: 50.00 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.805 min
Delta R.T.: 0.000 min
Response: 307122792
Conc: 50.00 ng/ml



#11 AR-1232-1


R.T.: 5.018 min
Delta R.T.: 0.000 min
Response: 16057787
Conc: 500.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1232ICC500




#11 AR-1232-1

R.T.: 4.168 min
Delta R.T.: 0.000 min
Response: 73283142
Conc: 500.00 ng/ml



#12 AR-1232-2


R.T.: 5.544 min
Delta R.T.: 0.000 min
Response: 8387436
Conc: 500.00 ng/ml



#12 AR-1232-2

R.T.: 4.894 min
Delta R.T.: 0.000 min
Response: 126098814
Conc: 500.00 ng/ml

Response_ Signal: PP075253.D\ECD1A.ch




#13 AR-1232-3

R.T.: 5.831 min
Delta R.T.: 0.000 min
Response: 16720373
Conc: 500.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1232ICC500


Time Response_ Signal: PP075253.D\ECD2B.ch



#13 AR-1232-3

R.T.: 5.071 min
Delta R.T.: 0.000 min
Response: 31507455
Conc: 500.00 ng/ml


Time Response_ Signal: PP075253.D\ECD1A.ch



#14 AR-1232-4


R.T.: 5.990 min
Delta R.T.: 0.000 min
Response: 8649563
Conc: 500.00 ng/ml

Time Response_ Signal: PP075253.D\ECD2B.ch



#14 AR-1232-4


R.T.: 5.154 min
Delta R.T.: 0.000 min
Response: 51581797
Conc: 500.00 ng/ml



#15 AR-1232-5

R.T.: 6.080 min
Delta R.T.: 0.000 min
Response: 6002207
Conc: 500.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1232ICC500



#15 AR-1232-5

R.T.: 5.326 min
Delta R.T.: 0.000 min
Response: 32023272
Conc: 500.00 ng/ml

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP092325\
 Data File : PP075254.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 23 Sep 2025 11:00
 Operator : YP\AJ
 Sample : AR1242ICC1000
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1242ICC1000

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 09/24/2025
 Supervised By :mohammad ahmed 09/25/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 23 13:48:35 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP092325.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Sep 23 13:43:35 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.660	3.799	123.6E6	637.1E6	96.711	102.369
2) SA Decachlor...	10.443	8.807	92662703	621.0E6	96.022	98.992
Target Compounds						
16) L4 AR-1242-1	5.812	4.896	36886529	472.2E6	944.472	985.391
17) L4 AR-1242-2	5.834	4.954	55040068	217.5E6	954.815	1002.295
18) L4 AR-1242-3	5.896	5.073	34279044	123.3E6	940.039	1006.547
19) L4 AR-1242-4	5.993	5.155	28461551	159.3E6	952.139	958.832m
20) L4 AR-1242-5	6.723	5.677	31568865	192.1E6	939.198	994.431

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP092325\
Data File : PP075254.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Sep 2025 11:00
Operator : YP\AJ
Sample : AR1242ICC1000
Misc :
ALS Vial : 10 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
AR1242ICC1000


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 09/24/2025
Supervised By :mohammad ahmed 09/25/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 23 13:48:35 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP092325.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Sep 23 13:43:35 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm





#1 Tetrachloro-m-xylene

R.T.: 4.660 min
Delta R.T.: 0.000 min
Response: 123623850
Conc: 96.71 ng/ml

Instrument :

ECD_P

ClientSampleId :


AR1242ICC1000

Manual Integrations

APPROVED


Reviewed By :Yogesh Patel 09/24/2025

Supervised By :mohammad ahmed 09/25/2025




#1 Tetrachloro-m-xylene

R.T.: 3.799 min
Delta R.T.: 0.000 min
Response: 637105784
Conc: 102.37 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.443 min
Delta R.T.: 0.000 min
Response: 92662703
Conc: 96.02 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.807 min
Delta R.T.: 0.000 min
Response: 621028389
Conc: 98.99 ng/ml



R.T.: 5.812 min
Delta R.T.: 0.000 min
Response: 36886529
Conc: 944.47 ng/ml


Instrument :
ECD_P
ClientSampleId :
AR1242ICC1000

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 09/24/2025
Supervised By :mohammad ahmed 09/25/2025

Time
Response_ Signal: PP075254.D\ECD2B.ch


#16 AR-1242-1



R.T.: 4.896 min
Delta R.T.: 0.000 min
Response: 472236904
Conc: 985.39 ng/ml

Time
Response_ Signal: PP075254.D\ECD1A.ch


#17 AR-1242-2



R.T.: 5.834 min
Delta R.T.: 0.000 min
Response: 55040068
Conc: 954.81 ng/ml

Time
Response_ Signal: PP075254.D\ECD2B.ch

#17 AR-1242-2



R.T.: 4.954 min
Delta R.T.: 0.000 min
Response: 217510086
Conc: 1002.30 ng/ml

Response_ Signal: PP075254.D\ECD1A.ch

#18 AR-1242-3

R.T.: 5.896 min
Delta R.T.: 0.000 min
Response: 34279044
Conc: 940.04 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242ICC1000

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 09/24/2025
Supervised By :mohammad ahmed 09/25/2025

Time

Response_ Signal: PP075254.D\ECD2B.ch

#18 AR-1242-3

R.T.: 5.073 min
Delta R.T.: 0.000 min
Response: 123250580
Conc: 1006.55 ng/ml

Time

Response_ Signal: PP075254.D\ECD1A.ch

#19 AR-1242-4

R.T.: 5.993 min
Delta R.T.: 0.000 min
Response: 28461551
Conc: 952.14 ng/ml

Time

Response_ Signal: PP075254.D\ECD2B.ch

#19 AR-1242-4

R.T.: 5.155 min
Delta R.T.: 0.000 min
Response: 159270704
Conc: 958.83 ng/ml m

Time

Response_ Signal: PP075254.D\ECD1A.ch

#19 AR-1242-4

R.T.: 5.155 min
Delta R.T.: 0.000 min
Response: 159270704
Conc: 958.83 ng/ml m

Time

Response_ Signal: PP075254.D\ECD2B.ch

#19 AR-1242-4

R.T.: 5.155 min
Delta R.T.: 0.000 min
Response: 159270704
Conc: 958.83 ng/ml m

Time

Response_ Signal: PP075254.D\ECD1A.ch

#19 AR-1242-4

R.T.: 5.155 min
Delta R.T.: 0.000 min
Response: 159270704
Conc: 958.83 ng/ml m

Time

Response_ Signal: PP075254.D\ECD2B.ch

#19 AR-1242-4

R.T.: 5.155 min
Delta R.T.: 0.000 min
Response: 159270704
Conc: 958.83 ng/ml m

Time

Response_ Signal: PP075254.D\ECD1A.ch

#19 AR-1242-4

R.T.: 5.155 min
Delta R.T.: 0.000 min
Response: 159270704
Conc: 958.83 ng/ml m

Time

Response_ Signal: PP075254.D\ECD2B.ch

#19 AR-1242-4

R.T.: 5.155 min
Delta R.T.: 0.000 min
Response: 159270704
Conc: 958.83 ng/ml m

Time

Response_ Signal: PP075254.D\ECD1A.ch

#19 AR-1242-4

R.T.: 5.155 min
Delta R.T.: 0.000 min
Response: 159270704
Conc: 958.83 ng/ml m

Time

Response_ Signal: PP075254.D\ECD2B.ch

#19 AR-1242-4

R.T.: 5.155 min
Delta R.T.: 0.000 min
Response: 159270704
Conc: 958.83 ng/ml m

Time

Response_ Signal: PP075254.D\ECD1A.ch

#19 AR-1242-4

R.T.: 5.155 min
Delta R.T.: 0.000 min
Response: 159270704
Conc: 958.83 ng/ml m

Time

Response_ Signal: PP075254.D\ECD2B.ch

#19 AR-1242-4

R.T.: 5.155 min
Delta R.T.: 0.000 min
Response: 159270704
Conc: 958.83 ng/ml m

Time

Response_ Signal: PP075254.D\ECD1A.ch

#19 AR-1242-4

R.T.: 5.155 min
Delta R.T.: 0.000 min
Response: 159270704
Conc: 958.83 ng/ml m

Time

Response_ Signal: PP075254.D\ECD2B.ch

#19 AR-1242-4

R.T.: 5.155 min
Delta R.T.: 0.000 min
Response: 159270704
Conc: 958.83 ng/ml m

Time

Response_ Signal: PP075254.D\ECD1A.ch

#19 AR-1242-4

R.T.: 5.155 min
Delta R.T.: 0.000 min
Response: 159270704
Conc: 958.83 ng/ml m

Time

Response_ Signal: PP075254.D\ECD2B.ch

#19 AR-1242-4

R.T.: 5.155 min
Delta R.T.: 0.000 min
Response: 159270704
Conc: 958.83 ng/ml m

Time

Response_ Signal: PP075254.D\ECD1A.ch

#19 AR-1242-4

R.T.: 5.155 min
Delta R.T.: 0.000 min
Response: 159270704
Conc: 958.83 ng/ml m

Time

Response_ Signal: PP075254.D\ECD2B.ch

#19 AR-1242-4

R.T.: 5.155 min
Delta R.T.: 0.000 min
Response: 159270704
Conc: 958.83 ng/ml m

Time

Response_ Signal: PP075254.D\ECD1A.ch

#19 AR-1242-4

R.T.: 5.155 min
Delta R.T.: 0.000 min
Response: 159270704
Conc: 958.83 ng/ml m

Time

Response_ Signal: PP075254.D\ECD2B.ch

#19 AR-1242-4

R.T.: 5.155 min
Delta R.T.: 0.000 min
Response: 159270704
Conc: 958.83 ng/ml m

Time

Response_ Signal: PP075254.D\ECD1A.ch

#19 AR-1242-4

R.T.: 5.155 min
Delta R.T.: 0.000 min
Response: 159270704
Conc: 958.83 ng/ml m

Time

Response_ Signal: PP075254.D\ECD2B.ch

#19 AR-1242-4

R.T.: 5.155 min
Delta R.T.: 0.000 min
Response: 159270704
Conc: 958.83 ng/ml m

Time

Response_ Signal: PP075254.D\ECD1A.ch

#19 AR-1242-4

R.T.: 5.155 min
Delta R.T.: 0.000 min
Response: 159270704
Conc: 958.83 ng/ml m

Time

Response_ Signal: PP075254.D\ECD2B.ch

#19 AR-1242-4

R.T.: 5.155 min
Delta R.T.: 0.000 min
Response: 159270704
Conc: 958.83 ng/ml m

Time

Response_ Signal: PP075254.D\ECD1A.ch

#19 AR-1242-4

R.T.: 5.155 min
Delta R.T.: 0.000 min
Response: 159270704
Conc: 958.83 ng/ml m

Time

Response_ Signal: PP075254.D\ECD2B.ch

#19 AR-1242-4

R.T.: 5.155 min
Delta R.T.: 0.000 min
Response: 159270704
Conc: 958.83 ng/ml m

Time

Response_ Signal: PP075254.D\ECD1A.ch

#19 AR-1242-4

R.T.: 5.155 min
Delta R.T.: 0.000 min
Response: 159270704
Conc: 958.83 ng/ml m

Time

Response_ Signal: PP075254.D\ECD2B.ch

#19 AR-1242-4

R.T.: 5.155 min
Delta R.T.: 0.000 min
Response: 159270704
Conc: 958.83 ng/ml m

Time

Response_ Signal: PP075254.D\ECD1A.ch


#19 AR-1242-4

R.T.: 5.155 min
Delta R.T.: 0.000 min
Response: 159270704
Conc: 958.83 ng/ml m

Time

Response_ Signal: PP075254.D\ECD2B.ch

#19 AR-1242-4




#20 AR-1242-5

R.T.: 6.723 min
Delta R.T.: 0.000 min
Response: 31568865
Conc: 939.20 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242ICC1000

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 09/24/2025
Supervised By :mohammad ahmed 09/25/2025



#20 AR-1242-5

R.T.: 5.677 min
Delta R.T.: 0.000 min
Response: 192075351
Conc: 994.43 ng/ml

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP092325\
 Data File : PP075255.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 23 Sep 2025 11:16
 Operator : YP\AJ
 Sample : AR1242ICC750
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1242ICC750

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 09/24/2025
 Supervised By :mohammad ahmed 09/25/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 23 13:51:29 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP092325.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Sep 23 13:43:35 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.656	3.800	97168171	469.6E6	75.674	75.300
2) SA Decachlor...	10.439	8.807	72638787	466.8E6	75.181	74.600
Target Compounds						
16) L4 AR-1242-1	5.808	4.897	29413504	364.4E6	752.082	756.863
17) L4 AR-1242-2	5.830	4.955	43483200	162.3E6	752.881	748.698
18) L4 AR-1242-3	5.892	5.073	27466678	92235555	752.145	752.168
19) L4 AR-1242-4	5.989	5.156	22479541	121.9E6	751.345	763.610m
20) L4 AR-1242-5	6.719	5.678	25081847	147.5E6	747.465	759.181

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP092325\
Data File : PP075255.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Sep 2025 11:16
Operator : YP\AJ
Sample : AR1242ICC750
Misc :
ALS Vial : 11 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
AR1242ICC750


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 09/24/2025
Supervised By :mohammad ahmed 09/25/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 23 13:51:29 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP092325.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Sep 23 13:43:35 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.656 min
Delta R.T.: 0.000 min
Response: 97168171
Conc: 75.67 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242ICC750


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 09/24/2025
Supervised By :mohammad ahmed 09/25/2025




#1 Tetrachloro-m-xylene

R.T.: 3.800 min
Delta R.T.: 0.000 min
Response: 469578212
Conc: 75.30 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.439 min
Delta R.T.: 0.000 min
Response: 72638787
Conc: 75.18 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.807 min
Delta R.T.: 0.000 min
Response: 466755427
Conc: 74.60 ng/ml

Response_ Signal: PP075255.D\ECD1A.ch

#16 AR-1242-1

R.T.: 5.808 min
Delta R.T.: 0.000 min
Response: 29413504
Conc: 752.08 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242ICC750

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 09/24/2025
Supervised By :mohammad ahmed 09/25/2025

Time

Response_ Signal: PP075255.D\ECD2B.ch

#16 AR-1242-1

R.T.: 4.897 min
Delta R.T.: 0.000 min
Response: 364384624
Conc: 756.86 ng/ml

Time

Response_ Signal: PP075255.D\ECD1A.ch

#17 AR-1242-2

R.T.: 5.830 min
Delta R.T.: 0.000 min
Response: 43483200
Conc: 752.88 ng/ml


Time

Response_ Signal: PP075255.D\ECD2B.ch

#17 AR-1242-2

R.T.: 4.955 min
Delta R.T.: 0.000 min
Response: 162335588
Conc: 748.70 ng/ml

Time




#18 AR-1242-3

R.T.: 5.892 min
Delta R.T.: 0.000 min
Response: 27466678
Conc: 752.15 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242ICC750


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 09/24/2025
Supervised By :mohammad ahmed 09/25/2025




#18 AR-1242-3

R.T.: 5.073 min
Delta R.T.: 0.000 min
Response: 92235555
Conc: 752.17 ng/ml




#19 AR-1242-4

R.T.: 5.989 min
Delta R.T.: 0.000 min
Response: 22479541
Conc: 751.35 ng/ml



#19 AR-1242-4

R.T.: 5.156 min
Delta R.T.: 0.000 min
Response: 121935692
Conc: 763.61 ng/ml m




#20 AR-1242-5

R.T.: 6.719 min
Delta R.T.: 0.000 min
Response: 25081847
Conc: 747.47 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242ICC750

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 09/24/2025
Supervised By :mohammad ahmed 09/25/2025



#20 AR-1242-5

R.T.: 5.678 min
Delta R.T.: 0.000 min
Response: 147539663
Conc: 759.18 ng/ml

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP092325\
 Data File : PP075256.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 23 Sep 2025 11:32
 Operator : YP\AJ
 Sample : AR1242ICC500
 Misc :
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1242ICC500

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 09/24/2025
 Supervised By :mohammad ahmed 09/25/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 23 13:45:22 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP092325.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Sep 23 13:43:35 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.659	3.799	66015633	303.8E6	50.000	50.000
2) SA Decachlor...	10.443	8.808	50169729	316.8E6	50.000	50.000
Target Compounds						
16) L4 AR-1242-1	5.811	4.895	20611924	243.1E6	500.000	500.000
17) L4 AR-1242-2	5.832	4.954	30124732	108.3E6	500.000	500.000
18) L4 AR-1242-3	5.895	5.073	19326019	60823625	500.000	500.000
19) L4 AR-1242-4	5.992	5.155	15661455	81817884	500.000	481.178m
20) L4 AR-1242-5	6.722	5.676	17828128	97113427	500.000	500.000

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP092325\
Data File : PP075256.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Sep 2025 11:32
Operator : YP\AJ
Sample : AR1242ICC500
Misc :
ALS Vial : 12 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
AR1242ICC500


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 09/24/2025
Supervised By :mohammad ahmed 09/25/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 23 13:45:22 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP092325.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Sep 23 13:43:35 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.659 min
Delta R.T.: 0.000 min
Response: 66015633
Conc: 50.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242ICC500


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 09/24/2025
Supervised By :mohammad ahmed 09/25/2025




#1 Tetrachloro-m-xylene

R.T.: 3.799 min
Delta R.T.: 0.000 min
Response: 303811171
Conc: 50.00 ng/ml



#2 Decachlorobiphenyl


R.T.: 10.443 min
Delta R.T.: 0.000 min
Response: 50169729
Conc: 50.00 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.808 min
Delta R.T.: 0.000 min
Response: 316835092
Conc: 50.00 ng/ml

Response_ Signal: PP075256.D\ECD1A.ch



#16 AR-1242-1


R.T.: 5.811 min
Delta R.T.: 0.000 min
Response: 20611924
Conc: 500.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242ICC500

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 09/24/2025
Supervised By :mohammad ahmed 09/25/2025


Response_ Signal: PP075256.D\ECD2B.ch



#16 AR-1242-1

R.T.: 4.895 min
Delta R.T.: 0.000 min
Response: 243119630
Conc: 500.00 ng/ml


Response_ Signal: PP075256.D\ECD1A.ch



#17 AR-1242-2


R.T.: 5.832 min
Delta R.T.: 0.000 min
Response: 30124732
Conc: 500.00 ng/ml

Response_ Signal: PP075256.D\ECD2B.ch



#17 AR-1242-2

R.T.: 4.954 min
Delta R.T.: 0.000 min
Response: 108256965
Conc: 500.00 ng/ml




#18 AR-1242-3

R.T.: 5.895 min
Delta R.T.: 0.000 min
Response: 19326019
Conc: 500.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242ICC500


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 09/24/2025
Supervised By :mohammad ahmed 09/25/2025




#18 AR-1242-3

R.T.: 5.073 min
Delta R.T.: 0.000 min
Response: 60823625
Conc: 500.00 ng/ml




#19 AR-1242-4

R.T.: 5.992 min
Delta R.T.: 0.000 min
Response: 15661455
Conc: 500.00 ng/ml



#19 AR-1242-4

R.T.: 5.155 min
Delta R.T.: 0.000 min
Response: 81817884
Conc: 481.18 ng/ml m




#20 AR-1242-5

R.T.: 6.722 min
Delta R.T.: 0.000 min
Response: 17828128
Conc: 500.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242ICC500

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 09/24/2025
Supervised By :mohammad ahmed 09/25/2025



#20 AR-1242-5

R.T.: 5.676 min
Delta R.T.: 0.000 min
Response: 97113427
Conc: 500.00 ng/ml

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP092325\
 Data File : PP075257.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 23 Sep 2025 11:48
 Operator : YP\AJ
 Sample : AR1242ICC250
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1242ICC250

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 09/24/2025
 Supervised By :mohammad ahmed 09/25/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 23 13:55:28 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP092325.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Sep 23 13:54:24 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.661	3.797	34555716	150.8E6	26.407	24.382
2) SA Decachlor...	10.444	8.805	26178805	152.5E6	26.539	24.523
Target Compounds						
16) L4 AR-1242-1	5.813	4.894	11111719	121.8E6	274.745	252.283
17) L4 AR-1242-2	5.834	4.952	16022261	54845990	270.012	252.207
18) L4 AR-1242-3	5.897	5.072	10474071	31286183	276.635	253.831
19) L4 AR-1242-4	5.994	5.153	8389612	43909943	272.135	267.938m
20) L4 AR-1242-5	6.724	5.675	9880995	50425039	281.928	257.034

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP092325\
Data File : PP075257.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Sep 2025 11:48
Operator : YP\AJ
Sample : AR1242ICC250
Misc :
ALS Vial : 13 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
AR1242ICC250


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 09/24/2025
Supervised By :mohammad ahmed 09/25/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 23 13:55:28 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP092325.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Sep 23 13:54:24 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.661 min
Delta R.T.: 0.000 min
Response: 34555716
Conc: 26.41 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242ICC250


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 09/24/2025
Supervised By :mohammad ahmed 09/25/2025




#1 Tetrachloro-m-xylene

R.T.: 3.797 min
Delta R.T.: 0.000 min
Response: 150809845
Conc: 24.38 ng/ml



#2 Decachlorobiphenyl


R.T.: 10.444 min
Delta R.T.: 0.000 min
Response: 26178805
Conc: 26.54 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.805 min
Delta R.T.: 0.000 min
Response: 152462792
Conc: 24.52 ng/ml

Response_ Signal: PP075257.D\ECD1A.ch



#16 AR-1242-1


R.T.: 5.813 min
Delta R.T.: 0.000 min
Response: 11111719
Conc: 274.74 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242ICC250

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 09/24/2025
Supervised By :mohammad ahmed 09/25/2025


Time Response_ Signal: PP075257.D\ECD2B.ch



#16 AR-1242-1

R.T.: 4.894 min
Delta R.T.: 0.000 min
Response: 121830487
Conc: 252.28 ng/ml


Time Response_ Signal: PP075257.D\ECD1A.ch



#17 AR-1242-2


R.T.: 5.834 min
Delta R.T.: 0.000 min
Response: 16022261
Conc: 270.01 ng/ml

Time Response_ Signal: PP075257.D\ECD2B.ch



#17 AR-1242-2

R.T.: 4.952 min
Delta R.T.: 0.000 min
Response: 54845990
Conc: 252.21 ng/ml



#18 AR-1242-3

R.T.: 5.897 min
Delta R.T.: 0.000 min
Response: 10474071
Conc: 276.64 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242ICC250

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 09/24/2025
Supervised By :mohammad ahmed 09/25/2025

#18 AR-1242-3


R.T.: 5.072 min
Delta R.T.: 0.000 min
Response: 31286183
Conc: 253.83 ng/ml

#19 AR-1242-4

R.T.: 5.994 min
Delta R.T.: 0.000 min
Response: 8389612
Conc: 272.13 ng/ml

#19 AR-1242-4

R.T.: 5.153 min
Delta R.T.: 0.000 min
Response: 43909943
Conc: 267.94 ng/ml m



#20 AR-1242-5

R.T.: 6.724 min
Delta R.T.: 0.000 min
Response: 9880995
Conc: 281.93 ng/ml

Instrument :

ECD_P

ClientSampleId :


AR1242ICC250

Manual Integrations

APPROVED

Reviewed By :Yogesh Patel 09/24/2025

Supervised By :mohammad ahmed 09/25/2025



#20 AR-1242-5

R.T.: 5.675 min
Delta R.T.: 0.000 min
Response: 50425039
Conc: 257.03 ng/ml

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP092325\
 Data File : PP075258.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 23 Sep 2025 12:05
 Operator : YP\AJ
 Sample : AR1242ICC050
 Misc :
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1242ICC050

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 09/24/2025
 Supervised By :mohammad ahmed 09/25/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 23 14:00:26 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP092325.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Sep 23 13:59:11 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.657	3.796	6258278	24170300	4.827m	4.086
2) SA Decachlor...	10.437	8.803	4689985	25196266	4.802	4.212
Target Compounds						
16) L4 AR-1242-1	5.810	4.893	2326163	20197768	55.837	43.239
17) L4 AR-1242-2	5.831	4.952	3041537	8746690	51.001	41.859
18) L4 AR-1242-3	5.895	5.076	2534074	5839738	62.684	47.881
19) L4 AR-1242-4	5.991	5.155	1692732	6716503	53.850	42.218
20) L4 AR-1242-5	6.719	5.676	1779340	9434797	50.613m	48.462

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP092325\
Data File : PP075258.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Sep 2025 12:05
Operator : YP\AJ
Sample : AR1242ICC050
Misc :
ALS Vial : 14 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
AR1242ICC050


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 09/24/2025
Supervised By :mohammad ahmed 09/25/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 23 14:00:26 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP092325.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Sep 23 13:59:11 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.657 min
Delta R.T.: 0.000 min
Response: 6258278
Conc: 4.83 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242IC050


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 09/24/2025
Supervised By :mohammad ahmed 09/25/2025




#1 Tetrachloro-m-xylene

R.T.: 3.796 min
Delta R.T.: 0.000 min
Response: 24170300
Conc: 4.09 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.437 min
Delta R.T.: 0.000 min
Response: 4689985
Conc: 4.80 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.803 min
Delta R.T.: 0.000 min
Response: 25196266
Conc: 4.21 ng/ml




#16 AR-1242-1

R.T.: 5.810 min
Delta R.T.: 0.000 min
Response: 2326163
Conc: 55.84 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242ICC050


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 09/24/2025
Supervised By :mohammad ahmed 09/25/2025




#16 AR-1242-1

R.T.: 4.893 min
Delta R.T.: 0.000 min
Response: 20197768
Conc: 43.24 ng/ml




#17 AR-1242-2

R.T.: 5.831 min
Delta R.T.: 0.000 min
Response: 3041537
Conc: 51.00 ng/ml



#17 AR-1242-2

R.T.: 4.952 min
Delta R.T.: 0.000 min
Response: 8746690
Conc: 41.86 ng/ml




#18 AR-1242-3

R.T.: 5.895 min
Delta R.T.: 0.000 min
Response: 2534074
Conc: 62.68 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242ICC050


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 09/24/2025
Supervised By :mohammad ahmed 09/25/2025




#18 AR-1242-3

R.T.: 5.076 min
Delta R.T.: 0.000 min
Response: 5839738
Conc: 47.88 ng/ml




#19 AR-1242-4

R.T.: 5.991 min
Delta R.T.: 0.000 min
Response: 1692732
Conc: 53.85 ng/ml



#19 AR-1242-4

R.T.: 5.155 min
Delta R.T.: 0.000 min
Response: 6716503
Conc: 42.22 ng/ml



#20 AR-1242-5

R.T.: 6.719 min
Delta R.T.: 0.000 min
Response: 1779340
Conc: 50.61 ng/ml

ICAL Form

Instrument :

ECD_P

ClientSampleId :


AR1242ICC050

Manual Integrations

APPROVED

Reviewed By :Yogesh Patel 09/24/2025

Supervised By :mohammad ahmed 09/25/2025



#20 AR-1242-5

R.T.: 5.676 min
Delta R.T.: 0.000 min
Response: 9434797
Conc: 48.46 ng/ml

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP092325\
 Data File : PP075259.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 23 Sep 2025 12:37
 Operator : YP\AJ
 Sample : AR1248ICC1000
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1248ICC1000

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 09/24/2025
 Supervised By :mohammad ahmed 09/25/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 23 14:08:15 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP092325.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Sep 23 14:04:14 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.656	3.797	123.5E6	634.2E6	96.737	101.347
2) SA Decachlor...	10.430	8.803	93621437	628.1E6	96.237	99.121
Target Compounds						
21) L5 AR-1248-1	5.807	4.893	29172565	313.0E6	949.898	994.107
22) L5 AR-1248-2	6.078	5.113	39359802	194.6E6	950.495	985.073
23) L5 AR-1248-3	6.281	5.153	44196144	238.8E6	953.794	934.027m
24) L5 AR-1248-4	6.678	5.326	53562132	233.4E6	949.444	1004.611
25) L5 AR-1248-5	6.718	5.716	51867384	417.5E6	947.013	958.730

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP092325\
Data File : PP075259.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Sep 2025 12:37
Operator : YP\AJ
Sample : AR1248ICC1000
Misc :
ALS Vial : 15 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
AR1248ICC1000


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 09/24/2025
Supervised By :mohammad ahmed 09/25/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 23 14:08:15 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP092325.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Sep 23 14:04:14 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm





#1 Tetrachloro-m-xylene

R.T.: 4.656 min
Delta R.T.: 0.000 min
Response: 123522262
Conc: 96.74 ng/ml

Instrument :

ECD_P

ClientSampleId :


AR1248ICC1000

Manual Integrations

APPROVED


Reviewed By :Yogesh Patel 09/24/2025

Supervised By :mohammad ahmed 09/25/2025




#1 Tetrachloro-m-xylene

R.T.: 3.797 min
Delta R.T.: 0.000 min
Response: 634213576
Conc: 101.35 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.430 min
Delta R.T.: 0.000 min
Response: 93621437
Conc: 96.24 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.803 min
Delta R.T.: 0.000 min
Response: 628061988
Conc: 99.12 ng/ml




#21 AR-1248-1

R.T.: 5.807 min
Delta R.T.: 0.000 min
Response: 29172565
Conc: 949.90 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248ICC1000


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 09/24/2025
Supervised By :mohammad ahmed 09/25/2025




#21 AR-1248-1

R.T.: 4.893 min
Delta R.T.: 0.000 min
Response: 313001525
Conc: 994.11 ng/ml



#22 AR-1248-2

R.T.: 6.078 min
Delta R.T.: 0.000 min
Response: 39359802
Conc: 950.49 ng/ml



#22 AR-1248-2

R.T.: 5.113 min
Delta R.T.: 0.000 min
Response: 194621063
Conc: 985.07 ng/ml

Response_ Signal: PP075259.D\ECD1A.ch

#23 AR-1248-3

R.T.: 6.281 min
Delta R.T.: 0.000 min
Response: 44196144
Conc: 953.79 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248ICC1000

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 09/24/2025
Supervised By :mohammad ahmed 09/25/2025

Time

Response_ Signal: PP075259.D\ECD2B.ch

#23 AR-1248-3

R.T.: 5.153 min
Delta R.T.: 0.000 min
Response: 238833016
Conc: 934.03 ng/ml m

Time

Response_ Signal: PP075259.D\ECD1A.ch

#24 AR-1248-4

R.T.: 6.678 min
Delta R.T.: 0.000 min
Response: 53562132
Conc: 949.44 ng/ml


Time

Response_ Signal: PP075259.D\ECD2B.ch

#24 AR-1248-4

R.T.: 5.326 min
Delta R.T.: 0.000 min
Response: 233390044
Conc: 1004.61 ng/ml

Time




#25 AR-1248-5

R.T.: 6.718 min
Delta R.T.: 0.000 min
Response: 51867384
Conc: 947.01 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248ICC1000

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 09/24/2025
Supervised By :mohammad ahmed 09/25/2025



#25 AR-1248-5

R.T.: 5.716 min
Delta R.T.: 0.000 min
Response: 417516536
Conc: 958.73 ng/ml

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP092325\
 Data File : PP075260.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 23 Sep 2025 12:53
 Operator : YP\AJ
 Sample : AR1248ICC750
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1248ICC750

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 09/24/2025
 Supervised By :mohammad ahmed 09/25/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 23 14:11:30 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP092325.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Sep 23 14:04:14 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.658	3.801	96387949	476.3E6	75.324	75.742
2) SA Decachlor...	10.437	8.808	73513004	475.1E6	75.377	74.984
Target Compounds						
21) L5 AR-1248-1	5.810	4.897	23548183	235.7E6	761.091	749.095
22) L5 AR-1248-2	6.081	5.117	31337463	153.7E6	754.496	768.427
23) L5 AR-1248-3	6.284	5.157	35028270	183.3E6	753.951	722.769m
24) L5 AR-1248-4	6.681	5.331	42467941	177.0E6	751.857	757.895
25) L5 AR-1248-5	6.721	5.719	41109339	325.2E6	750.393	747.880

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP092325\
Data File : PP075260.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Sep 2025 12:53
Operator : YP\AJ
Sample : AR1248ICC750
Misc :
ALS Vial : 16 Sample Multiplier: 1

Instrument :

ECD_P

ClientSampleId :

AR1248ICC750

Manual Integrations



APPROVED


Reviewed By :Yogesh Patel 09/24/2025

Supervised By :mohammad ahmed 09/25/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 23 14:11:30 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP092325.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Sep 23 14:04:14 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm





#1 Tetrachloro-m-xylene

R.T.: 4.658 min
Delta R.T.: 0.000 min
Response: 96387949
Conc: 75.32 ng/ml

Instrument :

ECD_P

ClientSampleId :


AR1248ICC750

Manual Integrations

APPROVED


Reviewed By :Yogesh Patel 09/24/2025

Supervised By :mohammad ahmed 09/25/2025




#1 Tetrachloro-m-xylene

R.T.: 3.801 min
Delta R.T.: 0.000 min
Response: 476334942
Conc: 75.74 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.437 min
Delta R.T.: 0.000 min
Response: 73513004
Conc: 75.38 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.808 min
Delta R.T.: 0.000 min
Response: 475073939
Conc: 74.98 ng/ml




#21 AR-1248-1

R.T.: 5.810 min
Delta R.T.: 0.000 min
Response: 23548183
Conc: 761.09 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248ICC750


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 09/24/2025
Supervised By :mohammad ahmed 09/25/2025




#21 AR-1248-1

R.T.: 4.897 min
Delta R.T.: 0.000 min
Response: 235715637
Conc: 749.10 ng/ml




#22 AR-1248-2

R.T.: 6.081 min
Delta R.T.: 0.000 min
Response: 31337463
Conc: 754.50 ng/ml



#22 AR-1248-2

R.T.: 5.117 min
Delta R.T.: 0.000 min
Response: 153706618
Conc: 768.43 ng/ml




#23 AR-1248-3

R.T.: 6.284 min
Delta R.T.: 0.000 min
Response: 35028270
Conc: 753.95 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248ICC750


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 09/24/2025
Supervised By :mohammad ahmed 09/25/2025




#23 AR-1248-3

R.T.: 5.157 min
Delta R.T.: 0.000 min
Response: 183253915
Conc: 722.77 ng/ml m




#24 AR-1248-4

R.T.: 6.681 min
Delta R.T.: 0.000 min
Response: 42467941
Conc: 751.86 ng/ml



#24 AR-1248-4

R.T.: 5.331 min
Delta R.T.: 0.000 min
Response: 177004743
Conc: 757.89 ng/ml




#25 AR-1248-5

R.T.: 6.721 min
Delta R.T.: 0.000 min
Response: 41109339
Conc: 750.39 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248ICC750

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 09/24/2025
Supervised By :mohammad ahmed 09/25/2025



#25 AR-1248-5

R.T.: 5.719 min
Delta R.T.: 0.000 min
Response: 325234126
Conc: 747.88 ng/ml

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP092325\
Data File : PP075261.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Sep 2025 13:10
Operator : YP\AJ
Sample : AR1248ICC500
Misc :
ALS Vial : 17 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1248ICC500

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 09/24/2025
Supervised By :mohammad ahmed 09/25/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 23 14:04:29 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP092325.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Sep 23 14:04:14 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.660	3.796	65927642	308.7E6	50.000	50.000
2) SA Decachlor...	10.439	8.803	50471290	319.6E6	50.000	50.000
Target Compounds						
21) L5 AR-1248-1	5.811	4.892	16124985	158.4E6	500.000	500.000
22) L5 AR-1248-2	6.083	5.112	21729896	100.3E6	500.000	500.000
23) L5 AR-1248-3	6.285	5.152	24239118	131.8E6	500.000	515.550m
24) L5 AR-1248-4	6.683	5.325	29633124	115.6E6	500.000	500.000
25) L5 AR-1248-5	6.723	5.714	28835751	226.7E6	500.000	500.000

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP092325\
Data File : PP075261.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Sep 2025 13:10
Operator : YP\AJ
Sample : AR1248ICC500
Misc :
ALS Vial : 17 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
AR1248ICC500

Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 09/24/2025
Supervised By :mohammad ahmed 09/25/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 23 14:04:29 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP092325.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Sep 23 14:04:14 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Response_ Signal: PP075261.D\ECD1A.ch



#1 Tetrachloro-m-xylene


R.T.: 4.660 min
Delta R.T.: 0.000 min
Response: 65927642
Conc: 50.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248ICC500

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 09/24/2025
Supervised By :mohammad ahmed 09/25/2025


Time Response_ Signal: PP075261.D\ECD2B.ch



#1 Tetrachloro-m-xylene

R.T.: 3.796 min
Delta R.T.: 0.000 min
Response: 308675147
Conc: 50.00 ng/ml


Time Response_ Signal: PP075261.D\ECD1A.ch



#2 Decachlorobiphenyl


R.T.: 10.439 min
Delta R.T.: 0.000 min
Response: 50471290
Conc: 50.00 ng/ml

Time Response_ Signal: PP075261.D\ECD2B.ch



#2 Decachlorobiphenyl

R.T.: 8.803 min
Delta R.T.: 0.000 min
Response: 319600115
Conc: 50.00 ng/ml




#21 AR-1248-1

R.T.: 5.811 min
Delta R.T.: 0.000 min
Response: 16124985
Conc: 500.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248ICC500


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 09/24/2025
Supervised By :mohammad ahmed 09/25/2025




#21 AR-1248-1

R.T.: 4.892 min
Delta R.T.: 0.000 min
Response: 158356149
Conc: 500.00 ng/ml




#22 AR-1248-2

R.T.: 6.083 min
Delta R.T.: 0.000 min
Response: 21729896
Conc: 500.00 ng/ml



#22 AR-1248-2

R.T.: 5.112 min
Delta R.T.: 0.000 min
Response: 100259637
Conc: 500.00 ng/ml




#23 AR-1248-3

R.T.: 6.285 min
Delta R.T.: 0.000 min
Response: 24239118
Conc: 500.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248ICC500


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 09/24/2025
Supervised By :mohammad ahmed 09/25/2025




#23 AR-1248-3

R.T.: 5.152 min
Delta R.T.: 0.000 min
Response: 131811046
Conc: 515.55 ng/ml m




#24 AR-1248-4

R.T.: 6.683 min
Delta R.T.: 0.000 min
Response: 29633124
Conc: 500.00 ng/ml



#24 AR-1248-4

R.T.: 5.325 min
Delta R.T.: 0.000 min
Response: 115623755
Conc: 500.00 ng/ml



#25 AR-1248-5

R.T.: 6.723 min
Delta R.T.: 0.000 min
Response: 28835751
Conc: 500.00 ng/ml

Instrument :

ECD_P

ClientSampleId :


AR1248ICC500

Manual Integrations

APPROVED

Reviewed By :Yogesh Patel 09/24/2025

Supervised By :mohammad ahmed 09/25/2025



#25 AR-1248-5

R.T.: 5.714 min
Delta R.T.: 0.000 min
Response: 226731113
Conc: 500.00 ng/ml

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP092325\
 Data File : PP075262.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 23 Sep 2025 13:26
 Operator : YP\AJ
 Sample : AR1248ICC250
 Misc :
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1248ICC250

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 09/24/2025
 Supervised By :mohammad ahmed 09/25/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 23 14:14:44 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP092325.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Sep 23 14:04:14 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.657	3.801	33217069	138.8E6	25.712	22.741
2) SA Decachlor...	10.437	8.806	25462185	150.8E6	25.822	24.087
Target Compounds						
21) L5 AR-1248-1	5.809	4.895	8438835	78304958	266.682	249.137
22) L5 AR-1248-2	6.080	5.116	11269776	51510660	265.668	255.596
23) L5 AR-1248-3	6.282	5.157	12543499	64693904	264.697	262.977m
24) L5 AR-1248-4	6.681	5.329	15335475	57505517	265.786	247.159
25) L5 AR-1248-5	6.720	5.719	15386978	109.3E6	272.457	250.932

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP092325\
Data File : PP075262.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Sep 2025 13:26
Operator : YP\AJ
Sample : AR1248ICC250
Misc :
ALS Vial : 18 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
AR1248ICC250


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 09/24/2025
Supervised By :mohammad ahmed 09/25/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 23 14:14:44 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP092325.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Sep 23 14:04:14 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.657 min
Delta R.T.: 0.000 min
Response: 33217069
Conc: 25.71 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248ICC250


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 09/24/2025
Supervised By :mohammad ahmed 09/25/2025




#1 Tetrachloro-m-xylene

R.T.: 3.801 min
Delta R.T.: 0.000 min
Response: 138838609
Conc: 22.74 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.437 min
Delta R.T.: 0.000 min
Response: 25462185
Conc: 25.82 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.806 min
Delta R.T.: 0.000 min
Response: 150774763
Conc: 24.09 ng/ml




#21 AR-1248-1

R.T.: 5.809 min
Delta R.T.: 0.000 min
Response: 8438835
Conc: 266.68 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248ICC250


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 09/24/2025
Supervised By :mohammad ahmed 09/25/2025




#21 AR-1248-1

R.T.: 4.895 min
Delta R.T.: 0.000 min
Response: 78304958
Conc: 249.14 ng/ml




#22 AR-1248-2

R.T.: 6.080 min
Delta R.T.: 0.000 min
Response: 11269776
Conc: 265.67 ng/ml



#22 AR-1248-2

R.T.: 5.116 min
Delta R.T.: 0.000 min
Response: 51510660
Conc: 255.60 ng/ml




#23 AR-1248-3

R.T.: 6.282 min
Delta R.T.: 0.000 min
Response: 12543499
Conc: 264.70 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248ICC250


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 09/24/2025
Supervised By :mohammad ahmed 09/25/2025




#23 AR-1248-3

R.T.: 5.157 min
Delta R.T.: 0.000 min
Response: 64693904
Conc: 262.98 ng/ml m




#24 AR-1248-4

R.T.: 6.681 min
Delta R.T.: 0.000 min
Response: 15335475
Conc: 265.79 ng/ml



#24 AR-1248-4

R.T.: 5.329 min
Delta R.T.: 0.000 min
Response: 57505517
Conc: 247.16 ng/ml




#25 AR-1248-5

R.T.: 6.720 min
Delta R.T.: 0.000 min
Response: 15386978
Conc: 272.46 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248ICC250

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 09/24/2025
Supervised By :mohammad ahmed 09/25/2025



#25 AR-1248-5

R.T.: 5.719 min
Delta R.T.: 0.000 min
Response: 109260049
Conc: 250.93 ng/ml

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP092325\
 Data File : PP075263.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 23 Sep 2025 13:42
 Operator : YP\AJ
 Sample : AR1248ICC050
 Misc :
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1248ICC050

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 09/24/2025
 Supervised By :mohammad ahmed 09/25/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 23 14:19:30 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP092325.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Sep 23 14:04:14 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.655	3.797	6566364	21995820	5.074m	3.816
2) SA Decachlor...	10.437	8.804	4912073	26887807	4.985	4.420
Target Compounds						
21) L5 AR-1248-1	5.809	4.897	1561167	12255520	49.467	40.788
22) L5 AR-1248-2	6.080	5.114	2357910	10313506	54.370	50.936
23) L5 AR-1248-3	6.282	5.154	2462035	12158074	51.552	45.736m
24) L5 AR-1248-4	6.681	5.326	3238253	11652585	54.782	50.066
25) L5 AR-1248-5	6.719	5.728	3479899	15059630	59.114m	36.859 #

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP092325\
Data File : PP075263.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Sep 2025 13:42
Operator : YP\AJ
Sample : AR1248ICC050
Misc :
ALS Vial : 19 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
AR1248ICC050


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 09/24/2025
Supervised By :mohammad ahmed 09/25/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 23 14:19:30 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP092325.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Sep 23 14:04:14 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm





#1 Tetrachloro-m-xylene

R.T.: 4.655 min
Delta R.T.: 0.000 min
Response: 6566364
Conc: 5.07 ng/ml

Instrument :

ECD_P

ClientSampleId :


AR1248ICC050

Manual Integrations

APPROVED


Reviewed By :Yogesh Patel 09/24/2025

Supervised By :mohammad ahmed 09/25/2025




#1 Tetrachloro-m-xylene

R.T.: 3.797 min
Delta R.T.: 0.000 min
Response: 21995820
Conc: 3.82 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.437 min
Delta R.T.: 0.000 min
Response: 4912073
Conc: 4.99 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.804 min
Delta R.T.: 0.000 min
Response: 26887807
Conc: 4.42 ng/ml




#21 AR-1248-1

R.T.: 5.809 min
Delta R.T.: 0.000 min
Response: 1561167
Conc: 49.47 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248ICC050


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 09/24/2025
Supervised By :mohammad ahmed 09/25/2025




#21 AR-1248-1

R.T.: 4.897 min
Delta R.T.: 0.000 min
Response: 12255520
Conc: 40.79 ng/ml




#22 AR-1248-2

R.T.: 6.080 min
Delta R.T.: 0.000 min
Response: 2357910
Conc: 54.37 ng/ml



#22 AR-1248-2

R.T.: 5.114 min
Delta R.T.: 0.000 min
Response: 10313506
Conc: 50.94 ng/ml




#23 AR-1248-3

R.T.: 6.282 min
Delta R.T.: 0.000 min
Response: 2462035
Conc: 51.55 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248ICC050


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 09/24/2025
Supervised By :mohammad ahmed 09/25/2025




#23 AR-1248-3

R.T.: 5.154 min
Delta R.T.: -0.001 min
Response: 12158074
Conc: 45.74 ng/ml m




#24 AR-1248-4

R.T.: 6.681 min
Delta R.T.: 0.000 min
Response: 3238253
Conc: 54.78 ng/ml



#24 AR-1248-4

R.T.: 5.326 min
Delta R.T.: 0.000 min
Response: 11652585
Conc: 50.07 ng/ml




#25 AR-1248-5

R.T.: 6.719 min
Delta R.T.: 0.000 min
Response: 3479899
Conc: 59.11 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248ICC050

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 09/24/2025
Supervised By :mohammad ahmed 09/25/2025



#25 AR-1248-5

R.T.: 5.728 min
Delta R.T.: 0.000 min
Response: 15059630
Conc: 36.86 ng/ml

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP092325\
Data File : PP075264.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Sep 2025 14:15
Operator : YP\AJ
Sample : AR1254ICC1000
Misc :
ALS Vial : 20 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1254ICC1000

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 09/24/2025
Supervised By :mohammad ahmed 09/25/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 23 15:27:05 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP092325.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Sep 23 15:25:14 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.660	3.799	124.3E6	624.1E6	96.412	98.637
2) SA Decachlor...	10.440	8.805	94019612	623.8E6	94.956	97.883
Target Compounds						
26) L6 AR-1254-1	6.659	5.678	61475917	503.0E6	968.019m	971.997
27) L6 AR-1254-2	6.876	5.825	77552440	378.2E6	944.080	972.887
28) L6 AR-1254-3	7.238	6.227	82845377	715.2E6	946.428	997.656
29) L6 AR-1254-4	7.520	6.453	64880438	505.6E6	943.395	962.588
30) L6 AR-1254-5	7.936	6.870	77310574	551.8E6	947.519	921.721

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP092325\
Data File : PP075264.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Sep 2025 14:15
Operator : YP\AJ
Sample : AR1254ICC1000
Misc :
ALS Vial : 20 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
AR1254ICC1000


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 09/24/2025
Supervised By :mohammad ahmed 09/25/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 23 15:27:05 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP092325.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Sep 23 15:25:14 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm





#1 Tetrachloro-m-xylene

R.T.: 4.660 min
Delta R.T.: 0.000 min
Response: 124326548
Conc: 96.41 ng/ml

Instrument :

ECD_P


ClientSampleId :

AR1254ICC1000

Manual Integrations


APPROVED

Reviewed By :Yogesh Patel 09/24/2025
Supervised By :mohammad ahmed 09/25/2025




#1 Tetrachloro-m-xylene

R.T.: 3.799 min
Delta R.T.: 0.000 min
Response: 624084086
Conc: 98.64 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.440 min
Delta R.T.: 0.000 min
Response: 94019612
Conc: 94.96 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.805 min
Delta R.T.: 0.000 min
Response: 623770605
Conc: 97.88 ng/ml

Response_ Signal: PP075264.D\ECD1A.ch

#26 AR-1254-1

R.T.: 6.659 min
Delta R.T.: 0.000 min
Response: 61475917
Conc: 968.02 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254ICC1000

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 09/24/2025
Supervised By :mohammad ahmed 09/25/2025

Time 6.50 6.55 6.60 6.65 6.70 6.75

Response_ Signal: PP075264.D\ECD2B.ch

#26 AR-1254-1

R.T.: 5.678 min
Delta R.T.: 0.000 min
Response: 503004292
Conc: 972.00 ng/ml

Time 5.40 5.50 5.60 5.70 5.80 5.90 6.00

Response_ Signal: PP075264.D\ECD1A.ch

#27 AR-1254-2

R.T.: 6.876 min
Delta R.T.: 0.000 min
Response: 77552440
Conc: 944.08 ng/ml


Time 6.70 6.75 6.80 6.85 6.90 6.95 7.00

Response_ Signal: PP075264.D\ECD2B.ch

#27 AR-1254-2

R.T.: 5.825 min
Delta R.T.: 0.000 min
Response: 378219544
Conc: 972.89 ng/ml

Time 5.65 5.70 5.75 5.80 5.85 5.90 5.95




#28 AR-1254-3

R.T.: 7.238 min
Delta R.T.: 0.000 min
Response: 82845377
Conc: 946.43 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254ICC1000


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 09/24/2025
Supervised By :mohammad ahmed 09/25/2025




#28 AR-1254-3

R.T.: 6.227 min
Delta R.T.: 0.000 min
Response: 715208754
Conc: 997.66 ng/ml




#29 AR-1254-4

R.T.: 7.520 min
Delta R.T.: 0.000 min
Response: 64880438
Conc: 943.40 ng/ml



#29 AR-1254-4

R.T.: 6.453 min
Delta R.T.: 0.000 min
Response: 505627774
Conc: 962.59 ng/ml




#30 AR-1254-5

R.T.: 7.936 min
Delta R.T.: 0.000 min
Response: 77310574
Conc: 947.52 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254ICC1000

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 09/24/2025
Supervised By :mohammad ahmed 09/25/2025



#30 AR-1254-5

R.T.: 6.870 min
Delta R.T.: 0.000 min
Response: 551845651
Conc: 921.72 ng/ml

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP092325\
 Data File : PP075265.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 23 Sep 2025 14:31
 Operator : YP\AJ
 Sample : AR1254ICC750
 Misc :
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1254ICC750

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 09/24/2025
 Supervised By :mohammad ahmed 09/25/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 23 15:30:58 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP092325.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Sep 23 15:25:14 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.656	3.799	97801585	485.2E6	75.559	76.113
2) SA Decachlor...	10.437	8.806	74646283	470.9E6	75.259	74.263
Target Compounds						
26) L6 AR-1254-1	6.656	5.678	47050728	401.3E6	729.090m	766.735
27) L6 AR-1254-2	6.873	5.825	61746016	302.7E6	751.107	768.826
28) L6 AR-1254-3	7.235	6.227	65961946	542.1E6	752.364	754.081
29) L6 AR-1254-4	7.517	6.454	51519817	403.9E6	749.416	762.571
30) L6 AR-1254-5	7.933	6.871	61456126	473.6E6	752.135	776.903

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP092325\
Data File : PP075265.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Sep 2025 14:31
Operator : YP\AJ
Sample : AR1254ICC750
Misc :
ALS Vial : 21 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
AR1254ICC750

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 09/24/2025
Supervised By :mohammad ahmed 09/25/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 23 15:30:58 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP092325.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Sep 23 15:25:14 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm



Response_ Signal: PP075265.D\ECD1A.ch

#1 Tetrachloro-m-xylene

R.T.: 4.656 min
Delta R.T.: 0.000 min
Response: 97801585
Conc: 75.56 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254ICC750

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 09/24/2025
Supervised By :mohammad ahmed 09/25/2025

Time

Response_ Signal: PP075265.D\ECD2B.ch

#1 Tetrachloro-m-xylene

R.T.: 3.799 min
Delta R.T.: 0.000 min
Response: 485175176
Conc: 76.11 ng/ml

Time

Response_ Signal: PP075265.D\ECD1A.ch

#2 Decachlorobiphenyl

R.T.: 10.437 min
Delta R.T.: 0.000 min
Response: 74646283
Conc: 75.26 ng/ml

Time

Response_ Signal: PP075265.D\ECD2B.ch

#2 Decachlorobiphenyl

R.T.: 8.806 min
Delta R.T.: 0.000 min
Response: 470934373
Conc: 74.26 ng/ml

Time

Response_ Signal: PP075265.D\ECD1A.ch

#2 Decachlorobiphenyl

R.T.: 10.437 min
Delta R.T.: 0.000 min
Response: 74646283
Conc: 75.26 ng/ml

Time

Response_ Signal: PP075265.D\ECD2B.ch

#2 Decachlorobiphenyl

R.T.: 8.806 min
Delta R.T.: 0.000 min
Response: 470934373
Conc: 74.26 ng/ml

Time

Response_ Signal: PP075265.D\ECD1A.ch

#2 Decachlorobiphenyl

R.T.: 10.437 min
Delta R.T.: 0.000 min
Response: 74646283
Conc: 75.26 ng/ml

Time

Response_ Signal: PP075265.D\ECD2B.ch

#2 Decachlorobiphenyl

R.T.: 8.806 min
Delta R.T.: 0.000 min
Response: 470934373
Conc: 74.26 ng/ml

Time

Response_ Signal: PP075265.D\ECD1A.ch

#2 Decachlorobiphenyl

R.T.: 10.437 min
Delta R.T.: 0.000 min
Response: 74646283
Conc: 75.26 ng/ml

Time

Response_ Signal: PP075265.D\ECD2B.ch

#2 Decachlorobiphenyl

R.T.: 8.806 min
Delta R.T.: 0.000 min
Response: 470934373
Conc: 74.26 ng/ml

Time

Response_ Signal: PP075265.D\ECD1A.ch

#2 Decachlorobiphenyl

R.T.: 10.437 min
Delta R.T.: 0.000 min
Response: 74646283
Conc: 75.26 ng/ml

Time

Response_ Signal: PP075265.D\ECD2B.ch

#2 Decachlorobiphenyl

R.T.: 8.806 min
Delta R.T.: 0.000 min
Response: 470934373
Conc: 74.26 ng/ml

Time

Response_ Signal: PP075265.D\ECD1A.ch

#2 Decachlorobiphenyl

R.T.: 10.437 min
Delta R.T.: 0.000 min
Response: 74646283
Conc: 75.26 ng/ml

Time

Response_ Signal: PP075265.D\ECD2B.ch

#2 Decachlorobiphenyl

R.T.: 8.806 min
Delta R.T.: 0.000 min
Response: 470934373
Conc: 74.26 ng/ml

Time

Response_ Signal: PP075265.D\ECD1A.ch

#2 Decachlorobiphenyl

R.T.: 10.437 min
Delta R.T.: 0.000 min
Response: 74646283
Conc: 75.26 ng/ml

Time

Response_ Signal: PP075265.D\ECD2B.ch

#2 Decachlorobiphenyl

R.T.: 8.806 min
Delta R.T.: 0.000 min
Response: 470934373
Conc: 74.26 ng/ml

Time

Response_ Signal: PP075265.D\ECD1A.ch

#2 Decachlorobiphenyl

R.T.: 10.437 min
Delta R.T.: 0.000 min
Response: 74646283
Conc: 75.26 ng/ml

Time

Response_ Signal: PP075265.D\ECD2B.ch

#2 Decachlorobiphenyl

R.T.: 8.806 min
Delta R.T.: 0.000 min
Response: 470934373
Conc: 74.26 ng/ml

Time

Response_ Signal: PP075265.D\ECD1A.ch

#2 Decachlorobiphenyl

R.T.: 10.437 min
Delta R.T.: 0.000 min
Response: 74646283
Conc: 75.26 ng/ml

Time

Response_ Signal: PP075265.D\ECD2B.ch

#2 Decachlorobiphenyl

R.T.: 8.806 min
Delta R.T.: 0.000 min
Response: 470934373
Conc: 74.26 ng/ml

Time

Response_ Signal: PP075265.D\ECD1A.ch


#2 Decachlorobiphenyl

R.T.: 10.437 min
Delta R.T.: 0.000 min
Response: 74646283
Conc: 75.26 ng/ml

Time

Response_ Signal: PP075265.D\ECD2B.ch

#2 Decachlorobiphenyl




R.T.: 6.656 min
Delta R.T.: 0.000 min
Response: 47050728
Conc: 729.09 ng/ml


Instrument :
ECD_P
ClientSampleId :
AR1254ICC750

Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 09/24/2025
Supervised By :mohammad ahmed 09/25/2025




R.T.: 5.678 min
Delta R.T.: 0.000 min
Response: 401258825
Conc: 766.74 ng/ml



R.T.: 6.873 min
Delta R.T.: 0.000 min
Response: 61746016
Conc: 751.11 ng/ml



R.T.: 5.825 min
Delta R.T.: 0.000 min
Response: 302687531
Conc: 768.83 ng/ml




#28 AR-1254-3

R.T.: 7.235 min
Delta R.T.: 0.000 min
Response: 65961946
Conc: 752.36 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254ICC750


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 09/24/2025
Supervised By :mohammad ahmed 09/25/2025




#28 AR-1254-3

R.T.: 6.227 min
Delta R.T.: 0.000 min
Response: 542066993
Conc: 754.08 ng/ml




#29 AR-1254-4

R.T.: 7.517 min
Delta R.T.: 0.000 min
Response: 51519817
Conc: 749.42 ng/ml



#29 AR-1254-4

R.T.: 6.454 min
Delta R.T.: 0.000 min
Response: 403948394
Conc: 762.57 ng/ml



#30 AR-1254-5

R.T.: 7.933 min
Delta R.T.: 0.000 min
Response: 61456126
Conc: 752.13 ng/ml

Instrument :

ECD_P

ClientSampleId :


AR1254ICC750

Manual Integrations

APPROVED

Reviewed By :Yogesh Patel 09/24/2025

Supervised By :mohammad ahmed 09/25/2025



#30 AR-1254-5

R.T.: 6.871 min
Delta R.T.: 0.000 min
Response: 473636635
Conc: 776.90 ng/ml

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP092325\
Data File : PP075266.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Sep 2025 14:47
Operator : YP\AJ
Sample : AR1254ICC500
Misc :
ALS Vial : 22 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1254ICC500

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 09/24/2025
Supervised By :mohammad ahmed 09/25/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 23 15:14:33 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP092325.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Sep 23 15:13:57 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.660	3.799	66790785	320.7E6	50.000	50.000
2) SA Decachlor...	10.441	8.806	52004273	325.4E6	50.000	50.000
Target Compounds						
26) L6 AR-1254-1	6.659	5.679	32993341	266.0E6	519.163m	500.000
27) L6 AR-1254-2	6.876	5.826	43369825	199.7E6	500.000	500.000
28) L6 AR-1254-3	7.238	6.228	46112134	359.3E6	500.000	500.000
29) L6 AR-1254-4	7.520	6.455	36333103	272.5E6	500.000	500.000
30) L6 AR-1254-5	7.936	6.871	42937319	322.8E6	500.000	500.000

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP092325\
Data File : PP075266.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Sep 2025 14:47
Operator : YP\AJ
Sample : AR1254ICC500
Misc :
ALS Vial : 22 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
AR1254ICC500


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 09/24/2025
Supervised By :mohammad ahmed 09/25/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 23 15:14:33 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP092325.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Sep 23 15:13:57 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.660 min
Delta R.T.: 0.000 min
Response: 66790785
Conc: 50.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254ICC500


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 09/24/2025
Supervised By :mohammad ahmed 09/25/2025




#1 Tetrachloro-m-xylene

R.T.: 3.799 min
Delta R.T.: 0.000 min
Response: 320666699
Conc: 50.00 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.441 min
Delta R.T.: 0.000 min
Response: 52004273
Conc: 50.00 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.806 min
Delta R.T.: 0.000 min
Response: 325373732
Conc: 50.00 ng/ml




#26 AR-1254-1

R.T.: 6.659 min
Delta R.T.: 0.000 min
Response: 32993341
Conc: 519.16 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254ICC500


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 09/24/2025
Supervised By :mohammad ahmed 09/25/2025




#26 AR-1254-1

R.T.: 5.679 min
Delta R.T.: 0.000 min
Response: 265993528
Conc: 500.00 ng/ml




#27 AR-1254-2

R.T.: 6.876 min
Delta R.T.: 0.000 min
Response: 43369825
Conc: 500.00 ng/ml



#27 AR-1254-2

R.T.: 5.826 min
Delta R.T.: 0.000 min
Response: 199650338
Conc: 500.00 ng/ml




#28 AR-1254-3

R.T.: 7.238 min
Delta R.T.: 0.000 min
Response: 46112134
Conc: 500.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254ICC500


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 09/24/2025
Supervised By :mohammad ahmed 09/25/2025




#28 AR-1254-3

R.T.: 6.228 min
Delta R.T.: 0.000 min
Response: 359284618
Conc: 500.00 ng/ml




#29 AR-1254-4

R.T.: 7.520 min
Delta R.T.: 0.000 min
Response: 36333103
Conc: 500.00 ng/ml



#29 AR-1254-4

R.T.: 6.455 min
Delta R.T.: 0.000 min
Response: 272465654
Conc: 500.00 ng/ml



#30 AR-1254-5

R.T.: 7.936 min
Delta R.T.: 0.000 min
Response: 42937319
Conc: 500.00 ng/ml

Instrument :

ECD_P

ClientSampleId :


AR1254ICC500

Manual Integrations

APPROVED

Reviewed By :Yogesh Patel 09/24/2025

Supervised By :mohammad ahmed 09/25/2025



#30 AR-1254-5

R.T.: 6.871 min
Delta R.T.: 0.000 min
Response: 322789634
Conc: 500.00 ng/ml

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP092325\
 Data File : PP075267.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 23 Sep 2025 15:04
 Operator : YP\AJ
 Sample : AR1254ICC250
 Misc :
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1254ICC250

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 09/24/2025
 Supervised By :mohammad ahmed 09/25/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 23 15:34:51 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP092325.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Sep 23 15:25:14 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.659	3.798	33173271	137.6E6	25.469	22.344
2) SA Decachlor...	10.440	8.806	25068322	153.9E6	25.205	24.451
Target Compounds						
26) L6 AR-1254-1	6.658	5.678	15749633	130.5E6	242.688m	249.504
27) L6 AR-1254-2	6.875	5.825	22036705	98058544	263.308	249.301
28) L6 AR-1254-3	7.238	6.227	23518741	165.6E6	263.446	234.968
29) L6 AR-1254-4	7.519	6.453	18117404	123.1E6	260.019	236.508
30) L6 AR-1254-5	7.935	6.870	21487152	144.6E6	259.604	240.237

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP092325\
Data File : PP075267.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Sep 2025 15:04
Operator : YP\AJ
Sample : AR1254ICC250
Misc :
ALS Vial : 23 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
AR1254ICC250


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 09/24/2025
Supervised By :mohammad ahmed 09/25/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 23 15:34:51 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP092325.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Sep 23 15:25:14 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.659 min
Delta R.T.: 0.000 min
Response: 33173271
Conc: 25.47 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254ICC250


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 09/24/2025
Supervised By :mohammad ahmed 09/25/2025




#1 Tetrachloro-m-xylene

R.T.: 3.798 min
Delta R.T.: 0.000 min
Response: 137556954
Conc: 22.34 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.440 min
Delta R.T.: 0.000 min
Response: 25068322
Conc: 25.21 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.806 min
Delta R.T.: 0.000 min
Response: 153931032
Conc: 24.45 ng/ml




#26 AR-1254-1

R.T.: 6.658 min
Delta R.T.: 0.000 min
Response: 15749633
Conc: 242.69 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254ICC250


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 09/24/2025
Supervised By :mohammad ahmed 09/25/2025




#26 AR-1254-1

R.T.: 5.678 min
Delta R.T.: 0.000 min
Response: 130487950
Conc: 249.50 ng/ml




#27 AR-1254-2

R.T.: 6.875 min
Delta R.T.: 0.000 min
Response: 22036705
Conc: 263.31 ng/ml



#27 AR-1254-2

R.T.: 5.825 min
Delta R.T.: 0.000 min
Response: 98058544
Conc: 249.30 ng/ml




#28 AR-1254-3

R.T.: 7.238 min
Delta R.T.: 0.000 min
Response: 23518741
Conc: 263.45 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254ICC250


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 09/24/2025
Supervised By :mohammad ahmed 09/25/2025




#28 AR-1254-3

R.T.: 6.227 min
Delta R.T.: 0.000 min
Response: 165586814
Conc: 234.97 ng/ml




#29 AR-1254-4

R.T.: 7.519 min
Delta R.T.: 0.000 min
Response: 18117404
Conc: 260.02 ng/ml



#29 AR-1254-4

R.T.: 6.453 min
Delta R.T.: 0.000 min
Response: 123068946
Conc: 236.51 ng/ml




#30 AR-1254-5

R.T.: 7.935 min
Delta R.T.: 0.000 min
Response: 21487152
Conc: 259.60 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254ICC250

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 09/24/2025
Supervised By :mohammad ahmed 09/25/2025



#30 AR-1254-5

R.T.: 6.870 min
Delta R.T.: 0.000 min
Response: 144577806
Conc: 240.24 ng/ml

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP092325\
 Data File : PP075268.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 23 Sep 2025 15:20
 Operator : YP\AJ
 Sample : AR1254ICC050
 Misc :
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1254ICC050

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 09/24/2025
 Supervised By :mohammad ahmed 09/25/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 23 15:51:51 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP092325.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Sep 23 15:51:36 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.655	3.799	6696765	22881590	5.113	3.918
2) SA Decachlor...	10.433	8.807	4781103	27052038	4.845	4.421
Target Compounds						
26) L6 AR-1254-1	6.655	5.680	3396225	20102889	52.688m	40.302
27) L6 AR-1254-2	6.872	5.827	4745100	21129184	55.218	52.931
28) L6 AR-1254-3	7.234	6.228	5250883	25236892	56.814	37.966 #
29) L6 AR-1254-4	7.516	6.453	3901845	19340151	54.687	39.178 #
30) L6 AR-1254-5	7.932	6.871	3898066	22310596	47.649	39.094

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP092325\
Data File : PP075268.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Sep 2025 15:20
Operator : YP\AJ
Sample : AR1254ICC050
Misc :
ALS Vial : 24 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
AR1254ICC050


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 09/24/2025
Supervised By :mohammad ahmed 09/25/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 23 15:51:51 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP092325.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Sep 23 15:51:36 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.655 min
Delta R.T.: 0.000 min
Response: 6696765
Conc: 5.11 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254ICC050


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 09/24/2025
Supervised By :mohammad ahmed 09/25/2025




#1 Tetrachloro-m-xylene

R.T.: 3.799 min
Delta R.T.: 0.000 min
Response: 22881590
Conc: 3.92 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.433 min
Delta R.T.: 0.000 min
Response: 4781103
Conc: 4.84 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.807 min
Delta R.T.: 0.000 min
Response: 27052038
Conc: 4.42 ng/ml




#26 AR-1254-1

R.T.: 6.655 min
Delta R.T.: 0.000 min
Response: 3396225
Conc: 52.69 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254ICC050


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 09/24/2025
Supervised By :mohammad ahmed 09/25/2025




#26 AR-1254-1

R.T.: 5.680 min
Delta R.T.: 0.000 min
Response: 20102889
Conc: 40.30 ng/ml




#27 AR-1254-2

R.T.: 6.872 min
Delta R.T.: 0.000 min
Response: 4745100
Conc: 55.22 ng/ml



#27 AR-1254-2

R.T.: 5.827 min
Delta R.T.: 0.000 min
Response: 21129184
Conc: 52.93 ng/ml




#28 AR-1254-3

R.T.: 7.234 min
Delta R.T.: 0.000 min
Response: 5250883
Conc: 56.81 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254ICC050


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 09/24/2025
Supervised By :mohammad ahmed 09/25/2025




#28 AR-1254-3

R.T.: 6.228 min
Delta R.T.: 0.000 min
Response: 25236892
Conc: 37.97 ng/ml




#29 AR-1254-4

R.T.: 7.516 min
Delta R.T.: 0.000 min
Response: 3901845
Conc: 54.69 ng/ml



#29 AR-1254-4

R.T.: 6.453 min
Delta R.T.: 0.000 min
Response: 19340151
Conc: 39.18 ng/ml



#30 AR-1254-5

R.T.: 7.932 min
Delta R.T.: 0.000 min
Response: 3898066
Conc: 47.65 ng/ml

Instrument :

ECD_P

ClientSampleId :


AR1254ICC050

Manual Integrations

APPROVED

Reviewed By :Yogesh Patel 09/24/2025

Supervised By :mohammad ahmed 09/25/2025



#30 AR-1254-5

R.T.: 6.871 min
Delta R.T.: 0.000 min
Response: 22310596
Conc: 39.09 ng/ml

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP092325\
Data File : PP075269.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Sep 2025 15:52
Operator : YP\AJ
Sample : AR1262ICC500
Misc :
ALS Vial : 25 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1262ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 23 16:11:22 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP092325.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Sep 23 16:09:50 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.659	3.800	64525262	296.7E6	50.000	50.000
2) SA Decachlor...	10.443	8.807	49716283	321.5E6	50.000	50.000
Target Compounds						
36) L8 AR-1262-1	8.243	6.909	37977598	395.6E6	500.000	500.000
37) L8 AR-1262-2	8.569	7.168	67680488	276.2E6	500.000	500.000
38) L8 AR-1262-3	8.895	7.690	46725485	250.0E6	500.000	500.000
39) L8 AR-1262-4	8.984	7.753	36118617	454.0E6	500.000	500.000
40) L8 AR-1262-5	9.656	8.247	26067736	215.0E6	500.000	500.000



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP092325\
Data File : PP075269.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Sep 2025 15:52
Operator : YP\AJ
Sample : AR1262ICC500
Misc :
ALS Vial : 25 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1262ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 23 16:11:22 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP092325.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Sep 23 16:09:50 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.659 min
Delta R.T.: 0.000 min
Response: 64525262
Conc: 50.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1262ICC500




#1 Tetrachloro-m-xylene

R.T.: 3.800 min
Delta R.T.: 0.000 min
Response: 296718295
Conc: 50.00 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.443 min
Delta R.T.: 0.000 min
Response: 49716283
Conc: 50.00 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.807 min
Delta R.T.: 0.000 min
Response: 321471739
Conc: 50.00 ng/ml



#36 AR-1262-1


R.T.: 8.243 min
Delta R.T.: 0.000 min
Response: 37977598
Conc: 500.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1262ICC500




#36 AR-1262-1

R.T.: 6.909 min
Delta R.T.: 0.000 min
Response: 395558471
Conc: 500.00 ng/ml



#37 AR-1262-2

R.T.: 8.569 min
Delta R.T.: 0.000 min
Response: 67680488
Conc: 500.00 ng/ml



#37 AR-1262-2


R.T.: 7.168 min
Delta R.T.: 0.000 min
Response: 276184209
Conc: 500.00 ng/ml

Response_ Signal: PP075269.D\ECD1A.ch

#38 AR-1262-3

R.T.: 8.895 min
Delta R.T.: 0.000 min
Response: 46725485
Conc: 500.00 ng/ml


Instrument :
ECD_P
ClientSampleId :
AR1262ICC500



Time Response_ Signal: PP075269.D\ECD2B.ch

#38 AR-1262-3


R.T.: 7.690 min
Delta R.T.: 0.000 min
Response: 249986782
Conc: 500.00 ng/ml



Time Response_ Signal: PP075269.D\ECD1A.ch

#39 AR-1262-4


R.T.: 8.984 min
Delta R.T.: 0.000 min
Response: 36118617
Conc: 500.00 ng/ml




Time Response_ Signal: PP075269.D\ECD2B.ch

#39 AR-1262-4

R.T.: 7.753 min
Delta R.T.: 0.000 min
Response: 454041768
Conc: 500.00 ng/ml






#40 AR-1262-5

R.T.: 9.656 min
Delta R.T.: 0.000 min
Response: 26067736
Conc: 500.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1262ICC500



#40 AR-1262-5

R.T.: 8.247 min
Delta R.T.: 0.000 min
Response: 214953059
Conc: 500.00 ng/ml

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP092325\
 Data File : PP075270.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 23 Sep 2025 16:09
 Operator : YP\AJ
 Sample : AR1268ICC1000
 Misc :
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1268ICC1000

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 23 17:03:25 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP092325.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Sep 23 16:59:51 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.659	3.799	127.9E6	656.7E6	95.826	99.444
2) SA Decachlor...	10.442	8.807	165.7E6	1209.5E6	94.811	99.431
Target Compounds						
41) L9 AR-1268-1	8.890	7.690	152.9E6	1460.2E6	949.456	967.154
42) L9 AR-1268-2	8.986	7.754	141.1E6	1471.7E6	950.525	983.689
43) L9 AR-1268-3	9.226	7.961	118.9E6	1142.5E6	951.307	977.313
44) L9 AR-1268-4	9.655	8.247	55874289	470.5E6	948.208	987.003
45) L9 AR-1268-5	10.086	8.545	373.7E6	3043.3E6	957.165	962.827



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP092325\
Data File : PP075270.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Sep 2025 16:09
Operator : YP\AJ
Sample : AR1268ICC1000
Misc :
ALS Vial : 26 Sample Multiplier: 1


Instrument :
ECD_P
ClientSampleId :
AR1268ICC1000

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 23 17:03:25 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP092325.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Sep 23 16:59:51 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Response_ Signal: PP075270.D\ECD1A.ch




#1 Tetrachloro-m-xylene

R.T.: 4.659 min
Delta R.T.: 0.000 min
Response: 127913745
Conc: 95.83 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1268ICC1000


Time Response_ Signal: PP075270.D\ECD2B.ch



#1 Tetrachloro-m-xylene

R.T.: 3.799 min
Delta R.T.: 0.000 min
Response: 656732962
Conc: 99.44 ng/ml


Time Response_ Signal: PP075270.D\ECD1A.ch



#2 Decachlorobiphenyl


R.T.: 10.442 min
Delta R.T.: 0.000 min
Response: 165740658
Conc: 94.81 ng/ml

Time Response_ Signal: PP075270.D\ECD2B.ch



#2 Decachlorobiphenyl


R.T.: 8.807 min
Delta R.T.: 0.000 min
Response: 1209469094
Conc: 99.43 ng/ml



#41 AR-1268-1


R.T.: 8.890 min
Delta R.T.: 0.000 min
Response: 152883178
Conc: 949.46 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1268ICC1000




#41 AR-1268-1

R.T.: 7.690 min
Delta R.T.: 0.000 min
Response: 1460168699
Conc: 967.15 ng/ml



#42 AR-1268-2

R.T.: 8.986 min
Delta R.T.: 0.000 min
Response: 141133811
Conc: 950.53 ng/ml



#42 AR-1268-2

R.T.: 7.754 min
Delta R.T.: 0.000 min
Response: 1471677824
Conc: 983.69 ng/ml

Response_ Signal: PP075270.D\ECD1A.ch

#43 AR-1268-3

R.T.: 9.226 min
Delta R.T.: 0.000 min
Response: 118887581
Conc: 951.31 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1268ICC1000

Time

Response_ Signal: PP075270.D\ECD2B.ch

#43 AR-1268-3

R.T.: 7.961 min
Delta R.T.: 0.000 min
Response: 1142468076
Conc: 977.31 ng/ml

Time

Response_ Signal: PP075270.D\ECD1A.ch

#44 AR-1268-4

R.T.: 9.655 min
Delta R.T.: 0.000 min
Response: 55874289
Conc: 948.21 ng/ml


Time

Response_ Signal: PP075270.D\ECD2B.ch

#44 AR-1268-4

R.T.: 8.247 min
Delta R.T.: 0.000 min
Response: 470517486
Conc: 987.00 ng/ml


Time



#45 AR-1268-5

R.T.: 10.086 min
Delta R.T.: 0.000 min
Response: 373677572
Conc: 957.16 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1268ICC1000



#45 AR-1268-5

R.T.: 8.545 min
Delta R.T.: 0.000 min
Response: 3043336629
Conc: 962.83 ng/ml

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP092325\
Data File : PP075271.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Sep 2025 16:25
Operator : YP\AJ
Sample : AR1268ICC750
Misc :
ALS Vial : 27 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1268ICC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 23 17:05:59 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP092325.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Sep 23 16:59:51 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.660	3.799	100.1E6	497.2E6	74.976	75.187
2) SA Decachlor...	10.442	8.807	130.2E6	914.1E6	74.643	75.099
Target Compounds						
41) L9 AR-1268-1	8.892	7.690	119.8E6	1135.3E6	745.861	751.304
42) L9 AR-1268-2	8.988	7.754	110.2E6	1125.2E6	744.931	751.391
43) L9 AR-1268-3	9.228	7.962	92608920	865.0E6	743.998	743.275
44) L9 AR-1268-4	9.656	8.248	43848213	354.7E6	746.070	746.016
45) L9 AR-1268-5	10.088	8.546	291.3E6	2388.6E6	747.392	753.776



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP092325\
Data File : PP075271.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Sep 2025 16:25
Operator : YP\AJ
Sample : AR1268ICC750
Misc :
ALS Vial : 27 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1268ICC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 23 17:05:59 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP092325.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Sep 23 16:59:51 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm




Response_ Signal: PP075271.D\ECD1A.ch

#1 Tetrachloro-m-xylene

R.T.: 4.660 min
Delta R.T.: 0.000 min
Response: 100066687
Conc: 74.98 ng/ml


Instrument :
ECD_P
ClientSampleId :
AR1268ICC750



Time Response_ Signal: PP075271.D\ECD2B.ch

#1 Tetrachloro-m-xylene


R.T.: 3.799 min
Delta R.T.: 0.000 min
Response: 497158030
Conc: 75.19 ng/ml



Time Response_ Signal: PP075271.D\ECD1A.ch

#2 Decachlorobiphenyl


R.T.: 10.442 min
Delta R.T.: 0.000 min
Response: 130175685
Conc: 74.64 ng/ml



Time Response_ Signal: PP075271.D\ECD2B.ch

#2 Decachlorobiphenyl

R.T.: 8.807 min
Delta R.T.: 0.000 min
Response: 914090698
Conc: 75.10 ng/ml




Response_ Signal: PP075271.D\ECD1A.ch

#41 AR-1268-1

R.T.: 8.892 min
Delta R.T.: 0.000 min
Response: 119769528
Conc: 745.86 ng/ml


Instrument :
ECD_P
ClientSampleId :
AR1268ICC750



Time Response_ Signal: PP075271.D\ECD2B.ch

#41 AR-1268-1


R.T.: 7.690 min
Delta R.T.: 0.000 min
Response: 1135274856
Conc: 751.30 ng/ml



Time Response_ Signal: PP075271.D\ECD1A.ch

#42 AR-1268-2


R.T.: 8.988 min
Delta R.T.: 0.000 min
Response: 110234639
Conc: 744.93 ng/ml



Time Response_ Signal: PP075271.D\ECD2B.ch

#42 AR-1268-2

R.T.: 7.754 min
Delta R.T.: 0.000 min
Response: 1125183792
Conc: 751.39 ng/ml




Response_ Signal: PP075271.D\ECD1A.ch

#43 AR-1268-3

R.T.: 9.228 min
Delta R.T.: 0.000 min
Response: 92608920
Conc: 744.00 ng/ml


Instrument :
ECD_P
ClientSampleId :
AR1268ICC750



Response_ Signal: PP075271.D\ECD2B.ch

#43 AR-1268-3


R.T.: 7.962 min
Delta R.T.: 0.000 min
Response: 865001509
Conc: 743.27 ng/ml



Response_ Signal: PP075271.D\ECD1A.ch

#44 AR-1268-4


R.T.: 9.656 min
Delta R.T.: 0.000 min
Response: 43848213
Conc: 746.07 ng/ml



Response_ Signal: PP075271.D\ECD2B.ch

#44 AR-1268-4

R.T.: 8.248 min
Delta R.T.: 0.000 min
Response: 354693309
Conc: 746.02 ng/ml




Response_

Signal: PP075271.D\ECD1A.ch

#45 AR-1268-5

ICAL Form



R.T.: 10.088 min
Delta R.T.: 0.000 min
Response: 291275954
Conc: 747.39 ng/ml


Instrument :
ECD_P
ClientSampleId :
AR1268ICC750

Time

Response_

Signal: PP075271.D\ECD2B.ch

#45 AR-1268-5



R.T.: 8.546 min
Delta R.T.: 0.000 min
Response: 2388574998
Conc: 753.78 ng/ml

Time

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP092325\
Data File : PP075272.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Sep 2025 16:41
Operator : YP\AJ
Sample : AR1268ICC500
Misc :
ALS Vial : 28 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1268ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 23 17:00:09 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP092325.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Sep 23 16:59:51 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.660	3.799	69528834	332.0E6	50.000	50.000
2) SA Decachlor...	10.443	8.806	91941579	611.7E6	50.000	50.000
Target Compounds						
41) L9 AR-1268-1	8.891	7.689	84580336	779.7E6	500.000	500.000
42) L9 AR-1268-2	8.987	7.753	77912875	760.2E6	500.000	500.000
43) L9 AR-1268-3	9.228	7.960	65529058	597.8E6	500.000	500.000
44) L9 AR-1268-4	9.655	8.246	30989057	241.5E6	500.000	500.000
45) L9 AR-1268-5	10.087	8.545	203.6E6	1639.2E6	500.000	500.000



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP092325\
Data File : PP075272.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Sep 2025 16:41
Operator : YP\AJ
Sample : AR1268ICC500
Misc :
ALS Vial : 28 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1268ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 23 17:00:09 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP092325.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Sep 23 16:59:51 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm




Response_ Signal: PP075272.D\ECD1A.ch

#1 Tetrachloro-m-xylene

R.T.: 4.660 min
Delta R.T.: 0.000 min
Response: 69528834
Conc: 50.00 ng/ml


Instrument :
ECD_P
ClientSampleId :
AR1268ICC500



Time Response_ Signal: PP075272.D\ECD2B.ch

#1 Tetrachloro-m-xylene


R.T.: 3.799 min
Delta R.T.: 0.000 min
Response: 332036100
Conc: 50.00 ng/ml



Time Response_ Signal: PP075272.D\ECD1A.ch

#2 Decachlorobiphenyl


R.T.: 10.443 min
Delta R.T.: 0.000 min
Response: 91941579
Conc: 50.00 ng/ml




Time Response_ Signal: PP075272.D\ECD2B.ch

#2 Decachlorobiphenyl

R.T.: 8.806 min
Delta R.T.: 0.000 min
Response: 611654619
Conc: 50.00 ng/ml






#41 AR-1268-1


R.T.: 8.891 min
Delta R.T.: 0.000 min
Response: 84580336
Conc: 500.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1268ICC500




#41 AR-1268-1

R.T.: 7.689 min
Delta R.T.: 0.000 min
Response: 779674117
Conc: 500.00 ng/ml



#42 AR-1268-2

R.T.: 8.987 min
Delta R.T.: 0.000 min
Response: 77912875
Conc: 500.00 ng/ml



#42 AR-1268-2

R.T.: 7.753 min
Delta R.T.: 0.000 min
Response: 760240815
Conc: 500.00 ng/ml

Response_ Signal: PP075272.D\ECD1A.ch

#43 AR-1268-3

R.T.: 9.228 min
Delta R.T.: 0.000 min
Response: 65529058
Conc: 500.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1268ICC500

Time

Response_ Signal: PP075272.D\ECD2B.ch

#43 AR-1268-3

R.T.: 7.960 min
Delta R.T.: 0.000 min
Response: 597755130
Conc: 500.00 ng/ml

Time

Response_ Signal: PP075272.D\ECD1A.ch

#44 AR-1268-4

R.T.: 9.655 min
Delta R.T.: 0.000 min
Response: 30989057
Conc: 500.00 ng/ml

Time

Response_ Signal: PP075272.D\ECD2B.ch

#44 AR-1268-4

R.T.: 8.246 min
Delta R.T.: 0.000 min
Response: 241454470
Conc: 500.00 ng/ml


Time

Response_

Signal: PP075272.D\ECD1A.ch

#45 AR-1268-5

ICAL Form



R.T.: 10.087 min
Delta R.T.: 0.000 min
Response: 203561599
Conc: 500.00 ng/ml


Instrument :
ECD_P
ClientSampleId :
AR1268ICC500

Time

Response_

Signal: PP075272.D\ECD2B.ch

#45 AR-1268-5



R.T.: 8.545 min
Delta R.T.: 0.000 min
Response: 1639165491
Conc: 500.00 ng/ml

Time

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP092325\
Data File : PP075273.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Sep 2025 16:58
Operator : YP\AJ
Sample : AR1268ICC250
Misc :
ALS Vial : 29 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1268ICC250

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 23 17:12:32 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP092325.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Sep 23 17:12:18 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.658	3.796	34379611	151.4E6	25.565	23.393
2) SA Decachlor...	10.439	8.804	46278136	303.1E6	26.135	24.927
Target Compounds						
41) L9 AR-1268-1	8.889	7.687	42980575	365.2E6	263.015	243.698
42) L9 AR-1268-2	8.984	7.751	39128437	345.6E6	260.660	235.330
43) L9 AR-1268-3	9.225	7.958	32512368	273.5E6	258.304	238.576
44) L9 AR-1268-4	9.653	8.245	15247313	115.2E6	257.007	244.227
45) L9 AR-1268-5	10.085	8.542	101.3E6	821.3E6	257.365	256.819



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP092325\
Data File : PP075273.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Sep 2025 16:58
Operator : YP\AJ
Sample : AR1268ICC250
Misc :
ALS Vial : 29 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1268ICC250

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 23 17:12:32 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP092325.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Sep 23 17:12:18 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.658 min
Delta R.T.: 0.000 min
Response: 34379611
Conc: 25.57 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1268ICC250




#1 Tetrachloro-m-xylene

R.T.: 3.796 min
Delta R.T.: 0.000 min
Response: 151432939
Conc: 23.39 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.439 min
Delta R.T.: 0.000 min
Response: 46278136
Conc: 26.13 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.804 min
Delta R.T.: 0.000 min
Response: 303111015
Conc: 24.93 ng/ml



#41 AR-1268-1


R.T.: 8.889 min
Delta R.T.: 0.000 min
Response: 42980575
Conc: 263.02 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1268ICC250




#41 AR-1268-1

R.T.: 7.687 min
Delta R.T.: 0.000 min
Response: 365177329
Conc: 243.70 ng/ml




#42 AR-1268-2

R.T.: 8.984 min
Delta R.T.: 0.000 min
Response: 39128437
Conc: 260.66 ng/ml



#42 AR-1268-2


R.T.: 7.751 min
Delta R.T.: 0.000 min
Response: 345638504
Conc: 235.33 ng/ml



#43 AR-1268-3


R.T.: 9.225 min
Delta R.T.: 0.000 min
Response: 32512368
Conc: 258.30 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1268ICC250




#43 AR-1268-3

R.T.: 7.958 min
Delta R.T.: 0.000 min
Response: 273482322
Conc: 238.58 ng/ml




#44 AR-1268-4

R.T.: 9.653 min
Delta R.T.: 0.000 min
Response: 15247313
Conc: 257.01 ng/ml



#44 AR-1268-4


R.T.: 8.245 min
Delta R.T.: 0.000 min
Response: 115230714
Conc: 244.23 ng/ml



#45 AR-1268-5

R.T.: 10.085 min
Delta R.T.: 0.000 min
Response: 101295864
Conc: 257.37 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1268ICC250



#45 AR-1268-5

R.T.: 8.542 min
Delta R.T.: 0.000 min
Response: 821277670
Conc: 256.82 ng/ml

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP092325\
 Data File : PP075274.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 23 Sep 2025 17:14
 Operator : YP\AJ
 Sample : AR1268ICC050
 Misc :
 ALS Vial : 30 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1268ICC050

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 09/24/2025
 Supervised By :mohammad ahmed 09/25/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 23 17:35:52 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP092325.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Sep 23 17:35:25 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.656	3.796	6162995	20516394	4.661	3.375m#
2) SA Decachlor...	10.438	8.806	8054661	48956522	4.632	4.189
Target Compounds						
41) L9 AR-1268-1	8.887	7.689	7806652	60962197	48.257m	42.258
42) L9 AR-1268-2	8.984	7.753	7067403	54955257	47.637	39.400
43) L9 AR-1268-3	9.224	7.960	5848482	43468397	47.131	39.653m
44) L9 AR-1268-4	9.653	8.247	3041493	18125694	51.009	40.754m
45) L9 AR-1268-5	10.084	8.544	18331022	120.0E6	47.221	39.458m

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP092325\
Data File : PP075274.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Sep 2025 17:14
Operator : YP\AJ
Sample : AR1268ICC050
Misc :
ALS Vial : 30 Sample Multiplier: 1


Instrument :
ECD_P
ClientSampleId :
AR1268ICC050


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 09/24/2025
Supervised By :mohammad ahmed 09/25/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 23 17:35:52 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP092325.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Sep 23 17:35:25 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.656 min
Delta R.T.: 0.000 min
Response: 6162995
Conc: 4.66 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1268ICC050


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 09/24/2025
Supervised By :mohammad ahmed 09/25/2025




#1 Tetrachloro-m-xylene

R.T.: 3.796 min
Delta R.T.: -0.001 min
Response: 20516394
Conc: 3.38 ng/ml m




#2 Decachlorobiphenyl

R.T.: 10.438 min
Delta R.T.: 0.000 min
Response: 8054661
Conc: 4.63 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.806 min
Delta R.T.: 0.000 min
Response: 48956522
Conc: 4.19 ng/ml




#41 AR-1268-1

R.T.: 8.887 min
Delta R.T.: 0.000 min
Response: 7806652
Conc: 48.26 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1268ICC050


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 09/24/2025
Supervised By :mohammad ahmed 09/25/2025




#41 AR-1268-1

R.T.: 7.689 min
Delta R.T.: 0.000 min
Response: 60962197
Conc: 42.26 ng/ml




#42 AR-1268-2

R.T.: 8.984 min
Delta R.T.: 0.000 min
Response: 7067403
Conc: 47.64 ng/ml



#42 AR-1268-2

R.T.: 7.753 min
Delta R.T.: 0.000 min
Response: 54955257
Conc: 39.40 ng/ml




#43 AR-1268-3

R.T.: 9.224 min
Delta R.T.: 0.000 min
Response: 5848482
Conc: 47.13 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1268ICC050


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 09/24/2025
Supervised By :mohammad ahmed 09/25/2025




#43 AR-1268-3

R.T.: 7.960 min
Delta R.T.: 0.000 min
Response: 43468397
Conc: 39.65 ng/ml m




#44 AR-1268-4

R.T.: 9.653 min
Delta R.T.: 0.000 min
Response: 3041493
Conc: 51.01 ng/ml



#44 AR-1268-4

R.T.: 8.247 min
Delta R.T.: 0.000 min
Response: 18125694
Conc: 40.75 ng/ml m




#45 AR-1268-5

R.T.: 10.084 min
Delta R.T.: 0.000 min
Response: 18331022
Conc: 47.22 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1268ICC050

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 09/24/2025
Supervised By :mohammad ahmed 09/25/2025



#45 AR-1268-5

R.T.: 8.544 min
Delta R.T.: 0.000 min
Response: 119998424
Conc: 39.46 ng/ml m

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP092325\
 Data File : PP075275.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 23 Sep 2025 17:46
 Operator : YP\AJ
 Sample : PP092325ICV500
 Misc :
 ALS Vial : 31 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 ICVPP092325

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 09/24/2025
 Supervised By :mohammad ahmed 09/25/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 23 18:28:29 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP092325.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Sep 23 17:43:53 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.663	3.799	70190259	330.4E6	53.643	54.154
2) SA Decachlor...	10.445	8.804	53037838	358.3E6	53.168	59.392m
Target Compounds						
3) L1 AR-1016-1	5.815	4.896	25480532	304.2E6	529.688	542.185
4) L1 AR-1016-2	5.836	4.953	36891066	135.1E6	514.315	525.487
5) L1 AR-1016-3	5.899	5.073	23781479	76410296	511.091	492.118
6) L1 AR-1016-4	5.996	5.114	19445011	76331861	545.646	519.377
7) L1 AR-1016-5	6.289	5.328	18500980	85017958	548.644	520.362
31) L7 AR-1260-1	7.406	6.357	33314276	213.7E6	554.134	526.554
32) L7 AR-1260-2	7.659	6.543	39216271	298.7E6	538.700	518.643
33) L7 AR-1260-3	8.017	6.697	31022586	223.0E6	572.952	520.892
34) L7 AR-1260-4	8.245	7.166	33773955	216.2E6	597.376	501.440
35) L7 AR-1260-5	8.571	7.405	61446307	573.0E6	559.578	544.665

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP092325\
Data File : PP075275.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Sep 2025 17:46
Operator : YP\AJ
Sample : PP092325ICV500
Misc :
ALS Vial : 31 Sample Multiplier: 1

Instrument :

ECD_P

ClientSampleId :

ICVPP092325

Manual Integrations



APPROVED


Reviewed By :Yogesh Patel 09/24/2025

Supervised By :mohammad ahmed 09/25/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 23 18:28:29 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP092325.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Sep 23 17:43:53 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.663 min
Delta R.T.: 0.008 min
Response: 70190259
Conc: 53.64 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP092325


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 09/24/2025
Supervised By :mohammad ahmed 09/25/2025




#1 Tetrachloro-m-xylene

R.T.: 3.799 min
Delta R.T.: 0.003 min
Response: 330389764
Conc: 54.15 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.445 min
Delta R.T.: 0.010 min
Response: 53037838
Conc: 53.17 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.804 min
Delta R.T.: 0.000 min
Response: 358284034
Conc: 59.39 ng/ml m

Response_ Signal: PP075275.D\ECD1A.ch

#3 AR-1016-1

R.T.: 5.815 min
Delta R.T.: 0.007 min
Response: 25480532
Conc: 529.69 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP092325

Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 09/24/2025
Supervised By :mohammad ahmed 09/25/2025

Time

Response_ Signal: PP075275.D\ECD2B.ch

#3 AR-1016-1


R.T.: 4.896 min
Delta R.T.: 0.003 min
Response: 304173227
Conc: 542.19 ng/ml



Response_ Signal: PP075275.D\ECD1A.ch

#4 AR-1016-2


R.T.: 5.836 min
Delta R.T.: 0.008 min
Response: 36891066
Conc: 514.31 ng/ml




Response_ Signal: PP075275.D\ECD2B.ch

#4 AR-1016-2

R.T.: 4.953 min
Delta R.T.: 0.002 min
Response: 135051180
Conc: 525.49 ng/ml






#5 AR-1016-3

R.T.: 5.899 min
Delta R.T.: 0.007 min
Response: 23781479
Conc: 511.09 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP092325


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 09/24/2025
Supervised By :mohammad ahmed 09/25/2025




#5 AR-1016-3

R.T.: 5.073 min
Delta R.T.: 0.002 min
Response: 76410296
Conc: 492.12 ng/ml




#6 AR-1016-4

R.T.: 5.996 min
Delta R.T.: 0.007 min
Response: 19445011
Conc: 545.65 ng/ml



#6 AR-1016-4

R.T.: 5.114 min
Delta R.T.: 0.003 min
Response: 76331861
Conc: 519.38 ng/ml




#7 AR-1016-5

R.T.: 6.289 min
Delta R.T.: 0.007 min
Response: 18500980
Conc: 548.64 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP092325


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 09/24/2025
Supervised By :mohammad ahmed 09/25/2025




#7 AR-1016-5

R.T.: 5.328 min
Delta R.T.: 0.003 min
Response: 85017958
Conc: 520.36 ng/ml



#31 AR-1260-1


R.T.: 7.406 min
Delta R.T.: 0.007 min
Response: 33314276
Conc: 554.13 ng/ml



#31 AR-1260-1

R.T.: 6.357 min
Delta R.T.: 0.002 min
Response: 213746874
Conc: 526.55 ng/ml

Response_ Signal: PP075275.D\ECD1A.ch



#32 AR-1260-2


R.T.: 7.659 min
Delta R.T.: 0.007 min
Response: 39216271
Conc: 538.70 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP092325

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 09/24/2025
Supervised By :mohammad ahmed 09/25/2025


Time Response_ Signal: PP075275.D\ECD2B.ch



#32 AR-1260-2

R.T.: 6.543 min
Delta R.T.: 0.002 min
Response: 298650131
Conc: 518.64 ng/ml


Response_ Signal: PP075275.D\ECD1A.ch



#33 AR-1260-3

R.T.: 8.017 min
Delta R.T.: 0.007 min
Response: 31022586
Conc: 572.95 ng/ml


Time Response_ Signal: PP075275.D\ECD2B.ch



#33 AR-1260-3

R.T.: 6.697 min
Delta R.T.: 0.002 min
Response: 222985953
Conc: 520.89 ng/ml

Response_ Signal: PP075275.D\ECD1A.ch



#34 AR-1260-4


R.T.: 8.245 min
Delta R.T.: 0.008 min
Response: 33773955
Conc: 597.38 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP092325

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 09/24/2025
Supervised By :mohammad ahmed 09/25/2025


Time Response_ Signal: PP075275.D\ECD2B.ch



#34 AR-1260-4

R.T.: 7.166 min
Delta R.T.: 0.000 min
Response: 216232762
Conc: 501.44 ng/ml


Time Response_ Signal: PP075275.D\ECD1A.ch



#35 AR-1260-5

R.T.: 8.571 min
Delta R.T.: 0.008 min
Response: 61446307
Conc: 559.58 ng/ml

Time Response_ Signal: PP075275.D\ECD2B.ch



#35 AR-1260-5

R.T.: 7.405 min
Delta R.T.: 0.002 min
Response: 572980909
Conc: 544.66 ng/ml

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP092325\
Data File : PP075276.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Sep 2025 18:35
Operator : YP\AJ
Sample : AR1242ICV500
Misc :
ALS Vial : 12 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
ICVPP092325AR1242

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 09/24/2025
Supervised By :mohammad ahmed 09/25/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 24 04:35:07 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP092325.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Sep 23 17:43:53 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.660	3.799	69296262	322.8E6	52.959	52.905
2) SA Decachlor...	10.440	8.803	52620761	359.1E6	52.750	59.525m
Target Compounds						
16) L4 AR-1242-1	5.812	4.896	21512826	253.0E6	516.394	541.693
17) L4 AR-1242-2	5.834	4.954	31427273	115.0E6	526.973	550.426
18) L4 AR-1242-3	5.897	5.074	20323849	64997022	502.739	532.921
19) L4 AR-1242-4	5.993	5.155	16403438	83515061	521.837	524.950m
20) L4 AR-1242-5	6.723	5.676	18943732	103.0E6	538.853	528.983

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP092325\
Data File : PP075276.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Sep 2025 18:35
Operator : YP\AJ
Sample : AR1242ICV500
Misc :
ALS Vial : 12 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
ICVPP092325AR1242


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 09/24/2025
Supervised By :mohammad ahmed 09/25/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 24 04:35:07 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP092325.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Sep 23 17:43:53 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.660 min
Delta R.T.: 0.005 min
Response: 69296262
Conc: 52.96 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP092325AR1242


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 09/24/2025
Supervised By :mohammad ahmed 09/25/2025




#1 Tetrachloro-m-xylene

R.T.: 3.799 min
Delta R.T.: 0.003 min
Response: 322766613
Conc: 52.90 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.440 min
Delta R.T.: 0.006 min
Response: 52620761
Conc: 52.75 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.803 min
Delta R.T.: 0.000 min
Response: 359083335
Conc: 59.52 ng/ml m




#16 AR-1242-1

R.T.: 5.812 min
Delta R.T.: 0.000 min
Response: 21512826
Conc: 516.39 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP092325AR1242


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 09/24/2025
Supervised By :mohammad ahmed 09/25/2025




#16 AR-1242-1

R.T.: 4.896 min
Delta R.T.: 0.000 min
Response: 253035792
Conc: 541.69 ng/ml




#17 AR-1242-2

R.T.: 5.834 min
Delta R.T.: 0.000 min
Response: 31427273
Conc: 526.97 ng/ml



#17 AR-1242-2

R.T.: 4.954 min
Delta R.T.: 0.000 min
Response: 115015732
Conc: 550.43 ng/ml




#18 AR-1242-3

R.T.: 5.897 min
Delta R.T.: 0.000 min
Response: 20323849
Conc: 502.74 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP092325AR1242


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 09/24/2025
Supervised By :mohammad ahmed 09/25/2025




#18 AR-1242-3

R.T.: 5.074 min
Delta R.T.: 0.000 min
Response: 64997022
Conc: 532.92 ng/ml




#19 AR-1242-4

R.T.: 5.993 min
Delta R.T.: 0.000 min
Response: 16403438
Conc: 521.84 ng/ml



#19 AR-1242-4

R.T.: 5.155 min
Delta R.T.: 0.000 min
Response: 83515061
Conc: 524.95 ng/ml m




#20 AR-1242-5

R.T.: 6.723 min
Delta R.T.: 0.000 min
Response: 18943732
Conc: 538.85 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP092325AR1242

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 09/24/2025
Supervised By :mohammad ahmed 09/25/2025



#20 AR-1242-5

R.T.: 5.676 min
Delta R.T.: 0.000 min
Response: 102984369
Conc: 528.98 ng/ml

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP092325\
 Data File : PP075277.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 23 Sep 2025 19:08
 Operator : YP\AJ
 Sample : AR1248ICV500
 Misc :
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 ICVPP092325AR1248

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 09/24/2025
 Supervised By :mohammad ahmed 09/25/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 24 04:30:59 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP092325.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Sep 23 17:43:53 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.660	3.798	67690061	321.9E6	51.732	52.757
2) SA Decachlor...	10.441	8.804	52715950	361.4E6	52.846	59.916m
Target Compounds						
21) L5 AR-1248-1	5.811	4.894	16390127	167.0E6	519.336	555.702
22) L5 AR-1248-2	6.082	5.114	22503142	105.0E6	518.888	518.581
23) L5 AR-1248-3	6.285	5.154	25155093	135.3E6	526.712	541.550m
24) L5 AR-1248-4	6.683	5.327	30695144	119.4E6	519.272	513.196
25) L5 AR-1248-5	6.723	5.718	29807919	227.8E6	504.369	557.543

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP092325\
Data File : PP075277.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Sep 2025 19:08
Operator : YP\AJ
Sample : AR1248ICV500
Misc :
ALS Vial : 17 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
ICVPP092325AR1248


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 09/24/2025
Supervised By :mohammad ahmed 09/25/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 24 04:30:59 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP092325.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Sep 23 17:43:53 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.660 min
Delta R.T.: 0.005 min
Response: 67690061
Conc: 51.73 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP092325AR1248


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 09/24/2025
Supervised By :mohammad ahmed 09/25/2025




#1 Tetrachloro-m-xylene

R.T.: 3.798 min
Delta R.T.: 0.002 min
Response: 321866586
Conc: 52.76 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.441 min
Delta R.T.: 0.006 min
Response: 52715950
Conc: 52.85 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.804 min
Delta R.T.: 0.000 min
Response: 361445303
Conc: 59.92 ng/ml m




#21 AR-1248-1

R.T.: 5.811 min
Delta R.T.: 0.005 min
Response: 16390127
Conc: 519.34 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP092325AR1248


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 09/24/2025
Supervised By :mohammad ahmed 09/25/2025




#21 AR-1248-1

R.T.: 4.894 min
Delta R.T.: 0.002 min
Response: 166969869
Conc: 555.70 ng/ml



#22 AR-1248-2

R.T.: 6.082 min
Delta R.T.: 0.004 min
Response: 22503142
Conc: 518.89 ng/ml



#22 AR-1248-2

R.T.: 5.114 min
Delta R.T.: 0.001 min
Response: 105001722
Conc: 518.58 ng/ml

Response_ Signal: PP075277.D\ECD1A.ch

#23 AR-1248-3

R.T.: 6.285 min
Delta R.T.: 0.004 min
Response: 25155093
Conc: 526.71 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP092325AR1248

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 09/24/2025
Supervised By :mohammad ahmed 09/25/2025

Time

Response_ Signal: PP075277.D\ECD2B.ch

#23 AR-1248-3

R.T.: 5.154 min
Delta R.T.: 0.000 min
Response: 135250085
Conc: 541.55 ng/ml m

Time

Response_ Signal: PP075277.D\ECD1A.ch

#24 AR-1248-4

R.T.: 6.683 min
Delta R.T.: 0.005 min
Response: 30695144
Conc: 519.27 ng/ml

Time

Response_ Signal: PP075277.D\ECD2B.ch

#24 AR-1248-4

R.T.: 5.327 min
Delta R.T.: 0.001 min
Response: 119442983
Conc: 513.20 ng/ml

Time

Response_ Signal: PP075277.D\ECD1A.ch

#24 AR-1248-4

R.T.: 6.682 min
Delta R.T.: 0.005 min
Response: 30695144
Conc: 519.27 ng/ml

Time

Response_ Signal: PP075277.D\ECD2B.ch

#24 AR-1248-4

R.T.: 5.327 min
Delta R.T.: 0.001 min
Response: 119442983
Conc: 513.20 ng/ml

Time

Response_ Signal: PP075277.D\ECD1A.ch

#24 AR-1248-4

R.T.: 6.682 min
Delta R.T.: 0.005 min
Response: 30695144
Conc: 519.27 ng/ml

Time

Response_ Signal: PP075277.D\ECD2B.ch

#24 AR-1248-4

R.T.: 5.327 min
Delta R.T.: 0.001 min
Response: 119442983
Conc: 513.20 ng/ml

Time

Response_ Signal: PP075277.D\ECD1A.ch

#24 AR-1248-4


R.T.: 6.682 min
Delta R.T.: 0.005 min
Response: 30695144
Conc: 519.27 ng/ml

Time

Response_ Signal: PP075277.D\ECD2B.ch

#24 AR-1248-4

R.T.: 5.327 min
Delta R.T.: 0.001 min
Response: 119442983
Conc: 513.20 ng/ml




#25 AR-1248-5

R.T.: 6.723 min
Delta R.T.: 0.005 min
Response: 29807919
Conc: 504.37 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP092325AR1248

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 09/24/2025
Supervised By :mohammad ahmed 09/25/2025



#25 AR-1248-5

R.T.: 5.718 min
Delta R.T.: 0.002 min
Response: 227795946
Conc: 557.54 ng/ml

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP092325\
Data File : PP075278.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Sep 2025 19:40
Operator : YP\AJ
Sample : AR1254ICV500
Misc :
ALS Vial : 22 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
ICVPP092325AR1254

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 09/24/2025
Supervised By :mohammad ahmed 09/25/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 24 05:12:50 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP092325.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Sep 24 05:07:55 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.660	3.799	68095844	321.3E6	52.042	52.664
2) SA Decachlor...	10.441	8.804	51811352	358.5E6	51.939	59.429m
Target Compounds						
26) L6 AR-1254-1	6.659	5.678	28467743	276.6E6	443.257m	554.578 #
27) L6 AR-1254-2	6.876	5.826	43938129	209.1E6	511.302	523.737
28) L6 AR-1254-3	7.238	6.227	46807513	363.2E6	506.453	546.317
29) L6 AR-1254-4	7.520	6.454	36595686	272.7E6	512.909	552.441
30) L6 AR-1254-5	7.936	6.870	43699326	290.3E6	534.174	508.619

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP092325\
Data File : PP075278.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Sep 2025 19:40
Operator : YP\AJ
Sample : AR1254ICV500
Misc :
ALS Vial : 22 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
ICVPP092325AR1254


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 09/24/2025
Supervised By :mohammad ahmed 09/25/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 24 05:12:50 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP092325.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Sep 24 05:07:55 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.660 min
Delta R.T.: 0.005 min
Response: 68095844
Conc: 52.04 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP092325AR1254


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 09/24/2025
Supervised By :mohammad ahmed 09/25/2025




#1 Tetrachloro-m-xylene

R.T.: 3.799 min
Delta R.T.: 0.003 min
Response: 321296251
Conc: 52.66 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.441 min
Delta R.T.: 0.006 min
Response: 51811352
Conc: 51.94 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.804 min
Delta R.T.: 0.000 min
Response: 358504884
Conc: 59.43 ng/ml m




#26 AR-1254-1

R.T.: 6.659 min
Delta R.T.: 0.000 min
Response: 28467743
Conc: 443.26 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP092325AR1254


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 09/24/2025
Supervised By :mohammad ahmed 09/25/2025




#26 AR-1254-1

R.T.: 5.678 min
Delta R.T.: -0.001 min
Response: 276624835
Conc: 554.58 ng/ml




#27 AR-1254-2

R.T.: 6.876 min
Delta R.T.: 0.000 min
Response: 43938129
Conc: 511.30 ng/ml



#27 AR-1254-2

R.T.: 5.826 min
Delta R.T.: 0.000 min
Response: 209067406
Conc: 523.74 ng/ml




#28 AR-1254-3

R.T.: 7.238 min
Delta R.T.: 0.000 min
Response: 46807513
Conc: 506.45 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP092325AR1254


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 09/24/2025
Supervised By :mohammad ahmed 09/25/2025




#28 AR-1254-3

R.T.: 6.227 min
Delta R.T.: -0.001 min
Response: 363150092
Conc: 546.32 ng/ml




#29 AR-1254-4

R.T.: 7.520 min
Delta R.T.: 0.000 min
Response: 36595686
Conc: 512.91 ng/ml



#29 AR-1254-4

R.T.: 6.454 min
Delta R.T.: 0.000 min
Response: 272711111
Conc: 552.44 ng/ml




#30 AR-1254-5

R.T.: 7.936 min
Delta R.T.: 0.000 min
Response: 43699326
Conc: 534.17 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP092325AR1254

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 09/24/2025
Supervised By :mohammad ahmed 09/25/2025



#30 AR-1254-5

R.T.: 6.870 min
Delta R.T.: 0.000 min
Response: 290265404
Conc: 508.62 ng/ml

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP092325\
 Data File : PP075279.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 23 Sep 2025 19:57
 Operator : YP\AJ
 Sample : AR1268ICV500
 Misc :
 ALS Vial : 35 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 ICVPP092325AR1268

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 09/24/2025
 Supervised By :mohammad ahmed 09/25/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 24 04:49:21 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP092325.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Sep 24 04:47:34 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.659	3.799	71353406	332.9E6	53.960	55.485
2) SA Decachlor...	10.440	8.804	92871605	696.3E6	53.411	59.581m
Target Compounds						
41) L9 AR-1268-1	8.890	7.689	89479292	799.5E6	552.483	554.201
42) L9 AR-1268-2	8.985	7.753	81657272	782.2E6	550.399	560.767
43) L9 AR-1268-3	9.226	7.960	67147736	596.0E6	541.127	546.351
44) L9 AR-1268-4	9.653	8.246	31706205	236.0E6	531.741	524.416
45) L9 AR-1268-5	10.085	8.544	206.8E6	1813.0E6	532.693	596.714

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP092325\
Data File : PP075279.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Sep 2025 19:57
Operator : YP\AJ
Sample : AR1268ICV500
Misc :
ALS Vial : 35 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
ICVPP092325AR1268


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 09/24/2025
Supervised By :mohammad ahmed 09/25/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 24 04:49:21 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP092325.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Sep 24 04:47:34 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.659 min
Delta R.T.: 0.000 min
Response: 71353406
Conc: 53.96 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP092325AR1268


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 09/24/2025
Supervised By :mohammad ahmed 09/25/2025




#1 Tetrachloro-m-xylene

R.T.: 3.799 min
Delta R.T.: 0.000 min
Response: 332881910
Conc: 55.49 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.440 min
Delta R.T.: -0.003 min
Response: 92871605
Conc: 53.41 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.804 min
Delta R.T.: -0.003 min
Response: 696275481
Conc: 59.58 ng/ml m

Response_ Signal: PP075279.D\ECD1A.ch

#41 AR-1268-1

R.T.: 8.890 min
Delta R.T.: -0.001 min
Response: 89479292
Conc: 552.48 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP092325AR1268

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 09/24/2025
Supervised By :mohammad ahmed 09/25/2025

Time

Response_ Signal: PP075279.D\ECD2B.ch

#41 AR-1268-1

R.T.: 7.689 min
Delta R.T.: 0.000 min
Response: 799509513
Conc: 554.20 ng/ml

Time

Response_ Signal: PP075279.D\ECD1A.ch

#42 AR-1268-2

R.T.: 8.985 min
Delta R.T.: -0.002 min
Response: 81657272
Conc: 550.40 ng/ml

Time

Response_ Signal: PP075279.D\ECD2B.ch

#42 AR-1268-2

R.T.: 7.753 min
Delta R.T.: 0.000 min
Response: 782164395
Conc: 560.77 ng/ml

Time

Response_ Signal: PP075279.D\ECD1A.ch

Time

Response_ Signal: PP075279.D\ECD2B.ch

Time

Response_ Signal: PP075279.D\ECD1A.ch

Time

Response_ Signal: PP075279.D\ECD2B.ch

Time

Response_ Signal: PP075279.D\ECD1A.ch

Time

Response_ Signal: PP075279.D\ECD2B.ch

Time

Response_ Signal: PP075279.D\ECD1A.ch

Time

Response_ Signal: PP075279.D\ECD2B.ch

Time

Response_ Signal: PP075279.D\ECD1A.ch

Time

Response_ Signal: PP075279.D\ECD2B.ch

Time

Response_ Signal: PP075279.D\ECD1A.ch

Time

Response_ Signal: PP075279.D\ECD2B.ch

Time

Response_ Signal: PP075279.D\ECD1A.ch

Time

Response_ Signal: PP075279.D\ECD2B.ch

Time

Response_ Signal: PP075279.D\ECD1A.ch

Time

Response_ Signal: PP075279.D\ECD2B.ch

Time

Response_ Signal: PP075279.D\ECD1A.ch

Time

Response_ Signal: PP075279.D\ECD2B.ch

Time

Response_ Signal: PP075279.D\ECD1A.ch

Time

Response_ Signal: PP075279.D\ECD2B.ch

Time

Response_ Signal: PP075279.D\ECD1A.ch

Time

Response_ Signal: PP075279.D\ECD2B.ch

Time

Response_ Signal: PP075279.D\ECD1A.ch

Time

Response_ Signal: PP075279.D\ECD2B.ch

Time

Response_ Signal: PP075279.D\ECD1A.ch

Time

Response_ Signal: PP075279.D\ECD2B.ch

Time

Response_ Signal: PP075279.D\ECD1A.ch

Time

Response_ Signal: PP075279.D\ECD2B.ch

Time

Response_ Signal: PP075279.D\ECD1A.ch

#43 AR-1268-3

R.T.: 9.226 min
Delta R.T.: -0.002 min
Response: 67147736
Conc: 541.13 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP092325AR1268

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 09/24/2025
Supervised By :mohammad ahmed 09/25/2025

Time

Response_ Signal: PP075279.D\ECD2B.ch

#43 AR-1268-3

R.T.: 7.960 min
Delta R.T.: 0.000 min
Response: 596026528
Conc: 546.35 ng/ml

Time

Response_ Signal: PP075279.D\ECD1A.ch

#44 AR-1268-4

R.T.: 9.653 min
Delta R.T.: -0.002 min
Response: 31706205
Conc: 531.74 ng/ml

Time

Response_ Signal: PP075279.D\ECD2B.ch

#44 AR-1268-4

R.T.: 8.246 min
Delta R.T.: 0.000 min
Response: 235964707
Conc: 524.42 ng/ml

Time

Response_ Signal: PP075279.D\ECD1A.ch

#44 AR-1268-4

R.T.: 9.653 min
Delta R.T.: -0.002 min
Response: 31706205
Conc: 531.74 ng/ml

Time

Response_ Signal: PP075279.D\ECD2B.ch

#44 AR-1268-4

R.T.: 8.246 min
Delta R.T.: 0.000 min
Response: 235964707
Conc: 524.42 ng/ml

Time

Response_ Signal: PP075279.D\ECD1A.ch

#44 AR-1268-4

R.T.: 9.653 min
Delta R.T.: -0.002 min
Response: 31706205
Conc: 531.74 ng/ml

Time

Response_ Signal: PP075279.D\ECD2B.ch

#44 AR-1268-4

R.T.: 8.246 min
Delta R.T.: 0.000 min
Response: 235964707
Conc: 524.42 ng/ml

Time

Response_ Signal: PP075279.D\ECD1A.ch

#44 AR-1268-4

R.T.: 9.653 min
Delta R.T.: -0.002 min
Response: 31706205
Conc: 531.74 ng/ml

Time

Response_ Signal: PP075279.D\ECD2B.ch

#44 AR-1268-4

R.T.: 8.246 min
Delta R.T.: 0.000 min
Response: 235964707
Conc: 524.42 ng/ml

Time

Response_ Signal: PP075279.D\ECD1A.ch

#44 AR-1268-4

R.T.: 9.653 min
Delta R.T.: -0.002 min
Response: 31706205
Conc: 531.74 ng/ml

Time

Response_ Signal: PP075279.D\ECD2B.ch

#44 AR-1268-4

R.T.: 8.246 min
Delta R.T.: 0.000 min
Response: 235964707
Conc: 524.42 ng/ml

Time

Response_ Signal: PP075279.D\ECD1A.ch

#44 AR-1268-4

R.T.: 9.653 min
Delta R.T.: -0.002 min
Response: 31706205
Conc: 531.74 ng/ml

Time

Response_ Signal: PP075279.D\ECD2B.ch

#44 AR-1268-4

R.T.: 8.246 min
Delta R.T.: 0.000 min
Response: 235964707
Conc: 524.42 ng/ml

Time

Response_ Signal: PP075279.D\ECD1A.ch

#44 AR-1268-4

R.T.: 9.653 min
Delta R.T.: -0.002 min
Response: 31706205
Conc: 531.74 ng/ml

Time

Response_ Signal: PP075279.D\ECD2B.ch

#44 AR-1268-4

R.T.: 8.246 min
Delta R.T.: 0.000 min
Response: 235964707
Conc: 524.42 ng/ml

Time

Response_ Signal: PP075279.D\ECD1A.ch

#44 AR-1268-4

R.T.: 9.653 min
Delta R.T.: -0.002 min
Response: 31706205
Conc: 531.74 ng/ml

Time

Response_ Signal: PP075279.D\ECD2B.ch

#44 AR-1268-4

R.T.: 8.246 min
Delta R.T.: 0.000 min
Response: 235964707
Conc: 524.42 ng/ml

Time

Response_ Signal: PP075279.D\ECD1A.ch

#44 AR-1268-4

R.T.: 9.653 min
Delta R.T.: -0.002 min
Response: 31706205
Conc: 531.74 ng/ml

Time

Response_ Signal: PP075279.D\ECD2B.ch

#44 AR-1268-4

R.T.: 8.246 min
Delta R.T.: 0.000 min
Response: 235964707
Conc: 524.42 ng/ml

Time

Response_ Signal: PP075279.D\ECD1A.ch

#44 AR-1268-4

R.T.: 9.653 min
Delta R.T.: -0.002 min
Response: 31706205
Conc: 531.74 ng/ml

Time

Response_ Signal: PP075279.D\ECD2B.ch

#44 AR-1268-4

R.T.: 8.246 min
Delta R.T.: 0.000 min
Response: 235964707
Conc: 524.42 ng/ml

Time

Response_ Signal: PP075279.D\ECD1A.ch

#44 AR-1268-4

R.T.: 9.653 min
Delta R.T.: -0.002 min
Response: 31706205
Conc: 531.74 ng/ml

Time

Response_ Signal: PP075279.D\ECD2B.ch

#44 AR-1268-4

R.T.: 8.246 min
Delta R.T.: 0.000 min
Response: 235964707
Conc: 524.42 ng/ml

Time

Response_ Signal: PP075279.D\ECD1A.ch


#44 AR-1268-4

R.T.: 9.653 min
Delta R.T.: -0.002 min
Response: 31706205
Conc: 531.74 ng/ml

Time

Response_ Signal: PP075279.D\ECD2B.ch

#44 AR-1268-4



R.T.: 10.085 min
Delta R.T.: -0.002 min
Response: 206788671
Conc: 532.69 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP092325AR1268

Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 09/24/2025
Supervised By :mohammad ahmed 09/25/2025

Time

Response

Signal: PP075279.D\ECD2B.ch

#45 AR-1268-5



R.T.: 8.544 min
Delta R.T.: 0.000 min
Response: 1812996598
Conc: 596.71 ng/ml

Time



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: ROYF02

Lab Code: ACE

SDG NO.: Q3179

Continuing Calib Date: 09/29/2025

Initial Calibration Date(s): 09/23/2025

09/23/2025

Continuing Calib Time: 18:38

Initial Calibration Time(s): 08:45

17:14

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1016-1	(1)	5.81	5.81	5.71	5.91	0.00
Aroclor-1016-2	(2)	5.83	5.83	5.73	5.93	0.00
Aroclor-1016-3	(3)	5.89	5.89	5.79	5.99	0.00
Aroclor-1016-4	(4)	5.99	5.99	5.89	6.09	0.00
Aroclor-1016-5	(5)	6.28	6.28	6.18	6.38	0.00
Aroclor-1260-1	(1)	7.40	7.40	7.30	7.50	0.00
Aroclor-1260-2	(2)	7.65	7.65	7.55	7.75	0.00
Aroclor-1260-3	(3)	8.01	8.01	7.91	8.11	0.00
Aroclor-1260-4	(4)	8.24	8.24	8.14	8.34	0.01
Aroclor-1260-5	(5)	8.56	8.56	8.46	8.66	0.00
Tetrachloro-m-xylene		4.66	4.66	4.56	4.76	0.00
Decachlorobiphenyl		10.43	10.43	10.33	10.53	0.00



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: ROYF02

Lab Code: ACE

SDG NO.: Q3179

Continuing Calib Date: 09/29/2025

Initial Calibration Date(s): 09/23/2025

09/23/2025

Continuing Calib Time: 18:38

Initial Calibration Time(s): 08:45

17:14

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1016-1	(1)	4.89	4.89	4.79	4.99	0.00
Aroclor-1016-2	(2)	4.95	4.95	4.85	5.05	0.00
Aroclor-1016-3	(3)	5.07	5.07	4.97	5.17	0.00
Aroclor-1016-4	(4)	5.11	5.11	5.01	5.21	0.00
Aroclor-1016-5	(5)	5.32	5.33	5.23	5.43	0.01
Aroclor-1260-1	(1)	6.35	6.36	6.26	6.46	0.01
Aroclor-1260-2	(2)	6.54	6.54	6.44	6.64	0.00
Aroclor-1260-3	(3)	6.69	6.70	6.60	6.80	0.01
Aroclor-1260-4	(4)	7.16	7.17	7.07	7.27	0.01
Aroclor-1260-5	(5)	7.40	7.40	7.30	7.50	0.00
Tetrachloro-m-xylene		3.80	3.80	3.70	3.90	0.00
Decachlorobiphenyl		8.80	8.80	8.70	8.90	0.00



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ROYF02
Lab Code: ACE **SDG NO.:** Q3179
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 09/23/2025 09/23/2025

Client Sample No.: CCAL01 **Date Analyzed:** 09/29/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP075413.D **Time Analyzed:** 18:38

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	5.806	5.708	5.908	475.770	500.000	-4.8
Aroclor-1016-2	5.827	5.729	5.929	490.020	500.000	-2.0
Aroclor-1016-3	5.890	5.792	5.992	473.990	500.000	-5.2
Aroclor-1016-4	5.987	5.889	6.089	515.030	500.000	3.0
Aroclor-1016-5	6.280	6.181	6.381	518.410	500.000	3.7
Aroclor-1260-1	7.397	7.299	7.499	495.570	500.000	-0.9
Aroclor-1260-2	7.649	7.551	7.751	489.480	500.000	-2.1
Aroclor-1260-3	8.007	7.909	8.109	536.930	500.000	7.4
Aroclor-1260-4	8.235	8.137	8.337	561.600	500.000	12.3
Aroclor-1260-5	8.561	8.464	8.664	545.160	500.000	9.0
Decachlorobiphenyl	10.430	10.334	10.534	51.390	50.000	2.8
Tetrachloro-m-xylene	4.655	4.555	4.755	50.180	50.000	0.4



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ROYF02
Lab Code: ACE **SDG NO.:** Q3179
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 09/23/2025 09/23/2025

Client Sample No.: CCAL01 **Date Analyzed:** 09/29/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP075413.D **Time Analyzed:** 18:38

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.891	4.793	4.993	534.060	500.000	6.8
Aroclor-1016-2	4.950	4.851	5.051	533.950	500.000	6.8
Aroclor-1016-3	5.069	4.971	5.171	494.970	500.000	-1.0
Aroclor-1016-4	5.110	5.011	5.211	518.630	500.000	3.7
Aroclor-1016-5	5.324	5.225	5.425	511.900	500.000	2.4
Aroclor-1260-1	6.352	6.255	6.455	484.640	500.000	-3.1
Aroclor-1260-2	6.539	6.441	6.641	484.970	500.000	-3.0
Aroclor-1260-3	6.692	6.595	6.795	482.320	500.000	-3.5
Aroclor-1260-4	7.160	7.065	7.265	454.900	500.000	-9.0
Aroclor-1260-5	7.400	7.304	7.504	489.010	500.000	-2.2
Decachlorobiphenyl	8.798	8.703	8.903	58.950	50.000	17.9
Tetrachloro-m-xylene	3.796	3.696	3.896	53.900	50.000	7.8

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP092925\
Data File : PP075413.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 29 Sep 2025 18:38
Operator : YP\AJ
Sample : AR1660CCC500
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 30 03:01:40 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP092325.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Sep 24 10:55:35 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.655	3.796	65655152	328.9E6	50.177	53.905
2) SA Decachlor...	10.430	8.798	51266149	355.6E6	51.392	58.948
Target Compounds						
3) L1 AR-1016-1	5.806	4.891	22886727	299.6E6	475.768	534.064
4) L1 AR-1016-2	5.827	4.950	35148637	137.2E6	490.023	533.949
5) L1 AR-1016-3	5.890	5.069	22055318	76853062	473.994	494.969
6) L1 AR-1016-4	5.987	5.110	18353975	76222347	515.031	518.632
7) L1 AR-1016-5	6.280	5.324	17481440	83635201	518.409	511.899
31) L7 AR-1260-1	7.397	6.352	29793567	196.7E6	495.572	484.645
32) L7 AR-1260-2	7.649	6.539	35632931	279.3E6	489.477	484.967
33) L7 AR-1260-3	8.007	6.692	29071938	206.5E6	536.926	482.317
34) L7 AR-1260-4	8.235	7.160	31751412	196.2E6	561.603	454.898
35) L7 AR-1260-5	8.561	7.400	59862567	514.4E6	545.155	489.007



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP092925\
Data File : PP075413.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 29 Sep 2025 18:38
Operator : YP\AJ
Sample : AR1660CCC500
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 30 03:01:40 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP092325.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Sep 24 10:55:35 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.655 min
Delta R.T.: 0.000 min
Response: 65655152
Conc: 50.18 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.796 min
Delta R.T.: 0.000 min
Response: 328866910
Conc: 53.90 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.430 min
Delta R.T.: -0.004 min
Response: 51266149
Conc: 51.39 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.798 min
Delta R.T.: -0.005 min
Response: 355602140
Conc: 58.95 ng/ml



#3 AR-1016-1


R.T.: 5.806 min
Delta R.T.: -0.001 min
Response: 22886727
Conc: 475.77 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#3 AR-1016-1

R.T.: 4.891 min
Delta R.T.: -0.002 min
Response: 299616871
Conc: 534.06 ng/ml



#4 AR-1016-2

R.T.: 5.827 min
Delta R.T.: -0.001 min
Response: 35148637
Conc: 490.02 ng/ml




#4 AR-1016-2

R.T.: 4.950 min
Delta R.T.: -0.001 min
Response: 137226026
Conc: 533.95 ng/ml

#5 AR-1016-3


R.T.: 5.890 min
Delta R.T.: -0.001 min
Response: 22055318
Conc: 473.99 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#5 AR-1016-3

R.T.: 5.069 min
Delta R.T.: -0.002 min
Response: 76853062
Conc: 494.97 ng/ml




#6 AR-1016-4


R.T.: 5.987 min
Delta R.T.: -0.001 min
Response: 18353975
Conc: 515.03 ng/ml



#6 AR-1016-4

R.T.: 5.110 min
Delta R.T.: -0.002 min
Response: 76222347
Conc: 518.63 ng/ml






#7 AR-1016-5


R.T.: 6.280 min
Delta R.T.: -0.001 min
Response: 17481440
Conc: 518.41 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#7 AR-1016-5

R.T.: 5.324 min
Delta R.T.: -0.002 min
Response: 83635201
Conc: 511.90 ng/ml




#31 AR-1260-1

R.T.: 7.397 min
Delta R.T.: -0.002 min
Response: 29793567
Conc: 495.57 ng/ml



#31 AR-1260-1


R.T.: 6.352 min
Delta R.T.: -0.003 min
Response: 196734468
Conc: 484.64 ng/ml



#32 AR-1260-2


R.T.: 7.649 min
Delta R.T.: -0.002 min
Response: 35632931
Conc: 489.48 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#32 AR-1260-2

R.T.: 6.539 min
Delta R.T.: -0.002 min
Response: 279258506
Conc: 484.97 ng/ml




#33 AR-1260-3

R.T.: 8.007 min
Delta R.T.: -0.002 min
Response: 29071938
Conc: 536.93 ng/ml



#33 AR-1260-3


R.T.: 6.692 min
Delta R.T.: -0.003 min
Response: 206472667
Conc: 482.32 ng/ml



#34 AR-1260-4


R.T.: 8.235 min
Delta R.T.: -0.002 min
Response: 31751412
Conc: 561.60 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#34 AR-1260-4

R.T.: 7.160 min
Delta R.T.: -0.005 min
Response: 196163012
Conc: 454.90 ng/ml



#35 AR-1260-5

R.T.: 8.561 min
Delta R.T.: -0.003 min
Response: 59862567
Conc: 545.16 ng/ml



#35 AR-1260-5

R.T.: 7.400 min
Delta R.T.: -0.004 min
Response: 514430006
Conc: 489.01 ng/ml



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: ROYF02

Lab Code: ACE

SDG NO.: Q3179

Continuing Calib Date: 09/29/2025

Initial Calibration Date(s): 09/23/2025 09/23/2025

Continuing Calib Time: 23:47

Initial Calibration Time(s): 08:45 17:14

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1016-1	(1)	5.81	5.81	5.71	5.91	0.00
Aroclor-1016-2	(2)	5.83	5.83	5.73	5.93	0.00
Aroclor-1016-3	(3)	5.89	5.89	5.79	5.99	0.00
Aroclor-1016-4	(4)	5.99	5.99	5.89	6.09	0.00
Aroclor-1016-5	(5)	6.28	6.28	6.18	6.38	0.00
Aroclor-1260-1	(1)	7.40	7.40	7.30	7.50	0.00
Aroclor-1260-2	(2)	7.65	7.65	7.55	7.75	0.00
Aroclor-1260-3	(3)	8.01	8.01	7.91	8.11	0.00
Aroclor-1260-4	(4)	8.24	8.24	8.14	8.34	0.00
Aroclor-1260-5	(5)	8.56	8.56	8.46	8.66	0.00
Tetrachloro-m-xylene		4.66	4.66	4.56	4.76	0.00
Decachlorobiphenyl		10.43	10.43	10.33	10.53	0.00



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: ROYF02

Lab Code: ACE

SDG NO.: Q3179

Continuing Calib Date: 09/29/2025

Initial Calibration Date(s): 09/23/2025

09/23/2025

Continuing Calib Time: 23:47

Initial Calibration Time(s): 08:45

17:14

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1016-1	(1)	4.89	4.89	4.79	4.99	0.00
Aroclor-1016-2	(2)	4.95	4.95	4.85	5.05	0.00
Aroclor-1016-3	(3)	5.07	5.07	4.97	5.17	0.00
Aroclor-1016-4	(4)	5.11	5.11	5.01	5.21	0.00
Aroclor-1016-5	(5)	5.32	5.33	5.23	5.43	0.01
Aroclor-1260-1	(1)	6.35	6.36	6.26	6.46	0.01
Aroclor-1260-2	(2)	6.54	6.54	6.44	6.64	0.00
Aroclor-1260-3	(3)	6.69	6.70	6.60	6.80	0.01
Aroclor-1260-4	(4)	7.16	7.17	7.07	7.27	0.01
Aroclor-1260-5	(5)	7.40	7.40	7.30	7.50	0.00
Tetrachloro-m-xylene		3.80	3.80	3.70	3.90	0.00
Decachlorobiphenyl		8.80	8.80	8.70	8.90	0.00



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ROYF02
Lab Code: ACE **SDG NO.:** Q3179
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 09/23/2025 09/23/2025

Client Sample No.: CCAL02 **Date Analyzed:** 09/29/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP075428.D **Time Analyzed:** 23:47

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.807	5.708	5.908	488.120	500.000	-2.4
Aroclor-1016-2	5.828	5.729	5.929	496.430	500.000	-0.7
Aroclor-1016-3	5.891	5.792	5.992	486.040	500.000	-2.8
Aroclor-1016-4	5.988	5.889	6.089	519.590	500.000	3.9
Aroclor-1016-5	6.281	6.181	6.381	527.540	500.000	5.5
Aroclor-1260-1	7.398	7.299	7.499	507.550	500.000	1.5
Aroclor-1260-2	7.650	7.551	7.751	502.340	500.000	0.5
Aroclor-1260-3	8.008	7.909	8.109	547.800	500.000	9.6
Aroclor-1260-4	8.237	8.137	8.337	575.600	500.000	15.1
Aroclor-1260-5	8.562	8.464	8.664	563.440	500.000	12.7
Decachlorobiphenyl	10.431	10.334	10.534	53.400	50.000	6.8
Tetrachloro-m-xylene	4.655	4.555	4.755	50.830	50.000	1.7



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ROYF02
Lab Code: ACE **SDG NO.:** Q3179
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 09/23/2025 09/23/2025

Client Sample No.: CCAL02 **Date Analyzed:** 09/29/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP075428.D **Time Analyzed:** 23:47

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.892	4.793	4.993	537.660	500.000	7.5
Aroclor-1016-2	4.950	4.851	5.051	549.500	500.000	9.9
Aroclor-1016-3	5.070	4.971	5.171	497.520	500.000	-0.5
Aroclor-1016-4	5.111	5.011	5.211	525.220	500.000	5.0
Aroclor-1016-5	5.324	5.225	5.425	505.020	500.000	1.0
Aroclor-1260-1	6.352	6.255	6.455	506.680	500.000	1.3
Aroclor-1260-2	6.539	6.441	6.641	496.240	500.000	-0.8
Aroclor-1260-3	6.692	6.595	6.795	511.100	500.000	2.2
Aroclor-1260-4	7.161	7.065	7.265	468.620	500.000	-6.3
Aroclor-1260-5	7.401	7.304	7.504	496.780	500.000	-0.6
Decachlorobiphenyl	8.796	8.703	8.903	61.070	50.000	22.1
Tetrachloro-m-xylene	3.797	3.696	3.896	54.520	50.000	9.0

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP092925\
 Data File : PP075428.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 29 Sep 2025 23:47
 Operator : YP\AJ
 Sample : AR1660CCC500
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1660CCC500

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 09/30/2025
 Supervised By :mohammad ahmed 10/01/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 30 03:05:07 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP092325.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Sep 24 10:55:35 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.655	3.797	66510643	332.7E6	50.830	54.525
2) SA Decachlor...	10.431	8.796	53269661	368.4E6	53.401	61.070m
Target Compounds						
3) L1 AR-1016-1	5.807	4.892	23480995	301.6E6	488.122	537.655
4) L1 AR-1016-2	5.828	4.950	35608074	141.2E6	496.428	549.498
5) L1 AR-1016-3	5.891	5.070	22615628	77249486	486.036	497.522
6) L1 AR-1016-4	5.988	5.111	18516537	77190450	519.592	525.220
7) L1 AR-1016-5	6.281	5.324	17789361	82511679	527.541	505.023
31) L7 AR-1260-1	7.398	6.352	30513864	205.7E6	507.554	506.684
32) L7 AR-1260-2	7.650	6.539	36569242	285.8E6	502.339	496.245
33) L7 AR-1260-3	8.008	6.692	29660582	218.8E6	547.798	511.095
34) L7 AR-1260-4	8.237	7.161	32543042	202.1E6	575.604	468.618
35) L7 AR-1260-5	8.562	7.401	61869920	522.6E6	563.436	496.785

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP092925\
Data File : PP075428.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 29 Sep 2025 23:47
Operator : YP\AJ
Sample : AR1660CCC500
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :

ECD_P

ClientSampleId :

AR1660CCC500

Manual Integrations



APPROVED


Reviewed By :Yogesh Patel 09/30/2025

Supervised By :mohammad ahmed 10/01/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 30 03:05:07 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP092325.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Sep 24 10:55:35 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.655 min
Delta R.T.: 0.000 min
Response: 66510643
Conc: 50.83 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 09/30/2025
Supervised By :mohammad ahmed 10/01/2025




#1 Tetrachloro-m-xylene

R.T.: 3.797 min
Delta R.T.: 0.000 min
Response: 332650210
Conc: 54.52 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.431 min
Delta R.T.: -0.004 min
Response: 53269661
Conc: 53.40 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.796 min
Delta R.T.: -0.007 min
Response: 368407387
Conc: 61.07 ng/ml m




#3 AR-1016-1

R.T.: 5.807 min
Delta R.T.: 0.000 min
Response: 23480995
Conc: 488.12 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 09/30/2025
Supervised By :mohammad ahmed 10/01/2025




#3 AR-1016-1

R.T.: 4.892 min
Delta R.T.: 0.000 min
Response: 301631883
Conc: 537.66 ng/ml




#4 AR-1016-2

R.T.: 5.828 min
Delta R.T.: 0.000 min
Response: 35608074
Conc: 496.43 ng/ml



#4 AR-1016-2

R.T.: 4.950 min
Delta R.T.: 0.000 min
Response: 141222152
Conc: 549.50 ng/ml




#5 AR-1016-3

R.T.: 5.891 min
Delta R.T.: 0.000 min
Response: 22615628
Conc: 486.04 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 09/30/2025
Supervised By :mohammad ahmed 10/01/2025




#5 AR-1016-3

R.T.: 5.070 min
Delta R.T.: -0.001 min
Response: 77249486
Conc: 497.52 ng/ml




#6 AR-1016-4

R.T.: 5.988 min
Delta R.T.: 0.000 min
Response: 18516537
Conc: 519.59 ng/ml



#6 AR-1016-4

R.T.: 5.111 min
Delta R.T.: 0.000 min
Response: 77190450
Conc: 525.22 ng/ml




#7 AR-1016-5

R.T.: 6.281 min
Delta R.T.: 0.000 min
Response: 17789361
Conc: 527.54 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 09/30/2025
Supervised By :mohammad ahmed 10/01/2025




#7 AR-1016-5

R.T.: 5.324 min
Delta R.T.: 0.000 min
Response: 82511679
Conc: 505.02 ng/ml




#31 AR-1260-1

R.T.: 7.398 min
Delta R.T.: 0.000 min
Response: 30513864
Conc: 507.55 ng/ml



#31 AR-1260-1

R.T.: 6.352 min
Delta R.T.: -0.003 min
Response: 205681153
Conc: 506.68 ng/ml




#32 AR-1260-2

R.T.: 7.650 min
Delta R.T.: 0.000 min
Response: 36569242
Conc: 502.34 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 09/30/2025
Supervised By :mohammad ahmed 10/01/2025




#32 AR-1260-2

R.T.: 6.539 min
Delta R.T.: -0.003 min
Response: 285752472
Conc: 496.24 ng/ml




#33 AR-1260-3

R.T.: 8.008 min
Delta R.T.: 0.000 min
Response: 29660582
Conc: 547.80 ng/ml



#33 AR-1260-3

R.T.: 6.692 min
Delta R.T.: -0.004 min
Response: 218792140
Conc: 511.10 ng/ml




#34 AR-1260-4

R.T.: 8.237 min
Delta R.T.: 0.000 min
Response: 32543042
Conc: 575.60 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 09/30/2025
Supervised By :mohammad ahmed 10/01/2025




#34 AR-1260-4

R.T.: 7.161 min
Delta R.T.: -0.004 min
Response: 202079464
Conc: 468.62 ng/ml



#35 AR-1260-5

R.T.: 8.562 min
Delta R.T.: -0.001 min
Response: 61869920
Conc: 563.44 ng/ml



#35 AR-1260-5

R.T.: 7.401 min
Delta R.T.: -0.003 min
Response: 522611557
Conc: 496.78 ng/ml



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: ROYF02

Lab Code: ACE

SDG NO.: Q3179

Continuing Calib Date: 09/30/2025

Initial Calibration Date(s): 09/23/2025

09/23/2025

Continuing Calib Time: 08:46

Initial Calibration Time(s): 08:45

17:14

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1016-1	(1)	5.81	5.81	5.71	5.91	0.00
Aroclor-1016-2	(2)	5.83	5.83	5.73	5.93	0.00
Aroclor-1016-3	(3)	5.89	5.89	5.79	5.99	0.00
Aroclor-1016-4	(4)	5.99	5.99	5.89	6.09	0.00
Aroclor-1016-5	(5)	6.28	6.28	6.18	6.38	0.00
Aroclor-1260-1	(1)	7.40	7.40	7.30	7.50	0.00
Aroclor-1260-2	(2)	7.65	7.65	7.55	7.75	0.00
Aroclor-1260-3	(3)	8.01	8.01	7.91	8.11	0.00
Aroclor-1260-4	(4)	8.24	8.24	8.14	8.34	0.00
Aroclor-1260-5	(5)	8.56	8.56	8.46	8.66	0.00
Tetrachloro-m-xylene		4.66	4.66	4.56	4.76	0.00
Decachlorobiphenyl		10.43	10.43	10.33	10.53	0.00



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: ROYF02

Lab Code: ACE

SDG NO.: Q3179

Continuing Calib Date: 09/30/2025

Initial Calibration Date(s): 09/23/2025 09/23/2025

Continuing Calib Time: 08:46

Initial Calibration Time(s): 08:45 17:14

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1016-1	(1)	4.89	4.89	4.79	4.99	0.00
Aroclor-1016-2	(2)	4.95	4.95	4.85	5.05	0.00
Aroclor-1016-3	(3)	5.07	5.07	4.97	5.17	0.00
Aroclor-1016-4	(4)	5.11	5.11	5.01	5.21	0.00
Aroclor-1016-5	(5)	5.32	5.33	5.23	5.43	0.01
Aroclor-1260-1	(1)	6.35	6.36	6.26	6.46	0.01
Aroclor-1260-2	(2)	6.54	6.54	6.44	6.64	0.00
Aroclor-1260-3	(3)	6.69	6.70	6.60	6.80	0.01
Aroclor-1260-4	(4)	7.16	7.17	7.07	7.27	0.01
Aroclor-1260-5	(5)	7.40	7.40	7.30	7.50	0.00
Tetrachloro-m-xylene		3.79	3.80	3.70	3.90	0.01
Decachlorobiphenyl		8.80	8.80	8.70	8.90	0.00



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ROYF02
Lab Code: ACE **SDG NO.:** Q3179
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 09/23/2025 09/23/2025

Client Sample No.: CCAL03 **Date Analyzed:** 09/30/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP075434.D **Time Analyzed:** 08:46

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	5.807	5.708	5.908	530.620	500.000	6.1
Aroclor-1016-2	5.828	5.729	5.929	525.210	500.000	5.0
Aroclor-1016-3	5.890	5.792	5.992	537.720	500.000	7.5
Aroclor-1016-4	5.988	5.889	6.089	583.920	500.000	16.8
Aroclor-1016-5	6.280	6.181	6.381	571.420	500.000	14.3
Aroclor-1260-1	7.397	7.299	7.499	522.290	500.000	4.5
Aroclor-1260-2	7.650	7.551	7.751	514.200	500.000	2.8
Aroclor-1260-3	8.008	7.909	8.109	563.460	500.000	12.7
Aroclor-1260-4	8.237	8.137	8.337	588.400	500.000	17.7
Aroclor-1260-5	8.562	8.464	8.664	573.040	500.000	14.6
Decachlorobiphenyl	10.432	10.334	10.534	54.160	50.000	8.3
Tetrachloro-m-xylene	4.655	4.555	4.755	52.100	50.000	4.2



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ROYF02
Lab Code: ACE **SDG NO.:** Q3179
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 09/23/2025 09/23/2025

Client Sample No.: CCAL03 **Date Analyzed:** 09/30/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP075434.D **Time Analyzed:** 08:46

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	4.890	4.793	4.993	546.990	500.000	9.4
Aroclor-1016-2	4.949	4.851	5.051	555.350	500.000	11.1
Aroclor-1016-3	5.068	4.971	5.171	493.970	500.000	-1.2
Aroclor-1016-4	5.108	5.011	5.211	555.270	500.000	11.1
Aroclor-1016-5	5.323	5.225	5.425	535.230	500.000	7.0
Aroclor-1260-1	6.351	6.255	6.455	522.520	500.000	4.5
Aroclor-1260-2	6.537	6.441	6.641	513.690	500.000	2.7
Aroclor-1260-3	6.691	6.595	6.795	510.050	500.000	2.0
Aroclor-1260-4	7.161	7.065	7.265	482.390	500.000	-3.5
Aroclor-1260-5	7.400	7.304	7.504	501.480	500.000	0.3
Decachlorobiphenyl	8.798	8.703	8.903	62.930	50.000	25.9
Tetrachloro-m-xylene	3.794	3.696	3.896	56.350	50.000	12.7

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP093025\
 Data File : PP075434.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 30 Sep 2025 08:46
 Operator : YP\AJ
 Sample : AR1660CCC500
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1660CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 01 01:55:12 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP092325.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Sep 24 10:55:35 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.655	3.794	68170819	343.8E6	52.099	56.351
2) SA Decachlor...	10.432	8.798	54023768	379.6E6	54.157	62.932
Target Compounds						
3) L1 AR-1016-1	5.807	4.890	25525426	306.9E6	530.621	546.988
4) L1 AR-1016-2	5.828	4.949	37672637	142.7E6	525.211	555.355
5) L1 AR-1016-3	5.890	5.068	25020525	76698528	537.720	493.974
6) L1 AR-1016-4	5.988	5.108	20809034	81606600	583.922	555.268
7) L1 AR-1016-5	6.280	5.323	19269167	87447268	571.424	535.231
31) L7 AR-1260-1	7.397	6.351	31399995	212.1E6	522.293	522.518
32) L7 AR-1260-2	7.650	6.537	37432702	295.8E6	514.200	513.687
33) L7 AR-1260-3	8.008	6.691	30508810	218.3E6	563.464	510.047
34) L7 AR-1260-4	8.237	7.161	33266747	208.0E6	588.405	482.395
35) L7 AR-1260-5	8.562	7.400	62924134	527.5E6	573.036	501.477



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP093025\
Data File : PP075434.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 30 Sep 2025 08:46
Operator : YP\AJ
Sample : AR1660CCC500
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 01 01:55:12 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP092325.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Sep 24 10:55:35 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.655 min
Delta R.T.: 0.000 min
Response: 68170819
Conc: 52.10 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.794 min
Delta R.T.: -0.001 min
Response: 343792956
Conc: 56.35 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.432 min
Delta R.T.: -0.002 min
Response: 54023768
Conc: 54.16 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.798 min
Delta R.T.: -0.005 min
Response: 379635233
Conc: 62.93 ng/ml



#3 AR-1016-1


R.T.: 5.807 min
Delta R.T.: 0.000 min
Response: 25525426
Conc: 530.62 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#3 AR-1016-1

R.T.: 4.890 min
Delta R.T.: -0.002 min
Response: 306867499
Conc: 546.99 ng/ml



#4 AR-1016-2

R.T.: 5.828 min
Delta R.T.: 0.000 min
Response: 37672637
Conc: 525.21 ng/ml




#4 AR-1016-2

R.T.: 4.949 min
Delta R.T.: -0.002 min
Response: 142727355
Conc: 555.35 ng/ml

#5 AR-1016-3


R.T.: 5.890 min
Delta R.T.: -0.002 min
Response: 25020525
Conc: 537.72 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#5 AR-1016-3

R.T.: 5.068 min
Delta R.T.: -0.003 min
Response: 76698528
Conc: 493.97 ng/ml




#6 AR-1016-4


R.T.: 5.988 min
Delta R.T.: 0.000 min
Response: 20809034
Conc: 583.92 ng/ml



#6 AR-1016-4

R.T.: 5.108 min
Delta R.T.: -0.003 min
Response: 81606600
Conc: 555.27 ng/ml






#7 AR-1016-5


R.T.: 6.280 min
Delta R.T.: -0.001 min
Response: 19269167
Conc: 571.42 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#7 AR-1016-5

R.T.: 5.323 min
Delta R.T.: -0.002 min
Response: 87447268
Conc: 535.23 ng/ml




#31 AR-1260-1

R.T.: 7.397 min
Delta R.T.: -0.002 min
Response: 31399995
Conc: 522.29 ng/ml



#31 AR-1260-1


R.T.: 6.351 min
Delta R.T.: -0.004 min
Response: 212108619
Conc: 522.52 ng/ml



#32 AR-1260-2


R.T.: 7.650 min
Delta R.T.: -0.001 min
Response: 37432702
Conc: 514.20 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#32 AR-1260-2

R.T.: 6.537 min
Delta R.T.: -0.004 min
Response: 295796129
Conc: 513.69 ng/ml




#33 AR-1260-3

R.T.: 8.008 min
Delta R.T.: -0.001 min
Response: 30508810
Conc: 563.46 ng/ml



#33 AR-1260-3


R.T.: 6.691 min
Delta R.T.: -0.004 min
Response: 218343672
Conc: 510.05 ng/ml



#34 AR-1260-4


R.T.: 8.237 min
Delta R.T.: 0.000 min
Response: 33266747
Conc: 588.40 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#34 AR-1260-4

R.T.: 7.161 min
Delta R.T.: -0.004 min
Response: 208020235
Conc: 482.39 ng/ml



#35 AR-1260-5

R.T.: 8.562 min
Delta R.T.: -0.001 min
Response: 62924134
Conc: 573.04 ng/ml



#35 AR-1260-5

R.T.: 7.400 min
Delta R.T.: -0.004 min
Response: 527547609
Conc: 501.48 ng/ml



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: ROYF02

Lab Code: ACE

SDG NO.: Q3179

Continuing Calib Date: 09/30/2025

Initial Calibration Date(s): 09/23/2025

09/23/2025

Continuing Calib Time: 16:14

Initial Calibration Time(s): 08:45

17:14

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1016-1	(1)	5.81	5.81	5.71	5.91	0.00
Aroclor-1016-2	(2)	5.83	5.83	5.73	5.93	0.00
Aroclor-1016-3	(3)	5.89	5.89	5.79	5.99	0.00
Aroclor-1016-4	(4)	5.99	5.99	5.89	6.09	0.00
Aroclor-1016-5	(5)	6.28	6.28	6.18	6.38	0.00
Aroclor-1260-1	(1)	7.40	7.40	7.30	7.50	0.00
Aroclor-1260-2	(2)	7.65	7.65	7.55	7.75	0.00
Aroclor-1260-3	(3)	8.01	8.01	7.91	8.11	0.00
Aroclor-1260-4	(4)	8.24	8.24	8.14	8.34	0.00
Aroclor-1260-5	(5)	8.56	8.56	8.46	8.66	0.00
Tetrachloro-m-xylene		4.66	4.66	4.56	4.76	0.00
Decachlorobiphenyl		10.43	10.43	10.33	10.53	0.00



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: ROYF02

Lab Code: ACE

SDG NO.: Q3179

Continuing Calib Date: 09/30/2025

Initial Calibration Date(s): 09/23/2025 09/23/2025

Continuing Calib Time: 16:14

Initial Calibration Time(s): 08:45 17:14

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1016-1	(1)	4.89	4.89	4.79	4.99	0.00
Aroclor-1016-2	(2)	4.95	4.95	4.85	5.05	0.00
Aroclor-1016-3	(3)	5.07	5.07	4.97	5.17	0.00
Aroclor-1016-4	(4)	5.11	5.11	5.01	5.21	0.00
Aroclor-1016-5	(5)	5.32	5.33	5.23	5.43	0.01
Aroclor-1260-1	(1)	6.35	6.36	6.26	6.46	0.01
Aroclor-1260-2	(2)	6.54	6.54	6.44	6.64	0.00
Aroclor-1260-3	(3)	6.69	6.70	6.60	6.80	0.01
Aroclor-1260-4	(4)	7.16	7.17	7.07	7.27	0.01
Aroclor-1260-5	(5)	7.40	7.40	7.30	7.50	0.00
Tetrachloro-m-xylene		3.80	3.80	3.70	3.90	0.00
Decachlorobiphenyl		8.80	8.80	8.70	8.90	0.00



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ROYF02
Lab Code: ACE **SDG NO.:** Q3179
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 09/23/2025 09/23/2025

Client Sample No.: CCAL04 **Date Analyzed:** 09/30/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP075449.D **Time Analyzed:** 16:14

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.808	5.708	5.908	473.340	500.000	-5.3
Aroclor-1016-2	5.829	5.729	5.929	477.380	500.000	-4.5
Aroclor-1016-3	5.892	5.792	5.992	463.180	500.000	-7.4
Aroclor-1016-4	5.989	5.889	6.089	498.870	500.000	-0.2
Aroclor-1016-5	6.281	6.181	6.381	501.220	500.000	0.2
Aroclor-1260-1	7.399	7.299	7.499	487.040	500.000	-2.6
Aroclor-1260-2	7.651	7.551	7.751	475.980	500.000	-4.8
Aroclor-1260-3	8.010	7.909	8.109	518.370	500.000	3.7
Aroclor-1260-4	8.238	8.137	8.337	545.630	500.000	9.1
Aroclor-1260-5	8.564	8.464	8.664	541.150	500.000	8.2
Decachlorobiphenyl	10.434	10.334	10.534	52.310	50.000	4.6
Tetrachloro-m-xylene	4.656	4.555	4.755	50.130	50.000	0.3



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ROYF02
Lab Code: ACE **SDG NO.:** Q3179
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 09/23/2025 09/23/2025

Client Sample No.: CCAL04 **Date Analyzed:** 09/30/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP075449.D **Time Analyzed:** 16:14

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.891	4.793	4.993	542.910	500.000	8.6
Aroclor-1016-2	4.950	4.851	5.051	569.090	500.000	13.8
Aroclor-1016-3	5.070	4.971	5.171	505.350	500.000	1.1
Aroclor-1016-4	5.111	5.011	5.211	533.050	500.000	6.6
Aroclor-1016-5	5.324	5.225	5.425	531.740	500.000	6.3
Aroclor-1260-1	6.353	6.255	6.455	423.560	500.000	-15.3
Aroclor-1260-2	6.539	6.441	6.641	428.070	500.000	-14.4
Aroclor-1260-3	6.692	6.595	6.795	439.130	500.000	-12.2
Aroclor-1260-4	7.161	7.065	7.265	415.220	500.000	-17.0
Aroclor-1260-5	7.401	7.304	7.504	432.220	500.000	-13.6
Decachlorobiphenyl	8.799	8.703	8.903	56.660	50.000	13.3
Tetrachloro-m-xylene	3.796	3.696	3.896	54.770	50.000	9.5

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP093025\
Data File : PP075449.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 30 Sep 2025 16:14
Operator : YP\AJ
Sample : AR1660CCC500
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 01 02:07:14 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP092325.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Sep 24 10:55:35 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.656	3.796	65592204	334.2E6	50.128	54.773
2) SA Decachlor...	10.434	8.799	52184946	341.8E6	52.313	56.656
Target Compounds						
3) L1 AR-1016-1	5.808	4.891	22769988	304.6E6	473.342	542.907
4) L1 AR-1016-2	5.829	4.950	34241466	146.3E6	477.375	569.092
5) L1 AR-1016-3	5.892	5.070	21552342	78464373	463.185	505.347
6) L1 AR-1016-4	5.989	5.111	17777933	78341493	498.866	533.051
7) L1 AR-1016-5	6.281	5.324	16901642	86876028	501.215	531.735
31) L7 AR-1260-1	7.399	6.353	29280863	171.9E6	487.044	423.558
32) L7 AR-1260-2	7.651	6.539	34650621	246.5E6	475.983	428.073
33) L7 AR-1260-3	8.010	6.692	28067195	188.0E6	518.370	439.126
34) L7 AR-1260-4	8.238	7.161	30848408	179.1E6	545.631	415.218
35) L7 AR-1260-5	8.564	7.401	59422948	454.7E6	541.152	432.222



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP093025\
Data File : PP075449.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 30 Sep 2025 16:14
Operator : YP\AJ
Sample : AR1660CCC500
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 01 02:07:14 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP092325.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Sep 24 10:55:35 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.656 min
Delta R.T.: 0.000 min
Response: 65592204
Conc: 50.13 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.796 min
Delta R.T.: 0.000 min
Response: 334164298
Conc: 54.77 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.434 min
Delta R.T.: 0.000 min
Response: 52184946
Conc: 52.31 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.799 min
Delta R.T.: -0.004 min
Response: 341778628
Conc: 56.66 ng/ml



#3 AR-1016-1


R.T.: 5.808 min
Delta R.T.: 0.000 min
Response: 22769988
Conc: 473.34 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#3 AR-1016-1

R.T.: 4.891 min
Delta R.T.: -0.002 min
Response: 304577903
Conc: 542.91 ng/ml



#4 AR-1016-2

R.T.: 5.829 min
Delta R.T.: 0.000 min
Response: 34241466
Conc: 477.38 ng/ml




#4 AR-1016-2

R.T.: 4.950 min
Delta R.T.: -0.001 min
Response: 146257903
Conc: 569.09 ng/ml

#5 AR-1016-3


R.T.: 5.892 min
Delta R.T.: 0.000 min
Response: 21552342
Conc: 463.18 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#5 AR-1016-3

R.T.: 5.070 min
Delta R.T.: -0.001 min
Response: 78464373
Conc: 505.35 ng/ml




#6 AR-1016-4


R.T.: 5.989 min
Delta R.T.: 0.000 min
Response: 17777933
Conc: 498.87 ng/ml



#6 AR-1016-4

R.T.: 5.111 min
Delta R.T.: 0.000 min
Response: 78341493
Conc: 533.05 ng/ml






#7 AR-1016-5


R.T.: 6.281 min
Delta R.T.: 0.000 min
Response: 16901642
Conc: 501.22 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#7 AR-1016-5

R.T.: 5.324 min
Delta R.T.: 0.000 min
Response: 86876028
Conc: 531.74 ng/ml




#31 AR-1260-1

R.T.: 7.399 min
Delta R.T.: 0.000 min
Response: 29280863
Conc: 487.04 ng/ml



#31 AR-1260-1


R.T.: 6.353 min
Delta R.T.: -0.002 min
Response: 171937282
Conc: 423.56 ng/ml



#32 AR-1260-2


R.T.: 7.651 min
Delta R.T.: 0.000 min
Response: 34650621
Conc: 475.98 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#32 AR-1260-2

R.T.: 6.539 min
Delta R.T.: -0.002 min
Response: 246497516
Conc: 428.07 ng/ml




#33 AR-1260-3

R.T.: 8.010 min
Delta R.T.: 0.000 min
Response: 28067195
Conc: 518.37 ng/ml



#33 AR-1260-3


R.T.: 6.692 min
Delta R.T.: -0.003 min
Response: 187983316
Conc: 439.13 ng/ml



#34 AR-1260-4


R.T.: 8.238 min
Delta R.T.: 0.000 min
Response: 30848408
Conc: 545.63 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#34 AR-1260-4

R.T.: 7.161 min
Delta R.T.: -0.003 min
Response: 179052177
Conc: 415.22 ng/ml



#35 AR-1260-5

R.T.: 8.564 min
Delta R.T.: 0.000 min
Response: 59422948
Conc: 541.15 ng/ml



#35 AR-1260-5

R.T.: 7.401 min
Delta R.T.: -0.003 min
Response: 454692924
Conc: 432.22 ng/ml

Analytical Sequence

Client: Weston Solutions, Inc.

SDG No.: Q3179

Project: RFP 916

Instrument ID: ECD_P

GC Column: ZB-MR1

ID: 0.32 (mm)

Inst. Calib. Date(s): 09/23/2025

09/23/2025

THE ANALYTICAL SEQUENCE OF PERFORMANCE EVALUATION MIXTURES, BLANKS, SAMPLES,
AND STANDARDS IS GIVEN BELOW:

CLIENT ID	LAB SAMPLE ID	DATE ANALYZED	TIME ANALYZED	DATAFILE	DCB RT #	TCX RT #
IBLK	IBLK	09/23/2025	08:29	PP075246.D	10.44	4.66
AR1660ICC1000	AR1660ICC1000	09/23/2025	08:45	PP075247.D	10.43	4.65
AR1660ICC750	AR1660ICC750	09/23/2025	09:01	PP075248.D	10.44	4.66
AR1660ICC500	AR1660ICC500	09/23/2025	09:18	PP075249.D	10.43	4.66
AR1660ICC250	AR1660ICC250	09/23/2025	09:34	PP075250.D	10.43	4.66
AR1660ICC050	AR1660ICC050	09/23/2025	10:11	PP075251.D	10.45	4.66
AR1221ICC500	AR1221ICC500	09/23/2025	10:27	PP075252.D	10.44	4.66
AR1232ICC500	AR1232ICC500	09/23/2025	10:43	PP075253.D	10.44	4.66
AR1242ICC1000	AR1242ICC1000	09/23/2025	11:00	PP075254.D	10.44	4.66
AR1242ICC750	AR1242ICC750	09/23/2025	11:16	PP075255.D	10.44	4.66
AR1242ICC500	AR1242ICC500	09/23/2025	11:32	PP075256.D	10.44	4.66
AR1242ICC250	AR1242ICC250	09/23/2025	11:48	PP075257.D	10.44	4.66
AR1242ICC050	AR1242ICC050	09/23/2025	12:05	PP075258.D	10.44	4.66
AR1248ICC1000	AR1248ICC1000	09/23/2025	12:37	PP075259.D	10.43	4.66
AR1248ICC750	AR1248ICC750	09/23/2025	12:53	PP075260.D	10.44	4.66
AR1248ICC500	AR1248ICC500	09/23/2025	13:10	PP075261.D	10.44	4.66
AR1248ICC250	AR1248ICC250	09/23/2025	13:26	PP075262.D	10.44	4.66
AR1248ICC050	AR1248ICC050	09/23/2025	13:42	PP075263.D	10.44	4.66
AR1254ICC1000	AR1254ICC1000	09/23/2025	14:15	PP075264.D	10.44	4.66
AR1254ICC750	AR1254ICC750	09/23/2025	14:31	PP075265.D	10.44	4.66
AR1254ICC500	AR1254ICC500	09/23/2025	14:47	PP075266.D	10.44	4.66
AR1254ICC250	AR1254ICC250	09/23/2025	15:04	PP075267.D	10.44	4.66
AR1254ICC050	AR1254ICC050	09/23/2025	15:20	PP075268.D	10.43	4.66
AR1262ICC500	AR1262ICC500	09/23/2025	15:52	PP075269.D	10.44	4.66
AR1268ICC1000	AR1268ICC1000	09/23/2025	16:09	PP075270.D	10.44	4.66
AR1268ICC750	AR1268ICC750	09/23/2025	16:25	PP075271.D	10.44	4.66
AR1268ICC500	AR1268ICC500	09/23/2025	16:41	PP075272.D	10.44	4.66
AR1268ICC250	AR1268ICC250	09/23/2025	16:58	PP075273.D	10.44	4.66
AR1268ICC050	AR1268ICC050	09/23/2025	17:14	PP075274.D	10.44	4.66
AR1660CCC500	AR1660CCC500	09/29/2025	18:38	PP075413.D	10.43	4.66
IBLK	IBLK	09/29/2025	19:59	PP075417.D	10.43	4.65
PB169881BL	PB169881BL	09/29/2025	20:15	PP075418.D	10.43	4.66
PB169881TB	PB169881TB	09/29/2025	20:48	PP075420.D	10.43	4.65
EME-TS14-01	Q3179-01	09/29/2025	21:04	PP075421.D	10.43	4.65
EME-TS14-01MS	Q3179-02MS	09/29/2025	21:20	PP075422.D	10.43	4.66
EME-TS14-01MSD	Q3179-03MSD	09/29/2025	21:37	PP075423.D	10.43	4.66
AR1660CCC500	AR1660CCC500	09/29/2025	23:47	PP075428.D	10.43	4.66
IBLK	IBLK	09/30/2025	01:08	PP075432.D	10.43	4.66
AR1660CCC500	AR1660CCC500	09/30/2025	08:46	PP075434.D	10.43	4.66
IBLK	IBLK	09/30/2025	09:51	PP075438.D	10.43	4.66
PB169881BS	PB169881BS	09/30/2025	10:08	PP075439.D	10.43	4.66
AR1660CCC500	AR1660CCC500	09/30/2025	16:14	PP075449.D	10.43	4.66

Analytical Sequence

I.BLK	I.BLK	09/30/2025	17:19	PP075453.D	10.43	4.66
-------	-------	------------	-------	------------	-------	------

Analytical Sequence

Client: Weston Solutions, Inc.

SDG No.: Q3179

Project: RFP 916

Instrument ID: ECD_P

GC Column: ZB-MR2

ID: 0.32 (mm)

Inst. Calib. Date(s): 09/23/2025

09/23/2025

THE ANALYTICAL SEQUENCE OF PERFORMANCE EVALUATION MIXTURES, BLANKS, SAMPLES,
AND STANDARDS IS GIVEN BELOW:

CLIENT ID	LAB SAMPLE ID	DATE ANALYZED	TIME ANALYZED	DATAFILE	DCB RT #	TCX RT #
IBLK	IBLK	09/23/2025	08:29	PP075246.D	8.81	3.80
AR1660ICC1000	AR1660ICC1000	09/23/2025	08:45	PP075247.D	8.81	3.80
AR1660ICC750	AR1660ICC750	09/23/2025	09:01	PP075248.D	8.80	3.80
AR1660ICC500	AR1660ICC500	09/23/2025	09:18	PP075249.D	8.80	3.80
AR1660ICC250	AR1660ICC250	09/23/2025	09:34	PP075250.D	8.80	3.80
AR1660ICC050	AR1660ICC050	09/23/2025	10:11	PP075251.D	8.81	3.80
AR1221ICC500	AR1221ICC500	09/23/2025	10:27	PP075252.D	8.81	3.80
AR1232ICC500	AR1232ICC500	09/23/2025	10:43	PP075253.D	8.81	3.80
AR1242ICC1000	AR1242ICC1000	09/23/2025	11:00	PP075254.D	8.81	3.80
AR1242ICC750	AR1242ICC750	09/23/2025	11:16	PP075255.D	8.81	3.80
AR1242ICC500	AR1242ICC500	09/23/2025	11:32	PP075256.D	8.81	3.80
AR1242ICC250	AR1242ICC250	09/23/2025	11:48	PP075257.D	8.81	3.80
AR1242ICC050	AR1242ICC050	09/23/2025	12:05	PP075258.D	8.80	3.80
AR1248ICC1000	AR1248ICC1000	09/23/2025	12:37	PP075259.D	8.80	3.80
AR1248ICC750	AR1248ICC750	09/23/2025	12:53	PP075260.D	8.81	3.80
AR1248ICC500	AR1248ICC500	09/23/2025	13:10	PP075261.D	8.80	3.80
AR1248ICC250	AR1248ICC250	09/23/2025	13:26	PP075262.D	8.81	3.80
AR1248ICC050	AR1248ICC050	09/23/2025	13:42	PP075263.D	8.80	3.80
AR1254ICC1000	AR1254ICC1000	09/23/2025	14:15	PP075264.D	8.81	3.80
AR1254ICC750	AR1254ICC750	09/23/2025	14:31	PP075265.D	8.81	3.80
AR1254ICC500	AR1254ICC500	09/23/2025	14:47	PP075266.D	8.81	3.80
AR1254ICC250	AR1254ICC250	09/23/2025	15:04	PP075267.D	8.81	3.80
AR1254ICC050	AR1254ICC050	09/23/2025	15:20	PP075268.D	8.81	3.80
AR1262ICC500	AR1262ICC500	09/23/2025	15:52	PP075269.D	8.81	3.80
AR1268ICC1000	AR1268ICC1000	09/23/2025	16:09	PP075270.D	8.81	3.80
AR1268ICC750	AR1268ICC750	09/23/2025	16:25	PP075271.D	8.81	3.80
AR1268ICC500	AR1268ICC500	09/23/2025	16:41	PP075272.D	8.81	3.80
AR1268ICC250	AR1268ICC250	09/23/2025	16:58	PP075273.D	8.80	3.80
AR1268ICC050	AR1268ICC050	09/23/2025	17:14	PP075274.D	8.81	3.80
AR1660CCC500	AR1660CCC500	09/29/2025	18:38	PP075413.D	8.80	3.80
IBLK	IBLK	09/29/2025	19:59	PP075417.D	8.80	3.80
PB169881BL	PB169881BL	09/29/2025	20:15	PP075418.D	8.80	3.80
PB169881TB	PB169881TB	09/29/2025	20:48	PP075420.D	8.80	3.80
EME-TS14-01	Q3179-01	09/29/2025	21:04	PP075421.D	8.80	3.79
EME-TS14-01MS	Q3179-02MS	09/29/2025	21:20	PP075422.D	8.80	3.80
EME-TS14-01MSD	Q3179-03MSD	09/29/2025	21:37	PP075423.D	8.80	3.79
AR1660CCC500	AR1660CCC500	09/29/2025	23:47	PP075428.D	8.80	3.80
IBLK	IBLK	09/30/2025	01:08	PP075432.D	8.80	3.80
AR1660CCC500	AR1660CCC500	09/30/2025	08:46	PP075434.D	8.80	3.79
IBLK	IBLK	09/30/2025	09:51	PP075438.D	8.80	3.80
PB169881BS	PB169881BS	09/30/2025	10:08	PP075439.D	8.80	3.80
AR1660CCC500	AR1660CCC500	09/30/2025	16:14	PP075449.D	8.80	3.80



I.BLK	I.BLK	09/30/2025	17:19	PP075453.D	8.80	3.80
-------	-------	------------	-------	------------	------	------

IDENTIFICATION SUMMARY
FOR MULTICOMPONENT ANALYTES

SAMPLE NO.

EME-TS14-01MS

Lab Name: Alliance

Contract: ROYF02

Lab Code: ACE

SDG NO.: Q3179

Lab Sample ID: Q3179-02MS

Date(s) Analyzed: 09/29/2025 09/29/2025

Instrument ID (1): ECD_P

Instrument ID (2): ECD_P

GC Column: (1): ZB-MR1 ID: 0.32 (mm)

GC Column: (2): ZB-MR2 ID: 0.32 (mm)

Data file PP075422.D

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	MEAN CONCENTRATION	%RPD
			FROM	TO			
Aroclor-1016	1	5.809	5.759	5.859	6.76	6.90	2.86
COLUMN 1	2	5.83	5.78	5.88	6.76		
	3	5.892	5.842	5.942	6.57		
	4	5.99	5.94	6.04	7.19		
	5	6.282	6.232	6.332	7.11		
	1	4.893	4.843	4.943	7.44		
COLUMN 2	2	4.951	4.901	5.001	7.17	7.10	
	3	5.072	5.022	5.122	6.84		
	4	5.111	5.061	5.161	6.76		
	5	5.326	5.276	5.376	7.06		
Aroclor-1260	1	7.399	7.349	7.449	7.37	7.10	2.86
COLUMN 1	2	7.652	7.602	7.702	7.15		
	3	8.009	7.959	8.059	6.53		
	4	8.236	8.186	8.286	7.43		
	5	8.563	8.513	8.613	6.92		
	1	6.354	6.304	6.404	7.31		
COLUMN 2	2	6.54	6.49	6.59	6.89	6.90	
	3	6.694	6.644	6.744	7.49		
	4	7.163	7.113	7.213	6.17		
	5	7.403	7.353	7.453	6.45		

IDENTIFICATION SUMMARY
FOR MULTICOMPONENT ANALYTES

SAMPLE NO.

EME-TS14-01MSD

Lab Name: Alliance

Contract: ROYF02

Lab Code: ACE

SDG NO.: Q3179

Lab Sample ID: Q3179-03MSD

Date(s) Analyzed: 09/29/2025 09/29/2025

Instrument ID (1): ECD_P

Instrument ID (2): ECD_P

GC Column: (1): ZB-MR1 ID: 0.32 (mm)

GC Column: (2): ZB-MR2 ID: 0.32 (mm)

Data file PP075423.D

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	MEAN CONCENTRATION	%RPD
			FROM	TO			
Aroclor-1016	1	5.806	5.756	5.856	6.51	6.70	1.48
COLUMN 1	2	5.827	5.777	5.877	6.59		
	3	5.89	5.84	5.94	6.38		
	4	5.987	5.937	6.037	6.92		
	5	6.28	6.23	6.33	6.87		
	1	4.89	4.84	4.94	7.24		
COLUMN 2	2	4.947	4.897	4.997	7.09	6.80	
	3	5.068	5.018	5.118	6.51		
	4	5.109	5.059	5.159	6.50		
	5	5.323	5.273	5.373	6.66		
Aroclor-1260	1	7.397	7.347	7.447	7.14	6.90	5.97
COLUMN 1	2	7.649	7.599	7.699	6.91		
	3	8.007	7.957	8.057	6.32		
	4	8.233	8.183	8.283	7.35		
	5	8.56	8.51	8.61	6.67		
	1	6.351	6.301	6.401	6.85	6.50	
COLUMN 2	2	6.538	6.488	6.588	6.60		
	3	6.691	6.641	6.741	7.15		
	4	7.16	7.11	7.21	5.77		
	5	7.399	7.349	7.449	6.30		

IDENTIFICATION SUMMARY
FOR MULTICOMPONENT ANALYTES

SAMPLE NO.

PB169881BS

Lab Name: Alliance

Contract: ROYF02

Lab Code: ACE

SDG NO.: Q3179

Lab Sample ID: PB169881BS

Date(s) Analyzed: 09/30/2025 09/30/2025

Instrument ID (1): ECD_P

Instrument ID (2): ECD_P

GC Column: (1): ZB-MR1 ID: 0.32 (mm)

GC Column: (2): ZB-MR2 ID: 0.32 (mm)

Data file PP075439.D

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	MEAN CONCENTRATION	%RPD
			FROM	TO			
Aroclor-1016	1	5.808	5.758	5.858	4.92	5.20	1.94
COLUMN 1	2	5.829	5.779	5.879	4.85		
	3	5.892	5.842	5.942	4.97		
	4	5.989	5.939	6.039	5.39		
	5	6.281	6.231	6.331	5.71		
	1	4.892	4.842	4.942	5.05		
COLUMN 2	2	4.95	4.9	5	5.77	5.10	
	3	5.07	5.02	5.12	5.48		
	4	5.11	5.06	5.16	4.41		
	5	5.325	5.275	5.375	4.71		
Aroclor-1260	1	7.398	7.348	7.448	5.42	5.00	2.02
COLUMN 1	2	7.651	7.601	7.701	5.10		
	3	8.009	7.959	8.059	4.60		
	4	8.235	8.185	8.285	5.01		
	5	8.562	8.512	8.612	4.92		
	1	6.352	6.302	6.402	5.43		
COLUMN 2	2	6.538	6.488	6.588	6.32	4.90	
	3	6.692	6.642	6.742	4.82		
	4	7.161	7.111	7.211	3.90		
	5	7.401	7.351	7.451	4.10		



QC SAMPLE DATA

- 1
- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13
- 14
- 15
- 16
- 17
- 18

Report of Analysis

Client:	Weston Solutions, Inc.	Date Collected:	
Project:	RFP 916	Date Received:	
Client Sample ID:	PB169881BL	SDG No.:	Q3179
Lab Sample ID:	PB169881BL	Matrix:	WATER
Analytical Method:	8082A	% Solid:	0
Sample Wt/Vol:	1000	Units:	mL
Soil Aliquot Vol:			uL
Extraction Type:		Test:	SPLP PCB
GPC Factor :	1.0	PH :	
Prep Method :	3510C	Injection Volume :	

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PP075418.D	1	09/29/25 09:15	09/29/25 20:15	PB169881

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
12674-11-2	Aroclor-1016	0.097	U	0.097	0.50	ug/L
11104-28-2	Aroclor-1221	0.13	U	0.13	0.50	ug/L
11141-16-5	Aroclor-1232	0.096	U	0.096	0.50	ug/L
53469-21-9	Aroclor-1242	0.12	U	0.12	0.50	ug/L
12672-29-6	Aroclor-1248	0.071	U	0.071	0.50	ug/L
11097-69-1	Aroclor-1254	0.094	U	0.094	0.50	ug/L
37324-23-5	Aroclor-1262	0.14	U	0.14	0.50	ug/L
11100-14-4	Aroclor-1268	0.11	U	0.11	0.50	ug/L
11096-82-5	Aroclor-1260	0.081	U	0.081	0.50	ug/L
SURROGATES						
877-09-8	Tetrachloro-m-xylene	19.5		30 - 173	97%	SPK: 20
2051-24-3	Decachlorobiphenyl	21.8		10 - 173	109%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP092925\
Data File : PP075418.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 29 Sep 2025 20:15
Operator : YP\AJ
Sample : PB169881BL
Misc :
ALS Vial : 25 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
PB169881BL

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 30 03:02:45 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP092325.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Sep 24 10:55:35 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.656	3.797	25473297	107.6E6	19.468	17.632
2) SA Decachlor...	10.431	8.799	21349930	131.5E6	21.402	21.804

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP092925\
Data File : PP075418.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 29 Sep 2025 20:15
Operator : YP\AJ
Sample : PB169881BL
Misc :
ALS Vial : 25 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
PB169881BL

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 30 03:02:45 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP092325.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Sep 24 10:55:35 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.656 min
Delta R.T.: 0.001 min
Response: 25473297
Conc: 19.47 ng/ml

Instrument :
ECD_P
ClientSampleId :
PB169881BL




#1 Tetrachloro-m-xylene

R.T.: 3.797 min
Delta R.T.: 0.002 min
Response: 107570061
Conc: 17.63 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.431 min
Delta R.T.: -0.004 min
Response: 21349930
Conc: 21.40 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.799 min
Delta R.T.: -0.005 min
Response: 131531031
Conc: 21.80 ng/ml

Report of Analysis

Client:	Weston Solutions, Inc.	Date Collected:	09/23/25
Project:	RFP 916	Date Received:	09/23/25
Client Sample ID:	PIBLK-PP075246.D	SDG No.:	Q3179
Lab Sample ID:	I.BLK-PP075246.D	Matrix:	WATER
Analytical Method:	8082A	% Solid:	0
Sample Wt/Vol:	1000	Units:	mL
Soil Aliquot Vol:			uL
Extraction Type:		Test:	SPLP PCB
GPC Factor :	1.0	PH :	
Prep Method :	5030	Injection Volume :	

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PP075246.D	1		09/23/25	pp092325

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
12674-11-2	Aroclor-1016	0.097	U	0.097	0.50	ug/L
11104-28-2	Aroclor-1221	0.13	U	0.13	0.50	ug/L
11141-16-5	Aroclor-1232	0.096	U	0.096	0.50	ug/L
53469-21-9	Aroclor-1242	0.12	U	0.12	0.50	ug/L
12672-29-6	Aroclor-1248	0.071	U	0.071	0.50	ug/L
11097-69-1	Aroclor-1254	0.094	U	0.094	0.50	ug/L
11096-82-5	Aroclor-1260	0.081	U	0.081	0.50	ug/L
37324-23-5	Aroclor-1262	0.14	U	0.14	0.50	ug/L
11100-14-4	Aroclor-1268	0.11	U	0.11	0.50	ug/L
SURROGATES						
877-09-8	Tetrachloro-m-xylene	17.9		60 - 140	90%	SPK: 20
2051-24-3	Decachlorobiphenyl	19.4		60 - 140	97%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP092325\
Data File : PP075246.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Sep 2025 08:29
Operator : YP\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 23 17:44:23 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP092325.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Sep 23 17:43:53 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.657	3.797	26080915	109.2E6	19.932	17.901
2) SA Decachlor...	10.437	8.805	19354075	120.9E6	19.402	20.045

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP092325\
Data File : PP075246.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Sep 2025 08:29
Operator : YP\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 23 17:44:23 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP092325.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Sep 23 17:43:53 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.657 min
Delta R.T.: 0.002 min
Response: 26080915
Conc: 19.93 ng/ml

Instrument :
ECD_P
ClientSampleId :
I.BLK




#1 Tetrachloro-m-xylene

R.T.: 3.797 min
Delta R.T.: 0.000 min
Response: 109213871
Conc: 17.90 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.437 min
Delta R.T.: 0.003 min
Response: 19354075
Conc: 19.40 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.805 min
Delta R.T.: 0.001 min
Response: 120921024
Conc: 20.04 ng/ml

Report of Analysis

Client:	Weston Solutions, Inc.	Date Collected:	09/29/25
Project:	RFP 916	Date Received:	09/29/25
Client Sample ID:	PIBLK-PP075417.D	SDG No.:	Q3179
Lab Sample ID:	I.BLK-PP075417.D	Matrix:	WATER
Analytical Method:	8082A	% Solid:	0
Sample Wt/Vol:	1000	Units:	mL
Soil Aliquot Vol:			uL
Extraction Type:		Test:	SPLP PCB
GPC Factor :	1.0	PH :	
Prep Method :	5030	Injection Volume :	

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PP075417.D	1		09/29/25	pp092925

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
12674-11-2	Aroclor-1016	0.097	U	0.097	0.50	ug/L
11104-28-2	Aroclor-1221	0.13	U	0.13	0.50	ug/L
11141-16-5	Aroclor-1232	0.096	U	0.096	0.50	ug/L
53469-21-9	Aroclor-1242	0.12	U	0.12	0.50	ug/L
12672-29-6	Aroclor-1248	0.071	U	0.071	0.50	ug/L
11097-69-1	Aroclor-1254	0.094	U	0.094	0.50	ug/L
11096-82-5	Aroclor-1260	0.081	U	0.081	0.50	ug/L
37324-23-5	Aroclor-1262	0.14	U	0.14	0.50	ug/L
11100-14-4	Aroclor-1268	0.11	U	0.11	0.50	ug/L
SURROGATES						
877-09-8	Tetrachloro-m-xylene	17.0		60 - 140	85%	SPK: 20
2051-24-3	Decachlorobiphenyl	21.1		60 - 140	106%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP092925\
Data File : PP075417.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 29 Sep 2025 19:59
Operator : YP\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 30 03:02:32 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP092325.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Sep 24 10:55:35 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
System Monitoring Compounds						
1) SA Tetrachlo...	4.653	3.795	25299183	103.9E6	19.335	17.030
2) SA Decachlor...	10.428	8.797	21093623	134.3E6	21.145	22.269

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP092925\
Data File : PP075417.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 29 Sep 2025 19:59
Operator : YP\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 30 03:02:32 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP092325.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Sep 24 10:55:35 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.653 min
Delta R.T.: -0.002 min
Response: 25299183
Conc: 19.33 ng/ml

Instrument :
ECD_P
ClientSampleId :
I.BLK




#1 Tetrachloro-m-xylene

R.T.: 3.795 min
Delta R.T.: 0.000 min
Response: 103897533
Conc: 17.03 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.428 min
Delta R.T.: -0.006 min
Response: 21093623
Conc: 21.15 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.797 min
Delta R.T.: -0.006 min
Response: 134338003
Conc: 22.27 ng/ml

Report of Analysis

Client:	Weston Solutions, Inc.	Date Collected:	09/30/25
Project:	RFP 916	Date Received:	09/30/25
Client Sample ID:	PIBLK-PP075432.D	SDG No.:	Q3179
Lab Sample ID:	I.BLK-PP075432.D	Matrix:	WATER
Analytical Method:	8082A	% Solid:	0
Sample Wt/Vol:	1000	Units:	mL
Soil Aliquot Vol:			uL
Extraction Type:		Test:	SPLP PCB
GPC Factor :	1.0	PH :	
Prep Method :	5030	Injection Volume :	

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PP075432.D	1		09/30/25	pp092925

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
12674-11-2	Aroclor-1016	0.097	U	0.097	0.50	ug/L
11104-28-2	Aroclor-1221	0.13	U	0.13	0.50	ug/L
11141-16-5	Aroclor-1232	0.096	U	0.096	0.50	ug/L
53469-21-9	Aroclor-1242	0.12	U	0.12	0.50	ug/L
12672-29-6	Aroclor-1248	0.071	U	0.071	0.50	ug/L
11097-69-1	Aroclor-1254	0.094	U	0.094	0.50	ug/L
11096-82-5	Aroclor-1260	0.081	U	0.081	0.50	ug/L
37324-23-5	Aroclor-1262	0.14	U	0.14	0.50	ug/L
11100-14-4	Aroclor-1268	0.11	U	0.11	0.50	ug/L
SURROGATES						
877-09-8	Tetrachloro-m-xylene	18.3		60 - 140	91%	SPK: 20
2051-24-3	Decachlorobiphenyl	21.6		60 - 140	108%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP092925\
Data File : PP075432.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 30 Sep 2025 01:08
Operator : YP\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 30 03:06:13 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP092325.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Sep 24 10:55:35 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
System Monitoring Compounds						
1) SA Tetrachlo...	4.655	3.795	25598525	111.4E6	19.564	18.263
2) SA Decachlor...	10.431	8.797	21592145	137.8E6	21.645	22.851

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP092925\
Data File : PP075432.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 30 Sep 2025 01:08
Operator : YP\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 30 03:06:13 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP092325.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Sep 24 10:55:35 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.655 min
Delta R.T.: 0.000 min
Response: 25598525
Conc: 19.56 ng/ml

Instrument :
ECD_P
ClientSampleId :
I.BLK




#1 Tetrachloro-m-xylene

R.T.: 3.795 min
Delta R.T.: 0.000 min
Response: 111418614
Conc: 18.26 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.431 min
Delta R.T.: -0.004 min
Response: 21592145
Conc: 21.65 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.797 min
Delta R.T.: -0.006 min
Response: 137849618
Conc: 22.85 ng/ml

Report of Analysis

Client:	Weston Solutions, Inc.	Date Collected:	09/30/25
Project:	RFP 916	Date Received:	09/30/25
Client Sample ID:	PIBLK-PP075438.D	SDG No.:	Q3179
Lab Sample ID:	I.BLK-PP075438.D	Matrix:	WATER
Analytical Method:	8082A	% Solid:	0
Sample Wt/Vol:	1000	Units:	mL
Soil Aliquot Vol:			uL
Extraction Type:		Test:	SPLP PCB
GPC Factor :	1.0	PH :	
Prep Method :	5030	Injection Volume :	

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PP075438.D	1		09/30/25	PP093025

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
12674-11-2	Aroclor-1016	0.097	U	0.097	0.50	ug/L
11104-28-2	Aroclor-1221	0.13	U	0.13	0.50	ug/L
11141-16-5	Aroclor-1232	0.096	U	0.096	0.50	ug/L
53469-21-9	Aroclor-1242	0.12	U	0.12	0.50	ug/L
12672-29-6	Aroclor-1248	0.071	U	0.071	0.50	ug/L
11097-69-1	Aroclor-1254	0.094	U	0.094	0.50	ug/L
11096-82-5	Aroclor-1260	0.081	U	0.081	0.50	ug/L
37324-23-5	Aroclor-1262	0.14	U	0.14	0.50	ug/L
11100-14-4	Aroclor-1268	0.11	U	0.11	0.50	ug/L
SURROGATES						
877-09-8	Tetrachloro-m-xylene	17.6		60 - 140	88%	SPK: 20
2051-24-3	Decachlorobiphenyl	22.4		60 - 140	112%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP093025\
Data File : PP075438.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 30 Sep 2025 09:51
Operator : YP\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 01 01:58:21 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP092325.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Sep 24 10:55:35 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.655	3.795	26816385	107.2E6	20.494	17.571
2) SA Decachlor...	10.430	8.797	22372300	145.9E6	22.427m	24.187

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP093025\
Data File : PP075438.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 30 Sep 2025 09:51
Operator : YP\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 01 01:58:21 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP092325.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Sep 24 10:55:35 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.655 min
Delta R.T.: 0.000 min
Response: 26816385
Conc: 20.49 ng/ml

Instrument :
ECD_P
ClientSampleId :
I.BLK




#1 Tetrachloro-m-xylene

R.T.: 3.795 min
Delta R.T.: 0.000 min
Response: 107198043
Conc: 17.57 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.430 min
Delta R.T.: -0.004 min
Response: 22372300
Conc: 22.43 ng/ml m



#2 Decachlorobiphenyl

R.T.: 8.797 min
Delta R.T.: -0.006 min
Response: 145907745
Conc: 24.19 ng/ml

Report of Analysis

Client:	Weston Solutions, Inc.	Date Collected:	09/30/25
Project:	RFP 916	Date Received:	09/30/25
Client Sample ID:	PIBLK-PP075453.D	SDG No.:	Q3179
Lab Sample ID:	I.BLK-PP075453.D	Matrix:	WATER
Analytical Method:	8082A	% Solid:	0
Sample Wt/Vol:	1000	Units:	mL
Soil Aliquot Vol:			uL
Extraction Type:		Test:	SPLP PCB
GPC Factor :	1.0	PH :	
Prep Method :	5030	Injection Volume :	

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PP075453.D	1		09/30/25	PP093025

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
12674-11-2	Aroclor-1016	0.097	U	0.097	0.50	ug/L
11104-28-2	Aroclor-1221	0.13	U	0.13	0.50	ug/L
11141-16-5	Aroclor-1232	0.096	U	0.096	0.50	ug/L
53469-21-9	Aroclor-1242	0.12	U	0.12	0.50	ug/L
12672-29-6	Aroclor-1248	0.071	U	0.071	0.50	ug/L
11097-69-1	Aroclor-1254	0.094	U	0.094	0.50	ug/L
11096-82-5	Aroclor-1260	0.081	U	0.081	0.50	ug/L
37324-23-5	Aroclor-1262	0.14	U	0.14	0.50	ug/L
11100-14-4	Aroclor-1268	0.11	U	0.11	0.50	ug/L
SURROGATES						
877-09-8	Tetrachloro-m-xylene	16.3		60 - 140	81%	SPK: 20
2051-24-3	Decachlorobiphenyl	18.8		60 - 140	94%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP093025\
Data File : PP075453.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 30 Sep 2025 17:19
Operator : YP\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 01 02:10:47 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP092325.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Sep 24 10:55:35 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.656	3.796	23125132	99369981	17.673	16.288
2) SA Decachlor...	10.434	8.799	19504828	113.2E6	19.553	18.757

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP093025\
Data File : PP075453.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 30 Sep 2025 17:19
Operator : YP\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 01 02:10:47 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP092325.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Sep 24 10:55:35 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.656 min
Delta R.T.: 0.000 min
Response: 23125132
Conc: 17.67 ng/ml

Instrument :
ECD_P
ClientSampleId :
I.BLK




#1 Tetrachloro-m-xylene

R.T.: 3.796 min
Delta R.T.: 0.000 min
Response: 99369981
Conc: 16.29 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.434 min
Delta R.T.: 0.000 min
Response: 19504828
Conc: 19.55 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.799 min
Delta R.T.: -0.005 min
Response: 113152955
Conc: 18.76 ng/ml

Report of Analysis

Client:	Weston Solutions, Inc.	Date Collected:	
Project:	RFP 916	Date Received:	
Client Sample ID:	PB169881BS	SDG No.:	Q3179
Lab Sample ID:	PB169881BS	Matrix:	WATER
Analytical Method:	8082A	% Solid:	0
Sample Wt/Vol:	1000	Units:	mL
Soil Aliquot Vol:			uL
Extraction Type:		Test:	SPLP PCB
GPC Factor :	1.0	PH :	
Prep Method :	3510C	Injection Volume :	

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PP075439.D	1	09/29/25 09:15	09/30/25 10:08	PB169881

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
12674-11-2	Aroclor-1016	5.20		0.097	0.50	ug/L
11104-28-2	Aroclor-1221	0.13	U	0.13	0.50	ug/L
11141-16-5	Aroclor-1232	0.096	U	0.096	0.50	ug/L
53469-21-9	Aroclor-1242	0.12	U	0.12	0.50	ug/L
12672-29-6	Aroclor-1248	0.071	U	0.071	0.50	ug/L
11097-69-1	Aroclor-1254	0.094	U	0.094	0.50	ug/L
37324-23-5	Aroclor-1262	0.14	U	0.14	0.50	ug/L
11100-14-4	Aroclor-1268	0.11	U	0.11	0.50	ug/L
11096-82-5	Aroclor-1260	5.00		0.081	0.50	ug/L
SURROGATES						
877-09-8	Tetrachloro-m-xylene	18.5		30 - 173	93%	SPK: 20
2051-24-3	Decachlorobiphenyl	20.2		10 - 173	101%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP093025\
 Data File : PP075439.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 30 Sep 2025 10:08
 Operator : YP\AJ
 Sample : PB169881BS
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 PB169881BS

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 01 01:59:10 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP092325.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Sep 24 10:55:35 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.656	3.795	24242568	79092577	18.527	12.964 #
2) SA Decachlor...	10.431	8.799	19309605	121.8E6	19.357	20.185
Target Compounds						
3) L1 AR-1016-1	5.808	4.892	23669693	283.3E6	492.045	505.017
4) L1 AR-1016-2	5.829	4.950	34822734	148.4E6	485.479	577.327
5) L1 AR-1016-3	5.892	5.070	23139378	85052263	497.292	547.776
6) L1 AR-1016-4	5.989	5.110	19218461	64792990	539.289	440.865
7) L1 AR-1016-5	6.281	5.325	19270211	76978671	571.455	471.157
31) L7 AR-1260-1	7.398	6.352	32564653	220.3E6	541.665	542.643
32) L7 AR-1260-2	7.651	6.538	37153841	363.9E6	510.369	632.020
33) L7 AR-1260-3	8.009	6.692	24889039	206.5E6	459.673	482.285
34) L7 AR-1260-4	8.235	7.161	28311289	168.1E6	500.755m	389.841
35) L7 AR-1260-5	8.562	7.401	54027132	431.6E6	492.013	410.264



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP093025\
Data File : PP075439.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 30 Sep 2025 10:08
Operator : YP\AJ
Sample : PB169881BS
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
PB169881BS

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 01 01:59:10 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP092325.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Sep 24 10:55:35 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.656 min
Delta R.T.: 0.001 min
Response: 24242568
Conc: 18.53 ng/ml

Instrument :
ECD_P
ClientSampleId :
PB169881BS




#1 Tetrachloro-m-xylene

R.T.: 3.795 min
Delta R.T.: 0.000 min
Response: 79092577
Conc: 12.96 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.431 min
Delta R.T.: -0.003 min
Response: 19309605
Conc: 19.36 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.799 min
Delta R.T.: -0.004 min
Response: 121764769
Conc: 20.18 ng/ml



#3 AR-1016-1


R.T.: 5.808 min
Delta R.T.: 0.000 min
Response: 23669693
Conc: 492.04 ng/ml

Instrument :
ECD_P
ClientSampleId :
PB169881BS




#3 AR-1016-1

R.T.: 4.892 min
Delta R.T.: 0.000 min
Response: 283321100
Conc: 505.02 ng/ml



#4 AR-1016-2

R.T.: 5.829 min
Delta R.T.: 0.000 min
Response: 34822734
Conc: 485.48 ng/ml




#4 AR-1016-2

R.T.: 4.950 min
Delta R.T.: 0.000 min
Response: 148374201
Conc: 577.33 ng/ml

#5 AR-1016-3


R.T.: 5.892 min
Delta R.T.: 0.000 min
Response: 23139378
Conc: 497.29 ng/ml

Instrument :
ECD_P
ClientSampleId :
PB169881BS




#5 AR-1016-3

R.T.: 5.070 min
Delta R.T.: -0.001 min
Response: 85052263
Conc: 547.78 ng/ml




#6 AR-1016-4


R.T.: 5.989 min
Delta R.T.: 0.000 min
Response: 19218461
Conc: 539.29 ng/ml



#6 AR-1016-4

R.T.: 5.110 min
Delta R.T.: -0.002 min
Response: 64792990
Conc: 440.86 ng/ml






#7 AR-1016-5


R.T.: 6.281 min
Delta R.T.: 0.000 min
Response: 19270211
Conc: 571.46 ng/ml

Instrument :
ECD_P
ClientSampleId :
PB169881BS




#7 AR-1016-5

R.T.: 5.325 min
Delta R.T.: 0.000 min
Response: 76978671
Conc: 471.16 ng/ml




#31 AR-1260-1

R.T.: 7.398 min
Delta R.T.: 0.000 min
Response: 32564653
Conc: 541.67 ng/ml



#31 AR-1260-1


R.T.: 6.352 min
Delta R.T.: -0.003 min
Response: 220277801
Conc: 542.64 ng/ml



#32 AR-1260-2


R.T.: 7.651 min
Delta R.T.: 0.000 min
Response: 37153841
Conc: 510.37 ng/ml

Instrument :
ECD_P
ClientSampleId :
PB169881BS




#32 AR-1260-2

R.T.: 6.538 min
Delta R.T.: -0.003 min
Response: 363935805
Conc: 632.02 ng/ml




#33 AR-1260-3

R.T.: 8.009 min
Delta R.T.: 0.000 min
Response: 24889039
Conc: 459.67 ng/ml



#33 AR-1260-3


R.T.: 6.692 min
Delta R.T.: -0.003 min
Response: 206458873
Conc: 482.28 ng/ml



#34 AR-1260-4


R.T.: 8.235 min
Delta R.T.: -0.002 min
Response: 28311289
Conc: 500.76 ng/ml

Instrument :
ECD_P
ClientSampleId :
PB169881BS




#34 AR-1260-4

R.T.: 7.161 min
Delta R.T.: -0.003 min
Response: 168108901
Conc: 389.84 ng/ml



#35 AR-1260-5

R.T.: 8.562 min
Delta R.T.: -0.001 min
Response: 54027132
Conc: 492.01 ng/ml



#35 AR-1260-5

R.T.: 7.401 min
Delta R.T.: -0.003 min
Response: 431592915
Conc: 410.26 ng/ml

Report of Analysis

Client:	Weston Solutions, Inc.	Date Collected:	09/23/25
Project:	RFP 916	Date Received:	09/23/25
Client Sample ID:	EME-TS14-01MS	SDG No.:	Q3179
Lab Sample ID:	Q3179-02MS	Matrix:	WATER
Analytical Method:	8082A	% Solid:	0
Sample Wt/Vol:	1000	Units:	mL
Soil Aliquot Vol:			uL
Extraction Type:		Test:	SPLP PCB
GPC Factor :	1.0	PH :	
Prep Method :	3510C	Injection Volume :	

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PP075422.D	1	09/29/25 09:15	09/29/25 21:20	PB169881

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
12674-11-2	Aroclor-1016	7.10		0.097	0.50	ug/L
11104-28-2	Aroclor-1221	0.13	U	0.13	0.50	ug/L
11141-16-5	Aroclor-1232	0.096	U	0.096	0.50	ug/L
53469-21-9	Aroclor-1242	0.12	U	0.12	0.50	ug/L
12672-29-6	Aroclor-1248	0.071	U	0.071	0.50	ug/L
11097-69-1	Aroclor-1254	0.094	U	0.094	0.50	ug/L
37324-23-5	Aroclor-1262	0.14	U	0.14	0.50	ug/L
11100-14-4	Aroclor-1268	0.11	U	0.11	0.50	ug/L
11096-82-5	Aroclor-1260	7.10		0.081	0.50	ug/L
SURROGATES						
877-09-8	Tetrachloro-m-xylene	23.4		30 - 173	117%	SPK: 20
2051-24-3	Decachlorobiphenyl	23.5		10 - 173	118%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP092925\
 Data File : PP075422.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 29 Sep 2025 21:20
 Operator : YP\AJ
 Sample : Q3179-02MS
 Misc :
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 EME-TS14-01MS

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 09/30/2025
 Supervised By :mohammad ahmed 10/01/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 30 03:03:37 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP092325.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Sep 24 10:55:35 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.658	3.798	30592438	88698059	23.380	14.539 #
2) SA Decachlor...	10.433	8.800	22459322	141.9E6	22.515	23.520
Target Compounds						
3) L1 AR-1016-1	5.809	4.893	32532707	417.7E6	676.288	744.478
4) L1 AR-1016-2	5.830	4.951	48493575	184.2E6	676.070	716.775
5) L1 AR-1016-3	5.892	5.072	30553929	106.2E6	656.639	683.764
6) L1 AR-1016-4	5.990	5.111	25639066	99395541	719.458	676.307
7) L1 AR-1016-5	6.282	5.326	23962861	115.4E6	710.615	706.253
31) L7 AR-1260-1	7.399	6.354	44287261	296.6E6	736.654	730.635
32) L7 AR-1260-2	7.652	6.540	52086275	396.9E6	715.490	689.234
33) L7 AR-1260-3	8.009	6.694	35335307	320.5E6	652.604	748.624
34) L7 AR-1260-4	8.236	7.163	41993608	266.3E6	742.761m	617.484
35) L7 AR-1260-5	8.563	7.403	75957125	678.7E6	691.725	645.164

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP092925\
Data File : PP075422.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 29 Sep 2025 21:20
Operator : YP\AJ
Sample : Q3179-02MS
Misc :
ALS Vial : 29 Sample Multiplier: 1

Instrument :

ECD_P

ClientSampleId :

EME-TS14-01MS

Manual Integrations



APPROVED


Reviewed By :Yogesh Patel 09/30/2025

Supervised By :mohammad ahmed 10/01/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 30 03:03:37 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP092325.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Sep 24 10:55:35 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.658 min
Delta R.T.: 0.003 min
Response: 30592438
Conc: 23.38 ng/ml

Instrument :
ECD_P
ClientSampleId :
EME-TS14-01MS


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 09/30/2025
Supervised By :mohammad ahmed 10/01/2025




#1 Tetrachloro-m-xylene

R.T.: 3.798 min
Delta R.T.: 0.002 min
Response: 88698059
Conc: 14.54 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.433 min
Delta R.T.: -0.002 min
Response: 22459322
Conc: 22.51 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.800 min
Delta R.T.: -0.004 min
Response: 141882514
Conc: 23.52 ng/ml




#3 AR-1016-1

R.T.: 5.809 min
Delta R.T.: 0.001 min
Response: 32532707
Conc: 676.29 ng/ml

Instrument :
ECD_P
ClientSampleId :
EME-TS14-01MS


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 09/30/2025
Supervised By :mohammad ahmed 10/01/2025




#3 AR-1016-1

R.T.: 4.893 min
Delta R.T.: 0.000 min
Response: 417661998
Conc: 744.48 ng/ml




#4 AR-1016-2

R.T.: 5.830 min
Delta R.T.: 0.001 min
Response: 48493575
Conc: 676.07 ng/ml



#4 AR-1016-2

R.T.: 4.951 min
Delta R.T.: 0.000 min
Response: 184212671
Conc: 716.77 ng/ml




#5 AR-1016-3

R.T.: 5.892 min
Delta R.T.: 0.000 min
Response: 30553929
Conc: 656.64 ng/ml

Instrument :
ECD_P
ClientSampleId :
EME-TS14-01MS


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 09/30/2025
Supervised By :mohammad ahmed 10/01/2025




#5 AR-1016-3

R.T.: 5.072 min
Delta R.T.: 0.000 min
Response: 106166870
Conc: 683.76 ng/ml




#6 AR-1016-4

R.T.: 5.990 min
Delta R.T.: 0.000 min
Response: 25639066
Conc: 719.46 ng/ml



#6 AR-1016-4

R.T.: 5.111 min
Delta R.T.: 0.000 min
Response: 99395541
Conc: 676.31 ng/ml




#7 AR-1016-5

R.T.: 6.282 min
Delta R.T.: 0.000 min
Response: 23962861
Conc: 710.61 ng/ml

Instrument :
ECD_P
ClientSampleId :
EME-TS14-01MS


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 09/30/2025
Supervised By :mohammad ahmed 10/01/2025




#7 AR-1016-5

R.T.: 5.326 min
Delta R.T.: 0.000 min
Response: 115389082
Conc: 706.25 ng/ml




#31 AR-1260-1

R.T.: 7.399 min
Delta R.T.: 0.000 min
Response: 44287261
Conc: 736.65 ng/ml



#31 AR-1260-1

R.T.: 6.354 min
Delta R.T.: 0.000 min
Response: 296590704
Conc: 730.64 ng/ml




#32 AR-1260-2

R.T.: 7.652 min
Delta R.T.: 0.000 min
Response: 52086275
Conc: 715.49 ng/ml

Instrument :
ECD_P
ClientSampleId :
EME-TS14-01MS


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 09/30/2025
Supervised By :mohammad ahmed 10/01/2025




#32 AR-1260-2

R.T.: 6.540 min
Delta R.T.: 0.000 min
Response: 396881648
Conc: 689.23 ng/ml




#33 AR-1260-3

R.T.: 8.009 min
Delta R.T.: 0.000 min
Response: 35335307
Conc: 652.60 ng/ml



#33 AR-1260-3

R.T.: 6.694 min
Delta R.T.: -0.001 min
Response: 320474859
Conc: 748.62 ng/ml




#34 AR-1260-4

R.T.: 8.236 min
Delta R.T.: -0.001 min
Response: 41993608
Conc: 742.76 ng/ml

Instrument :
ECD_P
ClientSampleId :
EME-TS14-01MS


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 09/30/2025
Supervised By :mohammad ahmed 10/01/2025




#34 AR-1260-4

R.T.: 7.163 min
Delta R.T.: -0.002 min
Response: 266273894
Conc: 617.48 ng/ml



#35 AR-1260-5

R.T.: 8.563 min
Delta R.T.: 0.000 min
Response: 75957125
Conc: 691.72 ng/ml



#35 AR-1260-5

R.T.: 7.403 min
Delta R.T.: 0.000 min
Response: 678705486
Conc: 645.16 ng/ml

Report of Analysis

Client:	Weston Solutions, Inc.	Date Collected:	09/23/25
Project:	RFP 916	Date Received:	09/23/25
Client Sample ID:	EME-TS14-01MSD	SDG No.:	Q3179
Lab Sample ID:	Q3179-03MSD	Matrix:	WATER
Analytical Method:	8082A	% Solid:	0
Sample Wt/Vol:	1000	Units:	mL
Soil Aliquot Vol:			uL
Extraction Type:		Test:	SPLP PCB
GPC Factor :	1.0	PH :	
Prep Method :	3510C	Injection Volume :	

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PP075423.D	1	09/29/25 09:15	09/29/25 21:37	PB169881

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
12674-11-2	Aroclor-1016	6.80		0.097	0.50	ug/L
11104-28-2	Aroclor-1221	0.13	U	0.13	0.50	ug/L
11141-16-5	Aroclor-1232	0.096	U	0.096	0.50	ug/L
53469-21-9	Aroclor-1242	0.12	U	0.12	0.50	ug/L
12672-29-6	Aroclor-1248	0.071	U	0.071	0.50	ug/L
11097-69-1	Aroclor-1254	0.094	U	0.094	0.50	ug/L
37324-23-5	Aroclor-1262	0.14	U	0.14	0.50	ug/L
11100-14-4	Aroclor-1268	0.11	U	0.11	0.50	ug/L
11096-82-5	Aroclor-1260	6.90		0.081	0.50	ug/L
SURROGATES						
877-09-8	Tetrachloro-m-xylene	23.1		30 - 173	116%	SPK: 20
2051-24-3	Decachlorobiphenyl	23.8		10 - 173	119%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP092925\
 Data File : PP075423.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 29 Sep 2025 21:37
 Operator : YP\AJ
 Sample : Q3179-03MSD
 Misc :
 ALS Vial : 30 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 EME-TS14-01MSD

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 09/30/2025
 Supervised By :mohammad ahmed 10/01/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 30 03:03:50 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP092325.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Sep 24 10:55:35 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.655	3.790	30273744	127.8E6	23.137	20.956m
2) SA Decachlor...	10.429	8.797	22665144	143.5E6	22.721	23.793
Target Compounds						
3) L1 AR-1016-1	5.806	4.890	31301113	406.2E6	650.686	723.978
4) L1 AR-1016-2	5.827	4.947	47234170	182.2E6	658.512	709.101
5) L1 AR-1016-3	5.890	5.068	29709338	101.1E6	638.488	651.423
6) L1 AR-1016-4	5.987	5.109	24676051	95582201	692.434	650.361
7) L1 AR-1016-5	6.280	5.323	23166712	108.7E6	687.005	665.547
31) L7 AR-1260-1	7.397	6.351	42911009	278.3E6	713.762	685.459
32) L7 AR-1260-2	7.649	6.538	50312246	380.2E6	691.121	660.180
33) L7 AR-1260-3	8.007	6.691	34232934	306.1E6	632.244	715.155
34) L7 AR-1260-4	8.233	7.160	41543748	249.0E6	734.804m	577.492
35) L7 AR-1260-5	8.560	7.399	73228215	662.5E6	666.873	629.800

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP092925\
Data File : PP075423.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 29 Sep 2025 21:37
Operator : YP\AJ
Sample : Q3179-03MSD
Misc :
ALS Vial : 30 Sample Multiplier: 1

Instrument :

ECD_P

ClientSampleId :

EME-TS14-01MSD

Manual Integrations



APPROVED


Reviewed By :Yogesh Patel 09/30/2025

Supervised By :mohammad ahmed 10/01/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 30 03:03:50 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP092325.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Sep 24 10:55:35 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.655 min
Delta R.T.: 0.000 min
Response: 30273744
Conc: 23.14 ng/ml

Instrument :
ECD_P
ClientSampleId :
EME-TS14-01MSD


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 09/30/2025
Supervised By :mohammad ahmed 10/01/2025




#1 Tetrachloro-m-xylene

R.T.: 3.790 min
Delta R.T.: -0.006 min
Response: 127848025
Conc: 20.96 ng/ml m




#2 Decachlorobiphenyl

R.T.: 10.429 min
Delta R.T.: -0.006 min
Response: 22665144
Conc: 22.72 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.797 min
Delta R.T.: -0.007 min
Response: 143533117
Conc: 23.79 ng/ml




#3 AR-1016-1

R.T.: 5.806 min
Delta R.T.: -0.002 min
Response: 31301113
Conc: 650.69 ng/ml

Instrument :
ECD_P
ClientSampleId :
EME-TS14-01MSD


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 09/30/2025
Supervised By :mohammad ahmed 10/01/2025




#3 AR-1016-1

R.T.: 4.890 min
Delta R.T.: -0.003 min
Response: 406161662
Conc: 723.98 ng/ml




#4 AR-1016-2

R.T.: 5.827 min
Delta R.T.: -0.001 min
Response: 47234170
Conc: 658.51 ng/ml



#4 AR-1016-2

R.T.: 4.947 min
Delta R.T.: -0.004 min
Response: 182240400
Conc: 709.10 ng/ml




#5 AR-1016-3

R.T.: 5.890 min
Delta R.T.: -0.001 min
Response: 29709338
Conc: 638.49 ng/ml

Instrument :
ECD_P
ClientSampleId :
EME-TS14-01MSD


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 09/30/2025
Supervised By :mohammad ahmed 10/01/2025




#5 AR-1016-3

R.T.: 5.068 min
Delta R.T.: -0.003 min
Response: 101145334
Conc: 651.42 ng/ml




#6 AR-1016-4

R.T.: 5.987 min
Delta R.T.: -0.002 min
Response: 24676051
Conc: 692.43 ng/ml



#6 AR-1016-4

R.T.: 5.109 min
Delta R.T.: -0.003 min
Response: 95582201
Conc: 650.36 ng/ml




#7 AR-1016-5

R.T.: 6.280 min
Delta R.T.: -0.002 min
Response: 23166712
Conc: 687.01 ng/ml

Instrument :
ECD_P
ClientSampleId :
EME-TS14-01MSD


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 09/30/2025
Supervised By :mohammad ahmed 10/01/2025




#7 AR-1016-5

R.T.: 5.323 min
Delta R.T.: -0.003 min
Response: 108738586
Conc: 665.55 ng/ml




#31 AR-1260-1

R.T.: 7.397 min
Delta R.T.: -0.002 min
Response: 42911009
Conc: 713.76 ng/ml



#31 AR-1260-1

R.T.: 6.351 min
Delta R.T.: -0.004 min
Response: 278251953
Conc: 685.46 ng/ml




#32 AR-1260-2

R.T.: 7.649 min
Delta R.T.: -0.002 min
Response: 50312246
Conc: 691.12 ng/ml

Instrument :
ECD_P
ClientSampleId :
EME-TS14-01MSD


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 09/30/2025
Supervised By :mohammad ahmed 10/01/2025




#32 AR-1260-2

R.T.: 6.538 min
Delta R.T.: -0.004 min
Response: 380151145
Conc: 660.18 ng/ml




#33 AR-1260-3

R.T.: 8.007 min
Delta R.T.: -0.002 min
Response: 34232934
Conc: 632.24 ng/ml



#33 AR-1260-3

R.T.: 6.691 min
Delta R.T.: -0.004 min
Response: 306147210
Conc: 715.16 ng/ml




#34 AR-1260-4

R.T.: 8.233 min
Delta R.T.: -0.004 min
Response: 41543748
Conc: 734.80 ng/ml

Instrument :
ECD_P
ClientSampleId :
EME-TS14-01MSD


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 09/30/2025
Supervised By :mohammad ahmed 10/01/2025




#34 AR-1260-4

R.T.: 7.160 min
Delta R.T.: -0.005 min
Response: 249028438
Conc: 577.49 ng/ml



#35 AR-1260-5

R.T.: 8.560 min
Delta R.T.: -0.003 min
Response: 73228215
Conc: 666.87 ng/ml



#35 AR-1260-5

R.T.: 7.399 min
Delta R.T.: -0.005 min
Response: 662542689
Conc: 629.80 ng/ml

Manual Integration Report

Sequence:	pp092325	Instrument	ECD_p
-----------	----------	------------	-------

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
AR1660ICC250	PP075250.D	AR-1260-4	yogesh	9/24/2025 8:09:44 AM	mohammad	9/25/2025 2:25:37	Peak Integrated by Software
AR1660ICC050	PP075251.D	AR-1016-3	yogesh	9/24/2025 8:09:46 AM	mohammad	9/25/2025 2:25:37	Peak Integrated by Software
AR1660ICC050	PP075251.D	AR-1016-5	yogesh	9/24/2025 8:09:46 AM	mohammad	9/25/2025 2:25:37	Peak Integrated by Software
AR1660ICC050	PP075251.D	AR-1260-1	yogesh	9/24/2025 8:09:46 AM	mohammad	9/25/2025 2:25:37	Peak Integrated by Software
AR1660ICC050	PP075251.D	AR-1260-3 #2	yogesh	9/24/2025 8:09:46 AM	mohammad	9/25/2025 2:25:37	Peak Integrated by Software
AR1660ICC050	PP075251.D	AR-1260-4	yogesh	9/24/2025 8:09:46 AM	mohammad	9/25/2025 2:25:37	Peak Integrated by Software
AR1660ICC050	PP075251.D	AR-1260-5 #2	yogesh	9/24/2025 8:09:46 AM	mohammad	9/25/2025 2:25:37	Peak Integrated by Software
AR1660ICC050	PP075251.D	Decachlorobiphenyl	yogesh	9/24/2025 8:09:46 AM	mohammad	9/25/2025 2:25:37	Peak Integrated by Software
AR1242ICC1000	PP075254.D	AR-1242-4 #2	yogesh	9/24/2025 8:09:49 AM	mohammad	9/25/2025 2:25:37	Peak Integrated by Software
AR1242ICC750	PP075255.D	AR-1242-4 #2	yogesh	9/24/2025 8:09:51 AM	mohammad	9/25/2025 2:25:37	Peak Integrated by Software
AR1242ICC500	PP075256.D	AR-1242-4 #2	yogesh	9/24/2025 8:09:53 AM	mohammad	9/25/2025 2:25:37	Peak Integrated by Software
AR1242ICC250	PP075257.D	AR-1242-4 #2	yogesh	9/24/2025 8:09:55 AM	mohammad	9/25/2025 2:25:37	Peak Integrated by Software
AR1242ICC050	PP075258.D	AR-1242-5	yogesh	9/24/2025 8:09:57 AM	mohammad	9/25/2025 2:25:37	Peak Integrated by Software

Manual Integration Report

Sequence:	pp092325	Instrument	ECD_p
-----------	----------	------------	-------

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
AR1242ICC050	PP075258.D	Tetrachloro-m-xylene	yogesh	9/24/2025 8:09:57 AM	mohammad	9/25/2025 2:25:37	Peak Integrated by Software
AR1248ICC1000	PP075259.D	AR-1248-3 #2	yogesh	9/24/2025 8:09:59 AM	mohammad	9/25/2025 2:25:37	Peak Integrated by Software
AR1248ICC750	PP075260.D	AR-1248-3 #2	yogesh	9/24/2025 8:10:01 AM	mohammad	9/25/2025 2:25:37	Peak Integrated by Software
AR1248ICC500	PP075261.D	AR-1248-3 #2	yogesh	9/24/2025 8:10:03 AM	mohammad	9/25/2025 2:25:37	Peak Integrated by Software
AR1248ICC250	PP075262.D	AR-1248-3 #2	yogesh	9/24/2025 8:10:05 AM	mohammad	9/25/2025 2:25:37	Peak Integrated by Software
AR1248ICC050	PP075263.D	AR-1248-3 #2	yogesh	9/24/2025 8:10:06 AM	mohammad	9/25/2025 2:25:37	Peak Integrated by Software
AR1248ICC050	PP075263.D	AR-1248-5	yogesh	9/24/2025 8:10:06 AM	mohammad	9/25/2025 2:25:37	Peak Integrated by Software
AR1248ICC050	PP075263.D	Tetrachloro-m-xylene	yogesh	9/24/2025 8:10:06 AM	mohammad	9/25/2025 2:25:37	Peak Integrated by Software
AR1254ICC1000	PP075264.D	AR-1254-1	yogesh	9/24/2025 8:10:09 AM	mohammad	9/25/2025 2:25:37	Peak Integrated by Software
AR1254ICC750	PP075265.D	AR-1254-1	yogesh	9/24/2025 8:10:11 AM	mohammad	9/25/2025 2:25:37	Peak Integrated by Software
AR1254ICC500	PP075266.D	AR-1254-1	yogesh	9/24/2025 8:10:12 AM	mohammad	9/25/2025 2:25:37	Peak Integrated by Software
AR1254ICC250	PP075267.D	AR-1254-1	yogesh	9/24/2025 8:10:15 AM	mohammad	9/25/2025 2:25:37	Peak Integrated by Software
AR1254ICC050	PP075268.D	AR-1254-1	yogesh	9/24/2025 8:10:16 AM	mohammad	9/25/2025 2:25:37	Peak Integrated by Software

Manual Integration Report

Sequence:	pp092325	Instrument	ECD_p
-----------	----------	------------	-------

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
AR1268ICC050	PP075274.D	AR-1268-1	yogesh	9/24/2025 8:10:18 AM	mohammad	9/25/2025 2:25:37	Peak Integrated by Software
AR1268ICC050	PP075274.D	AR-1268-3 #2	yogesh	9/24/2025 8:10:18 AM	mohammad	9/25/2025 2:25:37	Peak Integrated by Software
AR1268ICC050	PP075274.D	AR-1268-4 #2	yogesh	9/24/2025 8:10:18 AM	mohammad	9/25/2025 2:25:37	Peak Integrated by Software
AR1268ICC050	PP075274.D	AR-1268-5 #2	yogesh	9/24/2025 8:10:18 AM	mohammad	9/25/2025 2:25:37	Peak Integrated by Software
AR1268ICC050	PP075274.D	Tetrachloro-m-xylene #2	yogesh	9/24/2025 8:10:18 AM	mohammad	9/25/2025 2:25:37	Peak Integrated by Software
PP092325ICV500	PP075275.D	Decachlorobiphenyl #2	yogesh	9/24/2025 8:10:21 AM	mohammad	9/25/2025 2:25:37	Peak Integrated by Software
AR1242ICV500	PP075276.D	AR-1242-4 #2	yogesh	9/24/2025 8:10:24 AM	mohammad	9/25/2025 2:25:37	Peak Integrated by Software
AR1242ICV500	PP075276.D	Decachlorobiphenyl #2	yogesh	9/24/2025 8:10:24 AM	mohammad	9/25/2025 2:25:37	Peak Integrated by Software
AR1248ICV500	PP075277.D	AR-1248-3 #2	yogesh	9/24/2025 8:10:26 AM	mohammad	9/25/2025 2:25:37	Peak Integrated by Software
AR1254ICV500	PP075278.D	AR-1254-1	yogesh	9/24/2025 8:10:28 AM	mohammad	9/25/2025 2:25:37	Peak Integrated by Software
AR1254ICV500	PP075278.D	Decachlorobiphenyl #2	yogesh	9/24/2025 8:10:28 AM	mohammad	9/25/2025 2:25:37	Peak Integrated by Software
AR1268ICV500	PP075279.D	Decachlorobiphenyl #2	yogesh	9/24/2025 8:10:30 AM	mohammad	9/25/2025 2:25:37	Peak Integrated by Software



Manual Integration Report

Sequence:	pp092325	Instrument	ECD_p
-----------	----------	------------	-------

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
-----------	---------	-----------	-----------	-----------	---------------	---------------	--------

Manual Integration Report

Sequence:	pp092925	Instrument	ECD_p
-----------	----------	------------	-------

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
AR1242CCC500	PP075386.D	AR-1242-4 #2	yogesh	9/30/2025 8:46:44 AM	mohammad	10/1/2025 1:40:34	Peak Integrated by Software
AR1248CCC500	PP075387.D	AR-1248-3 #2	yogesh	9/30/2025 8:46:46 AM	mohammad	10/1/2025 1:40:34	Peak Integrated by Software
AR1254CCC500	PP075388.D	AR-1254-1	yogesh	9/30/2025 8:46:47 AM	mohammad	10/1/2025 1:40:34	Peak Integrated by Software
AR1242CCC500	PP075399.D	AR-1242-4 #2	yogesh	9/30/2025 8:46:49 AM	mohammad	10/1/2025 1:40:34	Peak Integrated by Software
AR1248CCC500	PP075400.D	AR-1248-3 #2	yogesh	9/30/2025 8:46:52 AM	mohammad	10/1/2025 1:40:34	Peak Integrated by Software
AR1254CCC500	PP075401.D	AR-1254-1	yogesh	9/30/2025 8:46:53 AM	mohammad	10/1/2025 1:40:34	Peak Integrated by Software
I.BLK	PP075402.D	Decachlorobiphenyl	yogesh	9/30/2025 8:46:55 AM	mohammad	10/1/2025 1:40:34	Peak Integrated by Software
AR1242CCC500	PP075414.D	AR-1242-4 #2	yogesh	9/30/2025 8:46:59 AM	mohammad	10/1/2025 1:40:34	Peak Integrated by Software
AR1242CCC500	PP075414.D	Decachlorobiphenyl #2	yogesh	9/30/2025 8:46:59 AM	mohammad	10/1/2025 1:40:34	Peak Integrated by Software
AR1248CCC500	PP075415.D	AR-1248-3 #2	yogesh	9/30/2025 8:47:00 AM	mohammad	10/1/2025 1:40:34	Peak Integrated by Software
AR1248CCC500	PP075415.D	Decachlorobiphenyl #2	yogesh	9/30/2025 8:47:00 AM	mohammad	10/1/2025 1:40:34	Peak Integrated by Software
AR1254CCC500	PP075416.D	AR-1254-1	yogesh	9/30/2025 8:47:02 AM	mohammad	10/1/2025 1:40:34	Peak Integrated by Software
AR1254CCC500	PP075416.D	Decachlorobiphenyl #2	yogesh	9/30/2025 8:47:02 AM	mohammad	10/1/2025 1:40:34	Peak Integrated by Software

Manual Integration Report

Sequence:	pp092925	Instrument	ECD_p
-----------	----------	------------	-------

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
Q3179-01	PP075421.D	Tetrachloro-m-xylene #2	yogesh	9/30/2025 8:47:04 AM	mohammad	10/1/2025 1:40:34	Peak Integrated by Software
Q3179-02MS	PP075422.D	AR-1260-4	yogesh	9/30/2025 8:47:07 AM	mohammad	10/1/2025 1:40:34	Peak Integrated by Software
Q3179-03MSD	PP075423.D	AR-1260-4	yogesh	9/30/2025 8:47:08 AM	mohammad	10/1/2025 1:40:34	Peak Integrated by Software
Q3179-03MSD	PP075423.D	Tetrachloro-m-xylene #2	yogesh	9/30/2025 8:47:08 AM	mohammad	10/1/2025 1:40:34	Peak Integrated by Software
AR1660CCC500	PP075428.D	Decachlorobiphenyl #2	yogesh	9/30/2025 8:47:11 AM	mohammad	10/1/2025 1:40:34	Peak Integrated by Software
AR1242CCC500	PP075429.D	AR-1242-2 #2	yogesh	9/30/2025 8:47:13 AM	mohammad	10/1/2025 1:40:34	Peak Integrated by Software
AR1242CCC500	PP075429.D	AR-1242-4 #2	yogesh	9/30/2025 8:47:13 AM	mohammad	10/1/2025 1:40:34	Peak Integrated by Software
AR1248CCC500	PP075430.D	AR-1248-3 #2	yogesh	9/30/2025 8:47:14 AM	mohammad	10/1/2025 1:40:34	Peak Integrated by Software
AR1254CCC500	PP075431.D	AR-1254-1	yogesh	9/30/2025 8:47:16 AM	mohammad	10/1/2025 1:40:34	Peak Integrated by Software

Manual Integration Report

Sequence:	PP093025	Instrument	ECD_p
-----------	----------	------------	-------

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
AR1242CCC500	PP075435.D	AR-1242-4 #2	yogesh	10/1/2025 7:35:39 AM	Sohil	10/2/2025 3:01:54	Peak Integrated by Software
AR1248CCC500	PP075436.D	AR-1248-3 #2	yogesh	10/1/2025 7:35:41 AM	Sohil	10/2/2025 3:01:56	Peak Integrated by Software
AR1248CCC500	PP075436.D	Decachlorobiphenyl	yogesh	10/1/2025 7:35:41 AM	Sohil	10/2/2025 3:01:56	Peak Integrated by Software
AR1254CCC500	PP075437.D	AR-1254-1	yogesh	10/1/2025 7:35:42 AM	Sohil	10/2/2025 3:01:59	Peak Integrated by Software
I.BLK	PP075438.D	Decachlorobiphenyl	yogesh	10/1/2025 7:35:45 AM	Sohil	10/2/2025 3:02:04	Peak Integrated by Software
PB169881BS	PP075439.D	AR-1260-4	yogesh	10/1/2025 7:35:47 AM	Sohil	10/2/2025 3:02:07	Peak Integrated by Software
AR1242CCC500	PP075450.D	AR-1242-4 #2	yogesh	10/1/2025 7:36:04 AM	Sohil	10/2/2025 3:02:27	Peak Integrated by Software
AR1248CCC500	PP075451.D	AR-1248-3 #2	yogesh	10/1/2025 7:36:05 AM	Sohil	10/2/2025 3:02:29	Peak Integrated by Software
AR1254CCC500	PP075452.D	AR-1254-1	yogesh	10/1/2025 7:36:07 AM	Sohil	10/2/2025 3:02:31	Peak Integrated by Software
AR1242CCC500	PP075465.D	AR-1242-4 #2	yogesh	10/1/2025 7:36:18 AM	Sohil	10/2/2025 3:02:49	Peak Integrated by Software
AR1248CCC500	PP075466.D	AR-1248-3 #2	yogesh	10/1/2025 7:36:20 AM	Sohil	10/2/2025 3:02:44	Peak Integrated by Software
AR1254CCC500	PP075467.D	AR-1254-1	yogesh	10/1/2025 7:36:22 AM	Sohil	10/2/2025 3:02:46	Peak Integrated by Software
AR1242CCC500	PP075471.D	AR-1242-4 #2	yogesh	10/1/2025 7:36:24 AM	Sohil	10/2/2025 3:02:51	Peak Integrated by Software

Manual Integration Report

Sequence:

PP093025

Instrument

ECD_p

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
AR1248CCC500	PP075472.D	AR-1248-3 #2	yogesh	10/1/2025 7:36:26 AM	Sohil	10/2/2025 3:02:54	Peak Integrated by Software
AR1254CCC500	PP075473.D	AR-1254-1	yogesh	10/1/2025 7:36:28 AM	Sohil	10/2/2025 3:02:56	Peak Integrated by Software
AR1254CCC500	PP075473.D	AR-1254-1 #2	yogesh	10/1/2025 7:36:28 AM	Sohil	10/2/2025 3:02:56	Peak Integrated by Software

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP092325

Review By	yogesh	Review On	9/23/2025 12:35:13 PM
Supervise By	mohammad	Supervise On	9/25/2025 2:25:37 AM
SubDirectory	PP092325	HP Acquire Method	HP Processing Method PP092325
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24823,PP24824,PP24825,PP24826,PP24827,PP24828,PP24829,PP24830,PP24831,PP24832,PP24833,PP24834,PP24835,PP24836,PP24837,PP24838,PP24839,PP24840,PP24841,PP24842,PP24843,PP24844,PP24845,PP24846,PP24847,PP24848,PP24849,PP24850,PP24851,PP24852,PP24853,PP24854 PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24855,PP24856,PP24857,PP24858,PP24859,PP24860,PP24861,PP24862		

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	HEXANE	PP075245.D	23 Sep 2025 08:13	YP\AJ	Ok
2	I.BLK	PP075246.D	23 Sep 2025 08:29	YP\AJ	Ok
3	AR1660ICC1000	PP075247.D	23 Sep 2025 08:45	YP\AJ	Ok
4	AR1660ICC750	PP075248.D	23 Sep 2025 09:01	YP\AJ	Ok
5	AR1660ICC500	PP075249.D	23 Sep 2025 09:18	YP\AJ	Ok
6	AR1660ICC250	PP075250.D	23 Sep 2025 09:34	YP\AJ	Ok,M
7	AR1660ICC050	PP075251.D	23 Sep 2025 10:11	YP\AJ	Ok,M
8	AR1221ICC500	PP075252.D	23 Sep 2025 10:27	YP\AJ	Ok
9	AR1232ICC500	PP075253.D	23 Sep 2025 10:43	YP\AJ	Ok
10	AR1242ICC1000	PP075254.D	23 Sep 2025 11:00	YP\AJ	Ok,M
11	AR1242ICC750	PP075255.D	23 Sep 2025 11:16	YP\AJ	Ok,M
12	AR1242ICC500	PP075256.D	23 Sep 2025 11:32	YP\AJ	Ok,M
13	AR1242ICC250	PP075257.D	23 Sep 2025 11:48	YP\AJ	Ok,M
14	AR1242ICC050	PP075258.D	23 Sep 2025 12:05	YP\AJ	Ok,M
15	AR1248ICC1000	PP075259.D	23 Sep 2025 12:37	YP\AJ	Ok,M
16	AR1248ICC750	PP075260.D	23 Sep 2025 12:53	YP\AJ	Ok,M
17	AR1248ICC500	PP075261.D	23 Sep 2025 13:10	YP\AJ	Ok,M
18	AR1248ICC250	PP075262.D	23 Sep 2025 13:26	YP\AJ	Ok,M
19	AR1248ICC050	PP075263.D	23 Sep 2025 13:42	YP\AJ	Ok,M
20	AR1254ICC1000	PP075264.D	23 Sep 2025 14:15	YP\AJ	Ok,M
21	AR1254ICC750	PP075265.D	23 Sep 2025 14:31	YP\AJ	Ok,M

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP092325

Review By	yogesh	Review On	9/23/2025 12:35:13 PM
Supervise By	mohammad	Supervise On	9/25/2025 2:25:37 AM
SubDirectory	PP092325	HP Acquire Method	HP Processing Method PP092325
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24823,PP24824,PP24825,PP24826,PP24827,PP24828,PP24829,PP24830,PP24831,PP24832,PP24833,PP24834,PP24835,PP24836,PP24837,PP24838,PP24839,PP24840,PP24841,PP24842,PP24843,PP24844,PP24845,PP24846,PP24847,PP24848,PP24849,PP24850,PP24851,PP24852,PP24853,PP24854		
CCC			
Internal Standard/PEM			
ICV/I.BLK	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24855,PP24856,PP24857,PP24858,PP24859,PP24860,PP24861,PP24862		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

22	AR1254ICC500	PP075266.D	23 Sep 2025 14:47	YP\AJ	Ok,M
23	AR1254ICC250	PP075267.D	23 Sep 2025 15:04	YP\AJ	Ok,M
24	AR1254ICC050	PP075268.D	23 Sep 2025 15:20	YP\AJ	Ok,M
25	AR1262ICC500	PP075269.D	23 Sep 2025 15:52	YP\AJ	Ok
26	AR1268ICC1000	PP075270.D	23 Sep 2025 16:09	YP\AJ	Ok
27	AR1268ICC750	PP075271.D	23 Sep 2025 16:25	YP\AJ	Ok
28	AR1268ICC500	PP075272.D	23 Sep 2025 16:41	YP\AJ	Ok
29	AR1268ICC250	PP075273.D	23 Sep 2025 16:58	YP\AJ	Ok
30	AR1268ICC050	PP075274.D	23 Sep 2025 17:14	YP\AJ	Ok,M
31	PP092325ICV500	PP075275.D	23 Sep 2025 17:46	YP\AJ	Ok,M
32	AR1242ICV500	PP075276.D	23 Sep 2025 18:35	YP\AJ	Ok,M
33	AR1248ICV500	PP075277.D	23 Sep 2025 19:08	YP\AJ	Ok,M
34	AR1254ICV500	PP075278.D	23 Sep 2025 19:40	YP\AJ	Ok,M
35	AR1268ICV500	PP075279.D	23 Sep 2025 19:57	YP\AJ	Ok,M
36	DDT ANALOG	PP075280.D	23 Sep 2025 20:29	YP\AJ	Ok

M : Manual Integration

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP092925

Review By	yogesh	Review On	9/29/2025 10:08:10 AM
Supervise By	mohammad	Supervise On	10/1/2025 1:40:34 AM
SubDirectory	PP092925	HP Acquire Method	HP Processing Method PP092325
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24823,PP24824,PP24825,PP24826,PP24827,PP24828,PP24829,PP24830,PP24831,PP24832,PP24833,PP24834,PP24835,PP24836,PP24837,PP24838,PP24839,PP24840,PP24841,PP24842,PP24843,PP24844,PP24845,PP24846,PP24847,PP24848,PP24849,PP24850,PP24851,PP24852,PP24853,PP24854		
CCC	PP24824,PP24828,PP24832,PP24836,PP24840,PP24844,PP24852		
Internal Standard/PEM			
ICV/I.BLK	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24855,PP24856,PP24857,PP24858,PP24859,PP24860,PP24861,PP24862		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	HEXANE	PP075384.D	29 Sep 2025 09:09	YP\AJ	Ok
2	AR1660CCC500	PP075385.D	29 Sep 2025 09:25	YP\AJ	Ok
3	AR1242CCC500	PP075386.D	29 Sep 2025 09:41	YP\AJ	Ok,M
4	AR1248CCC500	PP075387.D	29 Sep 2025 09:58	YP\AJ	Ok,M
5	AR1254CCC500	PP075388.D	29 Sep 2025 10:14	YP\AJ	Ok,M
6	I.BLK	PP075389.D	29 Sep 2025 10:30	YP\AJ	Ok
7	PB169876BL	PP075390.D	29 Sep 2025 10:46	YP\AJ	Ok
8	PB169876BS	PP075391.D	29 Sep 2025 11:03	YP\AJ	Ok
9	Q3199-06	PP075392.D	29 Sep 2025 11:19	YP\AJ	Ok
10	Q3196-03	PP075393.D	29 Sep 2025 11:35	YP\AJ	Ok
11	Q3196-05	PP075394.D	29 Sep 2025 11:51	YP\AJ	Ok
12	Q3199-05	PP075395.D	29 Sep 2025 12:08	YP\AJ	Ok
13	Q3198-01	PP075396.D	29 Sep 2025 12:24	YP\AJ	Ok
14	Q3198-02	PP075397.D	29 Sep 2025 12:40	YP\AJ	Ok
15	AR1660CCC500	PP075398.D	29 Sep 2025 13:45	YP\AJ	Ok
16	AR1242CCC500	PP075399.D	29 Sep 2025 14:01	YP\AJ	Ok,M
17	AR1248CCC500	PP075400.D	29 Sep 2025 14:17	YP\AJ	Ok,M
18	AR1254CCC500	PP075401.D	29 Sep 2025 14:34	YP\AJ	Ok,M
19	I.BLK	PP075402.D	29 Sep 2025 14:50	YP\AJ	Ok,M
20	PB169887BL	PP075403.D	29 Sep 2025 15:06	YP\AJ	Ok,M
21	PB169887BS	PP075404.D	29 Sep 2025 15:23	YP\AJ	Not Ok

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP092925

Review By	yogesh	Review On	9/29/2025 10:08:10 AM
Supervise By	mohammad	Supervise On	10/1/2025 1:40:34 AM
SubDirectory	PP092925	HP Acquire Method	HP Processing Method PP092325
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24823,PP24824,PP24825,PP24826,PP24827,PP24828,PP24829,PP24830,PP24831,PP24832,PP24833,PP24834,PP24835,PP24836,PP24837,PP24838,PP24839,PP24840,PP24841,PP24842,PP24843,PP24844,PP24845,PP24846,PP24847,PP24848,PP24849,PP24850,PP24851,PP24852,PP24853,PP24854		
CCC	PP24824,PP24828,PP24832,PP24836,PP24840,PP24844,PP24852		
Internal Standard/PEM			
ICV/I.BLK	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24855,PP24856,PP24857,PP24858,PP24859,PP24860,PP24861,PP24862		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

22	Q3203-01	PP075405.D	29 Sep 2025 15:39	YPIAJ	Ok
23	Q3203-02	PP075406.D	29 Sep 2025 15:55	YPIAJ	Ok
24	Q3203-03	PP075407.D	29 Sep 2025 16:11	YPIAJ	Ok
25	Q3203-04	PP075408.D	29 Sep 2025 16:28	YPIAJ	Ok
26	Q3203-05	PP075409.D	29 Sep 2025 16:44	YPIAJ	Ok
27	Q3204-01	PP075410.D	29 Sep 2025 17:00	YPIAJ	Ok
28	Q3204-02	PP075411.D	29 Sep 2025 17:16	YPIAJ	Ok
29	Q3204-03	PP075412.D	29 Sep 2025 17:33	YPIAJ	Ok
30	AR1660CCC500	PP075413.D	29 Sep 2025 18:38	YPIAJ	Ok
31	AR1242CCC500	PP075414.D	29 Sep 2025 19:10	YPIAJ	Ok,M
32	AR1248CCC500	PP075415.D	29 Sep 2025 19:27	YPIAJ	Ok,M
33	AR1254CCC500	PP075416.D	29 Sep 2025 19:43	YPIAJ	Ok,M
34	I.BLK	PP075417.D	29 Sep 2025 19:59	YPIAJ	Ok
35	PB169881BL	PP075418.D	29 Sep 2025 20:15	YPIAJ	Ok
36	PB169881BS	PP075419.D	29 Sep 2025 20:32	YPIAJ	Not Ok
37	PB169881TB	PP075420.D	29 Sep 2025 20:48	YPIAJ	Ok
38	Q3179-01	PP075421.D	29 Sep 2025 21:04	YPIAJ	Ok,M
39	Q3179-02MS	PP075422.D	29 Sep 2025 21:20	YPIAJ	Ok,M
40	Q3179-03MSD	PP075423.D	29 Sep 2025 21:37	YPIAJ	Ok,M
41	Q3223-01	PP075424.D	29 Sep 2025 21:53	YPIAJ	Ok,NS
42	Q3223-02	PP075425.D	29 Sep 2025 22:09	YPIAJ	Ok
43	Q3223-03	PP075426.D	29 Sep 2025 22:25	YPIAJ	Ok
44	Q3223-04	PP075427.D	29 Sep 2025 22:42	YPIAJ	Ok

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP092925

Review By	yogesh	Review On	9/29/2025 10:08:10 AM
Supervise By	mohammad	Supervise On	10/1/2025 1:40:34 AM
SubDirectory	PP092925	HP Acquire Method	HP Processing Method PP092325
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24823,PP24824,PP24825,PP24826,PP24827,PP24828,PP24829,PP24830,PP24831,PP24832,PP24833,PP24834,PP24835,PP24836,PP24837,PP24838,PP24839,PP24840,PP24841,PP24842,PP24843,PP24844,PP24845,PP24846,PP24847,PP24848,PP24849,PP24850,PP24851,PP24852,PP24853,PP24854		
CCC	PP24824,PP24828,PP24832,PP24836,PP24840,PP24844,PP24852		
Internal Standard/PEM			
ICV/I.BLK	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24855,PP24856,PP24857,PP24858,PP24859,PP24860,PP24861,PP24862		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

45	AR1660CCC500	PP075428.D	29 Sep 2025 23:47	YP\AJ	Ok,M
46	AR1242CCC500	PP075429.D	30 Sep 2025 00:19	YP\AJ	Ok,M
47	AR1248CCC500	PP075430.D	30 Sep 2025 00:36	YP\AJ	Ok,M
48	AR1254CCC500	PP075431.D	30 Sep 2025 00:52	YP\AJ	Ok,M
49	I.BLK	PP075432.D	30 Sep 2025 01:08	YP\AJ	Ok

M : Manual Integration

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP093025

Review By	yogesh	Review On	9/30/2025 10:20:14 AM
Supervise By	Sohil	Supervise On	10/2/2025 3:03:07 PM
SubDirectory	PP093025	HP Acquire Method	HP Processing Method PP092325
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24823,PP24824,PP24825,PP24826,PP24827,PP24828,PP24829,PP24830,PP24831,PP24832,PP24833,PP24834,PP24835,PP24836,PP24837,PP24838,PP24839,PP24840,PP24841,PP24842,PP24843,PP24844,PP24845,PP24846,PP24847,PP24848,PP24849,PP24850,PP24851,PP24852,PP24853,PP24854		
CCC	PP24824,PP24828,PP24832,PP24836,PP24840,PP24844,PP24852		
Internal Standard/PEM			
ICV/I.BLK	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24855,PP24856,PP24857,PP24858,PP24859,PP24860,PP24861,PP24862		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	HEXANE	PP075433.D	30 Sep 2025 08:30	YP\AJ	Ok
2	AR1660CCC500	PP075434.D	30 Sep 2025 08:46	YP\AJ	Ok
3	AR1242CCC500	PP075435.D	30 Sep 2025 09:03	YP\AJ	Ok,M
4	AR1248CCC500	PP075436.D	30 Sep 2025 09:19	YP\AJ	Ok,M
5	AR1254CCC500	PP075437.D	30 Sep 2025 09:35	YP\AJ	Ok,M
6	I.BLK	PP075438.D	30 Sep 2025 09:51	YP\AJ	Ok,M
7	PB169881BS	PP075439.D	30 Sep 2025 10:08	YP\AJ	Ok,M
8	PB169887BS	PP075440.D	30 Sep 2025 10:31	YP\AJ	Ok,M
9	Q3224-01	PP075441.D	30 Sep 2025 10:50	YP\AJ	Ok,M
10	PB169907BL	PP075442.D	30 Sep 2025 13:32	YP\AJ	Ok,M
11	PB169907BS	PP075443.D	30 Sep 2025 13:48	YP\AJ	Ok,M
12	Q3238-01	PP075444.D	30 Sep 2025 14:04	YP\AJ	Ok
13	Q3240-01	PP075445.D	30 Sep 2025 14:20	YP\AJ	Ok
14	Q3240-01MS	PP075446.D	30 Sep 2025 14:36	YP\AJ	Ok,M
15	Q3240-01MSD	PP075447.D	30 Sep 2025 14:53	YP\AJ	Ok,M
16	Q3243-01	PP075448.D	30 Sep 2025 15:09	YP\AJ	ReRun
17	AR1660CCC500	PP075449.D	30 Sep 2025 16:14	YP\AJ	Ok
18	AR1242CCC500	PP075450.D	30 Sep 2025 16:30	YP\AJ	Ok,M
19	AR1248CCC500	PP075451.D	30 Sep 2025 16:47	YP\AJ	Ok,M
20	AR1254CCC500	PP075452.D	30 Sep 2025 17:03	YP\AJ	Ok,M
21	I.BLK	PP075453.D	30 Sep 2025 17:19	YP\AJ	Ok

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP093025

Review By	yogesh	Review On	9/30/2025 10:20:14 AM
Supervise By	Sohil	Supervise On	10/2/2025 3:03:07 PM
SubDirectory	PP093025	HP Acquire Method	HP Processing Method PP092325
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24823,PP24824,PP24825,PP24826,PP24827,PP24828,PP24829,PP24830,PP24831,PP24832,PP24833,PP24834,PP24835,PP24836,PP24837,PP24838,PP24839,PP24840,PP24841,PP24842,PP24843,PP24844,PP24845,PP24846,PP24847,PP24848,PP24849,PP24850,PP24851,PP24852,PP24853,PP24854		
CCC	PP24824,PP24828,PP24832,PP24836,PP24840,PP24844,PP24852		
Internal Standard/PEM			
ICV/I.BLK	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24855,PP24856,PP24857,PP24858,PP24859,PP24860,PP24861,PP24862		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

22	PB169922BL	PP075454.D	30 Sep 2025 17:36	YP\AJ	Ok
23	PB169922BS	PP075455.D	30 Sep 2025 17:52	YP\AJ	Ok,M
24	Q3249-02	PP075456.D	30 Sep 2025 18:08	YP\AJ	Ok,M
25	Q3249-07	PP075457.D	30 Sep 2025 18:24	YP\AJ	Ok
26	Q3249-10	PP075458.D	30 Sep 2025 18:41	YP\AJ	Ok,M
27	Q3249-13	PP075459.D	30 Sep 2025 18:57	YP\AJ	Ok,M
28	Q3249-15	PP075460.D	30 Sep 2025 19:13	YP\AJ	Ok
29	Q3249-18	PP075461.D	30 Sep 2025 19:29	YP\AJ	Dilution
30	Q3249-21	PP075462.D	30 Sep 2025 19:46	YP\AJ	Dilution
31	Q3249-23	PP075463.D	30 Sep 2025 20:02	YP\AJ	Dilution
32	AR1660CCC500	PP075464.D	30 Sep 2025 21:07	YP\AJ	Ok
33	AR1242CCC500	PP075465.D	30 Sep 2025 21:39	YP\AJ	Ok,M
34	AR1248CCC500	PP075466.D	30 Sep 2025 21:56	YP\AJ	Ok,M
35	AR1254CCC500	PP075467.D	30 Sep 2025 22:12	YP\AJ	Ok,M
36	I.BLK	PP075468.D	30 Sep 2025 22:28	YP\AJ	Ok
37	Q3249-26	PP075469.D	30 Sep 2025 22:45	YP\AJ	Dilution
38	AR1660CCC500	PP075470.D	30 Sep 2025 23:50	YP\AJ	Ok
39	AR1242CCC500	PP075471.D	01 Oct 2025 00:22	YP\AJ	Ok,M
40	AR1248CCC500	PP075472.D	01 Oct 2025 00:38	YP\AJ	Ok,M
41	AR1254CCC500	PP075473.D	01 Oct 2025 00:55	YP\AJ	Ok,M
42	I.BLK	PP075474.D	01 Oct 2025 01:11	YP\AJ	Ok

M : Manual Integration

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP092325

Review By	yogesh	Review On	9/23/2025 12:35:13 PM
Supervise By	mohammad	Supervise On	9/25/2025 2:25:37 AM
SubDirectory	PP092325	HP Acquire Method	HP Processing Method PP092325

STD. NAME	STD REF.#
Tune/Reschk	
Initial Calibration Stds	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24823,PP24824,PP24825,PP24826,PP24827,PP24828,PP24829,P24830,PP24831,PP24832,PP24833,PP24834,PP24835,PP24836,PP24837,PP24838,PP24839,PP24840,PP24841,PP24842,PP24843,PP24844,PP24845,PP24846,PP24847,PP24848,PP24849,PP24850,PP24851,PP24852,PP24853,PP24854
CCC	
Internal Standard/PEM	
ICV/I.BLK	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24855,PP24856,PP24857,PP24858,PP24859,PP24860,PP24861,PP24862
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PP075245.D	23 Sep 2025 08:13		YPIAJ	Ok
2	I.BLK	I.BLK	PP075246.D	23 Sep 2025 08:29		YPIAJ	Ok
3	AR1660ICC1000	AR1660ICC1000	PP075247.D	23 Sep 2025 08:45		YPIAJ	Ok
4	AR1660ICC750	AR1660ICC750	PP075248.D	23 Sep 2025 09:01		YPIAJ	Ok
5	AR1660ICC500	AR1660ICC500	PP075249.D	23 Sep 2025 09:18		YPIAJ	Ok
6	AR1660ICC250	AR1660ICC250	PP075250.D	23 Sep 2025 09:34		YPIAJ	Ok,M
7	AR1660ICC050	AR1660ICC050	PP075251.D	23 Sep 2025 10:11		YPIAJ	Ok,M
8	AR1221ICC500	AR1221ICC500	PP075252.D	23 Sep 2025 10:27		YPIAJ	Ok
9	AR1232ICC500	AR1232ICC500	PP075253.D	23 Sep 2025 10:43		YPIAJ	Ok
10	AR1242ICC1000	AR1242ICC1000	PP075254.D	23 Sep 2025 11:00		YPIAJ	Ok,M
11	AR1242ICC750	AR1242ICC750	PP075255.D	23 Sep 2025 11:16		YPIAJ	Ok,M
12	AR1242ICC500	AR1242ICC500	PP075256.D	23 Sep 2025 11:32		YPIAJ	Ok,M
13	AR1242ICC250	AR1242ICC250	PP075257.D	23 Sep 2025 11:48		YPIAJ	Ok,M
14	AR1242ICC050	AR1242ICC050	PP075258.D	23 Sep 2025 12:05		YPIAJ	Ok,M
15	AR1248ICC1000	AR1248ICC1000	PP075259.D	23 Sep 2025 12:37		YPIAJ	Ok,M
16	AR1248ICC750	AR1248ICC750	PP075260.D	23 Sep 2025 12:53		YPIAJ	Ok,M
17	AR1248ICC500	AR1248ICC500	PP075261.D	23 Sep 2025 13:10		YPIAJ	Ok,M
18	AR1248ICC250	AR1248ICC250	PP075262.D	23 Sep 2025 13:26		YPIAJ	Ok,M

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP092325

Review By	yogesh	Review On	9/23/2025 12:35:13 PM
Supervise By	mohammad	Supervise On	9/25/2025 2:25:37 AM
SubDirectory	PP092325	HP Acquire Method	HP Processing Method PP092325
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24823,PP24824,PP24825,PP24826,PP24827,PP24828,PP24829,P P24830,PP24831,PP24832,PP24833,PP24834,PP24835,PP24836,PP24837,PP24838,PP24839,PP24840,PP24841,PP24842,PP24843,PP24844,PP 24845,PP24846,PP24847,PP24848,PP24849,PP24850,PP24851,PP24852,PP24853,PP24854 PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24855,PP24856,PP24857,PP24858,PP24859,PP24860,PP24861,PP24862		

19	AR1248ICC050	AR1248ICC050	PP075263.D	23 Sep 2025 13:42		YPIAJ	Ok,M
20	AR1254ICC1000	AR1254ICC1000	PP075264.D	23 Sep 2025 14:15		YPIAJ	Ok,M
21	AR1254ICC750	AR1254ICC750	PP075265.D	23 Sep 2025 14:31		YPIAJ	Ok,M
22	AR1254ICC500	AR1254ICC500	PP075266.D	23 Sep 2025 14:47		YPIAJ	Ok,M
23	AR1254ICC250	AR1254ICC250	PP075267.D	23 Sep 2025 15:04		YPIAJ	Ok,M
24	AR1254ICC050	AR1254ICC050	PP075268.D	23 Sep 2025 15:20		YPIAJ	Ok,M
25	AR1262ICC500	AR1262ICC500	PP075269.D	23 Sep 2025 15:52		YPIAJ	Ok
26	AR1268ICC1000	AR1268ICC1000	PP075270.D	23 Sep 2025 16:09		YPIAJ	Ok
27	AR1268ICC750	AR1268ICC750	PP075271.D	23 Sep 2025 16:25		YPIAJ	Ok
28	AR1268ICC500	AR1268ICC500	PP075272.D	23 Sep 2025 16:41		YPIAJ	Ok
29	AR1268ICC250	AR1268ICC250	PP075273.D	23 Sep 2025 16:58		YPIAJ	Ok
30	AR1268ICC050	AR1268ICC050	PP075274.D	23 Sep 2025 17:14		YPIAJ	Ok,M
31	PP092325ICV500	ICVPP092325	PP075275.D	23 Sep 2025 17:46		YPIAJ	Ok,M
32	AR1242ICV500	ICVPP092325AR1242	PP075276.D	23 Sep 2025 18:35		YPIAJ	Ok,M
33	AR1248ICV500	ICVPP092325AR1248	PP075277.D	23 Sep 2025 19:08		YPIAJ	Ok,M
34	AR1254ICV500	ICVPP092325AR1254	PP075278.D	23 Sep 2025 19:40		YPIAJ	Ok,M
35	AR1268ICV500	ICVPP092325AR1268	PP075279.D	23 Sep 2025 19:57		YPIAJ	Ok,M
36	DDT ANALOG	DDT ANALOG	PP075280.D	23 Sep 2025 20:29		YPIAJ	Ok

M : Manual Integration

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP092925

Review By	yogesh	Review On	9/29/2025 10:08:10 AM
Supervise By	mohammad	Supervise On	10/1/2025 1:40:34 AM
SubDirectory	PP092925	HP Acquire Method	HP Processing Method PP092325

STD. NAME	STD REF.#
Tune/Reschk	
Initial Calibration Stds	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24823,PP24824,PP24825,PP24826,PP24827,PP24828,PP24829,P P24830,PP24831,PP24832,PP24833,PP24834,PP24835,PP24836,PP24837,PP24838,PP24839,PP24840,PP24841,PP24842,PP24843,PP24844,PP 24845,PP24846,PP24847,PP24848,PP24849,PP24850,PP24851,PP24852,PP24853,PP24854
CCC	PP24824,PP24828,PP24832,PP24836,PP24840,PP24844,PP24852
Internal Standard/PEM	
ICV/I.BLK	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24855,PP24856,PP24857,PP24858,PP24859,PP24860,PP24861,PP24862
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PP075384.D	29 Sep 2025 09:09		YP\AJ	Ok
2	AR1660CCC500	AR1660CCC500	PP075385.D	29 Sep 2025 09:25		YP\AJ	Ok
3	AR1242CCC500	AR1242CCC500	PP075386.D	29 Sep 2025 09:41		YP\AJ	Ok,M
4	AR1248CCC500	AR1248CCC500	PP075387.D	29 Sep 2025 09:58		YP\AJ	Ok,M
5	AR1254CCC500	AR1254CCC500	PP075388.D	29 Sep 2025 10:14		YP\AJ	Ok,M
6	I.BLK	I.BLK	PP075389.D	29 Sep 2025 10:30		YP\AJ	Ok
7	PB169876BL	PB169876BL	PP075390.D	29 Sep 2025 10:46		YP\AJ	Ok
8	PB169876BS	PB169876BS	PP075391.D	29 Sep 2025 11:03		YP\AJ	Ok
9	Q3199-06	ELZ-25-000125	PP075392.D	29 Sep 2025 11:19		YP\AJ	Ok
10	Q3196-03	FMC-9-24-25-3	PP075393.D	29 Sep 2025 11:35		YP\AJ	Ok
11	Q3196-05	FMC-9-24-25-5	PP075394.D	29 Sep 2025 11:51		YP\AJ	Ok
12	Q3199-05	ELZ-25-000124	PP075395.D	29 Sep 2025 12:08	DCB high in 1st column	YP\AJ	Ok
13	Q3198-01	KZZ356V-END-1-1	PP075396.D	29 Sep 2025 12:24		YP\AJ	Ok
14	Q3198-02	KZZ356V-END-1-2	PP075397.D	29 Sep 2025 12:40		YP\AJ	Ok
15	AR1660CCC500	AR1660CCC500	PP075398.D	29 Sep 2025 13:45		YP\AJ	Ok
16	AR1242CCC500	AR1242CCC500	PP075399.D	29 Sep 2025 14:01		YP\AJ	Ok,M
17	AR1248CCC500	AR1248CCC500	PP075400.D	29 Sep 2025 14:17		YP\AJ	Ok,M
18	AR1254CCC500	AR1254CCC500	PP075401.D	29 Sep 2025 14:34		YP\AJ	Ok,M

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP092925

Review By	yogesh	Review On	9/29/2025 10:08:10 AM
Supervise By	mohammad	Supervise On	10/1/2025 1:40:34 AM
SubDirectory	PP092925	HP Acquire Method	HP Processing Method PP092325
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24823,PP24824,PP24825,PP24826,PP24827,PP24828,PP24829,PP24830,PP24831,PP24832,PP24833,PP24834,PP24835,PP24836,PP24837,PP24838,PP24839,PP24840,PP24841,PP24842,PP24843,PP24844,PP24845,PP24846,PP24847,PP24848,PP24849,PP24850,PP24851,PP24852,PP24853,PP24854		
CCC	PP24824,PP24828,PP24832,PP24836,PP24840,PP24844,PP24852		
Internal Standard/PEM			
ICV/I.BLK	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24855,PP24856,PP24857,PP24858,PP24859,PP24860,PP24861,PP24862		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

19	I.BLK	I.BLK	PP075402.D	29 Sep 2025 14:50		YPIAJ	Ok,M
20	PB169887BL	PB169887BL	PP075403.D	29 Sep 2025 15:06		YPIAJ	Ok,M
21	PB169887BS	PB169887BS	PP075404.D	29 Sep 2025 15:23	AR1016 recovery fail	YPIAJ	Not Ok
22	Q3203-01	1	PP075405.D	29 Sep 2025 15:39		YPIAJ	Ok
23	Q3203-02	2	PP075406.D	29 Sep 2025 15:55		YPIAJ	Ok
24	Q3203-03	3	PP075407.D	29 Sep 2025 16:11		YPIAJ	Ok
25	Q3203-04	4	PP075408.D	29 Sep 2025 16:28		YPIAJ	Ok
26	Q3203-05	5	PP075409.D	29 Sep 2025 16:44		YPIAJ	Ok
27	Q3204-01	TRE-25-0103	PP075410.D	29 Sep 2025 17:00		YPIAJ	Ok
28	Q3204-02	TRE-25-0104	PP075411.D	29 Sep 2025 17:16		YPIAJ	Ok
29	Q3204-03	TRE-25-0105	PP075412.D	29 Sep 2025 17:33		YPIAJ	Ok
30	AR1660CCC500	AR1660CCC500	PP075413.D	29 Sep 2025 18:38		YPIAJ	Ok
31	AR1242CCC500	AR1242CCC500	PP075414.D	29 Sep 2025 19:10		YPIAJ	Ok,M
32	AR1248CCC500	AR1248CCC500	PP075415.D	29 Sep 2025 19:27		YPIAJ	Ok,M
33	AR1254CCC500	AR1254CCC500	PP075416.D	29 Sep 2025 19:43		YPIAJ	Ok,M
34	I.BLK	I.BLK	PP075417.D	29 Sep 2025 19:59		YPIAJ	Ok
35	PB169881BL	PB169881BL	PP075418.D	29 Sep 2025 20:15		YPIAJ	Ok
36	PB169881BS	PB169881BS	PP075419.D	29 Sep 2025 20:32	AR1016 & 1260 Recovery fail	YPIAJ	Not Ok
37	PB169881TB	PB169881TB	PP075420.D	29 Sep 2025 20:48		YPIAJ	Ok

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP092925

Review By	yogesh	Review On	9/29/2025 10:08:10 AM
Supervise By	mohammad	Supervise On	10/1/2025 1:40:34 AM
SubDirectory	PP092925	HP Acquire Method	HP Processing Method PP092325
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24823,PP24824,PP24825,PP24826,PP24827,PP24828,PP24829,P P24830,PP24831,PP24832,PP24833,PP24834,PP24835,PP24836,PP24837,PP24838,PP24839,PP24840,PP24841,PP24842,PP24843,PP24844,PP 24845,PP24846,PP24847,PP24848,PP24849,PP24850,PP24851,PP24852,PP24853,PP24854		
CCC	PP24824,PP24828,PP24832,PP24836,PP24840,PP24844,PP24852		
Internal Standard/PEM			
ICV/I.BLK	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24855,PP24856,PP24857,PP24858,PP24859,PP24860,PP24861,PP24862		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

38	Q3179-01	EME-TS14-01	PP075421.D	29 Sep 2025 21:04		YPIAJ	Ok,M
39	Q3179-02MS	EME-TS14-01MS	PP075422.D	29 Sep 2025 21:20		YPIAJ	Ok,M
40	Q3179-03MSD	EME-TS14-01MSD	PP075423.D	29 Sep 2025 21:37		YPIAJ	Ok,M
41	Q3223-01	917	PP075424.D	29 Sep 2025 21:53	AR1242 + 1254 Hit	YPIAJ	Ok,NS
42	Q3223-02	917-A	PP075425.D	29 Sep 2025 22:09	AR1242 + 1254 Hit	YPIAJ	Ok
43	Q3223-03	2027	PP075426.D	29 Sep 2025 22:25		YPIAJ	Ok
44	Q3223-04	2028	PP075427.D	29 Sep 2025 22:42		YPIAJ	Ok
45	AR1660CCC500	AR1660CCC500	PP075428.D	29 Sep 2025 23:47	DCB high in 2nd column	YPIAJ	Ok,M
46	AR1242CCC500	AR1242CCC500	PP075429.D	30 Sep 2025 00:19	DCB high in 2nd column	YPIAJ	Ok,M
47	AR1248CCC500	AR1248CCC500	PP075430.D	30 Sep 2025 00:36	DCB high in 2nd column	YPIAJ	Ok,M
48	AR1254CCC500	AR1254CCC500	PP075431.D	30 Sep 2025 00:52		YPIAJ	Ok,M
49	I.BLK	I.BLK	PP075432.D	30 Sep 2025 01:08		YPIAJ	Ok

M : Manual Integration

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP093025

Review By	yogesh	Review On	9/30/2025 10:20:14 AM
Supervise By	Sohil	Supervise On	10/2/2025 3:03:07 PM
SubDirectory	PP093025	HP Acquire Method	HP Processing Method PP092325
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24823,PP24824,PP24825,PP24826,PP24827,PP24828,PP24829,PP24830,PP24831,PP24832,PP24833,PP24834,PP24835,PP24836,PP24837,PP24838,PP24839,PP24840,PP24841,PP24842,PP24843,PP24844,PP24845,PP24846,PP24847,PP24848,PP24849,PP24850,PP24851,PP24852,PP24853,PP24854		
CCC	PP24824,PP24828,PP24832,PP24836,PP24840,PP24844,PP24852		
Internal Standard/PEM			
ICV/I.BLK	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24855,PP24856,PP24857,PP24858,PP24859,PP24860,PP24861,PP24862		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PP075433.D	30 Sep 2025 08:30		YP\AJ	Ok
2	AR1660CCC500	AR1660CCC500	PP075434.D	30 Sep 2025 08:46	DCB HIGH in second column	YP\AJ	Ok
3	AR1242CCC500	AR1242CCC500	PP075435.D	30 Sep 2025 09:03	DCB HIGH in second column	YP\AJ	Ok,M
4	AR1248CCC500	AR1248CCC500	PP075436.D	30 Sep 2025 09:19	DCB HIGH in second column	YP\AJ	Ok,M
5	AR1254CCC500	AR1254CCC500	PP075437.D	30 Sep 2025 09:35	DCB HIGH in second column	YP\AJ	Ok,M
6	I.BLK	I.BLK	PP075438.D	30 Sep 2025 09:51		YP\AJ	Ok,M
7	PB169881BS	PB169881BS	PP075439.D	30 Sep 2025 10:08		YP\AJ	Ok,M
8	PB169887BS	PB169887BS	PP075440.D	30 Sep 2025 10:31		YP\AJ	Ok,M
9	Q3224-01	60159	PP075441.D	30 Sep 2025 10:50		YP\AJ	Ok,M
10	PB169907BL	PB169907BL	PP075442.D	30 Sep 2025 13:32		YP\AJ	Ok,M
11	PB169907BS	PB169907BS	PP075443.D	30 Sep 2025 13:48		YP\AJ	Ok,M
12	Q3238-01	RCA	PP075444.D	30 Sep 2025 14:04		YP\AJ	Ok
13	Q3240-01	OK-02-092925	PP075445.D	30 Sep 2025 14:20		YP\AJ	Ok
14	Q3240-01MS	OK-02-092925MS	PP075446.D	30 Sep 2025 14:36	Recovery fail	YP\AJ	Ok,M
15	Q3240-01MSD	OK-02-092925MSD	PP075447.D	30 Sep 2025 14:53	Recovery fail	YP\AJ	Ok,M
16	Q3243-01	91825AB	PP075448.D	30 Sep 2025 15:09	all surrogate fail	YP\AJ	ReRun
17	AR1660CCC500	AR1660CCC500	PP075449.D	30 Sep 2025 16:14		YP\AJ	Ok
18	AR1242CCC500	AR1242CCC500	PP075450.D	30 Sep 2025 16:30		YP\AJ	Ok,M

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP093025

Review By	yogesh	Review On	9/30/2025 10:20:14 AM
Supervise By	Sohil	Supervise On	10/2/2025 3:03:07 PM
SubDirectory	PP093025	HP Acquire Method	HP Processing Method PP092325
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24823,PP24824,PP24825,PP24826,PP24827,PP24828,PP24829,PP24830,PP24831,PP24832,PP24833,PP24834,PP24835,PP24836,PP24837,PP24838,PP24839,PP24840,PP24841,PP24842,PP24843,PP24844,PP24845,PP24846,PP24847,PP24848,PP24849,PP24850,PP24851,PP24852,PP24853,PP24854		
CCC	PP24824,PP24828,PP24832,PP24836,PP24840,PP24844,PP24852		
Internal Standard/PEM			
ICV/I.BLK	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24855,PP24856,PP24857,PP24858,PP24859,PP24860,PP24861,PP24862		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

19	AR1248CCC500	AR1248CCC500	PP075451.D	30 Sep 2025 16:47		YPIAJ	Ok,M
20	AR1254CCC500	AR1254CCC500	PP075452.D	30 Sep 2025 17:03		YPIAJ	Ok,M
21	I.BLK	I.BLK	PP075453.D	30 Sep 2025 17:19		YPIAJ	Ok
22	PB169922BL	PB169922BL	PP075454.D	30 Sep 2025 17:36		YPIAJ	Ok
23	PB169922BS	PB169922BS	PP075455.D	30 Sep 2025 17:52		YPIAJ	Ok,M
24	Q3249-02	S-3(3-3.5)	PP075456.D	30 Sep 2025 18:08	AR1254 & AR1260 hits	YPIAJ	Ok,M
25	Q3249-07	S-8(3-3.5)	PP075457.D	30 Sep 2025 18:24	AR1254 hits	YPIAJ	Ok
26	Q3249-10	S-9(3-3.5)	PP075458.D	30 Sep 2025 18:41	AR1254 & AR1260 hits	YPIAJ	Ok,M
27	Q3249-13	S-12(3-3.5)	PP075459.D	30 Sep 2025 18:57	AR1260 hits	YPIAJ	Ok,M
28	Q3249-15	S-14(3-3.5)	PP075460.D	30 Sep 2025 19:13	AR1254 & AR1260 hits	YPIAJ	Ok
29	Q3249-18	S-18(3-3.5)	PP075461.D	30 Sep 2025 19:29	AR1254 & AR1260 hits need 4x	YPIAJ	Dilution
30	Q3249-21	S-20(3-3.5)	PP075462.D	30 Sep 2025 19:46	AR1254 & AR1260 hits need 10x	YPIAJ	Dilution
31	Q3249-23	S-22(3-3.5)	PP075463.D	30 Sep 2025 20:02	AR1254 & AR1260 hits need 5x	YPIAJ	Dilution
32	AR1660CCC500	AR1660CCC500	PP075464.D	30 Sep 2025 21:07		YPIAJ	Ok
33	AR1242CCC500	AR1242CCC500	PP075465.D	30 Sep 2025 21:39		YPIAJ	Ok,M
34	AR1248CCC500	AR1248CCC500	PP075466.D	30 Sep 2025 21:56		YPIAJ	Ok,M
35	AR1254CCC500	AR1254CCC500	PP075467.D	30 Sep 2025 22:12		YPIAJ	Ok,M
36	I.BLK	I.BLK	PP075468.D	30 Sep 2025 22:28		YPIAJ	Ok
37	Q3249-26	S-24(3-3.5)	PP075469.D	30 Sep 2025 22:45	AR1254 & AR1260 hits need 10x	YPIAJ	Dilution

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP093025

Review By	yogesh	Review On	9/30/2025 10:20:14 AM
Supervise By	Sohil	Supervise On	10/2/2025 3:03:07 PM
SubDirectory	PP093025	HP Acquire Method	HP Processing Method PP092325
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24823,PP24824,PP24825,PP24826,PP24827,PP24828,PP24829,P P24830,PP24831,PP24832,PP24833,PP24834,PP24835,PP24836,PP24837,PP24838,PP24839,PP24840,PP24841,PP24842,PP24843,PP24844,PP 24845,PP24846,PP24847,PP24848,PP24849,PP24850,PP24851,PP24852,PP24853,PP24854		
CCC	PP24824,PP24828,PP24832,PP24836,PP24840,PP24844,PP24852		
Internal Standard/PEM			
ICV/I.BLK	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24855,PP24856,PP24857,PP24858,PP24859,PP24860,PP24861,PP24862		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

38	AR1660CCC500	AR1660CCC500	PP075470.D	30 Sep 2025 23:50		YPIAJ	Ok
39	AR1242CCC500	AR1242CCC500	PP075471.D	01 Oct 2025 00:22		YPIAJ	Ok,M
40	AR1248CCC500	AR1248CCC500	PP075472.D	01 Oct 2025 00:38		YPIAJ	Ok,M
41	AR1254CCC500	AR1254CCC500	PP075473.D	01 Oct 2025 00:55		YPIAJ	Ok,M
42	I.BLK	I.BLK	PP075474.D	01 Oct 2025 01:11		YPIAJ	Ok

M : Manual Integration

SOP ID : M1312-SPLP-10
SDG No : N/A
Weigh By : JP
Balance ID : WC SC-7
pH Meter ID : WC PH METER-1
Extraction By : JP
Filter By : JP
Pipette ID : WC
Tumbler ID : T-2
TCLP Filter ID : 40819719

Start Prep Date : 09/24/2025 Time : 14:30
End Prep Date : 09/25/2025 Time : 08:10
Combination Ratio : 20
ZHE Cleaning Batch : N/A 71
Initial Room Temperature : 24 °C
Final Room Temperature : 22 °C
TCLP Technician Signature : *JP*
Supervisor By : *12*

Standard Name	MLS USED	STD REF. # FROM LOG
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A

Chemical Used	ML/SAMPLE U	Lot Number
SPLP FLUID	WP114530	N/A
N/A	N/A	N/A
HNO3-TCLP,1N	N/A	WP114528
pH Strips	N/A	W1931,W1934,W3171,W3172
pH Strips	W1940,W1941,W1942	W3166,W1938,W1939,
1 Liter Amber	N/A	90924-08
120ml Plastic bottle	N/A	2738
1:1 HNO3	N/A	MP86978

Extraction Conformance/Non-Conformance Comments:

TUMBLER T-2 checked,30 rpm. Matrix spikes are added after filtration and before preservation.

Date / Time	Prepped Sample Relinquished By/Location	Received By/Location
09/25/25 10:00	<i>JP</i> / TCY 120 cm	<i>SKJ.</i> RJ 1E-X
	Preparation Group	Analysis Group

met nity

Sample ID	ClientID	TCLP Vessel ID	Sample Wt (g)	Volume Extraction Fluid #1 (mL)	Multi phasic	Phase Miscible	Phases Combined	Final Leachate PH	Metals Leachate Adj. PH	Prep. Pos
PB169818TB	LEB818	15	N/A	2000	N/A	N/A	N/A	4.24	1.5	T-2
Q3177-01	EME-TS12-01	11	100.02	2000	N/A	N/A	N/A	7.2	1.0	T-2
Q3177-02	EME-TS13-01	12	100.03	2000	N/A	N/A	N/A	7.6	1.5	T-2
Q3177-03	EME-TS14-02	13	100.02	2000	N/A	N/A	N/A	7.2	1.5	T-2
Q3179-01	EME-TS14-01	14	100.02	2000	N/A	N/A	N/A	7.6	1.0	T-2
Q3179-02	EME-TS14-01MS	14	100.02	2000	N/A	N/A	N/A	7.6	1.5	T-2
Q3179-03	EME-TS14-01MSD	14	100.02	2000	N/A	N/A	N/A	7.6	1.0	T-2

SampleID	ClientID	Sample Weight (g)	Filter Weight (g)	Filtrate (mL)	Filter + Solid (After 100°C)	% solids	% Dry Solids
PB169818TB	LEB818	N/A	N/A	N/A	N/A	N/A	N/A
Q3177-01	EME-TS12-01	N/A	N/A	N/A	N/A	100	N/A
Q3177-02	EME-TS13-01	N/A	N/A	N/A	N/A	100	N/A
Q3177-03	EME-TS14-02	N/A	N/A	N/A	N/A	100	N/A
Q3179-01	EME-TS14-01	N/A	N/A	N/A	N/A	100	N/A
Q3179-02	EME-TS14-01MS	N/A	N/A	N/A	N/A	100	N/A
Q3179-03	EME-TS14-01MSD	N/A	N/A	N/A	N/A	100	N/A

Hot Block ID : WC S-1 / WC S-2

Thermometer ID : FLASHPOINT

SampleID	ClientID	Sample Weight (g)	Volume DI Water (mL)	PH after 5 min stir	PH after 10 min stir	Extraction Fluid 1 or 2	pH Extraction Fluid
PB169818TB	LEB818	N/A	N/A	N/A	N/A	#1	4.24
Q3177-01	EME-TS12-01	N/A	N/A	N/A	N/A	#1	4.24
Q3177-02	EME-TS13-01	N/A	N/A	N/A	N/A	#1	4.24
Q3177-03	EME-TS14-02	N/A	N/A	N/A	N/A	#1	4.24
Q3179-01	EME-TS14-01	N/A	N/A	N/A	N/A	#1	4.24
Q3179-02	EME-TS14-01MS	N/A	N/A	N/A	N/A	#1	4.24
Q3179-03	EME-TS14-01MSD	N/A	N/A	N/A	N/A	#1	4.24

WORKLIST(Hardcopy Internal Chain)

WorkList Name : splp q3177

WorkList ID : 192051

Department : TCLP Extraction

Date : 09-24-2025 13:40:41

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q3177-01	EME-TS12-01	Solid	SPLP Extraction	Cool 4 deg C	ROYF02	J23	09/23/2025	1312
Q3177-02	EME-TS13-01	Solid	SPLP Extraction	Cool 4 deg C	ROYF02	J23	09/23/2025	1312
Q3177-03	EME-TS14-02	Solid	SPLP Extraction	Cool 4 deg C	ROYF02	J23	09/23/2025	1312
Q3179-01	EME-TS14-01	Solid	SPLP Extraction	Cool 4 deg C	ROYF02	J42	09/23/2025	1312
Q3179-02	EME-TS14-01MS	Solid	SPLP Extraction	Cool 4 deg C	ROYF02	J42	09/23/2025	1312
Q3179-03	EME-TS14-01MSD	Solid	SPLP Extraction	Cool 4 deg C	ROYF02	J42	09/23/2025	1312

Date/Time

09/24/25 13:45

Raw Sample Received by:

JH WRC

Raw Sample Relinquished by:

JTCsm

Date/Time

09/24/25

Raw Sample Received by:

JTCsm

Raw Sample Relinquished by:

JH WRC

SOP ID: M3510C,3580A-Extraction PCB-15

Clean Up SOP #: Acid Cleanup

Matrix : Water

Extraction Start Date : 09/29/2025

Extraction Start Time : 09:15

Extraction By: RJ

Extraction End Date : 09/29/2025

Balance check: N/A

Filter By: RJ

Extraction End Time : 13:50

Balance ID: N/A

pH Meter ID: N/A

Concentration By: EH

pH Strip Lot#: E3953

Hood ID: 3,7

Supervisor By : ritesh

Extraction Method: ☒ Separatory Funnel ☐ Continuous Liquid/Liquid ☐ Sonication ☐ Waste Dilution ☐ Soxhlet

Standard Name	MLS USED	Concentration ug/mL	STD REF. # FROM LOG
Spike Sol 1	1.0ML	5000 PPB	PP24930
Surrogate	1.0ML	200 PPB	PP24663
N/A	N/A	N/A	N/A
N/A	N/A	N/A	N/A
N/A	N/A	N/A	N/A

Chemical Used	ML/SAMPLE USED	Lot Number
Methylene Chloride	N/A	E3973
Baked Na2SO4	N/A	EP2646
Hexane	N/A	E3974
H2SO4 1:1	N/A	EP2610
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A

Extraction Conformance/Non-Conformance Comments:

40ML Vial Lot # 03-40BTS721.

KD Bath ID: WATER BATH-1,2

Envap ID: NEVAP-02

KD Bath Temperature: 60 °C

Envap Temperature: 40 °C

Date / Time	Prepped Sample Relinquished By/Location	Received By/Location
9/29/25	RJ (EXT-1006)	Y-P-PegHPCB
13:55	Preparation Group	Analysis Group

Analytical Method: M3510C,3580A-Extraction PCB-15

Concentration Date: 09/29/2025

Sample ID	Client Sample ID	Test	g / ml	PH	Surr/Spike By:		Final Vol. (mL)	JarID	Comments	Prep Pos
					AddedBy	VerifiedBy				
PB169881BL	ABLK881	SPLP PCB	1000	6	ritesh	Evelyn	10			SEP-1
PB169881BS	ALCS881	SPLP PCB	1000	6	ritesh	Evelyn	10			2
PB169881TB	PB169881TB	SPLP PCB	1000	6	ritesh	Evelyn	10			3
Q3177-01	EME-TS12-01	SPLP PCB	1000	6	ritesh	Evelyn	10	A		4
Q3177-02	EME-TS13-01	SPLP PCB	1000	6	ritesh	Evelyn	10	A		5
Q3177-03	EME-TS14-02	SPLP PCB	1000	6	ritesh	Evelyn	10	A		6
Q3179-01	EME-TS14-01	SPLP PCB	1000	6	ritesh	Evelyn	10	A		7
Q3179-02	EME-TS14-01MS	SPLP PCB	1000	6	ritesh	Evelyn	10	A		8
Q3179-03	EME-TS14-01MSD	SPLP PCB	1000	6	ritesh	Evelyn	10	A		9

RJ
9/29

* Extracts relinquished on the same date as received.

Sample ID	ClientID	TCLP Vessel ID	Sample Wt (g)	Volume Extraction Fluid #1 (mL)	Multi phasic	Phase Miscible	Phases Combined	Final Leachate PH	Metals Leachate Adj. PH	Prep Po
PB169818TB	LEB818	15	N/A	2000	N/A	N/A	N/A	4.24	1.5	T-2
Q3177-01	EME-TS12-01	11	100.02	2000	N/A	N/A	N/A	7.2	1.0	T-2
Q3177-02	EME-TS13-01	12	100.03	2000	N/A	N/A	N/A	7.6	1.5	T-2
Q3177-03	EME-TS14-02	13	100.02	2000	N/A	N/A	N/A	7.2	1.5	T-2
Q3179-01	EME-TS14-01	14	100.02	2000	N/A	N/A	N/A	7.6	1.0	T-2
Q3179-02	EME-TS14-01MS	14	100.02	2000	N/A	N/A	N/A	7.6	1.5	T-2
Q3179-03	EME-TS14-01MSD	14	100.02	2000	N/A	N/A	N/A	7.6	1.0	T-2

6/9/25/25
10:00

Prep Standard - Chemical Standard Summary

Order ID : Q3179

Test : SPLP PCB

Prepbatch ID : PB169881,

Sequence ID/Qc Batch ID: pp092925,PP093025,

Standard ID :

EP2610,EP2646,PP24329,PP24330,PP24381,PP24663,PP24805,PP24806,PP24807,PP24808,PP24809,PP24810,PP24811,PP24812,PP24814,PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24823,PP24824,PP24825,PP24826,PP24827,PP24828,PP24829,PP24830,PP24831,PP24832,PP24833,PP24834,PP24835,PP24836,PP24837,PP24838,PP24839,PP24840,PP24841,PP24842,PP24843,PP24844,PP24845,PP24846,PP24847,PP24848,PP24849,PP24850,PP24851,PP24852,PP24853,PP24854,PP24855,PP24856,PP24857,PP24858,PP24859,PP24860,PP24861,PP24862,PP24930,

Chemical ID :

E3804,E3875,E3944,E3962,E3963,E3972,E3973,E3974,M6157,P 11522,P11592,P12700,P12703,P12952,P12953,P12957,P13356,P13373,P13375,P13590,P13592,P13695,P13697,P13706,P13786,P13789,P13833,P13879,P13881,P13885,W3112,W3177,W3234,

Extractions STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
314	1.1 H2SO4 SOLN	EP2610	05/07/2025	11/07/2025	RUPESHKUMAR SHAH	Extraction_SC ALE_2 (EX-SC-2)	None	Riteshkumar Patel 05/07/2025

FROM 1000.00000ml of M6157 + 1000.00000ml of W3112 = Final Quantity: 2000.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3923	Baked Sodium Sulfate	EP2646	09/26/2025	01/28/2026	Evelyn Huang	Extraction_SC ALE_2 (EX-SC-2)	None	Riteshkumar Patel 09/26/2025

FROM 4000.00000gram of E3875 = Final Quantity: 4000.000 gram

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
84	Pest/PCB Surrogate Stock 20 PPM	PP24329	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 1.00000ml of P13356 + 9.00000ml of W3177 = Final Quantity: 10.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
202	AR1660 1000/100 ppb working solution 1st source	PP24330	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.10000ml of P13697 + 99.40000ml of W3177 + 0.50000ml of PP24329 = Final Quantity: 100.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1893	AR1254 1000 PPB Working Sol. 2nd Source	PP24381	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza 04/03/2025

FROM 1.00000ml of P12957 + 98.50000ml of W3177 + 0.50000ml of PP24329 = Final Quantity: 100.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
465	200 PPB Pest/PCB Surrogate Spike	PP24663	06/24/2025	12/24/2025	Abdul Mirza	None	None	Yogesh Patel 07/21/2025

FROM 1.00000ml of P13786 + 999.00000ml of E3944 = Final Quantity: 1000.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
84	Pest/PCB Surrogate Stock 20 PPM	PP24805	08/18/2025	02/18/2026	Yogesh Patel	None	None	Abdul Mirza
								08/19/2025

FROM 1.00000ml of P13789 + 9.00000ml of E3962 = Final Quantity: 10.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
213	AR1221 1000 PPB WORKING SOLUTION	PP24806	08/18/2025	02/18/2026	Yogesh Patel	None	None	Abdul Mirza
								08/19/2025

FROM 0.10000ml of P13375 + 99.40000ml of E3962 + 0.50000ml of PP24805 = Final Quantity: 100.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
214	AR1232 1000 PPB WORKING SOLUTION	PP24807	08/18/2025	02/18/2026	Yogesh Patel	None	None	Abdul Mirza
								08/19/2025

FROM 0.10000ml of P13879 + 99.40000ml of E3962 + 0.50000ml of PP24805 = Final Quantity: 100.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
215	AR1242 1000 PPB WORKING STD	PP24808	08/18/2025	02/18/2026	Yogesh Patel	None	None	Abdul Mirza
								08/19/2025

FROM 0.10000ml of P13706 + 99.40000ml of E3962 + 0.50000ml of PP24805 = Final Quantity: 100.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
216	AR1248 1000 PPB WORKING STD	PP24809	08/18/2025	02/18/2026	Yogesh Patel	None	None	Abdul Mirza
								08/19/2025

FROM 0.10000ml of P13695 + 99.40000ml of E3962 + 0.50000ml of PP24805 = Final Quantity: 100.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
217	AR1254 1000 PPB WORKING STD	PP24810	08/18/2025	02/18/2026	Yogesh Patel	None	None	Abdul Mirza
								08/19/2025

FROM 0.10000ml of P11592 + 99.40000ml of E3962 + 0.50000ml of PP24805 = Final Quantity: 100.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1529	AR1262 1000 PPB Working Solution	PP24811	08/18/2025	02/18/2026	Yogesh Patel	None	None	Abdul Mirza
								08/19/2025

FROM 0.10000ml of P13881 + 99.40000ml of E3962 + 0.50000ml of PP24805 = Final Quantity: 100.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1532	AR1268 1000 PPB Working Solution	PP24812	08/18/2025	02/18/2026	Yogesh Patel	None	None	Abdul Mirza
								08/19/2025

FROM 0.10000ml of P13885 + 99.40000ml of E3962 + 0.50000ml of PP24805 = Final Quantity: 100.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
404	AR1660 100 PPM Stock Solution 2nd Source	PP24814	08/18/2025	02/18/2026	Yogesh Patel	None	None	Abdul Mirza 08/19/2025

FROM 1.00000ml of P12952 + 9.00000ml of E3804 = Final Quantity: 10.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
405	AR1660 1000/100 PPB ICV STD	PP24815	08/18/2025	02/18/2026	Yogesh Patel	None	None	Abdul Mirza 08/19/2025

FROM 98.50000ml of E3962 + 0.50000ml of PP24805 + 1.00000ml of PP24814 = Final Quantity: 100.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3789	AR1221 1000 PPB WORKING SOL.2ND SOURCE(AGILENT)	PP24816	08/18/2025	09/18/2025	Yogesh Patel	None	None	Abdul Mirza
								08/19/2025

FROM 1.00000ml of P13373 + 98.50000ml of E3962 + 0.50000ml of PP24805 = Final Quantity: 100.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1887	AR1232 1000 PPB Working Sol. 2nd Source	PP24817	08/18/2025	02/18/2026	Yogesh Patel	None	None	Abdul Mirza
								08/19/2025

FROM 1.00000ml of P12700 + 98.50000ml of E3962 + 0.50000ml of PP24805 = Final Quantity: 100.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1889	AR1242 1000 PPB Working Sol. 2nd Source	PP24818	08/18/2025	02/18/2026	Yogesh Patel	None	None	Abdul Mirza 08/19/2025

FROM 1.00000ml of P13590 + 98.50000ml of E3962 + 0.50000ml of PP24805 = Final Quantity: 100.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1890	AR1248 1000 PPB Working Sol. 2nd Source	PP24819	08/18/2025	02/18/2026	Yogesh Patel	None	None	Abdul Mirza 08/19/2025

FROM 1.00000ml of P13592 + 98.50000ml of E3962 + 0.50000ml of PP24805 = Final Quantity: 100.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
2193	AR1254 100 PPM STOCK SOLUTION	PP24820	08/18/2025	02/15/2026	Yogesh Patel	None	None	Abdul Mirza 08/19/2025

FROM 1.00000ml of P13833 + 9.00000ml of E3963 = Final Quantity: 10.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3757	AR1262 1000 PPB Working Solution second source	PP24821	08/18/2025	02/18/2026	Yogesh Patel	None	None	Abdul Mirza 08/19/2025

FROM 1.00000ml of P12703 + 98.50000ml of E3962 + 0.50000ml of PP24805 = Final Quantity: 100.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3817	AR1268 1000 ppb Working Soln. 2nd source	PP24822	08/18/2025	09/18/2025	Yogesh Patel	None	None	Abdul Mirza 08/19/2025

FROM 1.00000ml of P11522 + 98.50000ml of E3962 + 0.50000ml of PP24805 = Final Quantity: 100.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
203	AR1660 750 PPB STD	PP24823	08/18/2025	02/18/2026	Yogesh Patel	None	None	Abdul Mirza 08/19/2025

FROM 0.25000ml of W3234 + 0.75000ml of PP24330 = Final Quantity: 1.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
204	AR1660 500 PPB STD	PP24824	08/18/2025	02/18/2026	Yogesh Patel	None	None	Abdul Mirza
								08/19/2025

FROM 0.50000ml of W3234 + 0.50000ml of PP24330 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
205	AR1660 250 PPB STD	PP24825	08/18/2025	02/18/2026	Yogesh Patel	None	None	Abdul Mirza
								08/19/2025

FROM 0.75000ml of W3234 + 0.25000ml of PP24330 = Final Quantity: 1.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
206	AR1660 50 PPB STD	PP24826	08/18/2025	02/18/2026	Yogesh Patel	None	None	Abdul Mirza
08/19/2025								

FROM 0.90000ml of W3234 + 0.10000ml of PP24824 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1079	AR1221 750 PPB STD	PP24827	08/18/2025	02/18/2026	Yogesh Patel	None	None	Abdul Mirza
08/19/2025								

FROM 0.25000ml of W3234 + 0.75000ml of PP24806 = Final Quantity: 1.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
222	AR1221 500 PPB STD	PP24828	08/18/2025	02/18/2026	Yogesh Patel	None	None	Abdul Mirza
								08/19/2025

FROM 0.50000ml of W3234 + 0.50000ml of PP24806 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1080	AR1221 250 PPB STD	PP24829	08/18/2025	02/18/2026	Yogesh Patel	None	None	Abdul Mirza
								08/19/2025

FROM 0.75000ml of W3234 + 0.25000ml of PP24806 = Final Quantity: 1.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1081	AR1221 50 PPB STD	PP24830	08/18/2025	02/18/2026	Yogesh Patel	None	None	Abdul Mirza
08/19/2025								

FROM 0.90000ml of W3234 + 0.10000ml of PP24828 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1063	AR1232 750 PPB STD	PP24831	08/18/2025	02/18/2026	Yogesh Patel	None	None	Abdul Mirza
08/19/2025								

FROM 0.25000ml of W3234 + 0.75000ml of PP24807 = Final Quantity: 1.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
223	AR1232 500 PPB STD	PP24832	08/18/2025	02/18/2026	Yogesh Patel	None	None	Abdul Mirza
								08/19/2025

FROM 0.50000ml of W3234 + 0.50000ml of PP24807 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1064	AR1232 250 PPB STD	PP24833	08/18/2025	02/18/2026	Yogesh Patel	None	None	Abdul Mirza
								08/19/2025

FROM 0.75000ml of W3234 + 0.25000ml of PP24807 = Final Quantity: 1.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1065	AR1232 50 PPB STD	PP24834	08/18/2025	02/18/2026	Yogesh Patel	None	None	Abdul Mirza
								08/19/2025

FROM 0.90000ml of W3234 + 0.10000ml of PP24832 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1067	AR1242 750 PPB STD	PP24835	08/18/2025	02/18/2026	Yogesh Patel	None	None	Abdul Mirza
								08/19/2025

FROM 0.25000ml of W3234 + 0.75000ml of PP24808 = Final Quantity: 1.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
224	AR1242 500 PPB STD	PP24836	08/18/2025	02/18/2026	Yogesh Patel	None	None	Abdul Mirza
								08/19/2025

FROM 0.50000ml of W3234 + 0.50000ml of PP24808 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1068	AR1242 250 PPB STD	PP24837	08/18/2025	02/18/2026	Yogesh Patel	None	None	Abdul Mirza
								08/19/2025

FROM 0.75000ml of W3234 + 0.25000ml of PP24808 = Final Quantity: 1.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1069	AR1242 50 PPB STD	PP24838	08/18/2025	02/18/2026	Yogesh Patel	None	None	Abdul Mirza
								08/19/2025

FROM 0.90000ml of W3234 + 0.10000ml of PP24836 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1075	AR1248 750 PPB STD	PP24839	08/18/2025	02/18/2026	Yogesh Patel	None	None	Abdul Mirza
								08/19/2025

FROM 0.25000ml of W3234 + 0.75000ml of PP24809 = Final Quantity: 1.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
225	AR1248 500 PPB STD	PP24840	08/18/2025	02/18/2026	Yogesh Patel	None	None	Abdul Mirza
								08/19/2025

FROM 0.50000ml of W3234 + 0.50000ml of PP24809 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1076	AR1248 250 PPB STD	PP24841	08/18/2025	02/18/2026	Yogesh Patel	None	None	Abdul Mirza
								08/19/2025

FROM 0.75000ml of W3234 + 0.25000ml of PP24809 = Final Quantity: 1.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1077	AR1248 50 PPB STD	PP24842	08/18/2025	02/18/2026	Yogesh Patel	None	None	Abdul Mirza
								08/19/2025

FROM 0.90000ml of W3234 + 0.10000ml of PP24840 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1071	AR1254 750 PPB STD	PP24843	08/18/2025	02/18/2026	Yogesh Patel	None	None	Abdul Mirza
								08/19/2025

FROM 0.25000ml of W3234 + 0.75000ml of PP24810 = Final Quantity: 1.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
226	AR1254 500 PPB STD	PP24844	08/18/2025	02/18/2026	Yogesh Patel	None	None	Abdul Mirza
								08/19/2025

FROM 0.50000ml of W3234 + 0.50000ml of PP24810 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1072	AR1254 250 PPB STD	PP24845	08/18/2025	02/18/2026	Yogesh Patel	None	None	Abdul Mirza
								08/19/2025

FROM 0.75000ml of W3234 + 0.25000ml of PP24810 = Final Quantity: 1.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1073	AR1254 50 PPB STD	PP24846	08/18/2025	02/18/2026	Yogesh Patel	None	None	Abdul Mirza
								08/19/2025

FROM 0.90000ml of W3234 + 0.10000ml of PP24844 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3753	AR1262 750 PPB STD	PP24847	08/18/2025	02/18/2026	Yogesh Patel	None	None	Abdul Mirza
								08/19/2025

FROM 0.25000ml of W3234 + 0.75000ml of PP24811 = Final Quantity: 1.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1530	AR1262 500 PPB STD	PP24848	08/18/2025	02/18/2026	Yogesh Patel	None	None	Abdul Mirza
								08/19/2025

FROM 0.50000ml of W3234 + 0.50000ml of PP24811 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3754	AR1262 250 PPB STD	PP24849	08/18/2025	02/18/2026	Yogesh Patel	None	None	Abdul Mirza
								08/19/2025

FROM 0.75000ml of W3234 + 0.25000ml of PP24811 = Final Quantity: 1.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3755	AR1262 50 PPB STD	PP24850	08/18/2025	02/18/2026	Yogesh Patel	None	None	Abdul Mirza
								08/19/2025

FROM 0.90000ml of W3234 + 0.10000ml of PP24848 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3820	AR1268 750 PPB STD	PP24851	08/18/2025	02/18/2026	Yogesh Patel	None	None	Abdul Mirza
								08/19/2025

FROM 0.25000ml of W3234 + 0.75000ml of PP24812 = Final Quantity: 1.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1533	AR1268 500 PPB STD	PP24852	08/18/2025	02/18/2026	Yogesh Patel	None	None	Abdul Mirza
								08/19/2025

FROM 0.50000ml of W3234 + 0.50000ml of PP24812 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3821	AR1268 250 PPB STD	PP24853	08/18/2025	02/18/2026	Yogesh Patel	None	None	Abdul Mirza
								08/19/2025

FROM 0.75000ml of W3234 + 0.25000ml of PP24812 = Final Quantity: 1.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3822	AR1268 50 PPB STD	PP24854	08/18/2025	02/18/2026	Yogesh Patel	None	None	Abdul Mirza
								08/19/2025

FROM 0.90000ml of W3234 + 0.10000ml of PP24852 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
406	AR1660 500 PPB ICV	PP24855	08/18/2025	02/18/2026	Yogesh Patel	None	None	Abdul Mirza
								08/19/2025

FROM 0.50000ml of W3234 + 0.50000ml of PP24815 = Final Quantity: 1.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3790	AR1221 500 PPB ICV(AGILENT)	PP24856	08/18/2025	09/18/2025	Yogesh Patel	None	None	Abdul Mirza
								08/19/2025

FROM 0.50000ml of W3234 + 0.50000ml of PP24816 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1888	AR1232 500 PPB ICV	PP24857	08/18/2025	02/18/2026	Yogesh Patel	None	None	Abdul Mirza
								08/19/2025

FROM 0.50000ml of W3234 + 0.50000ml of PP24817 = Final Quantity: 1.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1891	AR1242 500 PPB ICV	PP24858	08/18/2025	02/18/2026	Yogesh Patel	None	None	Abdul Mirza
								08/19/2025

FROM 0.50000ml of W3234 + 0.50000ml of PP24818 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1892	AR1248 500 PPB ICV	PP24859	08/18/2025	02/18/2026	Yogesh Patel	None	None	Abdul Mirza
								08/19/2025

FROM 0.50000ml of W3234 + 0.50000ml of PP24819 = Final Quantity: 1.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1894	AR1254 500 PPB ICV	PP24860	08/18/2025	02/18/2026	Yogesh Patel	None	None	Abdul Mirza
								08/19/2025

FROM 0.50000ml of W3234 + 0.50000ml of PP24381 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3758	AR1262 500 PPB STD ICV	PP24861	08/18/2025	02/18/2026	Yogesh Patel	None	None	Abdul Mirza
								08/19/2025

FROM 0.50000ml of W3234 + 0.50000ml of PP24821 = Final Quantity: 1.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3823	AR1268 500 PPB STD ICV	PP24862	08/18/2025	09/18/2025	Yogesh Patel	None	None	Abdul Mirza
								08/19/2025

FROM 0.50000ml of W3234 + 0.50000ml of PP24822 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3857	5000 PPB PCB SPIKE SOLUTION 2ND SOURCE	PP24930	09/18/2025	03/18/2026	Abdul Mirza	None	None	Yogesh Patel
								09/19/2025

FROM 0.50000ml of P12953 + 99.50000ml of E3972 = Final Quantity: 100.000 ml

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	9005-05 / Acetone Ultra (cs/4x4L)	24E0761004	02/18/2026	10/01/2024 / Rajesh	09/25/2024 / Rajesh	E3804

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
PCI Scientific Supply, Inc.	PC19631-100 / SODIUM SULFATE, ANHYDROUS, PEST GRADE, 1	417203	01/28/2026	07/28/2025 / RUPESH	01/29/2025 / Rajesh	E3875

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-9254-03 / Acetone, Ultra Resi (cs/4x4L)	24H1462005	05/24/2027	06/20/2025 / RUPESH	05/14/2025 / RUPESH	E3944

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-9262-03 / Hexane, Ultra-Resi (cs/4x4L)	25C0362005	04/30/2026	08/05/2025 / RUPESH	07/30/2025 / RUPESH	E3962

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-9254-03 / Acetone, Ultra Resi (cs/4x4L)	24H2762008	02/15/2026	08/15/2025 / RUPESH	08/07/2025 / RUPESH	E3963

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-9254-03 / Acetone, Ultra Resi (cs/4x4L)	24H1462005	05/24/2027	09/16/2025 / Evelyn	09/04/2025 / Riteshkumar	E3972

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-9644-A4 / Methylene Chloride,U-Resi, Cycle-Tainer (215L)	25C1262005	04/16/2026	09/15/2025 / Riteshkumar	09/15/2025 / Riteshkumar	E3973

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-9262-03 / Hexane, Ultra-Resi (cs/4x4L)	25C0362006	04/10/2027	09/26/2025 / Riteshkumar	09/26/2025 / Riteshkumar	E3974

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-9673-33 / Sulfuric Acid, Instra-Analyzed (cs/6c2.5L)	24i1262013	11/07/2025	05/07/2025 / RUPESH	02/18/2025 / Mohan	M6157

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Agilent Technologies	PP-382-1 / Aroclor 1268	0006587800	09/18/2025	03/18/2025 / yogesh	02/21/2022 / Ankita	P11522

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	32011 / PCB Mix, Aroclor 1254, 1000ug/mL, Hexane, 1mL/ampul	A0175403	02/18/2026	08/18/2025 / yogesh	03/18/2022 / Abdul	P11592

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards,Inc	91867 / Aroclor 1232 100 ug/mL	020823	02/18/2026	08/18/2025 / yogesh	08/07/2023 / Ankita	P12700

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards, Inc	x9166 / Aroclor 1262 100 ug/mL	060523	02/18/2026	08/18/2025 / yogesh	08/07/2023 / Ankita	P12703

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards, Inc.	20064 / Aroclor 1016/1260	022023	02/18/2026	08/18/2025 / yogesh	12/20/2023 / Yogesh	P12952

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards, Inc.	20064 / Aroclor 1016/1260	022023	03/18/2026	09/18/2025 / Abdul	12/20/2023 / Yogesh	P12953

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards, Inc.	/ Arochlor 1254	121823	04/03/2025	10/03/2024 / Ankita	12/20/2023 / Yogesh	P12957

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	32000 / Pesticide Mix, CLP method, Pesticide Surrogate Mix, 200ug/mL, Acetone, 1mL	A0206810	09/18/2025	03/18/2025 / yogesh	04/22/2024 / Abdul	P13356

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Agilent Technologies	PP-292-1 / Aroclor 1221	0006783205	09/18/2025	03/18/2025 / yogesh	05/02/2024 / Ankita	P13373

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	32007 / PCB Mix, Aroclor 1221, 1000ug/mL, Hexane, 1mL/ampul	A0207511	02/18/2026	08/18/2025 / yogesh	05/03/2024 / Abdul	P13375

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Agilent Technologies	PP-312-1 / Aroclor 1242	0006665550	02/18/2026	08/18/2025 / yogesh	10/14/2024 / Ankita	P13590

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Agilent Technologies	PP-342-1 / Aroclor 1248	0006726317	02/18/2026	08/18/2025 / yogesh	10/14/2024 / Ankita	P13592

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	32010 / PCB Mix, Aroclor 1248, 1000ug/mL, Hexane, 1mL/ampul	A0214733	02/18/2026	08/18/2025 / yogesh	10/17/2024 / yogesh	P13695

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	32039 / PCB Mix, Aroclor 1016/1260, 1000ug/mL, hexane, 1mL/ampul	A0210629	09/18/2025	03/18/2025 / yogesh	10/17/2024 / yogesh	P13697

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	32009 / PCB Mix, Aroclor 1242, 1000ug/mL, Hexane, 1mL/ampul	A0210232	02/18/2026	08/18/2025 / yogesh	10/17/2024 / yogesh	P13706

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	32000 / Pesticide Mix, CLP method, Pesticide Surrogate Mix, 200ug/mL, Acetone, 1mL	A0214495	12/24/2025	06/24/2025 / Abdul	11/19/2024 / Ankita	P13786

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	32000 / Pesticide Mix, CLP method, Pesticide Surrogate Mix, 200ug/mL, Acetone, 1mL	A0214495	02/18/2026	08/18/2025 / yogesh	11/19/2024 / Ankita	P13789

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	32011 / PCB Mix, Aroclor 1254, 1000ug/mL, Hexane, 1mL/ampul	a0217391	02/18/2026	08/18/2025 / yogesh	12/09/2024 / Ankita	P13833

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	32008 / PCB Mix, Aroclor 1232, 1000ug/mL, Hexane, 1mL/ampul	A0219655	02/18/2026	08/18/2025 / yogesh	01/23/2025 / Ankita	P13879

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	32409 / PCB Stock Solution, Aroclor 1262 Std, 1mL, Hexane	A0217041	02/18/2026	08/18/2025 / yogesh	01/23/2025 / Ankita	P13881

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	32410 / PCB Stock Solution, Aroclor 1268 Std, 1mL, Hexane	A0217264	02/18/2026	08/18/2025 / yogesh	01/23/2025 / Ankita	P13885

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	DIW / DI Water	Daily Lab-Certified	07/03/2029	07/03/2024 / lwona	07/03/2024 / lwona	W3112

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-9262-03 / Hexane, Ultra-Resi (cs/4x4L)	24G1962003	08/22/2025	02/03/2025 / jignesh	01/31/2025 / jignesh	W3177

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-9262-03 / Hexane, Ultra-Resi (cs/4x4L)	25C0362005	04/30/2026	07/28/2025 / jignesh	07/25/2025 / jignesh	W3234

Material No.: 9005-05
Batch No.: 24E0761004
Manufactured Date: 2024-05-02
Retest Date: 2029-05-01
Revision No.: 0

Certificate of Analysis

Test	Specification	Result
Assay ((CH ₃) ₂ CO) (by GC, corrected for water)	≥ 99.5 %	99.8 %
Color (APHA)	≤ 10	< 5
Residue after Evaporation	≤ 5 ppm	< 1 ppm
Titration Acid (μeq/g)	≤ 0.3	0.1
Titration Base (μeq/g)	≤ 0.5	0.1
Water (H ₂ O)	≤ 0.5 %	0.1 %
Solubility in H ₂ O	Passes Test	Passes Test
Chloride (Cl)	≤ 0.2 ppm	< 0.2 ppm
Phosphate (PO ₄)	≤ 0.05 ppm	< 0.05 ppm
Trace Impurities – Aluminum (Al)	≤ 50.0 ppb	< 5.0 ppb
Arsenic and Antimony (as As)	≤ 5.0 ppb	< 5.0 ppb
Trace Impurities – Barium (Ba)	≤ 20.0 ppb	< 1.0 ppb
Trace Impurities – Beryllium (Be)	≤ 10.0 ppb	< 1.0 ppb
Trace Impurities – Bismuth (Bi)	≤ 20.0 ppb	< 10.0 ppb
Trace Impurities – Boron (B)	≤ 10.0 ppb	< 5.0 ppb
Trace Impurities – Cadmium (Cd)	≤ 10.0 ppb	< 1.0 ppb
Trace Impurities – Calcium (Ca)	≤ 25.0 ppb	3.6 ppb
Trace Impurities – Chromium (Cr)	≤ 10.0 ppb	< 1.0 ppb
Trace Impurities – Cobalt (Co)	≤ 10.0 ppb	< 1.0 ppb
Trace Impurities – Copper (Cu)	≤ 10.0 ppb	< 1.0 ppb
Trace Impurities – Gallium (Ga)	≤ 10.0 ppb	< 1.0 ppb
Trace Impurities – Germanium (Ge)	≤ 10.0 ppb	< 10.0 ppb
Trace Impurities – Gold (Au)	≤ 20 ppb	< 5 ppb
Trace Impurities – Iron (Fe)	≤ 20.0 ppb	< 1.0 ppb
Trace Impurities – Lead (Pb)	≤ 10.0 ppb	< 10.0 ppb
Trace Impurities – Lithium (Li)	≤ 10.0 ppb	< 1.0 ppb
Trace Impurities – Magnesium (Mg)	≤ 20 ppb	< 1 ppb
Trace Impurities – Manganese (Mn)	≤ 10.0 ppb	< 1.0 ppb

Recd by RP on 9/25/24

E 3804

>>> Continued on page 2 >>>

Acetone
CMOS

avantor™



Material No.: 9005-05
Batch No.: 24E0761004

Test	Specification	Result
Trace Impurities – Molybdenum (Mo)	≤ 10.0 ppb	< 5.0 ppb
Trace Impurities – Nickel (Ni)	≤ 10.0 ppb	< 5.0 ppb
Trace Impurities – Niobium (Nb)	≤ 50.0 ppb	< 1.0 ppb
Trace Impurities – Potassium (K)	≤ 10.0 ppb	< 10.0 ppb
Trace Impurities – Silicon (Si)	≤ 50 ppb	< 10 ppb
Trace Impurities – Silver (Ag)	≤ 10.0 ppb	< 1.0 ppb
Trace Impurities – Sodium (Na)	≤ 10.0 ppb	< 1.0 ppb
Trace Impurities – Strontium (Sr)	≤ 10.0 ppb	< 1.0 ppb
Trace Impurities – Tantalum (Ta)	≤ 50.0 ppb	< 5.0 ppb
Trace Impurities – Thallium (Tl)	≤ 10.0 ppb	< 5.0 ppb
Trace Impurities – Tin (Sn)	≤ 20.0 ppb	< 10.0 ppb
Trace Impurities – Titanium (Ti)	≤ 10.0 ppb	< 1.0 ppb
Trace Impurities – Vanadium (V)	≤ 10.0 ppb	< 1.0 ppb
Trace Impurities – Zinc (Zn)	≤ 20.0 ppb	7.9 ppb
Trace Impurities – Zirconium (Zr)	≤ 10.0 ppb	< 1.0 ppb
Particle Count – 0.5 µm and greater (Rion KS42AF)	≤ 100 par/ml	8 par/ml
Particle Count – 1.0 µm and greater (Rion KS42AF)	≤ 8 par/ml	2 par/ml

>>> Continued on page 3 >>>

n-Hexane 95%
ULTRA RESI-ANALYZED
For Organic Residue Analysis



Material No.: 9262-03
Batch No.: 25C0362006
Manufactured Date: 2025-01-29
Expiration Date: 2026-04-30
Revision No.: 0

Certificate of Analysis

Test	Specification	Result
FID-Sensitive Impurities (as 2-Octanol) Single Impurity Peak (ng/mL)	≤ 5	1
ECD Sensitive Impurities (as Heptachlor Epoxide) Single Peak (pg/mL)	≤ 10	6
ECD-Sensitive Impurities (as Ethylene Dibromide) - Single Impurity Peak (ng/mL)	≤ 5	4
Assay (Total Saturated C ₆ Isomers) (by GC, corrected for water)	$\geq 99.5\%$	100.0 %
Assay (as n-Hexane) (by GC, corrected for water)	$\geq 95\%$	100 %
Color (APHA)	≤ 10	10
Residue after Evaporation	≤ 1.0 ppm	0.2 ppm
Substances Darkened by H ₂ SO ₄	Passes Test	Passes Test
Water (by KF, coulometric)	$\leq 0.05\%$	$< 0.01\%$

For Laboratory, Research, or Manufacturing Use
MEETS SPECIFICATIONS WITHIN THE EXPIRATION PERIOD

Country of Origin: United States
Packaging Site: Phillipsburg Mfg Ctr & DC

E 3974

Jamie Croak
Director Quality Operations, Bioscience Production

For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.386.1700



**PRODUCTOS
QUÍMICOS
MONTERREY, S.A. DE C.V.**

MIRADOR 201, COL. MIRADOR
MONTERREY, N.L. MÉXICO
CP 64070
TEL +52 81 13 52 67 67
www.pqm.com.mx

CERTIFICATE OF ANALYSIS

PRODUCT :	SODIUM SULFATE CRYSTALS ANHYDROUS		
QUALITY :	ACS (CODE RMB3375)	FORMULA :	Na ₂ SO ₄
SPECIFICATION NUMBER:	6399	RELEASE DATE:	MAY/23/2024
LOT NUMBER :	417203		

TEST	SPECIFICATIONS	LOT VALUES
Assay (Na ₂ SO ₄)	Min. 99.0%	99.8 %
pH of a 5% solution at 25°C	5.2 - 9.2	6.2
Insoluble matter	Max. 0.01%	0.001 %
Loss on ignition	Max. 0.5%	0.1 %
Chloride (Cl)	Max. 0.001%	<0.001 %
Nitrogen compounds (as N)	Max. 5 ppm	<5 ppm
Phosphate (PO ₄)	Max. 0.001%	<0.001 %
Heavy metals (as Pb)	Max. 5 ppm	<5 ppm
Iron (Fe)	Max. 0.001%	<0.001 %
Calcium (Ca)	Max. 0.01%	0.001 %
Magnesium (Mg)	Max. 0.005%	0.001 %
Potassium (K)	Max. 0.008%	0.001 %
Extraction-concentration suitability	Passes test	Passes test
Appearance	Passes test	Passes test
Identification	Passes test	Passes test
Solubility and foreign matter	Passes test	Passes test
Retained on US Standard No. 10 sieve	Max. 1%	0.2 %
Retained on US Standard No. 60 sieve	Min. 94%	96.2 %
Through US Standard No. 60 sieve	Max. 5%	3.5 %
Through US Standard No. 100 sieve	Max. 10%	0.1 %

COMMENTS

QC: PhC Irma Belmares

If you need further details, please call our factory or contact our local distributor.

RE-02-01, Ed. 3

E 3875

Acetone
BAKER RESI-ANALYZED® Reagent
For Organic Residue Analysis

avantor™



Material No.: 9254-03
Batch No.: 24H1462005
Manufactured Date: 2024-05-24
Expiration Date: 2027-05-24
Revision No.: 0

Certificate of Analysis

Test	Specification	Result
Assay ((CH ₃) ₂ CO) (by GC, corrected for water)	>= 99.4 %	99.8 %
Color (APHA)	<= 10	5
Residue after Evaporation	<= 1.0 ppm	0.2 ppm
Substances Reducing Permanganate	Passes Test	Passes Test
Titration Acid (μeq/g)	<= 0.3	0.2
Titration Base (μeq/g)	<= 0.6	<0.1
Water (H ₂ O)	<= 0.5 %	0.2 %
FID-Sensitive Impurities (as 2-Octanol) Single Impurity Peak (ng/mL)	<= 5	<1
ECD Sensitive Impurities (as Heptachlor Epoxide) Single Peak (pg/mL)	<= 10	1

For Laboratory, Research, or Manufacturing Use
MEETS SPECIFICATIONS WITHIN THE EXPIRATION PERIOD

Country of Origin: United States
Packaging Site: Phillipsburg Mfg Ctr & DC

E3944

Jamie Croak
Director Quality Operations, Bioscience Production

For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.386.1700

n-Hexane 95%
ULTRA RESI-ANALYZED
For Organic Residue Analysis



Material No.: 9262-03
Batch No.: 25C0362005
Manufactured Date: 2025-01-29
Expiration Date: 2026-04-30
Revision No.: 0

Certificate of Analysis

Test	Specification	Result
FID-Sensitive Impurities (as 2-Octanol) Single Impurity Peak (ng/mL)	≤ 5	1
ECD Sensitive Impurities (as HeptachlorEpoxide) Single Peak (pg/mL)	≤ 10	6
ECD-Sensitive Impurities (as EthyleneDibromide) - Single Impurity Peak (ng/mL)	≤ 5	5
Assay (Total Saturated C ₆ Isomers) (by GC, corrected for water)	$\geq 99.5 \%$	100.0 %
Assay (as n-Hexane) (by GC, corrected for water)	$\geq 95 \%$	100 %
Color (APHA)	≤ 10	10
Residue after Evaporation	≤ 1.0 ppm	0.1 ppm
Substances Darkened by H ₂ SO ₄	Passes Test	Passes Test
Water (by KF, coulometric)	$\leq 0.05 \%$	$< 0.01 \%$

For Laboratory, Research, or Manufacturing Use
MEETS SPECIFICATIONS WITHIN THE EXPIRATION PERIOD

Country of Origin: United States
Packaging Site: Phillipsburg Mfg Ctr & DC

Received on 7/30/25

E3962

Jamie Croak
Director Quality Operations, Bioscience Production

For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.386.1700

Acetone.
BAKER RESI-ANALYZED® Reagent
For Organic Residue Analysis

Avantor™



Material No.: 9254-03
Batch No.: 24H2762008
Manufactured Date: 2024-04-18
Expiration Date: 2027-04-18
Revision No.: 0

Certificate of Analysis

Test	Specification	Result
Assay ((CH ₃) ₂ CO) (by GC, corrected for water)	>= 99.4 %	100.0 %
Color (APHA)	<= 10	5
Residue after Evaporation	<= 1.0 ppm	0.0 ppm
Substances Reducing Permanganate	Passes Test	Passes Test
Titration Acid (μeq/g)	<= 0.3	0.2
Titration Base (μeq/g)	<= 0.6	<0.1
Water (H ₂ O)	<= 0.5 %	<0.1 %
FID-Sensitive Impurities (as 2-Octanol) Single Impurity Peak (ng/mL)	<= 5	1
ECD Sensitive Impurities (as Heptachlor Epoxide) Single Peak (pg/mL)	<= 10	1

For Laboratory, Research, or Manufacturing Use
MEETS SPECIFICATIONS WITHIN THE EXPIRATION PERIOD

Country of Origin: United States
Packaging Site: Phillipsburg Mfg Ctr & DC

Received on 8/6/25

E 3963

J. Croak

Jamie Croak
Director Quality Operations, Bioscience Production

For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.386.1700

Avantor Performance Materials, LLC

100 Matsonford Rd, Suite 200, Radnor, PA, 19087, U.S.A. Phone 610.386.1700

Acetone
BAKER RESI-ANALYZED® Reagent
For Organic Residue Analysis

avantor™



Material No.: 9254-03
Batch No.: 24H1462005
Manufactured Date: 2024-05-24
Expiration Date: 2027-05-24
Revision No.: 0

Certificate of Analysis

Test	Specification	Result
Assay ((CH ₃) ₂ CO) (by GC, corrected for water)	>= 99.4 %	99.8 %
Color (APHA)	<= 10	5
Residue after Evaporation	<= 1.0 ppm	0.2 ppm
Substances Reducing Permanganate	Passes Test	Passes Test
Titration Acid (µeq/g)	<= 0.3	0.2
Titration Base (µeq/g)	<= 0.6	<0.1
Water (H ₂ O)	<= 0.5 %	0.2 %
FID-Sensitive Impurities (as 2-Octanol) Single Impurity Peak (ng/mL)	<= 5	<1
ECD Sensitive Impurities (as Heptachlor Epoxide) Single Peak (pg/mL)	<= 10	1

For Laboratory, Research, or Manufacturing Use
MEETS SPECIFICATIONS WITHIN THE EXPIRATION PERIOD

Country of Origin: United States
Packaging Site: Phillipsburg Mfg Ctr & DC

E3972

Jamie Croak
Director Quality Operations, Bioscience Production

For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.386.1700

Sulfuric Acid
BAKER INSTRA-ANALYZED® Reagent
For Trace Metal Analysis
Low Selenium

avantor™



MG157
MS

Material No.: 9673-33

Batch No.: 24I1262013

Manufactured Date: 2024-08-07

Retest Date: 2029-08-06

Revision No.: 0

Certificate of Analysis

Test	Specification	Result
ACS – Assay (H ₂ SO ₄)	95.0 – 98.0 %	96.2 %
Appearance	Passes Test	Passes Test
ACS – Color (APHA)	<= 10	5
ACS – Residue after Ignition	<= 3 ppm	<1 ppm
ACS – Substances Reducing Permanganate(as SO ₂)	<= 2 ppm	<2 ppm
Ammonium (NH ₄)	<= 1 ppm	<1 ppm
Chloride (Cl)	<= 0.1 ppm	<0.1 ppm
Nitrate (NO ₃)	<= 0.2 ppm	0.1 ppm
Phosphate (PO ₄)	<= 0.5 ppm	<0.1 ppm
Trace Impurities – Aluminum (Al)	<= 30.0 ppb	<5.0 ppb
Arsenic & Antimony (as As)	<= 4.0 ppb	<2.0 ppb
Trace Impurities – Boron (B)	<= 10.0 ppb	<5.0 ppb
Trace Impurities – Cadmium (Cd)	<= 2.0 ppb	<1.0 ppb
Trace Impurities – Chromium (Cr)	<= 6.0 ppb	<1.0 ppb
Trace Impurities – Cobalt (Co)	<= 0.5 ppb	<0.3 ppb
Trace Impurities – Copper (Cu)	<= 1.0 ppb	<1.0 ppb
Trace Impurities – Gold (Au)	<= 10.0 ppb	<5.0 ppb
Heavy Metals (as Pb)	<= 500.0 ppb	<100.0 ppb
Trace Impurities – Iron (Fe)	<= 50.0 ppb	<1.0 ppb
Trace Impurities – Lead (Pb)	<= 0.5 ppb	<0.5 ppb
Trace Impurities – Magnesium (Mg)	<= 7.0 ppb	<1.0 ppb
Trace Impurities – Manganese (Mn)	<= 1.0 ppb	<1.0 ppb
Trace Impurities – Mercury (Hg)	<= 0.5 ppb	<0.1 ppb
Trace Impurities – Nickel (Ni)	<= 2.0 ppb	<0.3 ppb
Trace Impurities – Potassium (K)	<= 500.0 ppb	<10.0 ppb
Trace Impurities – Selenium (Se)	<= 50.0 ppb	7.2 ppb
Trace Impurities – Silicon (Si)	<= 100.0 ppb	12.8 ppb
Trace Impurities – Silver (Ag)	<= 1.0 ppb	<1.0 ppb

For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.386.1705

Sulfuric Acid

BAKER INSTRA-ANALYZED® Reagent

For Trace Metal Analysis

Low Selenium

 **avantor**™



Material No.: 9673-33

Batch No.: 24I1262013

Test	Specification	Result
Trace Impurities – Sodium (Na)	≤ 500.0 ppb	< 5.0 ppb
Trace Impurities – Strontium (Sr)	≤ 5.0 ppb	< 1.0 ppb
Trace Impurities – Tin (Sn)	≤ 5.0 ppb	1.1 ppb
Trace Impurities – Zinc (Zn)	≤ 5.0 ppb	< 1.0 ppb

For Laboratory, Research, or Manufacturing Use

Country of Origin: United States

Packaging Site: Phillipsburg Mfg Ctr & DC



Jamie Croak

Director Quality Operations, Bioscience Production



Certificate of Analysis

P11518
↓
P11522
AJ
02/21/22

Product Name: Aroclor 1268 Standard

Product Number: PP-382-1

Lot Issue Date: 09-Feb-2021

Lot Number: 0006587800

Expiration Date: 31-Mar-2029

Description:

This analytical reference material (RM) was manufactured and verified in accordance with an ISO 9001 registered quality system and analyte concentrations were verified by an ISO 17025 accredited laboratory. The concentration and uncertainty value at the 95% confidence level for each analyte, determined gravimetrically, is listed below.

Analyte	CAS#	Analyte Lot	Concentration ± Uncertainty
Aroclor 1268	011100-14-4	RM00937	100.0 ± 0.5 µg/mL

Matrix: isooctane (2,2,4-trimethylpentane)

Storage Conditions: Store at Room Temperature (15° to 30°C).

Traceability:

The balances used for these measurements are calibrated with weights traceable to NIST in compliance with ANSI/NCSL Z540.3, ISO 9001, ISO 17025, and ISO 17034. Calibrated Class A glassware is used for volumetric measurements. Thermometers are calibrated against a NIST traceable thermometer in accordance with NIST Special Publication 1088.

Homogeneity:

This RM was unitized according to an in-house procedure and is guaranteed to be homogeneous. There is no minimum sub-sample size required.

Intended Use:

This RM is intended for the preparation of working reference samples for use in routine laboratory analyses, calibration of instruments, validation of analytical methods, assessments of measurement methods, and continuing calibration verification.

Instructions for Use:

Sample aliquots for analysis should be withdrawn at 20°C to 25°C immediately after opening the container and should be processed without delay for the certified values to be valid within the stated uncertainties.

Hazards:

Refer to the Safety Data Sheet on www.agilent.com for information regarding this RM.


Expiration of Certification:

The certification of this RM is valid until the expiration date specified above, provided the RM is handled and stored in accordance with the instructions given in this certificate. This certification is nullified if the RM is damaged, contaminated, or otherwise modified.

Maintenance of Certification:

If substantive changes are noted that affect the certification before the expiration of this certificate, Agilent will notify the purchaser.

Sample lot approver:


Monica Bourgeois
QMS Representative



ISO 17034 Cert
No. AR-1936

RM was produced in accordance with TUV USA Inc registered ISO 9001 Quality Management System. Cert # 56 100 18560026
Page: 1 of 1

www.agilent.com/quality/
CSD-QA-015.1



ISO 17025 Cert
No. AT-1937



CERTIFIED REFERENCE MATERIAL

110 Benner Circle
Bellefonte, PA 16823-8812
Tel: (800)356-1688
Fax: (814)353-1309

www.restek.com

Certificate of Analysis



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 32011 **Lot No.:** A0175403

Description : Aroclor® 1254 Standard

Aroclor® 1254 Standard 1,000 µg/mL, Hexane, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : November 30, 2027 **Storage:** 25°C nominal

Handling: This product contains PCBs. **Ship:** Ambient

CERTIFIED VALUES

Elution Order	Compound	Grav. Conc. (weight/volume)	Expanded Uncertainty (95% C.L.; K=2)
1	Aroclor 1254 CAS # 11097-69-1 (Lot 124-191-B) Purity ----%	1,000.7 µg/mL	+/- 5.9437 µg/mL Gravimetric +/- 31.7284 µg/mL Unstressed +/- 41.4406 µg/mL Stressed

Solvent: Hexane
CAS # 110-54-3
Purity 99%

P11588
↓
P11592 / (S)

AR
04/30/2022

Column:
30m x .25mm x .2um
Rtx-CLP II (cat.# 11323)


Carrier Gas:
helium-constant pressure 20 psi.

Temp. Program:
200°C to 300°C
@ 25°C/min. (hold 10 min.)


Inj. Temp:
250°C

Det. Temp:
300°C


Det. Type:
ECD



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.


Cathleen Soltis - Mix Technician

Date Mixed: 15-Aug-2021 Balance: 1128360905


Alexis Shelow - Operations Tech I

Date Passed: 17-Aug-2021

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

P11588
↓
P11592

AR
04/30/22



CERTIFIED WEIGHT REPORT

Part Number: 91867
Lot Number: 020823
Description: WP 037 - Aroclor 1232
PCB Technical Mixture
Expiration Date: 020833
Recommended Storage: Ambient (20 °C)
Nominal Concentration (µg/mL): 100
NIST Test ID#: 6UTB

Solvent/
Acet


Weight(s) shown below were combined and diluted to (mL): 100.0
5E-05 Balance Uncertainty
0.057 Flask Uncertainty

Compound	RM#	Lot Number	Nominal Conc (µg/mL)	Purity (%)	Uncertainty Purity	Target Weight (g)
1. Aroclor 1232	17	45-6A	100	100	0.5	0.01000

- The certified value is the concentration calculated from gravimetric and volumetric measurements unless otherwise stated.
- Standards are prepared gravimetrically using balances that are calibrated with weights traceable to NIST (see above).
- Standards are certified (+/-) 0.5% of the stated value, unless otherwise stated.
- All Standards, after opening ampule, should be stored with caps tight and under appropriate laboratory conditions.
- Uncertainty Reference: Taylor, B.N. and Kuyat, C.E., "Guidelines for Evaluating and Expressing the Uncertainty of NIST Measure Technical Note 1297, U.S. Government Printing Office, Washington, DC, (1994).

Comments

GC3-M1 Analysis by Melissa Stonier
Column ID SPB-608 30 meter X 0.53mm X 5µm film thickness
Flow rates: Helium (carrier) = 5mL/min, Helium (make-up) = 25mL/min
Hydrogen (make-up) = 30mL/min, Air (make-up) = 350mL/min
Oven Profile: Temp 1 = 150°C (Time 1 = 4 min), Temp 2 = 290°C (Time 2 = 13.5 min)
Rate = 8°C/min, Total run time = 35 min
Injector temp. = 200°C, FID Temp. = 300°C. FID Signal = Edaq Channel 1
Standard injection = 1.5µL, Range=3






CERTIFIED WEIGHT REPORT

Part Number: X9166 Solvent(s):
Lot Number: 060523 Methanol
Description: Aroclor 1262
Appendix IX Compound
Expiration Date: 060533
Recommended Storage: Ambient (20 °C)
Nominal Concentration (µg/mL): 100
NIST Test ID#: 6UTB 5E-05 Balance Uncertainty
Volume(s) shown below were combined and diluted to (mL): 20.0 0.002 Flask Uncertainty

Compound	Part Number	Lot Number	Dil. Factor	Initial Vol. (mL)	Uncertainty Pipette (mL)
1. Aroclor 1262	70444	080322	0.10	2.00	0.017

Method GC7MSD-7.M: Column:(30m X 0.25mm ID X 0.25µm film thickness min.), Rate = 8°C/min., Injector B= 200°C, Detector B = 290°C.



- The certified value is the concentration calculated from gravimetric and volumetric measurements.
- Standards are prepared gravimetrically using balances that are calibrated with weights.
- Standards are certified (+/-) 0.5% of the stated value, unless otherwise stated.
- All Standards, after opening ampule, should be stored with caps tight and under appropriate conditions.
- Uncertainty Reference: Taylor, B.N. and Kuyat, C.E., "Guidelines for Evaluating and Expressing the Uncertainty of Measurement Results," NIST Technical Note 1297, U.S. Government Printing Office, Washington, DC, (1994).



Certified Reference Material CRM

CERTIFIED WEIGHT REPORT

Part Number:
Lot Number:
Description:
Expiration Date:
Recommended Storage:
Nominal Concentration (µg/mL):
NIST Test ID#:

20064
022023
CLP PCB'S - Arador Mix
Aroclors 1016 & 1260
022033
Ambient (20 °C)
1000
6UTB

Solvent(s):	Lot#
Hexane	273615

022023	DATE
<i>Benson Chan</i>	
Formulated By:	Benson Chan
022023	DATE
<i>Pedro L. Rentas</i>	
Reviewed By:	Pedro L. Rentas


5E-05	Balance Uncertainty
0.010	Flask Uncertainty

Weight(s) shown below were combined and diluted to (mL):

SDS information											
(Solvent Safety Info. On Attached pg.)											
Compound	Lot Number	Nominal Conc (µg/mL)	Purity (%)	Uncertainty	Target Weight(g)	Actual Weight(g)	Actual Conc (µg/mL)	Expanded Uncertainty (+/-) (µg/mL)	CAS#	OSHA PEL (TWA)	
											LD50
1. Aroclor 1016	15	020491JC	1000	100	0.2	0.20004	0.20060	1002.8	4.0	12674-11-2	N/A
2. Aroclor 1260	21	020491JC	1000	100	0.2	0.20004	0.20081	1003.9	4.0	11096-82-5	0.5mg/m3 ori-rat 1315mg/kg

- The certified value is the concentration calculated from gravimetric and volumetric measurements unless otherwise stated.
- Standards are prepared gravimetrically using balances that are calibrated with weights traceable to NIST (see above).
- Standards are certified (+/-) 0.5% of the stated value, unless otherwise stated.
- All Standards, after opening sample, should be stored with caps tight and under appropriate laboratory conditions.
- Uncertainty Reference: Taylor, B.N. and Kuyat, C.E., "Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Result," NIST Technical Note 1297, U.S. Government Printing Office, Washington, DC, (1994).

Comments:
GC3-A11 Analysis by Melissa Sznitser
Column ID: SFB-608 30 meter X 0.53mm X5um film thickness
Flow rates: Helium (Carrier) = 5mL/min, Helium (make-up) = 25mL/min
Hydrogen (make-up) = 30mL/min, Air (make-up) = 350mL/min
Oven Profile: Temp 1 = 150 C (Time 1 = 4 min), Temp 2 = 280 C (Time 2 = 13.5 min)
Rate = 8 C/min, Total run time = 35 min
Injector temp = 200 C, FID Temp = 300 C, FID Signal = Etdaq Channel 1
Standard Injection = 1.5uL, Range=3



- 1
- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13
- 14
- 15
- 16
- 17
- 18

- 1
- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13
- 14
- 15
- 16
- 17
- 18



CERTIFIED WEIGHT REPORT

Part Number: 99139
Lot Number: 121823
Description: Aroclor 1254

Solvent(s): Iso-octane
Lot#: 82227

Formulated By:	Anthony Mahoney	121823	DATE
Reviewed By:	Pedro L. Rentas	121823	DATE

P12956 / Y.P.
L
P12957 / Z/PB/PB

Expiration Date: 121833
Recommended Storage: Ambient (20 °C)
Nominal Concentration (µg/mL): 100
NIST Test ID#: 6UTB
Volume(s) shown below were combined and diluted to (mL): 20.0
Note: Aroclor 1254 is a mix of isomers.


5E-05 Balance Uncertainty
0.003 Flask Uncertainty

Compound	Part Number	Lot Number	Dilution Factor	Initial Vol. (mL)	Uncertainty Pipette (mL)	Initial Conc. (µg/mL)	Final Conc. (µg/mL)	Expanded Uncertainty (+/-) (µg/mL)	SDS Information		
									(Solvent Safety Info. On Attached pg.)	CAS#	OSHA PEL (TWA) LD50
1. Aroclor 1254	79100	121823	0.10	2.00	0.017	1003.3	100.1	1.8	11097-69-1	0.5mg/m3 (skin)	or-rat 1295mg/kg

* The certified value is the concentration calculated from gravimetric and volumetric measurements unless otherwise stated.
* Standards are prepared gravimetrically using balances that are calibrated with weights traceable to NIST (see above).
* Standards are certified (+/-) 0.5% of the stated value, unless otherwise stated.
* All Standards, after opening ampule, should be stored with caps tight and under appropriate laboratory conditions.
* Uncertainty Reference: Taylor, B.N. and Kuyat, C.E., "Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Result," NIST Technical Note 1297, U.S. Government Printing Office, Washington, DC, (1994).

Comments
GC3-M1 Analysis by Melissa Stonier
Column ID SPB-608 30 meter X 0.53mm X5µm film thickness
Flow rates: Helium (carrier) = 5mL/min, Helium (make-up) = 25mL/min
Hydrogen (make-up) = 30mL/min, Air (make-up) = 350mL/min
Oven Profile: Temp 1 = 150 °C (Time 1 = 4 min), Temp 2 = 250 °C (Time 2 = 13.5 min)
Rate = 8 °C/min, Total run time = 35 min
Injector temp. = 200 °C, FID Temp. = 300 °C, FID Signal = Edaq Channel 1
Standard Injection = 1.5µL, Range=3

Peak No.	Name	FID RT (min.)
1	Total Aroclor 1254	18.12



- 1
- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13
- 14
- 15
- 16
- 17
- 18



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 32000 **Lot No.:** A0206810

Description : Pesticide Surrogate Mix
Pesticide Surrogate Mix 200 µg/mL, Acetone, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : April 30, 2030 **Storage:** 10°C or colder

Handling: Contains PCBs - sonicate prior to use. **Ship:** Ambient

P13348
↓
P13357
10
DAUF
04/25/2024

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	2,4,5,6-Tetrachloro-m-xylene	877-09-8	RP220407	99%	200.3 µg/mL	+/- 11.1143
2	Decachlorobiphenyl (BZ# 209)	2051-24-3	30638	99%	200.6 µg/mL	+/- 11.1298

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Acetone
CAS # 67-64-1
Purity 99%

Tech Tips:

Decachlorobiphenyl has poor solubility in most organic solvents. The maximum concentration that can be prepared in acetone, hexane, or isooctane is 200µg/mL. Temperature will affect the solubility as well. Storing solutions at reduced temperatures will cause decachlorobiphenyl to precipitate.

Products containing decachlorobiphenyl must be sonicated for a minimum of 10 minutes prior to opening the ampul. Because each ultrasonic bath operates at a different energy level, 10 minutes is a guideline only. Longer sonication time will not affect product quality.

These precautions apply to working solutions prepared in your laboratory as well. The amount of compound that precipitates depends on concentration AND temperature. If you store your standards at a temperature lower than 4°C (even dilute solutions), allow extra sonication time.

Quality Confirmation Test

Column:

30m x .25mm x .2um
Rtx-CLP II (cat.# 11323)

Carrier Gas:

helium-constant pressure 20 psi.

Temp. Program:

200°C to 300°C
@ 25°C/min. (hold 10 min.)

Inj. Temp:

250°C

Det. Temp:

300°C

Det. Type:


ECD

Split Vent:


10 ml/min.

Inj. Vol

1µl




This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.


Laith Clemente - Operations Technician I

Date Mixed: 22-Jan-2024


Balance Serial # 1128360905


Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 24-Jan-2024

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

P 13348
↓
P 13357
10


SAUF
04/25/2025



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 32007 **Lot No.:** A0207511

Description : Aroclor® 1221 Standard

Aroclor® 1221 Standard 1,000 µg/mL, Hexane, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : May 31, 2030 **Storage:** 25°C nominal

Handling: This product contains PCBs. **Ship:** Ambient

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Aroclor 1221	11104-28-2	13192500	----%	1,002.7 µg/mL	+/- 55.6424

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Hexane
CAS # 110-54-3
Purity 99%

P 13374
↓
P 13375
/

[Signature]
05/6/2024

1
2
3
4
5
6
7
8
9
10
11
12
13
14
15
16
17
18

Quality Confirmation Test

Column:

30m x .25mm x .2um
Rtx-CLP II (cat.# 11323)

Carrier Gas:

helium-constant pressure 20 psi.

Temp. Program:

200°C to 300°C
@ 25°C/min. (hold 10 min.)

Inj. Temp:

250°C

Det. Temp:

300°C

Det. Type:


ECD

Split Vent:

10 ml/min.

Inj. Vol

1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.


Laith Clemente - Operations Technician I

Date Mixed: 06-Feb-2024


Balance Serial # 1128360905


Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 08-Feb-2024


Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

P 13374 } ②
↓
P 13375 }


05/6/2024

- 1
- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13
- 14
- 15
- 16
- 17
- 18

ISO 17034



Agilent

Trusted Answers

Reference Material Certificate

Product Name: Aroclor 1242 Standard

Lot Number: 0006665550

Product Number: PP-312-1

Lot Issue Date: 08-Feb-2022

Storage Conditions: Store at Room Temperature (15° to 30°C).

Expiration Date: 31-Jan-2027

Component Name	CERTIFIED VALUES			CAS#	Analyte Lot
	Concentration	Expanded Uncertainty			
Aroclor 1242	100.4	± 0.5 µg/mL		053469-21-9	NT01020

Matrix: isooctane (2,2,4-trimethylpentane)

Description:

This document is prepared in accordance with ISO 17034 and Guide 31. This analytical reference material standard was manufactured and verified in accordance with an ISO 9001 registered quality system and analyte concentrations were verified by an ISO 17025 accredited laboratory. The concentration and uncertainty value at the 95% confidence level for each analyte, determined gravimetrically, is listed above.

Traceability:

The balances used for these measurements are calibrated with weights traceable to NIST in compliance with ANSI/NCSL Z540.3, ISO 9001, ISO 17025, and ISO 17034. Calibrated Class A glassware is used for volumetric measurements. Thermometers are calibrated against a NIST traceable thermometer in accordance with NIST Special Publication 1088.

Homogeneity:

This analytical reference standard was unitized according to an in-house procedure and is guaranteed to be homogeneous. There is no minimum sub-sample size required.

Instructions for Use:

Sample aliquots for analysis should be withdrawn at 20°C to 25°C immediately after opening the container and should be processed without delay for the certified values to be valid within the stated uncertainties.

Safety:

Refer to the Safety Data Sheet on www.agilent.com for information regarding this analytical reference material.

Intended Use:

This analytical reference standard is intended for the preparation of working reference samples for use in routine laboratory analyses, calibration of instruments, validation of analytical methods, assessments of measurement methods, and continuing calibration verification.

Expiration of Certification:

The certification of this analytical reference standard is valid until the expiration date specified above, provided the material is handled and stored in accordance with the instructions given in this certificate. This certification is nullified if the material is damaged, contaminated, or otherwise modified.

P13589
↓
P13590AJ
10/14/24

Maintenance of Certification:

If substantive changes are noted that affect the certification before the expiration of this certificate, Agilent will notify the purchaser.

Sample lot approver:

Monica Bourgeois
QMS Representative



ISO 17034 Cert
No. AR-1936

RM was produced in accordance with the TUV/SUD registered ISO 9001:2015
Quality Management System. Cert# 95T215321

Page: 2 of 2

www.agilent.com/quality/

CSD-QA-015.1



ISO 17025
Cert No. AT-

Reference Material Certificate
Product Information Sheet

Product Name: Aroclor 1248 Standard

Lot Number: 0006726317

Product Number: PP-342-1

Lot Issue Date: 27-Jan-2023

Storage Conditions: Store at Room Temperature (15° to 30°C).

Expiration Date: 28-Feb-2031

Component Name	Concentration	Uncertainty	CAS#	Analyte Lot
Aroclor 1248	100.3 ±	0.5 µg/mL	012672-29-6	NT01582

Matrix: isooctane (2,2,4-trimethylpentane)

Description:

This document is prepared in accordance with ISO 17034 and Guide 31. This analytical reference material (RM) standard was manufactured and verified in accordance with an ISO 9001 registered quality system and analyte concentrations were verified by an ISO 17025 accredited laboratory. The concentration and uncertainty value at the 95% confidence level for each analyte, determined gravimetrically, is listed above. Purity values are taken from approved vendor raw material certificates.

Traceability:

The balances used for these measurements are calibrated with weights traceable to NIST in compliance with ANSI/NCSL Z540.3, ISO 9001, ISO 17025, and ISO 17034. Calibrated Class A glassware is used for volumetric measurements. Thermometers are calibrated against a NIST traceable thermometer in accordance with NIST Special Publication 1088.

Homogeneity:

This analytical reference (RM) standard was unitized according to an in-house procedure and is guaranteed to be homogeneous. There is no minimum sub-sample size required.

Instructions for Use:

Sample aliquots for analysis should be withdrawn at 20°C to 25°C immediately after opening the container and should be processed without delay for the certified values to be valid within the stated uncertainties.

Safety:

Refer to the Safety Data Sheet on www.agilent.com for information regarding this analytical reference material.

Intended Use:

This analytical reference (RM) standard is intended for the preparation of working reference samples for use in routine laboratory analyses, calibration of instruments, validation of analytical methods, assessments of measurement methods, and continuing calibration verification.

Expiration of Certification:

The certification of this analytical reference standard (RM) is valid until the expiration date specified above, provided the material is handled and stored in accordance with the instructions given in this certificate. This certification is nullified if the material is damaged, contaminated, or otherwise modified.

P13591
↓
P13592

AS
10/14/2024

Maintenance of Certification:

If substantive changes are noted that affect the certification before the expiration of this certificate, Agilent will notify the purchaser.

Sample lot approver:



Monica Bourgeois
QMS Representative



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 32010 **Lot No.:** A0214733

Description : Aroclor® 1248 Standard
Aroclor® 1248 Standard 1,000µg/mL, Hexane, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : November 30, 2030 **Storage:** 25°C nominal

Handling: This product contains PCBs. **Ship:** Ambient

P13695 } Y.P.
P13696 } 10/17/24

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Aroclor 1248	12672-29-6	15242800	----%	1,005.0 µg/mL	+/- 55.7700

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Hexane
CAS # 110-54-3
Purity 99%

Quality Confirmation Test

Column:

30m x .25mm x .2um
Rtx-CLP II (cat.# 11323)

Carrier Gas:

helium-constant pressure 20 psi.

Temp. Program:

200°C to 300°C
@ 25°C/min. (hold 10 min.)

Inj. Temp:

250°C

Det. Temp:

300°C

Det. Type:


ECD

Split Vent:

10 ml/min.

Inj. Vol

0.2µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Rebecca Gingerich - Operations Tech II

Date Mixed: 02-Aug-2024

Balance Serial # C322230531

Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 13-Aug-2024

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k \sqrt{u_{gravimetric}^2 + u_{homogeneity}^2 + u_{storage\ stability}^2 + u_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.

1
2
3
4
5
6
7
8
9
10
11
12
13
14
15
16
17
18



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 32039 **Lot No.:** A0210629

Description : Aroclor® 1016/1260 Mix
Aroclor® 1016/1260 Mix 1,000 µg/mL, Hexane, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : July 31, 2030 **Storage:** 25°C nominal

Handling: This product contains PCBs. **Ship:** Ambient

P13697
↓
P13701 } Y.P.
10/19/24

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Aroclor 1016	12674-11-2	07	----%	1,005.3 µg/mL	+/- 55.7809
2	Aroclor 1260	11096-82-5	1320657	----%	1,000.0 µg/mL	+/- 55.4850

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Hexane
CAS # 110-54-3
Purity 99%

Quality Confirmation Test

Column:

30m x .25mm x .2um
Rtx-CLP II (cat.# 11323)

Carrier Gas:

helium-constant pressure 20 psi.

Temp. Program:

200°C to 300°C
@ 25°C/min. { hold 10 min.}

Inj. Temp:

250°C

Det. Temp:

300°C

Det. Type:


ECD

Split Vent:


10 ml/min.

Inj. Vol

0.2µl




This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.


Laith Clemente - Operations Technician I

Date Mixed: 22-Apr-2024

Balance Serial # B442140311


Dillan Murphy - Operations Technician I

Date Passed: 24-Apr-2024

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k \sqrt{u_{gravimetric}^2 + u_{homogeneity}^2 + u_{storage\ stability}^2 + u_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.

1
2
3
4
5
6
7
8
9
10
11
12
13
14
15
16
17
18



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 32009 **Lot No.:** A0210232

Description : Aroclor® 1242 Standard
Aroclor® 1242 Standard 1,000 µg/mL, Hexane, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : July 31, 2030 **Storage:** 25°C nominal

Handling: This product contains PCBs. **Ship:** Ambient

P13706 } y.p.
P13707 } 10/17/24

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Aroclor 1242	53469-21-9	01141	----%	1,000.7 µg/mL	+/- 55.5295

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Hexane
CAS # 110-54-3
Purity 99%

Quality Confirmation Test

Column:

30m x .25mm x .2um
Rtx-CLP II (cat.# 11323)

Carrier Gas:

helium-constant pressure 20 psi.

Temp. Program:

200°C to 300°C
@ 25°C/min. (hold 10 min.)

Inj. Temp:

250°C

Det. Temp:

300°C

Det. Type:


ECD

Split Vent:

10 ml/min.

Inj. Vol

0.2µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Amanda Miller
Amanda Miller - Operations Tech III - ARM QC

Date Mixed: 11-Apr-2024

Balance Serial # B442140311

Christie Mills
Christie Mills - Operations Lead Tech - ARM QC

Date Passed: 17-Apr-2024

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k \sqrt{u_{gravimetric}^2 + u_{homogeneity}^2 + u_{storage\ stability}^2 + u_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.

1
2
3
4
5
6
7
8
9
10
11
12
13
14
15
16
17
18




110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 32000 **Lot No.:** A0214495

Description : Pesticide Surrogate Mix

Pesticide Surrogate Mix 200 µg/mL, Acetone, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : October 31, 2030 **Storage:** 10°C or colder

Handling: Contains PCBs - sonicate prior to use. **Ship:** Ambient

P19785
AJ
11/19/24
P19789

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	2,4,5,6-Tetrachloro-m-xylene	877-09-8	RP220407	99%	200.2 µg/mL	+/- 11.1087
2	Decachlorobiphenyl (BZ# 209)	2051-24-3	30679	99%	201.4 µg/mL	+/- 11.1753

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent:

Acetone

CAS # 67-64-1

Purity 99%

Tech Tips:

Decachlorobiphenyl has poor solubility in most organic solvents. The maximum concentration that can be prepared in acetone, hexane, or isooctane is 200µg/mL. Temperature will affect the solubility as well. Storing solutions at reduced temperatures will cause decachlorobiphenyl to precipitate.

Products containing decachlorobiphenyl must be sonicated for a minimum of 10 minutes prior to opening the ampul. Because each ultrasonic bath operates at a different energy level, 10 minutes is a guideline only. Longer sonication time will not affect product quality.

These precautions apply to working solutions prepared in your laboratory as well. The amount of compound that precipitates depends on concentration AND temperature. If you store your standards at a temperature lower than 4°C (even dilute solutions), allow extra sonication time.

Column:
DB30m x 25mm x 2um
Rtx-CLP II (cat.# 11323)

Carrier Gas:
Helium-constant pressure 20 psi.

Temp. Program:
200°C to 300°C
@ 25°C/min. (hold 10 min.)

Inj. Temp:
250°C


Det. Temp:
300°C

Det. Type:
ECD

Split Vent:
10 ml/min.

Inj. Vol
1µl

Quality Confirmation Test



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

A.O.P.

Aaron Enyart - Operations Tech I

Date Mixed: 29-Jul-2024 Balance Serial # B345965662

Jennifer Pollino

Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 01-Aug-2024

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397




110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 32000 **Lot No.:** A0214495

Description : Pesticide Surrogate Mix

Pesticide Surrogate Mix 200 µg/mL, Acetone, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : October 31, 2030 **Storage:** 10°C or colder

Handling: Contains PCBs - sonicate prior to use. **Ship:** Ambient

P19785
AJ
11/19/24
P19789

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	2,4,5,6-Tetrachloro-m-xylene	877-09-8	RP220407	99%	200.2 µg/mL	+/- 11.1087
2	Decachlorobiphenyl (BZ# 209)	2051-24-3	30679	99%	201.4 µg/mL	+/- 11.1753

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent:

Acetone

CAS # 67-64-1

Purity 99%

Tech Tips:

Decachlorobiphenyl has poor solubility in most organic solvents. The maximum concentration that can be prepared in acetone, hexane, or isooctane is 200µg/mL. Temperature will affect the solubility as well. Storing solutions at reduced temperatures will cause decachlorobiphenyl to precipitate.

Products containing decachlorobiphenyl must be sonicated for a minimum of 10 minutes prior to opening the ampul. Because each ultrasonic bath operates at a different energy level, 10 minutes is a guideline only. Longer sonication time will not affect product quality.

These precautions apply to working solutions prepared in your laboratory as well. The amount of compound that precipitates depends on concentration AND temperature. If you store your standards at a temperature lower than 4°C (even dilute solutions), allow extra sonication time.

Quality Confirmation Test

Column:
DB30m x 25mm x 2um
Rtx-CLP II (cat.# 11323)
Carrier Gas:
Helium-constant pressure 20 psi.

Temp. Program:
200°C to 300°C
@ 25°C/min. (hold 10 min.)


Inj. Temp:
250°C

Det. Temp:
300°C

Det. Type:
ECD

Split Vent:
10 ml/min.

Inj. Vol
1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

A. O. P.
Aaron Enyart - Operations Tech I
Jennifer Pollino
Jennifer Pollino - Operations Tech III - ARM QC

Date Mixed: 29-Jul-2024 **Balance Serial #** B345965662
Date Passed: 01-Aug-2024

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 32011 **Lot No.:** A0217391
Description : Aroclor® 1254 Standard
Aroclor® 1254 Standard 1,000 µg/mL, Hexane, 1mL/ampul
Container Size : 2 mL **Pkg Amt:** > 1 mL
Expiration Date : January 31, 2031 **Storage:** 25°C nominal
Handling: This product contains PCBs. **Ship:** Ambient

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Aroclor 1254	11097-69-1	124-191-B	----%	1,004.7 µg/mL	+/- 55.7515

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Hexane
CAS # 110-54-3
Purity 99%

P13830
↓
P13832
↓
P13834
AJ
12/09/24

Quality Confirmation Test

Column:
30m x .25mm x .2um
Rtx-CLP II (cat.# 11323)

Carrier Gas:
helium-constant pressure 20 psi.

Temp. Program:
200°C to 300°C
@ 25°C/min. (hold 10 min.)


Inj. Temp:
250°C

Det. Temp:
300°C

Det. Type:
ECD

Split Vent:
300 ml/min.

Inj. Vol
1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Michael Maye
Michael Maye - Operations Tech I

Date Mixed: 02-Oct-2024 **Balance Serial #** C322230531

Jennifer Pollino
Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 07-Oct-2024

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397




110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 32008

Lot No.: A0219655

Description : Aroclor® 1232 Standard

Aroclor® 1232 Standard 1,000 µg/mL, Hexane, 1mL/ampul

Container Size : 2 mL

Pkg Amt: > 1 mL

Expiration Date : March 31, 2031

Storage: 25°C nominal

Handling:

This product contains PCBs.

Ship: Ambient

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Aroclor 1232	11141-16-5	15665-01	---%	1,007.0 µg/mL	+/- 55.8810

Solvent: Hexane

CAS # 110-54-3

Purity 99%

* Expanded Uncertainty displayed in same units as Grav. Conc.

P13878



AJ
01/28/25

P13880

Column:
30m x 25mm x .2um
Ft-CLP II (cat.# 11323)
Carrier Gas:
helium-constant pressure 20 psi.

Temp. Program:
200°C to 300°C
@ 25°C/min. (hold 10 min.)

Inj. Temp:
250°C


Det. Temp:
300°C

Det. Type:
ECD

Split Vent:
10 ml/min.

Inj. Vol
1µl

Quality Confirmation Test



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Michael Maye
Michael Maye - Operations Tech I

Date Mixed: 02-Dec-2024 **Balance Serial #** C322230531

Brittany Federhik
Brittany Federhiko - Operations Tech I

Date Passed: 05-Dec-2024

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397




110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. :	32409	Lot No.:	A0217041
Description :	Aroclor® 1262 Standard		
	Aroclor® 1262 Standard 1,000 µg/mL, 1mL/ampul, Hexane		
Container Size :	2 mL	Pkg Amt:	> 1 mL
Expiration Date :	December 31, 2030	Storage:	25°C nominal
Handling:	This product contains PCBs.		
		Ship:	Ambient

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Aroclor 1262	37324-23-5	10849100	----%	1,000.0 µg/mL	+/- 55.4925


* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Hexane
CAS # 110-54-3
Purity 99%

P13881
AJ
01128125

Quality Confirmation Test

Column:
30m x .25mm x .2um
Rtx-CLP II (cat.# 11323)
Carrier Gas:
helium-constant pressure 20 psi.
Temp. Program:
200°C to 300°C
@ 25°C/min. (hold 10 min.)
Inj. Temp:
250°C
Det. Temp:
300°C
Det. Type:
ECD
Split Vent:
300 ml/min.
Inj. Vol
0.2µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.


Rebecca Gingerich - Operations Tech II


Jennifer Pollino - Operations Tech III - ARM QC

Date Mixed:	25-Sep-2024	Balance Serial #	C322230531
Date Passed:	30-Sep-2024		

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397





110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 32410

Lot No.: A0217264

Description : Aroclor® 1268 Standard

Aroclor® 1268 Standard 1,000 µg/mL, 1mL/ampul, Hexane

Container Size : 2 mL

Pkg Amt: > 1 mL

Expiration Date : January 31, 2031

Storage: 25°C nominal

Handling:

This product contains PCBs.

Ship: Ambient

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Aroclor 1268	11100-14-4	12353400	----%	1,009.8 µg/mL	+/- 56.0364

Solvent: Hexane

CAS # 110-54-3

Purity 99%

* Expanded Uncertainty displayed in same units as Grav. Conc.

P13884

+

AJ
01/28/25

P13886

Quality Confirmation Test

Column:
30m x .25mm x .2um
Rtx-CLP II (cat.# 11323)
Carrier Gas:
helium-constant pressure 20 psi.

Temp. Program:
200°C to 300°C
@ 25°C/min. (hold 10 min.)


Inj. Temp:
250°C

Det. Temp:
300°C

Det. Type:
ECD

Split Vent:
Split ratio 500:1

Inj. Vol
0.2µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Stacey Wanner

Stacey Wanner - Operations Technician I

Date Mixed: 01-Oct-2024 Balance Serial # 1128360905

Jennifer Pollino

Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 02-Oct-2024

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

n-Hexane 95%
ULTRA RESI-ANALYZED
For Organic Residue Analysis



Material No.: 9262-03
Batch No.: 24G1962003
Manufactured Date: 2024-05-23
Expiration Date: 2025-08-22
Revision No.: 0

W3147
W3147
COPATE. 02/03/2023
JP

Certificate of Analysis

Test	Specification	Result
FID-Sensitive Impurities (as 2-Octanol) Single Impurity Peak (ng/mL)	≤ 5	3
ECD Sensitive Impurities (as Heptachlor Epoxide) Single Peak (pg/mL)	≤ 10	1
ECD-Sensitive Impurities (as Ethylene Dibromide) - Single Impurity Peak (ng/mL)	≤ 5	1
Assay (Total Saturated C6 Isomers) (by GC, corrected for water)	≥ 99.5 %	99.7 %
Assay (as n-Hexane) (by GC, corrected for water)	≥ 95 %	98 %
Color (APHA)	≤ 10	5
Residue after Evaporation	≤ 1.0 ppm	0.1 ppm
Substances Darkened by H2SO4	Passes Test	Passes Test
Water (by KF, coulometric)	≤ 0.05 %	< 0.01 %

For Laboratory, Research, or Manufacturing Use
MEETS SPECIFICATIONS WITHIN THE EXPIRATION PERIOD
Country of Origin: USA
Packaging Site: Phillipsburg Mfg Ctr & DC

J. Croak
Jamie Croak
Director Quality Operations, Bioscience Products

n-Hexane 95%
ULTRA RESI-ANALYZED
For Organic Residue Analysis

avantor™



W3234 88
07/28/2025
07/28/2025

Material No.: 9262-03
Batch No.: 25C0362005
Manufactured Date: 2025-01-29
Expiration Date: 2026-04-30
Revision No.: 0

Certificate of Analysis

Test	Specification	Result
FID-Sensitive Impurities (as 2-Octanol) Single Impurity Peak (ng/mL)	<= 5	1
ECD Sensitive Impurities (as HeptachlorEpoxide) Single Peak (pg/mL)	<= 10	6
ECD-Sensitive Impurities (as EthyleneDibromide) - Single Impurity Peak (ng/mL)	<= 5	5
Assay (Total Saturated C ₆ Isomers) (by GC, corrected for water)	>= 99.5 %	100.0 %
Assay (as n-Hexane) (by GC, corrected for water)	>= 95 %	100 %
Color (APHA)	<= 10	10
Residue after Evaporation	<= 1.0 ppm	0.1 ppm
Substances Darkened by H ₂ SO ₄	Passes Test	Passes Test
Water (by KF, coulometric)	<= 0.05 %	<0.01 %

For Laboratory, Research, or Manufacturing Use
MEETS SPECIFICATIONS WITHIN THE EXPIRATION PERIOD

Country of Origin: United States
Packaging Site: Phillipsburg Mfg Ctr & DC

Jamie Croak
Director Quality Operations, Bioscience Production

For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.386.1700



SHIPPING DOCUMENTS

- 1
- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13
- 14
- 15
- 16
- 17
- 18

USEPA

DateShipped: 9/23/2025

CarrierName: Hand-Delivered

AirbillNo:

CHAIN OF CUSTODY RECORD

Case #:

Contact Name: Leticia Campiglia

Contact Phone: 732-570-4993

No: 2-092325-0004-0049-02

Cooler #: 2 of 2

Lab: Alliance Technical Group LLC (non-CLP)

Lab Phone: 908-789-8900

[illegible]

Special Instructions: RFP # 916. Please email results to leticia.campiglia@westonsolutions.com, smita.sumbaly@westonsolutions.com, ronan.clynch@westonsolutions.com and michael.lang@westonsolutions.com. 7 day preliminary TAT.

SAMPLES TRANSFERRED FROM	
1	2
3	4
5	6
7	8
9	10
11	12
13	14
15	16
17	18
19	20
21	22
23	24
25	26
27	28
29	30
31	32
33	34
35	36
37	38
39	40
41	42
43	44
45	46
47	48
49	50
51	52
53	54
55	56
57	58
59	60
61	62
63	64
65	66
67	68
69	70
71	72
73	74
75	76
77	78
79	80
81	82
83	84
85	86
87	88
89	90
91	92
93	94
95	96
97	98
99	100

CHAIN OF CUSTODY #

Items/Reason	Relinquished by (Signature and Organization)	Date/Time	Received by (Signature and Organization)	Date/Time	Sample Condition Upon Receipt
All Samples	LE / START VI Weston	9/23/25 1700		9-23-25 1700	1.3°C IR Gun #1 custody seals intact Temp. Bill present

Laboratory Certification

Certified By	License No.
Connecticut	PH-0830
DOD ELAP (ANAB)	L2219
Maine	2024021
Maryland	296
New Hampshire	255425
New Jersey	20012
New York	11376
Pennsylvania	68-00548
Soil Permit	525-24-234-08441
Texas	TX-C25-00189
Virginia	460312