



284 Sheffield Street, Mountainside, NJ 07092
Phone: 908 789 8900 Fax: 908 789 8922

CASE NARRATIVE

Weston Solutions, Inc.
Project Name: RFP 916
Project # N/A
Order ID # Q3179
Test Name: SPLP BNA

A. Number of Samples and Date of Receipt:

3 Solid samples were received on 09/23/2025.

B. Parameters

According to the Chain of Custody document, the following analyses were requested: EPH, SPLP BNA, SPLP Cyanide, SPLP Extraction, SPLP ICP Metals, SPLP Mercury, SPLP Metals, SPLP PCB, SPLP Pesticide, SPLP VOA and SPLP ZHE Ext. This data package contains results for SPLP BNA.

C. Analytical Techniques:

The samples were analyzed on instrument BNA_F using GC Column DB-UI 8270D which is 20 meters, 0.18 mm ID, 0.36 um df. The analysis of SPLP BNA was based on method 8270E and extraction was done based on method 3541.

D. QA/ QC Samples:

The Holding Times were met for all analysis.

The Surrogate recoveries met the acceptable criteria.

The Internal Standards Areas met the acceptable requirements.

The Retention Times were acceptable for all samples.

The MS {Q3179-02MS} with File ID: BF143839.D recoveries met the requirements for all compounds except for 2,3,4,6-Tetrachlorophenol[124%]. due to matrix interference, no corrective action is required.

The MSD {Q3179-03MSD} with File ID: BF143840.D recoveries met the acceptable requirements except for 2,3,4,6-Tetrachlorophenol[127%]. due to matrix interference, no corrective action is required.

The RPD met criteria.

The Blank Spike for {PB169880BS} with File ID: BF143832.D met requirements for all samples except for 2,6-Dinitrotoluene[116%], Benzoic acid[126%] and Benzyl Alcohol[94%]. The associate samples have no positive hit for these compounds therefore no corrective action was taken.

The Blank analysis did not indicate the presence of lab contamination.

The %RSD is greater than 20% in the Method 8270-BF092325.M for 2-Nitrophenol & 2,4-Dinitrotoluene & are passing on Linear regression, and 2,4-Dinitrophenol, 4,6-Dinitro-2-methylphenol, they are passing on Quadratic regression.

The Continuous Calibration File ID BF143830.D met the requirements except for 2-Nitrophenol and 4,6-Dinitro-2-methylphenol. The associate samples have no positive hit for this compound therefore no corrective action was taken.

The Tuning criteria met requirements.

E. Additional Comments:

Concentration of Water Sample:

Concentration ug/L = (Ax) (Is) (Vt) (DF) (GPC)

(Ais) ($\overline{\text{RRF}}$) (Vo) (Vi)

Where,

Ax = Area of the characteristic ion for the compound to be measured.

Ais = Area of the characteristic ion for the internal standard.

Is = Amount of internal standard injected in ng.

Vo = Volume of water extracted in mL.

Vi = Volume of extract injected in uL.

Vt = Volume of the concentrated extract in uL

$\overline{\text{RRF}}$ = Mean Relative Response Factor determined from the initial calibration standard.

GPC = $\frac{V_{in}}{V_{out}}$ = GPC factor (If no GPC is performed, GPC=1)

The Form 6 is not included in the data package because the Initial Calibration was performed using 7 points.

The soil samples results are based on a dry weight basis.

F. Manual Integration Comments:

Please refer to the Manual integration Report included with the Run Logs for information on the manual integrations performed.

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. The laboratory manager or his designee, as verified by the following signature has authorized release of the data contained in this hard copy data package.

Signature _____



284 Sheffield Street, Mountainside, NJ 07092
Phone: 908 789 8900 Fax: 908 789 8922