

Hit Summary Sheet
SW-846

SDG No.: Q3179

Client: Weston Solutions, Inc.

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	RDL	Units
Client ID:	EME-TS14-01							
Q3179-01	EME-TS14-01	WATER	Methylene Chloride	7.60		0.28	5.00	ug/L
			Total Voc :	7.60				
			Total Concentration:	7.60				



SAMPLE DATA

A

B

C

D

Report of Analysis

Client:	Weston Solutions, Inc.		Date Collected:	09/23/25	
Project:	RFP 916		Date Received:	09/23/25	
Client Sample ID:	EME-TS14-01		SDG No.:	Q3179	
Lab Sample ID:	Q3179-01		Matrix:	Water	
Analytical Method:	8260D		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	SPLP VOA	
GC Column:	RXI-624	ID : 0.25	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VN087956.D	1	09/25/25 15:45	VN092525

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
75-71-8	Dichlorodifluoromethane	0.22	U	0.22	5.00	ug/L
74-87-3	Chloromethane	0.32	U	0.32	5.00	ug/L
75-01-4	Vinyl Chloride	0.26	U	0.26	5.00	ug/L
141-78-6	Ethyl Acetate	0.31	U	0.31	5.00	ug/L
74-83-9	Bromomethane	1.40	U	1.40	5.00	ug/L
75-00-3	Chloroethane	0.47	U	0.47	5.00	ug/L
75-69-4	Trichlorofluoromethane	0.33	U	0.33	5.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	0.25	U	0.25	5.00	ug/L
75-35-4	1,1-Dichloroethene	0.23	U	0.23	5.00	ug/L
107-02-8	Acrolein	7.10	U	7.10	25.0	ug/L
107-13-1	Acrylonitrile	0.83	U	0.83	25.0	ug/L
67-64-1	Acetone	1.50	U	1.50	25.0	ug/L
75-15-0	Carbon Disulfide	0.21	U	0.21	5.00	ug/L
1634-04-4	Methyl tert-butyl Ether	0.16	U	0.16	5.00	ug/L
79-20-9	Methyl Acetate	0.27	U	0.27	5.00	ug/L
75-09-2	Methylene Chloride	7.60		0.28	5.00	ug/L
156-60-5	trans-1,2-Dichloroethene	0.23	U	0.23	5.00	ug/L
75-34-3	1,1-Dichloroethane	0.23	U	0.23	5.00	ug/L
110-82-7	Cyclohexane	1.50	U	1.50	5.00	ug/L
78-93-3	2-Butanone	0.98	U	0.98	25.0	ug/L
56-23-5	Carbon Tetrachloride	0.25	U	0.25	5.00	ug/L
594-20-7	2,2-Dichloropropane	0.21	U	0.21	5.00	ug/L
156-59-2	cis-1,2-Dichloroethene	0.19	U	0.19	5.00	ug/L
74-97-5	Bromochloromethane	0.22	U	0.22	5.00	ug/L
67-66-3	Chloroform	0.25	U	0.25	5.00	ug/L
71-55-6	1,1,1-Trichloroethane	0.20	U	0.20	5.00	ug/L
108-87-2	Methylcyclohexane	0.16	U	0.16	5.00	ug/L
563-58-6	1,1-Dichloropropene	0.15	U	0.15	5.00	ug/L
71-43-2	Benzene	0.15	U	0.15	5.00	ug/L
107-06-2	1,2-Dichloroethane	0.22	U	0.22	5.00	ug/L

Report of Analysis

Client:	Weston Solutions, Inc.		Date Collected:	09/23/25	
Project:	RFP 916		Date Received:	09/23/25	
Client Sample ID:	EME-TS14-01		SDG No.:	Q3179	
Lab Sample ID:	Q3179-01		Matrix:	Water	
Analytical Method:	8260D		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	SPLP VOA	
GC Column:	RXI-624	ID : 0.25	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VN087956.D	1	09/25/25 15:45	VN092525

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
79-01-6	Trichloroethene	0.090	U	0.090	5.00	ug/L
78-87-5	1,2-Dichloropropane	0.20	U	0.20	5.00	ug/L
74-95-3	Dibromomethane	0.25	U	0.25	5.00	ug/L
75-27-4	Bromodichloromethane	0.22	U	0.22	5.00	ug/L
108-10-1	4-Methyl-2-Pentanone	0.68	U	0.68	25.0	ug/L
108-88-3	Toluene	0.14	U	0.14	5.00	ug/L
10061-02-6	t-1,3-Dichloropropene	0.17	U	0.17	5.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	0.16	U	0.16	5.00	ug/L
79-00-5	1,1,2-Trichloroethane	0.21	U	0.21	5.00	ug/L
142-28-9	1,3-Dichloropropane	0.19	U	0.19	5.00	ug/L
110-75-8	2-Chloroethyl Vinyl ether	0.30	U	0.30	25.0	ug/L
591-78-6	2-Hexanone	0.89	U	0.89	25.0	ug/L
124-48-1	Dibromochloromethane	0.18	U	0.18	5.00	ug/L
106-93-4	1,2-Dibromoethane	0.15	U	0.15	5.00	ug/L
127-18-4	Tetrachloroethene	0.23	U	0.23	5.00	ug/L
108-90-7	Chlorobenzene	0.12	U	0.12	5.00	ug/L
630-20-6	1,1,1,2-Tetrachloroethane	0.19	U	0.19	5.00	ug/L
100-41-4	Ethyl Benzene	0.13	U	0.13	5.00	ug/L
179601-23-1	m/p-Xylenes	0.24	U	0.24	10.0	ug/L
95-47-6	o-Xylene	0.12	U	0.12	5.00	ug/L
100-42-5	Styrene	0.15	U	0.15	5.00	ug/L
75-25-2	Bromoform	0.19	U	0.19	5.00	ug/L
98-82-8	Isopropylbenzene	0.12	U	0.12	5.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	0.26	U	0.26	5.00	ug/L
96-18-4	1,2,3-Trichloropropane	0.35	U	0.35	5.00	ug/L
108-86-1	Bromobenzene	0.24	U	0.24	5.00	ug/L
95-49-8	2-Chlorotoluene	0.14	U	0.14	5.00	ug/L
108-67-8	1,3,5-Trimethylbenzene	0.15	U	0.15	5.00	ug/L
106-43-4	4-Chlorotoluene	0.13	U	0.13	5.00	ug/L
95-63-6	1,2,4-Trimethylbenzene	0.14	U	0.14	5.00	ug/L
541-73-1	1,3-Dichlorobenzene	0.16	U	0.16	5.00	ug/L

Report of Analysis

Client:	Weston Solutions, Inc.		Date Collected:	09/23/25	
Project:	RFP 916		Date Received:	09/23/25	
Client Sample ID:	EME-TS14-01		SDG No.:	Q3179	
Lab Sample ID:	Q3179-01		Matrix:	Water	
Analytical Method:	8260D		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	SPLP VOA	
GC Column:	RXI-624	ID : 0.25	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VN087956.D	1	09/25/25 15:45	VN092525

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
106-46-7	1,4-Dichlorobenzene	0.19	U	0.19	5.00	ug/L
95-50-1	1,2-Dichlorobenzene	0.16	U	0.16	5.00	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	0.53	U	0.53	5.00	ug/L
120-82-1	1,2,4-Trichlorobenzene	0.20	U	0.20	5.00	ug/L
87-61-6	1,2,3-Trichlorobenzene	0.20	U	0.20	5.00	ug/L
74-88-4	Methyl Iodide	0.83	U	0.83	5.00	ug/L
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	55.8		74 - 125	112%	SPK: 50
1868-53-7	Dibromofluoromethane	50.8		75 - 124	102%	SPK: 50
2037-26-5	Toluene-d8	47.6		86 - 113	95%	SPK: 50
460-00-4	4-Bromofluorobenzene	43.5		77 - 121	87%	SPK: 50
INTERNAL STANDARDS						
363-72-4	Pentafluorobenzene	165000	8.206			
540-36-3	1,4-Difluorobenzene	365000	9.082			
3114-55-4	Chlorobenzene-d5	357000	11.847			
3855-82-1	1,4-Dichlorobenzene-d4	167000	13.77			

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products



LAB CHRONICLE

OrderID:	Q3179	OrderDate:	9/24/2025 11:08:00 AM
Client:	Weston Solutions, Inc.	Project:	RFP 916
Contact:	Smita Sumbaly	Location:	J42

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
Q3179-01	EME-TS14-01	Water	SPLP VOA	8260D	09/23/25		09/25/25	09/23/25



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

A
B
C
D

Hit Summary Sheet
SW-846

SDG No.: Q3179
Client: Weston Solutions, Inc.

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	RDL	Units
Client ID :				0.000				
Total Svoc :					0.00			
Total Concentration:					0.00			



SAMPLE DATA

A

B

C

D

Report of Analysis

Client:	Weston Solutions, Inc.	Date Collected:	09/29/25
Project:	RFP 916	Date Received:	09/29/25
Client Sample ID:	PB169880TB	SDG No.:	Q3179
Lab Sample ID:	PB169880TB	Matrix:	Water
Analytical Method:	8270E	% Solid:	0
Sample Wt/Vol:	1000 Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	SPLP BNA
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF143833.D	1	09/29/25 08:42	09/30/25 15:04	PB169880

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
62-75-9	n-Nitrosodimethylamine	0.86	U	0.86	10.0	ug/L
110-86-1	Pyridine	1.30	U	1.30	5.00	ug/L
100-52-7	Benzaldehyde	3.90	U	3.90	10.0	ug/L
62-53-3	Aniline	1.50	U	1.50	5.00	ug/L
108-95-2	Phenol	0.91	U	0.91	5.00	ug/L
111-44-4	bis(2-Chloroethyl)ether	0.81	U	0.81	5.00	ug/L
95-57-8	2-Chlorophenol	0.58	U	0.58	5.00	ug/L
100-51-6	Benzyl Alcohol	1.60	UQ	1.60	10.0	ug/L
95-48-7	2-Methylphenol	1.10	U	1.10	5.00	ug/L
108-60-1	2,2-oxybis(1-Chloropropane)	1.30	U	1.30	5.00	ug/L
98-86-2	Acetophenone	0.74	U	0.74	5.00	ug/L
65794-96-9	3+4-Methylphenols	1.10	U	1.10	10.0	ug/L
621-64-7	n-Nitroso-di-n-propylamine	1.40	U	1.40	2.50	ug/L
67-72-1	Hexachloroethane	0.65	U	0.65	5.00	ug/L
98-95-3	Nitrobenzene	0.76	U	0.76	5.00	ug/L
78-59-1	Isophorone	0.75	U	0.75	5.00	ug/L
88-75-5	2-Nitrophenol	1.80	U	1.80	5.00	ug/L
105-67-9	2,4-Dimethylphenol	1.90	U	1.90	5.00	ug/L
111-91-1	bis(2-Chloroethoxy)methane	0.68	U	0.68	5.00	ug/L
120-83-2	2,4-Dichlorophenol	0.52	U	0.52	5.00	ug/L
65-85-0	Benzoic acid	4.20	UQ	4.20	10.0	ug/L
91-20-3	Naphthalene	0.50	U	0.50	5.00	ug/L
106-47-8	4-Chloroaniline	0.84	U	0.84	5.00	ug/L
87-68-3	Hexachlorobutadiene	0.54	U	0.54	5.00	ug/L
105-60-2	Caprolactam	1.10	U	1.10	10.0	ug/L
59-50-7	4-Chloro-3-methylphenol	0.59	U	0.59	5.00	ug/L
91-57-6	2-Methylnaphthalene	0.56	U	0.56	5.00	ug/L
77-47-4	Hexachlorocyclopentadiene	3.60	U	3.60	10.0	ug/L
88-06-2	2,4,6-Trichlorophenol	0.51	U	0.51	5.00	ug/L

Report of Analysis

Client:	Weston Solutions, Inc.	Date Collected:	09/29/25
Project:	RFP 916	Date Received:	09/29/25
Client Sample ID:	PB169880TB	SDG No.:	Q3179
Lab Sample ID:	PB169880TB	Matrix:	Water
Analytical Method:	8270E	% Solid:	0
Sample Wt/Vol:	1000 Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	SPLP BNA
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF143833.D	1	09/29/25 08:42	09/30/25 15:04	PB169880

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
95-95-4	2,4,5-Trichlorophenol	0.62	U	0.62	5.00	ug/L
92-52-4	1,1-Biphenyl	0.53	U	0.53	5.00	ug/L
91-58-7	2-Chloronaphthalene	0.61	U	0.61	5.00	ug/L
88-74-4	2-Nitroaniline	1.30	U	1.30	5.00	ug/L
131-11-3	Dimethylphthalate	0.61	U	0.61	5.00	ug/L
208-96-8	Acenaphthylene	0.75	U	0.75	5.00	ug/L
606-20-2	2,6-Dinitrotoluene	0.92	UQ	0.92	5.00	ug/L
99-09-2	3-Nitroaniline	1.10	U	1.10	5.00	ug/L
83-32-9	Acenaphthene	0.55	U	0.55	5.00	ug/L
51-28-5	2,4-Dinitrophenol	6.00	U	6.00	10.0	ug/L
100-02-7	4-Nitrophenol	2.40	U	2.40	10.0	ug/L
132-64-9	Dibenzofuran	0.61	U	0.61	5.00	ug/L
121-14-2	2,4-Dinitrotoluene	1.20	U	1.20	5.00	ug/L
84-66-2	Diethylphthalate	0.69	U	0.69	5.00	ug/L
7005-72-3	4-Chlorophenyl-phenylether	0.68	U	0.68	5.00	ug/L
86-73-7	Fluorene	0.63	U	0.63	5.00	ug/L
100-01-6	4-Nitroaniline	1.50	U	1.50	5.00	ug/L
534-52-1	4,6-Dinitro-2-methylphenol	2.90	U	2.90	10.0	ug/L
86-30-6	n-Nitrosodiphenylamine	0.58	U	0.58	5.00	ug/L
103-33-3	Azobenzene	0.81	U	0.81	5.00	ug/L
101-55-3	4-Bromophenyl-phenylether	0.40	U	0.40	5.00	ug/L
118-74-1	Hexachlorobenzene	0.52	U	0.52	5.00	ug/L
1912-24-9	Atrazine	1.00	U	1.00	5.00	ug/L
87-86-5	Pentachlorophenol	1.60	U	1.60	10.0	ug/L
85-01-8	Phenanthrene	0.50	U	0.50	5.00	ug/L
120-12-7	Anthracene	0.61	U	0.61	5.00	ug/L
86-74-8	Carbazole	0.72	U	0.72	5.00	ug/L
84-74-2	Di-n-butylphthalate	1.20	U	1.20	5.00	ug/L
206-44-0	Fluoranthene	0.82	U	0.82	5.00	ug/L
92-87-5	Benzidine	4.30	U	4.30	10.0	ug/L
129-00-0	Pyrene	0.50	U	0.50	5.00	ug/L

Report of Analysis

Client:	Weston Solutions, Inc.	Date Collected:	09/29/25
Project:	RFP 916	Date Received:	09/29/25
Client Sample ID:	PB169880TB	SDG No.:	Q3179
Lab Sample ID:	PB169880TB	Matrix:	Water
Analytical Method:	8270E	% Solid:	0
Sample Wt/Vol:	1000 Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	SPLP BNA
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF143833.D	1	09/29/25 08:42	09/30/25 15:04	PB169880

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
85-68-7	Butylbenzylphthalate	1.90	U	1.90	5.00	ug/L
91-94-1	3,3-Dichlorobenzidine	0.93	U	0.93	10.0	ug/L
56-55-3	Benzo(a)anthracene	0.45	U	0.45	5.00	ug/L
218-01-9	Chrysene	0.44	U	0.44	5.00	ug/L
117-81-7	Bis(2-ethylhexyl)phthalate	1.60	U	1.60	5.00	ug/L
117-84-0	Di-n-octyl phthalate	2.30	U	2.30	10.0	ug/L
205-99-2	Benzo(b)fluoranthene	0.49	U	0.49	5.00	ug/L
207-08-9	Benzo(k)fluoranthene	0.48	U	0.48	5.00	ug/L
50-32-8	Benzo(a)pyrene	0.55	U	0.55	5.00	ug/L
193-39-5	Indeno(1,2,3-cd)pyrene	0.59	U	0.59	5.00	ug/L
53-70-3	Dibenzo(a,h)anthracene	0.67	U	0.67	5.00	ug/L
191-24-2	Benzo(g,h,i)perylene	0.69	U	0.69	5.00	ug/L
95-94-3	1,2,4,5-Tetrachlorobenzene	0.52	U	0.52	5.00	ug/L
123-91-1	1,4-Dioxane	1.00	U	1.00	5.00	ug/L
58-90-2	2,3,4,6-Tetrachlorophenol	0.72	U	0.72	5.00	ug/L
90-12-0	1-Methylnaphthalene	0.66	U	0.66	5.00	ug/L

SURROGATES

367-12-4	2-Fluorophenol	109		23 - 138	73%	SPK: 150
13127-88-3	Phenol-d6	111		10 - 134	74%	SPK: 150
4165-60-0	Nitrobenzene-d5	82.1		67 - 132	82%	SPK: 100
321-60-8	2-Fluorobiphenyl	79.2		52 - 132	79%	SPK: 100
118-79-6	2,4,6-Tribromophenol	114		44 - 137	76%	SPK: 150
1718-51-0	Terphenyl-d14	94.6		42 - 152	95%	SPK: 100

INTERNAL STANDARDS

3855-82-1	1,4-Dichlorobenzene-d4	171000	6.887
1146-65-2	Naphthalene-d8	648000	8.163
15067-26-2	Acenaphthene-d10	342000	9.922
1517-22-2	Phenanthrene-d10	553000	11.41
1719-03-5	Chrysene-d12	479000	14.051
1520-96-3	Perylene-d12	680000	15.533

Report of Analysis

Client:	Weston Solutions, Inc.		Date Collected:	09/29/25	
Project:	RFP 916		Date Received:	09/29/25	
Client Sample ID:	PB169880TB		SDG No.:	Q3179	
Lab Sample ID:	PB169880TB		Matrix:	Water	
Analytical Method:	8270E		% Solid:	0	
Sample Wt/Vol:	1000	Units: mL	Final Vol:	1000	uL
Soil Aliquot Vol:		uL	Test:	SPLP BNA	
Extraction Type :		Decanted : N	Level :	LOW	
Injection Volume :		GPC Factor : 1.0	GPC Cleanup :	N	PH :
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF143833.D	1	09/29/25 08:42	09/30/25 15:04	PB169880

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 E = Value Exceeds Calibration Range
 Q = indicates LCS control criteria did not meet requirements
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 D = Dilution
 () = Laboratory InHouse Limit
 A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	Weston Solutions, Inc.	Date Collected:	09/23/25
Project:	RFP 916	Date Received:	09/23/25
Client Sample ID:	EME-TS14-01	SDG No.:	Q3179
Lab Sample ID:	Q3179-01	Matrix:	Water
Analytical Method:	8270E	% Solid:	0
Sample Wt/Vol:	1000 Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	SPLP BNA
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF143838.D	1	09/29/25 08:42	09/30/25 17:36	PB169880

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
62-75-9	n-Nitrosodimethylamine	0.86	U	0.86	10.0	ug/L
110-86-1	Pyridine	1.30	U	1.30	5.00	ug/L
100-52-7	Benzaldehyde	3.90	U	3.90	10.0	ug/L
62-53-3	Aniline	1.50	U	1.50	5.00	ug/L
108-95-2	Phenol	0.91	U	0.91	5.00	ug/L
111-44-4	bis(2-Chloroethyl)ether	0.81	U	0.81	5.00	ug/L
95-57-8	2-Chlorophenol	0.58	U	0.58	5.00	ug/L
100-51-6	Benzyl Alcohol	1.60	UQ	1.60	10.0	ug/L
95-48-7	2-Methylphenol	1.10	U	1.10	5.00	ug/L
108-60-1	2,2-oxybis(1-Chloropropane)	1.30	U	1.30	5.00	ug/L
98-86-2	Acetophenone	0.74	U	0.74	5.00	ug/L
65794-96-9	3+4-Methylphenols	1.10	U	1.10	10.0	ug/L
621-64-7	n-Nitroso-di-n-propylamine	1.40	U	1.40	2.50	ug/L
67-72-1	Hexachloroethane	0.65	U	0.65	5.00	ug/L
98-95-3	Nitrobenzene	0.76	U	0.76	5.00	ug/L
78-59-1	Isophorone	0.75	U	0.75	5.00	ug/L
88-75-5	2-Nitrophenol	1.80	U	1.80	5.00	ug/L
105-67-9	2,4-Dimethylphenol	1.90	U	1.90	5.00	ug/L
111-91-1	bis(2-Chloroethoxy)methane	0.68	U	0.68	5.00	ug/L
120-83-2	2,4-Dichlorophenol	0.52	U	0.52	5.00	ug/L
65-85-0	Benzoic acid	4.20	UQ	4.20	10.0	ug/L
91-20-3	Naphthalene	0.50	U	0.50	5.00	ug/L
106-47-8	4-Chloroaniline	0.84	U	0.84	5.00	ug/L
87-68-3	Hexachlorobutadiene	0.54	U	0.54	5.00	ug/L
105-60-2	Caprolactam	1.10	U	1.10	10.0	ug/L
59-50-7	4-Chloro-3-methylphenol	0.59	U	0.59	5.00	ug/L
91-57-6	2-Methylnaphthalene	0.56	U	0.56	5.00	ug/L
77-47-4	Hexachlorocyclopentadiene	3.60	U	3.60	10.0	ug/L
88-06-2	2,4,6-Trichlorophenol	0.51	U	0.51	5.00	ug/L

Report of Analysis

Client:	Weston Solutions, Inc.	Date Collected:	09/23/25
Project:	RFP 916	Date Received:	09/23/25
Client Sample ID:	EME-TS14-01	SDG No.:	Q3179
Lab Sample ID:	Q3179-01	Matrix:	Water
Analytical Method:	8270E	% Solid:	0
Sample Wt/Vol:	1000 Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	SPLP BNA
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF143838.D	1	09/29/25 08:42	09/30/25 17:36	PB169880

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
95-95-4	2,4,5-Trichlorophenol	0.62	U	0.62	5.00	ug/L
92-52-4	1,1-Biphenyl	0.53	U	0.53	5.00	ug/L
91-58-7	2-Chloronaphthalene	0.61	U	0.61	5.00	ug/L
88-74-4	2-Nitroaniline	1.30	U	1.30	5.00	ug/L
131-11-3	Dimethylphthalate	0.61	U	0.61	5.00	ug/L
208-96-8	Acenaphthylene	0.75	U	0.75	5.00	ug/L
606-20-2	2,6-Dinitrotoluene	0.92	UQ	0.92	5.00	ug/L
99-09-2	3-Nitroaniline	1.10	U	1.10	5.00	ug/L
83-32-9	Acenaphthene	0.55	U	0.55	5.00	ug/L
51-28-5	2,4-Dinitrophenol	6.00	U	6.00	10.0	ug/L
100-02-7	4-Nitrophenol	2.40	U	2.40	10.0	ug/L
132-64-9	Dibenzofuran	0.61	U	0.61	5.00	ug/L
121-14-2	2,4-Dinitrotoluene	1.20	U	1.20	5.00	ug/L
84-66-2	Diethylphthalate	0.69	U	0.69	5.00	ug/L
7005-72-3	4-Chlorophenyl-phenylether	0.68	U	0.68	5.00	ug/L
86-73-7	Fluorene	0.63	U	0.63	5.00	ug/L
100-01-6	4-Nitroaniline	1.50	U	1.50	5.00	ug/L
534-52-1	4,6-Dinitro-2-methylphenol	2.90	U	2.90	10.0	ug/L
86-30-6	n-Nitrosodiphenylamine	0.58	U	0.58	5.00	ug/L
103-33-3	Azobenzene	0.81	U	0.81	5.00	ug/L
101-55-3	4-Bromophenyl-phenylether	0.40	U	0.40	5.00	ug/L
118-74-1	Hexachlorobenzene	0.52	U	0.52	5.00	ug/L
1912-24-9	Atrazine	1.00	U	1.00	5.00	ug/L
87-86-5	Pentachlorophenol	1.60	U	1.60	10.0	ug/L
85-01-8	Phenanthrene	0.50	U	0.50	5.00	ug/L
120-12-7	Anthracene	0.61	U	0.61	5.00	ug/L
86-74-8	Carbazole	0.72	U	0.72	5.00	ug/L
84-74-2	Di-n-butylphthalate	1.20	U	1.20	5.00	ug/L
206-44-0	Fluoranthene	0.82	U	0.82	5.00	ug/L
92-87-5	Benzidine	4.30	U	4.30	10.0	ug/L
129-00-0	Pyrene	0.50	U	0.50	5.00	ug/L

Report of Analysis

Client:	Weston Solutions, Inc.	Date Collected:	09/23/25
Project:	RFP 916	Date Received:	09/23/25
Client Sample ID:	EME-TS14-01	SDG No.:	Q3179
Lab Sample ID:	Q3179-01	Matrix:	Water
Analytical Method:	8270E	% Solid:	0
Sample Wt/Vol:	1000 Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	SPLP BNA
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF143838.D	1	09/29/25 08:42	09/30/25 17:36	PB169880

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
85-68-7	Butylbenzylphthalate	1.90	U	1.90	5.00	ug/L
91-94-1	3,3-Dichlorobenzidine	0.93	U	0.93	10.0	ug/L
56-55-3	Benzo(a)anthracene	0.45	U	0.45	5.00	ug/L
218-01-9	Chrysene	0.44	U	0.44	5.00	ug/L
117-81-7	Bis(2-ethylhexyl)phthalate	1.60	U	1.60	5.00	ug/L
117-84-0	Di-n-octyl phthalate	2.30	U	2.30	10.0	ug/L
205-99-2	Benzo(b)fluoranthene	0.49	U	0.49	5.00	ug/L
207-08-9	Benzo(k)fluoranthene	0.48	U	0.48	5.00	ug/L
50-32-8	Benzo(a)pyrene	0.55	U	0.55	5.00	ug/L
193-39-5	Indeno(1,2,3-cd)pyrene	0.59	U	0.59	5.00	ug/L
53-70-3	Dibenzo(a,h)anthracene	0.67	U	0.67	5.00	ug/L
191-24-2	Benzo(g,h,i)perylene	0.69	U	0.69	5.00	ug/L
95-94-3	1,2,4,5-Tetrachlorobenzene	0.52	U	0.52	5.00	ug/L
123-91-1	1,4-Dioxane	1.00	U	1.00	5.00	ug/L
58-90-2	2,3,4,6-Tetrachlorophenol	0.72	U	0.72	5.00	ug/L
90-12-0	1-Methylnaphthalene	0.66	U	0.66	5.00	ug/L

SURROGATES

367-12-4	2-Fluorophenol	58.3		23 - 138	39%	SPK: 150
13127-88-3	Phenol-d6	38.4		10 - 134	26%	SPK: 150
4165-60-0	Nitrobenzene-d5	92.8		67 - 132	93%	SPK: 100
321-60-8	2-Fluorobiphenyl	85.7		52 - 132	86%	SPK: 100
118-79-6	2,4,6-Tribromophenol	139		44 - 137	92%	SPK: 150
1718-51-0	Terphenyl-d14	99.3		42 - 152	99%	SPK: 100

INTERNAL STANDARDS

3855-82-1	1,4-Dichlorobenzene-d4	143000	6.881
1146-65-2	Naphthalene-d8	546000	8.163
15067-26-2	Acenaphthene-d10	303000	9.922
1517-22-2	Phenanthrene-d10	531000	11.41
1719-03-5	Chrysene-d12	567000	14.057
1520-96-3	Perylene-d12	572000	15.533

Report of Analysis

Client:	Weston Solutions, Inc.		Date Collected:	09/23/25	
Project:	RFP 916		Date Received:	09/23/25	
Client Sample ID:	EME-TS14-01		SDG No.:	Q3179	
Lab Sample ID:	Q3179-01		Matrix:	Water	
Analytical Method:	8270E		% Solid:	0	
Sample Wt/Vol:	1000	Units: mL	Final Vol:	1000	uL
Soil Aliquot Vol:		uL	Test:	SPLP BNA	
Extraction Type :		Decanted : N	Level :	LOW	
Injection Volume :		GPC Factor : 1.0	GPC Cleanup :	N	PH :
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF143838.D	1	09/29/25 08:42	09/30/25 17:36	PB169880

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products



LAB CHRONICLE

OrderID:	Q3179	OrderDate:	9/24/2025 11:08:00 AM
Client:	Weston Solutions, Inc.	Project:	RFP 916
Contact:	Smita Sumbaly	Location:	J42

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
Q3179-01	EME-TS14-01	Water	SPLP BNA	8270E	09/23/25	09/29/25	09/30/25	09/23/25



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Hit Summary Sheet
SW-846

SDG No.: Q3179

Order ID: Q3179

Client: Weston Solutions, Inc.

Project ID: RFP 916

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	RDL	Units
-----------	-----------	--------	-----------	---------------	---	-----	-----	-------

Client ID :

Total Concentration: 0.000



SAMPLE DATA

A

B

C

D

Report of Analysis

Client:	Weston Solutions, Inc.		Date Collected:		
Project:	RFP 916		Date Received:	09/29/25	
Client Sample ID:	PB169882TB		SDG No.:	Q3179	
Lab Sample ID:	PB169882TB		Matrix:	WATER	
Analytical Method:	8081B		% Solid:	0	Decanted:
Sample Wt/Vol:	1000	Units: mL	Final Vol:	1000	uL
Soil Aliquot Vol:		uL	Test:	SPLP Pesticide	
Extraction Type:			Injection Volume :		
GPC Factor :	1.0	PH :			
Prep Method :	3510C				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PD090424.D	1	09/29/25 09:45	09/29/25 17:51	PB169882

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
319-84-6	alpha-BHC	0.00040	U	0.00040	0.0050	ug/L
319-85-7	beta-BHC	0.00050	U	0.00050	0.0050	ug/L
319-86-8	delta-BHC	0.0011	U	0.0011	0.0050	ug/L
58-89-9	gamma-BHC (Lindane)	0.00040	U	0.00040	0.0050	ug/L
76-44-8	Heptachlor	0.00030	U	0.00030	0.0050	ug/L
309-00-2	Aldrin	0.00040	U	0.00040	0.0050	ug/L
1024-57-3	Heptachlor epoxide	0.0010	U	0.0010	0.0050	ug/L
959-98-8	Endosulfan I	0.00030	U	0.00030	0.0050	ug/L
60-57-1	Dieldrin	0.00040	U	0.00040	0.0050	ug/L
72-55-9	4,4-DDE	0.00040	U	0.00040	0.0050	ug/L
72-20-8	Endrin	0.00030	U	0.00030	0.0050	ug/L
33213-65-9	Endosulfan II	0.00080	U	0.00080	0.0050	ug/L
72-54-8	4,4-DDD	0.00070	U	0.00070	0.0050	ug/L
1031-07-8	Endosulfan Sulfate	0.00040	U	0.00040	0.0050	ug/L
50-29-3	4,4-DDT	0.00040	U	0.00040	0.0050	ug/L
72-43-5	Methoxychlor	0.0011	U	0.0011	0.0050	ug/L
53494-70-5	Endrin ketone	0.00090	U	0.00090	0.0050	ug/L
7421-93-4	Endrin aldehyde	0.0011	U	0.0011	0.0050	ug/L
5103-71-9	alpha-Chlordane	0.00040	U	0.00040	0.0050	ug/L
5103-74-2	gamma-Chlordane	0.00040	U	0.00040	0.0050	ug/L
8001-35-2	Toxaphene	0.017	U	0.017	0.10	ug/L
57-74-9	Chlordane	0.0088	U	0.0088	0.050	ug/L
SURROGATES						
2051-24-3	Decachlorobiphenyl	14.5		57 - 171	72%	SPK: 20
877-09-8	Tetrachloro-m-xylene	15.5		61 - 148	78%	SPK: 20

Report of Analysis

Client:	Weston Solutions, Inc.		Date Collected:		
Project:	RFP 916		Date Received:	09/29/25	
Client Sample ID:	PB169882TB		SDG No.:	Q3179	
Lab Sample ID:	PB169882TB		Matrix:	WATER	
Analytical Method:	8081B		% Solid:	0	Decanted:
Sample Wt/Vol:	1000	Units: mL	Final Vol:	1000	uL
Soil Aliquot Vol:		uL	Test:	SPLP Pesticide	
Extraction Type:			Injection Volume :		
GPC Factor :	1.0	PH :			
Prep Method :	3510C				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PD090424.D	1	09/29/25 09:45	09/29/25 17:51	PB169882

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	Weston Solutions, Inc.		Date Collected:	09/23/25	
Project:	RFP 916		Date Received:	09/23/25	
Client Sample ID:	EME-TS14-01		SDG No.:	Q3179	
Lab Sample ID:	Q3179-01		Matrix:	WATER	
Analytical Method:	8081B		% Solid:	0	Decanted:
Sample Wt/Vol:	1000	Units: mL	Final Vol:	1000	uL
Soil Aliquot Vol:		uL	Test:	SPLP Pesticide	
Extraction Type:			Injection Volume :		
GPC Factor :	1.0	PH :			
Prep Method :	3510C				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PD090428.D	1	09/29/25 09:45	09/29/25 18:46	PB169882

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
319-84-6	alpha-BHC	0.00040	U	0.00040	0.0050	ug/L
319-85-7	beta-BHC	0.00050	U	0.00050	0.0050	ug/L
319-86-8	delta-BHC	0.0011	U	0.0011	0.0050	ug/L
58-89-9	gamma-BHC (Lindane)	0.00040	U	0.00040	0.0050	ug/L
76-44-8	Heptachlor	0.00030	U	0.00030	0.0050	ug/L
309-00-2	Aldrin	0.00040	U	0.00040	0.0050	ug/L
1024-57-3	Heptachlor epoxide	0.0010	U	0.0010	0.0050	ug/L
959-98-8	Endosulfan I	0.00030	U	0.00030	0.0050	ug/L
60-57-1	Dieldrin	0.00040	U	0.00040	0.0050	ug/L
72-55-9	4,4-DDE	0.00040	U	0.00040	0.0050	ug/L
72-20-8	Endrin	0.00030	U	0.00030	0.0050	ug/L
33213-65-9	Endosulfan II	0.00080	U	0.00080	0.0050	ug/L
72-54-8	4,4-DDD	0.00070	U	0.00070	0.0050	ug/L
1031-07-8	Endosulfan Sulfate	0.00040	U	0.00040	0.0050	ug/L
50-29-3	4,4-DDT	0.00040	U	0.00040	0.0050	ug/L
72-43-5	Methoxychlor	0.0011	U	0.0011	0.0050	ug/L
53494-70-5	Endrin ketone	0.00090	U	0.00090	0.0050	ug/L
7421-93-4	Endrin aldehyde	0.0011	U	0.0011	0.0050	ug/L
5103-71-9	alpha-Chlordane	0.00040	U	0.00040	0.0050	ug/L
5103-74-2	gamma-Chlordane	0.00040	U	0.00040	0.0050	ug/L
8001-35-2	Toxaphene	0.017	U	0.017	0.10	ug/L
57-74-9	Chlordane	0.0088	U	0.0088	0.050	ug/L
SURROGATES						
2051-24-3	Decachlorobiphenyl	14.9		57 - 171	74%	SPK: 20
877-09-8	Tetrachloro-m-xylene	22.0		61 - 148	110%	SPK: 20

Report of Analysis

Client:	Weston Solutions, Inc.		Date Collected:	09/23/25	
Project:	RFP 916		Date Received:	09/23/25	
Client Sample ID:	EME-TS14-01		SDG No.:	Q3179	
Lab Sample ID:	Q3179-01		Matrix:	WATER	
Analytical Method:	8081B		% Solid:	0	Decanted:
Sample Wt/Vol:	1000	Units: mL	Final Vol:	1000	uL
Soil Aliquot Vol:		uL	Test:	SPLP Pesticide	
Extraction Type:			Injection Volume :		
GPC Factor :	1.0	PH :			
Prep Method :	3510C				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PD090428.D	1	09/29/25 09:45	09/29/25 18:46	PB169882

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

LAB CHRONICLE

OrderID:	Q3179	OrderDate:	9/24/2025 11:08:00 AM
Client:	Weston Solutions, Inc.	Project:	RFP 916
Contact:	Smita Sumbaly	Location:	J42

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
Q3179-01	EME-TS14-01	Water			09/23/25			09/23/25
			EPH	NJEPH		09/29/25	09/29/25	
			SPLP PCB	8082A		09/29/25	09/29/25	
			SPLP Pesticide	8081B		09/29/25	09/29/25	



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Hit Summary Sheet
SW-846

SDG No.: Q3179

Order ID: Q3179

Client: Weston Solutions, Inc.

Project ID: RFP 916

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	RDL	Units
-----------	-----------	--------	-----------	---------------	---	-----	-----	-------

Client ID :

Total Concentration: 0.000

A

B

C

D



SAMPLE DATA

A

B

C

D

Report of Analysis

Client:	Weston Solutions, Inc.		Date Collected:		
Project:	RFP 916		Date Received:	09/29/25	
Client Sample ID:	PB169881TB		SDG No.:	Q3179	
Lab Sample ID:	PB169881TB		Matrix:	WATER	
Analytical Method:	8082A		% Solid:	0	Decanted:
Sample Wt/Vol:	1000	Units: mL	Final Vol:	10000	uL
Soil Aliquot Vol:		uL	Test:	SPLP PCB	
Extraction Type:			Injection Volume :		
GPC Factor :	1.0	PH :			
Prep Method :	3510C				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PP075420.D	1	09/29/25 09:15	09/29/25 20:48	PB169881

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
12674-11-2	Aroclor-1016	0.097	U	0.097	0.50	ug/L
11104-28-2	Aroclor-1221	0.13	U	0.13	0.50	ug/L
11141-16-5	Aroclor-1232	0.096	U	0.096	0.50	ug/L
53469-21-9	Aroclor-1242	0.12	U	0.12	0.50	ug/L
12672-29-6	Aroclor-1248	0.071	U	0.071	0.50	ug/L
11097-69-1	Aroclor-1254	0.094	U	0.094	0.50	ug/L
37324-23-5	Aroclor-1262	0.14	U	0.14	0.50	ug/L
11100-14-4	Aroclor-1268	0.11	U	0.11	0.50	ug/L
11096-82-5	Aroclor-1260	0.081	U	0.081	0.50	ug/L
SURROGATES						
877-09-8	Tetrachloro-m-xylene	20.2		30 - 173	101%	SPK: 20
2051-24-3	Decachlorobiphenyl	23.1		10 - 173	115%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	Weston Solutions, Inc.		Date Collected:	09/23/25	
Project:	RFP 916		Date Received:	09/23/25	
Client Sample ID:	EME-TS14-01		SDG No.:	Q3179	
Lab Sample ID:	Q3179-01		Matrix:	WATER	
Analytical Method:	8082A		% Solid:	0	Decanted:
Sample Wt/Vol:	1000	Units: mL	Final Vol:	10000	uL
Soil Aliquot Vol:		uL	Test:	SPLP PCB	
Extraction Type:			Injection Volume :		
GPC Factor :	1.0	PH :			
Prep Method :	3510C				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PP075421.D	1	09/29/25 09:15	09/29/25 21:04	PB169881

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
12674-11-2	Aroclor-1016	0.097	U	0.097	0.50	ug/L
11104-28-2	Aroclor-1221	0.13	U	0.13	0.50	ug/L
11141-16-5	Aroclor-1232	0.096	U	0.096	0.50	ug/L
53469-21-9	Aroclor-1242	0.12	U	0.12	0.50	ug/L
12672-29-6	Aroclor-1248	0.071	U	0.071	0.50	ug/L
11097-69-1	Aroclor-1254	0.094	U	0.094	0.50	ug/L
37324-23-5	Aroclor-1262	0.14	U	0.14	0.50	ug/L
11100-14-4	Aroclor-1268	0.11	U	0.11	0.50	ug/L
11096-82-5	Aroclor-1260	0.081	U	0.081	0.50	ug/L
SURROGATES						
877-09-8	Tetrachloro-m-xylene	29.6		30 - 173	148%	SPK: 20
2051-24-3	Decachlorobiphenyl	24.7		10 - 173	123%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

LAB CHRONICLE

OrderID:	Q3179	OrderDate:	9/24/2025 11:08:00 AM
Client:	Weston Solutions, Inc.	Project:	RFP 916
Contact:	Smita Sumbaly	Location:	J42

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
Q3179-01	EME-TS14-01	Water			09/23/25			09/23/25
			EPH	NJEPH		09/29/25	09/29/25	
			SPLP PCB	8082A		09/29/25	09/29/25	
			SPLP Pesticide	8081B		09/29/25	09/29/25	



SAMPLE DATA

A

B

C

Report of Analysis

Client:	Weston Solutions, Inc.	Date Collected:	09/23/25
Project:	RFP 916	Date Received:	09/23/25
Client Sample ID:	EME-TS14-01	SDG No.:	Q3179
Lab Sample ID:	Q3179-01	Matrix:	Water
Analytical Method:	NJEPH	% Solid:	0
Sample Wt/Vol:	1000 Units: mL	Final Vol:	2000 uL
Soil Aliquot Vol:	uL	Test:	EPH
Prep Method :	SW3510		

Prep Date :	Date Analyzed :	Prep Batch ID
09/29/25 08:30	09/29/25 19:56	PB169879

Datafile

CAS Number	Parameter	Conc.	Qualifier	Dilution	MDL	LOQ / CRQL	Units	
TARGETS								
Aliphatic C9-C12	Aliphatic C9-C12	8.68	J	1	8.28	30.0	ug/l	FE056055.D
Aliphatic C12-C16	Aliphatic C12-C16	17.2	J	1	13.5	20.0	ug/l	FE056055.D
Aliphatic C16-C21	Aliphatic C16-C21	100		1	8.34	30.0	ug/l	FE056055.D
Aliphatic C21-C28	Aliphatic C21-C28	12.5	U	1	12.5	40.0	ug/l	FE056055.D
Aliphatic C28-C40	Aliphatic C28-C40	403		1	23.8	60.0	ug/l	FE056055.D
Aromatic C10-C12	Aromatic C10-C12	3.71	J	1	2.07	20.0	ug/l	FF016476.D
Aromatic C12-C16	Aromatic C12-C16	7.91	J	1	4.26	30.0	ug/l	FF016476.D
Aromatic C16-C21	Aromatic C16-C21	43.8	J	1	8.27	50.0	ug/l	FF016476.D
Aromatic C21-C36	Aromatic C21-C36	23.8	U	1	23.8	80.0	ug/l	FF016476.D
Total AliphaticEPH	Total AliphaticEPH	529			66.4	180	ug/l	
Total AromaticEPH	Total AromaticEPH	55.4	J		38.4	180	ug/l	
Total EPH	Total EPH	584			105	360	ug/l	

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

Report of Analysis

Client:	Weston Solutions, Inc.	Date Collected:	09/23/25
Project:	RFP 916	Date Received:	09/23/25
Client Sample ID:	EME-TS14-01	SDG No.:	Q3179
Lab Sample ID:	Q3179-01	Matrix:	Water
Analytical Method:	NJEPH	% Solid:	0
Sample Wt/Vol:	1000 Units: mL	Final Vol:	2000 uL
Soil Aliquot Vol:	uL	Test:	EPH
Prep Method :	SW3510		

File ID :	Dilution:	Prep Date :	Date Analyzed :	Prep Batch ID
FE056055.D	1	09/29/25	09/29/25	PB169879

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
Aliphatic C9-C12	Aliphatic C9-C12	8.68	J	8.28	30.0	ug/l
Aliphatic C12-C16	Aliphatic C12-C16	17.2	J	13.5	20.0	ug/l
Aliphatic C16-C21	Aliphatic C16-C21	100		8.34	30.0	ug/l
Aliphatic C21-C28	Aliphatic C21-C28	40.0	U	12.5	40.0	ug/l
Aliphatic C28-C40	Aliphatic C28-C40	403		23.8	60.0	ug/l
SURROGATES						
3383-33-2	1-chlorooctadecane (SURR)	48.9		40 - 140	98%	SPK: 50
84-15-1	ortho-Terphenyl (SURR)	0.00		40 - 140	0%	SPK: 50

Quantitation Report For Aliphatic EPH Range.

Lab Sample ID:	Q3179-01	Acq On:	29 Sep 2025 19:56
Client Sample ID:	EME-TS14-01	Operator:	YP\AJ
Data file:	FE056055.D	Misc:	
Instrument:	FID_E	ALS Vial:	28
Dilution Factor:	1	Sample Multiplier:	1.00

Compound	R.T.		Response	Conc	highest_standard	Units
Aliphatic C9-C12	3.315	6.946	607540	4.339	300	ug/ml
Aliphatic C12-C16	6.947	10.396	1320678	8.611	200	ug/ml
Aliphatic C16-C21	10.397	13.773	8356375	50.376	300	ug/ml
Aliphatic C21-C28	13.774	17.443	920988	6.201	400	ug/ml
Aliphatic C28-C40	17.444	22.449	27507813	201.889	600	ug/ml
Aliphatic EPH	3.315	22.449	38713394	271.416		ug/ml
ortho-Terphenyl (SURR)	0.000	0.000	0	0		ug/ml
1-chlorooctadecane (SURR)	13.512	13.512	6818463	48.85		ug/ml
Aliphatic C9-C28	3.315	17.443	11205581	69.527	1200	ug/ml

Report of Analysis

Client:	Weston Solutions, Inc.		Date Collected:	09/23/25	
Project:	RFP 916		Date Received:	09/23/25	
Client Sample ID:	EME-TS14-01		SDG No.:	Q3179	
Lab Sample ID:	Q3179-01		Matrix:	Water	
Analytical Method:	NJEPH		% Solid:	0	
Sample Wt/Vol:	1000	Units: mL	Final Vol:	2000	uL
Soil Aliquot Vol:		uL	Test:	EPH	
Prep Method :	SW3510				

File ID :	Dilution:	Prep Date :	Date Analyzed :	Prep Batch ID
FF016476.D	1	09/29/25	09/29/25	PB169879

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
Aromatic C10-C12	Aromatic C10-C12	3.71	J	2.07	20.0	ug/l
Aromatic C12-C16	Aromatic C12-C16	7.91	J	4.26	30.0	ug/l
Aromatic C16-C21	Aromatic C16-C21	43.8	J	8.27	50.0	ug/l
Aromatic C21-C36	Aromatic C21-C36	23.8	U	23.8	80.0	ug/l
SURROGATES						
580-13-2	2-Bromonaphthalene (SURR)	49.2		40 - 140	98%	SPK: 50
321-60-8	2-Fluorobiphenyl (SURR)	48.6		40 - 140	97%	SPK: 50
84-15-1	ortho-Terphenyl (SURR)	47.1		40 - 140	94%	SPK: 50

Quantitation Report For Aromatic EPH Range.

Lab Sample ID:	Q3179-01	Acq On:	29 Sep 2025 19:33
Client Sample ID:	EME-TS14-01	Operator:	YP\AJ
Data file:	FF016476.D	Misc:	
Instrument:	FID_F	ALS Vial:	87
Dilution Factor:	1	Sample Multiplier:	1.00

Compound	R.T.		Response	Conc	highest_standard	Units
Aromatic C10-C12	4.521	6.397	266219	1.857	200	ug/ml
Aromatic C12-C16	6.398	9.101	574540	3.957	300	ug/ml
Aromatic C16-C21	9.102	13.420	3159636	21.916	500	ug/ml
Aromatic C21-C36	13.421	18.876	852645	6.517	800	ug/ml
Aromatic EPH	4.521	18.876	4853040	34.247		ug/ml
ortho-Terphenyl (SURR)	11.972	11.972	7183345	47.12		ug/ml
2-Bromonaphthalene (SURR)	8.027	8.027	6163999	49.21		ug/ml
2-Fluorobiphenyl (SURR)	8.902	8.902	4098461	48.59		ug/ml

Report of Analysis

Client:	Weston Solutions, Inc.		Date Collected:	
Project:	RFP 916		Date Received:	
Client Sample ID:	PB169879TB		SDG No.:	Q3179
Lab Sample ID:	PB169879TB		Matrix:	water
Analytical Method:	NJEPH		% Solid:	0
Sample Wt/Vol:	1000	Units: mL	Final Vol:	2000 uL
Soil Aliquot Vol:		uL	Test:	EPH
Prep Method :	SW3510			

File ID :	Dilution:	Prep Date :	Date Analyzed :	Prep Batch ID
FE056075.D	1		09/30/25	FE092925AL

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
Aliphatic C9-C12	Aliphatic C9-C12	8.28	U	8.28	30.0	ug/l
Aliphatic C12-C16	Aliphatic C12-C16	13.5	U	13.5	20.0	ug/l
Aliphatic C16-C21	Aliphatic C16-C21	8.34	U	8.34	30.0	ug/l
Aliphatic C21-C28	Aliphatic C21-C28	12.5	U	12.5	40.0	ug/l
Aliphatic C28-C40	Aliphatic C28-C40	23.8	U	23.8	60.0	ug/l
SURROGATES						
3383-33-2	1-chlorooctadecane (SURR)	33.6		40 - 140	67%	SPK: 50
84-15-1	ortho-Terphenyl (SURR)	0.00		40 - 140	0%	SPK: 50

Quantitation Report For Aliphatic EPH Range.

Lab Sample ID:	PB169879TB	Acq On:	30 Sep 2025 07:06
Client Sample ID:	PB169879TB	Operator:	YP\AJ
Data file:	FE056075.D	Misc:	
Instrument:	FID_E	ALS Vial:	44
Dilution Factor:	1	Sample Multiplier:	1.00

Compound	R.T.		Response	Conc	highest_standard	Units
Aliphatic C9-C12	3.315	6.946	0	0	300	ug/ml
Aliphatic C12-C16	6.947	10.396	0	0	200	ug/ml
Aliphatic C16-C21	10.397	13.773	0	0	300	ug/ml
Aliphatic C21-C28	13.774	17.443	0	0	400	ug/ml
Aliphatic C28-C40	17.444	22.449	0	0	600	ug/ml
Aliphatic EPH	3.315	22.449	0	0		ug/ml
ortho-Terphenyl (SURR)	0.000	0.000	0	0		ug/ml
1-chlorooctadecane (SURR)	13.511	13.511	4690622	33.61		ug/ml
Aliphatic C9-C28	3.315	17.443	0	0	1200	ug/ml

Report of Analysis

Client:	Weston Solutions, Inc.		Date Collected:	
Project:	RFP 916		Date Received:	
Client Sample ID:	PB169879TB		SDG No.:	Q3179
Lab Sample ID:	PB169879TB		Matrix:	water
Analytical Method:	NJEPH		% Solid:	0
Sample Wt/Vol:	1000	Units: mL	Final Vol:	2000 uL
Soil Aliquot Vol:		uL	Test:	EPH
Prep Method :	SW3510			

File ID :	Dilution:	Prep Date :	Date Analyzed :	Prep Batch ID
FF016472.D	1		09/29/25	FF092925AR

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
Aromatic C10-C12	Aromatic C10-C12	2.07	U	2.07	20.0	ug/l
Aromatic C12-C16	Aromatic C12-C16	4.26	U	4.26	30.0	ug/l
Aromatic C16-C21	Aromatic C16-C21	8.27	U	8.27	50.0	ug/l
Aromatic C21-C36	Aromatic C21-C36	23.8	U	23.8	80.0	ug/l
SURROGATES						
580-13-2	2-Bromonaphthalene (SURR)	48.0		40 - 140	96%	SPK: 50
321-60-8	2-Fluorobiphenyl (SURR)	43.2		40 - 140	86%	SPK: 50
84-15-1	ortho-Terphenyl (SURR)	38.9		40 - 140	78%	SPK: 50

Quantitation Report For Aromatic EPH Range.

Lab Sample ID:	PB169879TB	Acq On:	29 Sep 2025 17:35
Client Sample ID:	PB169879TB	Operator:	YP\AJ
Data file:	FF016472.D	Misc:	
Instrument:	FID_F	ALS Vial:	83
Dilution Factor:	1	Sample Multiplier:	1.00

Compound	R.T.		Response	Conc	highest_standard	Units
Aromatic C10-C12	4.521	6.397	0	0	200	ug/ml
Aromatic C12-C16	6.398	9.101	0	0	300	ug/ml
Aromatic C16-C21	9.102	13.420	0	0	500	ug/ml
Aromatic C21-C36	13.421	18.876	0	0	800	ug/ml
Aromatic EPH	4.521	18.876	0	0		ug/ml
2-Bromonaphthalene (SURR)	8.027	8.027	6015712	48.02		ug/ml
2-Fluorobiphenyl (SURR)	8.902	8.902	3646317	43.23		ug/ml
ortho-Terphenyl (SURR)	11.971	11.971	5934502	38.93		ug/ml

LAB CHRONICLE

OrderID:	Q3179	OrderDate:	9/24/2025 11:08:00 AM
Client:	Weston Solutions, Inc.	Project:	RFP 916
Contact:	Smita Sumbaly	Location:	J42

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
Q3179-01	EME-TS14-01	Water			09/23/25			09/23/25
			EPH	NJEPH		09/29/25	09/29/25	
			SPLP PCB	8082A		09/29/25	09/29/25	
			SPLP Pesticide	8081B		09/29/25	09/29/25	

Hit Summary Sheet
SW-846

SDG No.: Q3179
Client: Weston Solutions, Inc.

Order ID: Q3179
Project ID: RFP 916

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	RDL	Units
Client ID : EME-TS14-01								
Q3179-01	EME-TS14-01	Water	Aluminum	283		5.67	50.0	ug/L
Q3179-01	EME-TS14-01	Water	Barium	130		7.28	50.0	ug/L
Q3179-01	EME-TS14-01	Water	Cadmium	0.95	J	0.25	3.00	ug/L
Q3179-01	EME-TS14-01	Water	Calcium	152000		117	1000	ug/L
Q3179-01	EME-TS14-01	Water	Chromium	3.87	J	1.06	5.00	ug/L
Q3179-01	EME-TS14-01	Water	Copper	4.32	J	2.30	10.0	ug/L
Q3179-01	EME-TS14-01	Water	Iron	51.1		11.7	50.0	ug/L
Q3179-01	EME-TS14-01	Water	Magnesium	18500		122	1000	ug/L
Q3179-01	EME-TS14-01	Water	Manganese	3940		2.97	10.0	ug/L
Q3179-01	EME-TS14-01	Water	Nickel	2.26	J	1.53	20.0	ug/L
Q3179-01	EME-TS14-01	Water	Potassium	11200		459	1000	ug/L
Q3179-01	EME-TS14-01	Water	Sodium	5990		434	1000	ug/L
Q3179-01	EME-TS14-01	Water	Zinc	64.5		8.33	20.0	ug/L



SAMPLE DATA

A

B

C

D

Report of Analysis

Client:	Weston Solutions, Inc.	Date Collected:	09/23/25
Project:	RFP 916	Date Received:	09/23/25
Client Sample ID:	EME-TS14-01	SDG No.:	Q3179
Lab Sample ID:	Q3179-01	Matrix:	Water
Level (low/med):	low	% Solid:	0

Cas	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.	Prep Met.
7429-90-5	Aluminum	283		1	5.67	50.0	ug/L	09/25/25 12:30	09/29/25 18:22	6010D	SW3010
7440-36-0	Antimony	3.38	U	1	3.38	25.0	ug/L	09/25/25 12:30	09/29/25 18:22	6010D	SW3010
7440-38-2	Arsenic	2.56	U	1	2.56	10.0	ug/L	09/25/25 12:30	09/29/25 18:22	6010D	SW3010
7440-39-3	Barium	130		1	7.28	50.0	ug/L	09/25/25 12:30	09/29/25 18:22	6010D	SW3010
7440-41-7	Beryllium	0.28	U	1	0.28	3.00	ug/L	09/25/25 12:30	09/29/25 18:22	6010D	SW3010
7440-43-9	Cadmium	0.95	J	1	0.25	3.00	ug/L	09/25/25 12:30	09/29/25 18:22	6010D	SW3010
7440-70-2	Calcium	152000		1	117	1000	ug/L	09/25/25 12:30	09/29/25 18:22	6010D	SW3010
7440-47-3	Chromium	3.87	J	1	1.06	5.00	ug/L	09/25/25 12:30	09/29/25 18:22	6010D	SW3010
7440-48-4	Cobalt	1.13	U	1	1.13	15.0	ug/L	09/25/25 12:30	09/29/25 18:22	6010D	SW3010
7440-50-8	Copper	4.32	J	1	2.30	10.0	ug/L	09/25/25 12:30	09/29/25 18:22	6010D	SW3010
7439-89-6	Iron	51.1		1	11.7	50.0	ug/L	09/25/25 12:30	09/29/25 18:22	6010D	SW3010
7439-92-1	Lead	1.15	U	1	1.15	6.00	ug/L	09/25/25 12:30	09/29/25 18:22	6010D	SW3010
7439-95-4	Magnesium	18500		1	122	1000	ug/L	09/25/25 12:30	09/29/25 18:22	6010D	SW3010
7439-96-5	Manganese	3940		1	2.97	10.0	ug/L	09/25/25 12:30	09/29/25 18:22	6010D	SW3010
7439-97-6	Mercury	0.076	U	1	0.076	0.20	ug/L	09/25/25 12:15	09/25/25 15:50	7470A	
7440-02-0	Nickel	2.26	J	1	1.53	20.0	ug/L	09/25/25 12:30	09/29/25 18:22	6010D	SW3010
7440-09-7	Potassium	11200		1	459	1000	ug/L	09/25/25 12:30	09/29/25 18:22	6010D	SW3010
7782-49-2	Selenium	4.82	U	1	4.82	10.0	ug/L	09/25/25 12:30	09/29/25 18:22	6010D	SW3010
7440-22-4	Silver	0.81	U	1	0.81	5.00	ug/L	09/25/25 12:30	09/29/25 18:22	6010D	SW3010
7440-23-5	Sodium	5990	N	1	434	1000	ug/L	09/25/25 12:30	09/29/25 18:22	6010D	SW3010
7440-28-0	Thallium	2.19	U	1	2.19	20.0	ug/L	09/25/25 12:30	09/29/25 18:22	6010D	SW3010
7440-62-2	Vanadium	3.13	U	1	3.13	20.0	ug/L	09/25/25 12:30	09/29/25 18:22	6010D	SW3010
7440-66-6	Zinc	64.5		1	8.33	20.0	ug/L	09/25/25 12:30	09/29/25 18:22	6010D	SW3010

Color Before:	Colorless	Clarity Before:	Clear	Texture:
Color After:	Colorless	Clarity After:	Clear	Artifacts:
Comments:	SPLP Metals			

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 D = Dilution
 Q = indicates LCS control criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 * = indicates the duplicate analysis is not within control limits.
 E = Indicates the reported value is estimated because of the presence of interference.
 OR = Over Range
 N = Spiked sample recovery not within control limits

LAB CHRONICLE

OrderID:	Q3179	OrderDate:	9/24/2025 11:08:00 AM
Client:	Weston Solutions, Inc.	Project:	RFP 916
Contact:	Smita Sumbaly	Location:	J42

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
Q3179-01	EME-TS14-01	Water			09/23/25			09/23/25
			SPLP ICP Metals	6010D		09/25/25	09/29/25	
			SPLP Mercury	7470A		09/25/25	09/25/25	



SAMPLE DATA

A

B

C

Report of Analysis

Client:	Weston Solutions, Inc.	Date Collected:	09/23/25 12:45
Project:	RFP 916	Date Received:	09/23/25
Client Sample ID:	EME-TS14-01	SDG No.:	Q3179
Lab Sample ID:	Q3179-01	Matrix:	WATER
		% Solid:	0

Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
Cyanide	0.0032	J	1	0.00096	0.0050	mg/L	09/29/25 08:45	09/29/25 15:57	9012B

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

D = Dilution

Q = indicates LCS control criteria did not meet requirements

H = Sample Analysis Out Of Hold Time

J = Estimated Value

B = Analyte Found in Associated Method Blank

* = indicates the duplicate analysis is not within control limits.

E = Indicates the reported value is estimated because of the presence of interference.

OR = Over Range

N =Spiked sample recovery not within control limits

LAB CHRONICLE

OrderID:	Q3179	OrderDate:	9/24/2025 11:08:00 AM
Client:	Weston Solutions, Inc.	Project:	RFP 916
Contact:	Smita Sumbaly	Location:	J42

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
Q3179-01	EME-TS14-01	WATER			09/23/25 12:45			09/23/25
			SPLP Cyanide	9012B		09/29/25	09/29/25 15:57	